

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 5/30/2026 7:38:06 PM

JOB DESCRIPTION

RED-HILL
PFAS: Aiea Gulch Wells Pump 2
RUSH Weekly Red Hill

JOB NUMBER

380-215227-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
5/30/2026 7:38:06 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Qualifiers

LCMS

Qualifier	Qualifier Description
F1	MS and/or MSD recovery exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-215227-1

Job ID: 380-215227-1

Eurofins Pomona

Job Narrative 380-215227-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 5/20/2026 9:27 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.2°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-215227-1

No Detections.

Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-215227-2

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

This Detection Summary does not include radiochemical test results.

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-215227-1

Date Collected: 05/18/26 09:53

Matrix: Drinking Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:07	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	58		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C6 PFDA	74		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C5 PFHxA	76		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C4 PFHpA	77		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C8 PFOA	78		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C9 PFNA	82		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C7 PFUnA	80		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C2 PFDoA	83		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C4 PFBA	91		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C5 PFPeA	81		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C3 PFBS	109		50 - 200			05/29/26 07:57	05/30/26 00:07	1
13C3 PFHxS	105		50 - 200			05/29/26 07:57	05/30/26 00:07	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-215227-1

Date Collected: 05/18/26 09:53

Matrix: Drinking Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	112		50 - 200	05/29/26 07:57	05/30/26 00:07	1
13C2-4:2-FTS	123		50 - 200	05/29/26 07:57	05/30/26 00:07	1
13C2-6:2-FTS	109		50 - 200	05/29/26 07:57	05/30/26 00:07	1
13C2-8:2-FTS	108		50 - 200	05/29/26 07:57	05/30/26 00:07	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:14	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	103		70 - 130	05/22/26 16:13	05/23/26 15:14	1
13C2 PFHxA	110		70 - 130	05/22/26 16:13	05/23/26 15:14	1
13C2 PFDA	101		70 - 130	05/22/26 16:13	05/23/26 15:14	1
13C3-GenX	105		70 - 130	05/22/26 16:13	05/23/26 15:14	1

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-215227-2

Date Collected: 05/18/26 09:53

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-215227-2

Date Collected: 05/18/26 09:53

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		05/29/26 07:57	05/30/26 00:16	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	60		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C6 PFDA	67		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C5 PFHxA	74		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C4 PFHpA	74		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C8 PFOA	73		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C9 PFNA	73		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C7 PFUnA	71		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C2 PFDoA	74		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C4 PFBA	87		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C5 PFPeA	80		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C3 PFBS	101		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C3 PFHxS	103		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C8 PFOS	109		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C2-4:2-FTS	122		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C2-6:2-FTS	101		50 - 200	05/29/26 07:57	05/30/26 00:16	1
13C2-8:2-FTS	107		50 - 200	05/29/26 07:57	05/30/26 00:16	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-215227-2

Date Collected: 05/18/26 09:53

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
N-methylperfluorooctanesulfonamide cetic acid (NMeFOSAA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
N-ethylperfluorooctanesulfonamide cetic acid (NEtFOSAA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/22/26 16:13	05/23/26 15:34	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	99		70 - 130			05/22/26 16:13	05/23/26 15:34	1
13C2 PFHxA	109		70 - 130			05/22/26 16:13	05/23/26 15:34	1
13C2 PFDA	105		70 - 130			05/22/26 16:13	05/23/26 15:34	1
13C3-GenX	106		70 - 130			05/22/26 16:13	05/23/26 15:34	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-215227-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-215227-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-215227-1
 SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-215227-1	AIEA GULCH WELLS PUMP 2 (331	103	110	101	105

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-215204-B-1-A MS	Matrix Spike	98	108	103	105
380-215204-C-1-A MSD	Matrix Spike Duplicate	93	107	106	106
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	99	109	105	106
LCS 380-229099/21-A	Lab Control Sample	99	101	103	105
MBL 380-229099/19-A	Method Blank	100	103	104	99
MRL 380-229099/20-A	Lab Control Sample	98	103	100	101

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-215227-1	AIEA GULCH WELLS PUMP 2 (331	58	74	76	77	78	82	80	83

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-215227-1	AIEA GULCH WELLS PUMP 2 (331	91	81	109	105	112	123	109	108

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDaA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-214934-B-3-A MS	Matrix Spike	92	93	99	106	100	103	99	98
380-214934-C-3-B MSD	Matrix Spike Duplicate	88	88	96	92	94	98	91	92
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	60	67	74	74	73	73	71	74
LCS 380-230294/22-A	Lab Control Sample	79	93	96	93	96	101	95	96
MBL 380-230294/20-A	Method Blank	60	69	74	79	77	78	73	74
MRL 380-230294/21-A	Lab Control Sample	68	80	82	80	84	86	83	85

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-214934-B-3-A MS	Matrix Spike	118	114	116	110	111	127	113	109
380-214934-C-3-B MSD	Matrix Spike Duplicate	112	104	111	107	108	120	107	103
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	87	80	101	103	109	122	101	107
LCS 380-230294/22-A	Lab Control Sample	103	95	113	109	110	121	110	110
MBL 380-230294/20-A	Method Blank	93	87	104	110	115	125	117	112
MRL 380-230294/21-A	Lab Control Sample	93	87	111	111	109	130	111	115

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA

Isotope Dilution Summary

Client: City & County of Honolulu

Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-230294/20-A
Matrix: Water
Analysis Batch: 230525

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 230294

Analyte	MBL	MBL	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.25		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		05/29/26 07:57	05/29/26 22:12	1

Isotope Dilution	MBL	MBL	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C3 HFPO-DA	60		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C6 PFDA	69		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C5 PFHxA	74		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C4 PFHpA	79		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C8 PFOA	77		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C9 PFNA	78		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C7 PFUnA	73		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C2 PFDoA	74		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C4 PFBA	93		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C5 PFPeA	87		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C3 PFBS	104		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C3 PFHxS	110		50 - 200	05/29/26 07:57	05/29/26 22:12	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-230294/20-A
Matrix: Water
Analysis Batch: 230525

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 230294

Isotope Dilution	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C8 PFOS	115		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C2-4:2-FTS	125		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C2-6:2-FTS	117		50 - 200	05/29/26 07:57	05/29/26 22:12	1
13C2-8:2-FTS	112		50 - 200	05/29/26 07:57	05/29/26 22:12	1

Lab Sample ID: LCS 380-230294/22-A
Matrix: Water
Analysis Batch: 230525

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 230294

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	60.4	57.0		ng/L		94	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	60.4	60.0		ng/L		99	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	60.4	57.1		ng/L		95	70 - 130
Perfluorobutanesulfonic acid (PFBS)	60.4	57.1		ng/L		95	70 - 130
Perfluorodecanoic acid (PFDA)	60.4	64.7		ng/L		107	70 - 130
Perfluorododecanoic acid (PFDoA)	60.4	63.0		ng/L		104	70 - 130
Perfluoroheptanoic acid (PFHpA)	60.4	62.4		ng/L		103	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	60.4	61.3		ng/L		101	70 - 130
Perfluorohexanoic acid (PFHxA)	60.4	56.9		ng/L		94	70 - 130
Perfluorononanoic acid (PFNA)	60.4	59.5		ng/L		99	70 - 130
Perfluorooctanesulfonic acid (PFOS)	60.4	61.7		ng/L		102	70 - 130
Perfluorooctanoic acid (PFOA)	60.4	60.1		ng/L		100	70 - 130
Perfluoroundecanoic acid (PFUnA)	60.4	61.4		ng/L		102	70 - 130
Perfluorobutanoic acid (PFBA)	60.4	58.4		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	60.4	61.8		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	60.4	52.8		ng/L		87	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	60.4	61.0		ng/L		101	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	60.4	48.8		ng/L		81	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	60.4	58.9		ng/L		98	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	60.4	62.0		ng/L		103	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	60.4	58.0		ng/L		96	70 - 130
Perfluoropentanoic acid (PFPeA)	60.4	63.1		ng/L		105	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	60.4	63.8		ng/L		106	70 - 130

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-215227-1
 SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-230294/22-A
Matrix: Water
Analysis Batch: 230525

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 230294

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	60.4	59.1		ng/L		98	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	79		50 - 200				
13C6 PFDA	93		50 - 200				
13C5 PFHxA	96		50 - 200				
13C4 PFHpA	93		50 - 200				
13C8 PFOA	96		50 - 200				
13C9 PFNA	101		50 - 200				
13C7 PFUnA	95		50 - 200				
13C2 PFDoA	96		50 - 200				
13C4 PFBA	103		50 - 200				
13C5 PFPeA	95		50 - 200				
13C3 PFBS	113		50 - 200				
13C3 PFHxS	109		50 - 200				
13C8 PFOS	110		50 - 200				
13C2-4:2-FTS	121		50 - 200				
13C2-6:2-FTS	110		50 - 200				
13C2-8:2-FTS	110		50 - 200				

Lab Sample ID: MRL 380-230294/21-A
Matrix: Water
Analysis Batch: 230525

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 230294

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	2.28	J	ng/L		113	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	2.17	J	ng/L		108	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.87	J	ng/L		93	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.82	J	ng/L		91	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.92	J	ng/L		96	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.54	J	ng/L		127	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	2.51	J	ng/L		125	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.09	J	ng/L		104	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.10	J	ng/L		105	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.00	J	ng/L		99	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.22	J	ng/L		111	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	2.43	J	ng/L		121	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.00	J	ng/L		99	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	2.42	J	ng/L		121	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	2.13	J	ng/L		106	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-230294/21-A

Matrix: Water

Analysis Batch: 230525

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 230294

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	2.47	J	ng/L		123	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	1.84	J	ng/L		92	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	2.29	J	ng/L		114	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	1.78	J	ng/L		89	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.01	2.02	J	ng/L		100	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	2.03	J	ng/L		101	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.93	J	ng/L		96	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	2.20	J	ng/L		109	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	2.30	J	ng/L		114	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	1.79	J	ng/L		89	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	68		50 - 200
13C6 PFDA	80		50 - 200
13C5 PFHxA	82		50 - 200
13C4 PFHpA	80		50 - 200
13C8 PFOA	84		50 - 200
13C9 PFNA	86		50 - 200
13C7 PFUnA	83		50 - 200
13C2 PFDoA	85		50 - 200
13C4 PFBA	93		50 - 200
13C5 PFPeA	87		50 - 200
13C3 PFBS	111		50 - 200
13C3 PFHxS	111		50 - 200
13C8 PFOS	109		50 - 200
13C2-4:2-FTS	130		50 - 200
13C2-6:2-FTS	111		50 - 200
13C2-8:2-FTS	115		50 - 200

Lab Sample ID: 380-214934-B-3-A MS

Matrix: Water

Analysis Batch: 230525

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 230294

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	57.1		ng/L		95	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	55.3		ng/L		92	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	53.9		ng/L		90	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-214934-B-3-A MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 230525

Prep Batch: 230294

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Hexafluoropropylene Oxide	<2.0		60.2	55.4		ng/L		92	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	4.4		60.2	60.2		ng/L		93	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.2	67.5		ng/L		111	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	58.8		ng/L		98	70 - 130
Perfluoroheptanoic acid (PFHpA)	3.5		60.2	57.8		ng/L		90	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	3.6		60.2	64.1		ng/L		101	70 - 130
Perfluorohexanoic acid (PFHxA)	13		60.2	71.9		ng/L		98	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.2	59.7		ng/L		98	70 - 130
Perfluorooctanesulfonic acid (PFOS)	7.8		60.2	68.6		ng/L		101	70 - 130
Perfluorooctanoic acid (PFOA)	7.8		60.2	63.9		ng/L		93	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	57.3		ng/L		95	70 - 130
Perfluorobutanoic acid (PFBA)	5.5		60.2	59.0		ng/L		89	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	62.1		ng/L		103	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	51.9		ng/L		86	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	64.8		ng/L		108	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0	F1	60.2	41.7	F1	ng/L		69	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.2	54.2		ng/L		90	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	59.4		ng/L		99	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	58.2		ng/L		97	70 - 130
Perfluoropentanoic acid (PFPeA)	10		60.2	66.8		ng/L		94	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	58.9		ng/L		98	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	58.4		ng/L		96	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	92		50 - 200
13C6 PFDA	93		50 - 200
13C5 PFHxA	99		50 - 200
13C4 PFHpA	106		50 - 200
13C8 PFOA	100		50 - 200
13C9 PFNA	103		50 - 200
13C7 PFUnA	99		50 - 200
13C2 PFDoA	98		50 - 200
13C4 PFBA	118		50 - 200
13C5 PFPeA	114		50 - 200
13C3 PFBS	116		50 - 200
13C3 PFHxS	110		50 - 200
13C8 PFOS	111		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-214934-B-3-A MS

Matrix: Water

Analysis Batch: 230525

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 230294

<i>Isotope Dilution</i>	<i>MS</i>	<i>MS</i>	<i>Limits</i>
	<i>%Recovery</i>	<i>Qualifier</i>	
13C2-4:2-FTS	127		50 - 200
13C2-6:2-FTS	113		50 - 200
13C2-8:2-FTS	109		50 - 200

Lab Sample ID: 380-214934-C-3-B MSD

Matrix: Water

Analysis Batch: 230525

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 230294

<i>Analyte</i>	<i>Sample</i>	<i>Sample</i>	<i>Spike</i>	<i>MSD</i>	<i>MSD</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i>	<i>RPD</i>	<i>RPD</i>
	<i>Result</i>	<i>Qualifier</i>	<i>Added</i>	<i>Result</i>	<i>Qualifier</i>			<i>Limits</i>	<i>Limits</i>	<i>RPD</i>	<i>Limit</i>
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.1	55.9		ng/L		93	70 - 130	2	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.1	55.6		ng/L		92	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.1	62.5		ng/L		104	70 - 130	15	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.1	56.1		ng/L		93	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)	4.4		60.1	60.7		ng/L		94	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		60.1	66.4		ng/L		109	70 - 130	2	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.1	59.3		ng/L		99	70 - 130	1	30
Perfluoroheptanoic acid (PFHpA)	3.5		60.1	65.6		ng/L		103	70 - 130	13	30
Perfluorohexanesulfonic acid (PFHxS)	3.6		60.1	63.0		ng/L		99	70 - 130	2	30
Perfluorohexanoic acid (PFHxA)	13		60.1	72.7		ng/L		100	70 - 130	1	30
Perfluorononanoic acid (PFNA)	<2.0		60.1	60.3		ng/L		99	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	7.8		60.1	69.5		ng/L		103	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	7.8		60.1	67.3		ng/L		99	70 - 130	5	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.1	60.1		ng/L		100	70 - 130	5	30
Perfluorobutanoic acid (PFBA)	5.5		60.1	60.4		ng/L		91	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.1	65.2		ng/L		108	70 - 130	5	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.1	56.3		ng/L		94	70 - 130	8	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.1	57.9		ng/L		96	70 - 130	11	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0	F1	60.1	53.7		ng/L		89	70 - 130	25	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.1	59.3		ng/L		99	70 - 130	9	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.1	62.5		ng/L		104	70 - 130	5	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.1	62.0		ng/L		103	70 - 130	6	30
Perfluoropentanoic acid (PFPeA)	10		60.1	70.4		ng/L		100	70 - 130	5	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.1	62.3		ng/L		104	70 - 130	5	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.1	57.2		ng/L		94	70 - 130	2	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	88		50 - 200
13C6 PFDA	88		50 - 200
13C5 PFHxA	96		50 - 200
13C4 PFHpA	92		50 - 200
13C8 PFOA	94		50 - 200
13C9 PFNA	98		50 - 200
13C7 PFUnA	91		50 - 200
13C2 PFDoA	92		50 - 200
13C4 PFBA	112		50 - 200
13C5 PFPeA	104		50 - 200
13C3 PFBS	111		50 - 200
13C3 PFHxS	107		50 - 200
13C8 PFOS	108		50 - 200
13C2-4:2-FTS	120		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	103		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-229099/19-A
Matrix: Water
Analysis Batch: 229167

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 229099

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
Perfluorotridecanoic acid (PFTTrDA)	<0.36		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		05/22/26 16:13	05/23/26 13:01	1

Surrogate	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
d5-NEtFOSAA	100		70 - 130	05/22/26 16:13	05/23/26 13:01	1
13C2 PFHxA	103		70 - 130	05/22/26 16:13	05/23/26 13:01	1
13C2 PFDA	104		70 - 130	05/22/26 16:13	05/23/26 13:01	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-229099/19-A
Matrix: Water
Analysis Batch: 229167

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 229099

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	99		70 - 130	05/22/26 16:13	05/23/26 13:01	1

Lab Sample ID: LCS 380-229099/21-A
Matrix: Water
Analysis Batch: 229167

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 229099

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	25.0	25.6		ng/L		102	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	25.0	25.4		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.0	25.1		ng/L		100	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.0	24.4		ng/L		98	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.0	24.7		ng/L		99	70 - 130
Perfluorohexanoic acid (PFHxA)	25.0	26.8		ng/L		107	70 - 130
Perfluorododecanoic acid (PFDoA)	25.0	26.1		ng/L		104	70 - 130
Perfluorooctanoic acid (PFOA)	25.0	26.4		ng/L		106	70 - 130
Perfluorodecanoic acid (PFDA)	25.0	25.5		ng/L		102	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.0	27.6		ng/L		110	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.0	25.2		ng/L		101	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.0	26.4		ng/L		106	70 - 130
Perfluorononanoic acid (PFNA)	25.0	26.5		ng/L		106	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.0	20.9		ng/L		83	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.0	27.4		ng/L		110	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	25.0	25.8		ng/L		103	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.0	24.7		ng/L		99	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.0	25.9		ng/L		103	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	99		70 - 130
13C2 PFHxA	101		70 - 130
13C2 PFDA	103		70 - 130
13C3-GenX	105		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-229099/20-A
Matrix: Water
Analysis Batch: 229167

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 229099

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.03	J	ng/L		102	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.06	J	ng/L		103	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.09	J	ng/L		105	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	1.91	J	ng/L		96	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	2.11	J	ng/L		105	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	1.98	J	ng/L		99	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.00	J	ng/L		100	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.95	J	ng/L		97	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.19	J	ng/L		110	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	1.68	J	ng/L		84	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	2.10	J	ng/L		105	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	1.98	J	ng/L		99	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.91	J	ng/L		96	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.08	J	ng/L		104	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	98		70 - 130
13C2 PFHxA	103		70 - 130
13C2 PFDA	100		70 - 130
13C3-GenX	101		70 - 130

Lab Sample ID: 380-215204-B-1-A MS
Matrix: Water
Analysis Batch: 229167

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 229099

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	25.8		ng/L		103	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.1	24.9		ng/L		99	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	26.9		ng/L		108	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	24.3		ng/L		97	70 - 130

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

LCMS

Prep Batch: 229099

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215227-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	537.1 DW	
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	537.1 DW	
MBL 380-229099/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-229099/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-229099/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-215204-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-215204-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 229167

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215227-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	EPA 537.1 V2	229099
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	EPA 537.1 V2	229099
MBL 380-229099/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	229099
LCS 380-229099/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	229099
MRL 380-229099/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	229099
380-215204-B-1-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	229099
380-215204-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	229099

Prep Batch: 230294

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215227-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	533	
MBL 380-230294/20-A	Method Blank	Total/NA	Water	533	
LCS 380-230294/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-230294/21-A	Lab Control Sample	Total/NA	Water	533	
380-214934-B-3-A MS	Matrix Spike	Total/NA	Water	533	
380-214934-C-3-B MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 230525

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215227-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	230294
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	533	230294
MBL 380-230294/20-A	Method Blank	Total/NA	Water	533	230294
LCS 380-230294/22-A	Lab Control Sample	Total/NA	Water	533	230294
MRL 380-230294/21-A	Lab Control Sample	Total/NA	Water	533	230294
380-214934-B-3-A MS	Matrix Spike	Total/NA	Water	533	230294
380-214934-C-3-B MSD	Matrix Spike Duplicate	Total/NA	Water	533	230294

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-215227-1

Date Collected: 05/18/26 09:53

Matrix: Drinking Water

Date Received: 05/20/26 09:27

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			230294	XTD8	EA POM	05/29/26 07:57
Total/NA	Analysis	533		1	230525	Y5FM	EA POM	05/30/26 00:07
Total/NA	Prep	537.1 DW			229099	E2HD	EA POM	05/22/26 16:13
Total/NA	Analysis	EPA 537.1 V2		1	229167	SZ9R	EA POM	05/23/26 15:14

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-215227-2

Date Collected: 05/18/26 09:53

Matrix: Water

Date Received: 05/20/26 09:27

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			230294	XTD8	EA POM	05/29/26 07:57
Total/NA	Analysis	533		1	230525	Y5FM	EA POM	05/30/26 00:16
Total/NA	Prep	537.1 DW			229099	E2HD	EA POM	05/22/26 16:13
Total/NA	Analysis	EPA 537.1 V2		1	229167	SZ9R	EA POM	05/23/26 15:34

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-215227-1
SDG: PFAS: Aiea Gulch Wells Pump 2


Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-215227-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Drinking Water	05/18/26 09:53	05/20/26 09:27	HI0000331
380-215227-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Water	05/18/26 09:53	05/20/26 09:27	HI0000331

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Monrovia, CA (Suite 100)

750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone (626) 386-1100

Chain of Custody Record



Client Information		Sampler <u>Patrick Regan</u>		Lab PM Lopez, Maria		Carrier Tracking No(s)		COC No: 80-215227 COC					
Client Contact Kirk Iwamoto		Phone +1 808 748 5840		E-Mail Maria.Lopez@et.eurofinsus.com		State of Origin		Page Page 1 of 1					
Company City & County of Honolulu		PWSID		Analysis Requested						Job #.			
Address 630 South Beretania Street, Chemistry Lab		Due Date Requested		Field Filtered Sample (Yes or No) Perform MS/MSD (Yes or No) 537 1_DW_PREC - 537 1 Full List 533 - All Analytes						Total Number of Containers		Preservation Codes A HCL M Hexane B NaOH N None C Zn Acetate O - AsNaO2 D Nitric Acid P Na2O4S E NaHSO4 Q Na2SO3 F MeOH R Na2S2O3 G Amchlor S - H2SO4 H - Ascorbic Acid T TSP Dodecahydrate I Ice U Acetone J DI Water V MCAA K EDTA W pH 4-5 L EDA Y Trizma Z - other (specify) Other: NH4 Acetate	
City Honolulu		TAT Requested (days) RUSH											
State Zip HI, 96843		Compliance Project Δ No											
Phone 808-748-5840 (tel)		PO # C20525101 exp 05312023											
Email kiwamoto@hbws.org		WO #:											
Project Name RED-HILL/HBWS sites Event Desc. RUSH Weekly Red Hill		Project # 38001111											
Site		SSOW#											
Sample Identification		Sample Date		Sample Time		Sample Type (C=comp, G=grab)		Matrix (W=water, S=solid, O=wastefoil, BT=Tissue, A=Air)		Preservation Code:		Special Instructions/Note:	
AIEA GULCH WELLS PUMP 2 (331-202-TP072)		18-May-2026		09:53		G		Water		3 3			
FB AIEA GULCH WELLS PUMP 2 (331-202-TP072)		18-May-2026		09:53						1 1			
Possible Hazard Identification										Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)			
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological										<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months			
Deliverable Requested I, II, III, IV, Other (specify)										Special Instructions/QC Requirements			
Empty Kit Relinquished by		Date		Time		Method of Shipment: <u>FedEx 872003245647</u>							
Relin: [Redacted]		Date/Time: <u>5/19/26 12:00</u>		Company: <u>HBWS</u>		Received by: <u>[Signature]</u>		Date/Time: <u>5/20/26 9:27</u>		Company: <u>EBM</u>			
Relinquished by:		Date/Time:		Company:		Received by:		Date/Time:		Company:			
Relinquished by:		Date/Time:		Company:		Received by:		Date/Time:		Company:			
Custody Seals Intact. Δ Yes Δ No		Custody Seal No		Cooler Temperature(s) °C and Other Remarks: <u>(631A) 2 2 + 0.0 2 2 90(-) frozen</u>									



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-215227-1
SDG Number: PFAS: Aiea Gulch Wells Pump 2

Login Number: 215227

List Number: 1

Creator: Avila, Ivan

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

