

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 8/19/2026 6:04:37 PM

JOB DESCRIPTION

RED-HILL
PFAS: Halawa Shaft Viewing Pool

JOB NUMBER

380-230960-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1145

Generated
8/19/2026 6:04:37 PM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	14
QC Association Summary	25
Lab Chronicle	26
Certification Summary	27
Method Summary	28
Sample Summary	29
Chain of Custody	30
Receipt Checklists	31

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-230960-1

Job ID: 380-230960-1

Eurofins Pomona

Job Narrative 380-230960-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 8/13/2026 9:46 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.6°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	3.2		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.0		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.6		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.7		2.0	ng/L	1		EPA 537.1 V2	Total/NA

Client Sample ID: FB HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-1

Date Collected: 08/11/26 09:30

Matrix: Water

Date Received: 08/13/26 09:46

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorohexanesulfonic acid (PFHxS)	3.2		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorooctanesulfonic acid (PFOS)	3.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:48	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	81		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C6 PFDA	66		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C5 PFHxA	80		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C4 PFHpA	80		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C8 PFOA	75		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C9 PFNA	70		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C7 PFUnA	66		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C2 PFDoA	72		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C4 PFBA	93		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C5 PFPeA	87		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C3 PFBS	113		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C3 PFHxS	116		50 - 200	08/16/26 14:20	08/18/26 15:48	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-1

Date Collected: 08/11/26 09:30

Matrix: Water

Date Received: 08/13/26 09:46

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	114		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C2-4:2-FTS	129		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C2-6:2-FTS	116		50 - 200	08/16/26 14:20	08/18/26 15:48	1
13C2-8:2-FTS	110		50 - 200	08/16/26 14:20	08/18/26 15:48	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorooctanesulfonic acid (PFOS)	3.6		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorohexanesulfonic acid (PFHxS)	3.7		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
4,8-Dioxo-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 15:51	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	104		70 - 130			08/16/26 16:30	08/17/26 15:51	1
13C2 PFHxA	112		70 - 130			08/16/26 16:30	08/17/26 15:51	1
13C2 PFDA	110		70 - 130			08/16/26 16:30	08/17/26 15:51	1
13C3-GenX	106		70 - 130			08/16/26 16:30	08/17/26 15:51	1

Client Sample ID: FB HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-2

Date Collected: 08/11/26 09:30

Matrix: Water

Date Received: 08/13/26 09:46

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
4,8-Dioxo-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: FB HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-2

Date Collected: 08/11/26 09:30

Matrix: Water

Date Received: 08/13/26 09:46

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		08/16/26 14:20	08/18/26 15:58	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	70		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C6 PFDA	76		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C5 PFHxA	75		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C4 PFHpA	78		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C8 PFOA	76		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C9 PFNA	76		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C7 PFUnA	77		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C2 PFDoA	77		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C4 PFBA	88		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C5 PFPeA	90		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C3 PFBS	106		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C3 PFHxS	112		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C8 PFOS	108		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C2-4:2-FTS	117		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C2-6:2-FTS	107		50 - 200			08/16/26 14:20	08/18/26 15:58	1
13C2-8:2-FTS	107		50 - 200			08/16/26 14:20	08/18/26 15:58	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: FB HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-2

Date Collected: 08/11/26 09:30

Matrix: Water

Date Received: 08/13/26 09:46

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		08/16/26 16:30	08/17/26 16:02	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	105		70 - 130			08/16/26 16:30	08/17/26 16:02	1
13C2 PFHxA	106		70 - 130			08/16/26 16:30	08/17/26 16:02	1
13C2 PFDA	107		70 - 130			08/16/26 16:30	08/17/26 16:02	1
13C3-GenX	101		70 - 130			08/16/26 16:30	08/17/26 16:02	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.2		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.6		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.7		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-230654-E-1-A MS	Matrix Spike	102	105	106	105
380-230654-F-1-A MSD	Matrix Spike Duplicate	99	102	106	98
380-230960-1	HALAWA SHAFT VIEWING POOL	104	112	110	106
380-230960-2	FB HALAWA SHAFT VIEWING POOL	105	106	107	101
LCS 380-246690/21-A	Lab Control Sample	99	102	106	103
MBL 380-246690/19-A	Method Blank	101	103	102	101
MRL 380-246690/20-A	Lab Control Sample	103	97	102	95

Surrogate Legend

d5NEFOS = d5-NEtFOSAA

PFHxA = 13C2 PFHxA

PFDA = 13C2 PFDA

GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-230958-A-1-B MS	Matrix Spike	104	105	102	102	98	102	103	107
380-230958-A-1-C MSD	Matrix Spike Duplicate	97	106	103	99	102	101	103	107
380-230960-1	HALAWA SHAFT VIEWING POOL	81	66	80	80	75	70	66	72
380-230960-2	FB HALAWA SHAFT VIEWING POOL	70	76	75	78	76	76	77	77
LCS 380-246683/22-A	Lab Control Sample	84	99	92	92	90	94	99	105
MBL 380-246683/20-A	Method Blank	83	98	96	100	98	98	102	104
MRL 380-246683/21-A	Lab Control Sample	82	96	92	92	94	97	95	101
		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-230958-A-1-B MS	Matrix Spike	107	106	110	103	102	105	105	100
380-230958-A-1-C MSD	Matrix Spike Duplicate	105	108	109	102	104	108	100	105
380-230960-1	HALAWA SHAFT VIEWING POOL	93	87	113	116	114	129	116	110
380-230960-2	FB HALAWA SHAFT VIEWING POOL	88	90	106	112	108	117	107	107
LCS 380-246683/22-A	Lab Control Sample	97	96	101	115	114	112	104	105
MBL 380-246683/20-A	Method Blank	98	105	105	108	104	110	102	104
MRL 380-246683/21-A	Lab Control Sample	98	100	110	114	108	116	106	109
Surrogate Legend									
HFPODA = 13C3 HFPO-DA									
C6PFDA = 13C6 PFDA									
13C5PHA = 13C5 PFHxA									
C4PFHA = 13C4 PFHpA									
C8PFOA = 13C8 PFOA									
C9PFNA = 13C9 PFNA									
13C7PUA = 13C7 PFUnA									
PFDaA = 13C2 PFDaA									
PFBA = 13C4 PFBA									
PFPeA = 13C5 PFPeA									
C3PFBS = 13C3 PFBS									
C3PFHS = 13C3 PFHxS									
C8PFOS = 13C8 PFOS									
42FTS = 13C2-4:2-FTS									
62FTS = 13C2-6:2-FTS									
82FTS = 13C2-8:2-FTS									

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-246683/20-A

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 246683

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<0.29		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9CI-PF3ONS)	<0.30		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluorobutanoic acid (PFBA)	<0.68		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.36		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.47		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.46		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.24		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.45		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoropentanoic acid (PFPeA)	<0.37		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.35		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1
Perfluoropentanesulfonic acid (PFPeS)	<0.38		2.0	ng/L		08/16/26 14:20	08/17/26 11:13	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	83		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C6 PFDA	98		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C5 PFHxA	96		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C4 PFHpA	100		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C8 PFOA	98		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C9 PFNA	98		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C7 PFUnA	102		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C2 PFDoA	104		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C4 PFBA	98		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C5 PFPeA	105		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C3 PFBS	105		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C3 PFHxS	108		50 - 200	08/16/26 14:20	08/17/26 11:13	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-246683/20-A

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 246683

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	104		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C2-4:2-FTS	110		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C2-6:2-FTS	102		50 - 200	08/16/26 14:20	08/17/26 11:13	1
13C2-8:2-FTS	104		50 - 200	08/16/26 14:20	08/17/26 11:13	1

Lab Sample ID: LCS 380-246683/22-A

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 246683

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	59.1	54.5		ng/L		92	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	59.1	54.3		ng/L		92	70 - 130
4,8-Dioxo-3H-perfluorononanoic acid (ADONA)	59.1	54.4		ng/L		92	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	59.1	54.2		ng/L		92	70 - 130
Perfluorobutanesulfonic acid (PFBS)	59.1	62.4		ng/L		106	70 - 130
Perfluorodecanoic acid (PFDA)	59.1	56.7		ng/L		96	70 - 130
Perfluorododecanoic acid (PFDoA)	59.1	57.0		ng/L		96	70 - 130
Perfluoroheptanoic acid (PFHpA)	59.1	56.1		ng/L		95	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	59.1	57.1		ng/L		97	70 - 130
Perfluorohexanoic acid (PFHxA)	59.1	57.4		ng/L		97	70 - 130
Perfluorononanoic acid (PFNA)	59.1	58.7		ng/L		99	70 - 130
Perfluorooctanesulfonic acid (PFOS)	59.1	56.2		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	59.1	59.5		ng/L		101	70 - 130
Perfluoroundecanoic acid (PFUnA)	59.1	56.9		ng/L		96	70 - 130
Perfluorobutanoic acid (PFBA)	59.1	53.8		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	59.1	59.0		ng/L		100	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	59.1	58.8		ng/L		100	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	59.1	66.5		ng/L		113	70 - 130
Nonfluoro-3,6-dioxahexanoic acid (NFDHA)	59.1	51.7		ng/L		87	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	59.1	73.8		ng/L		125	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	59.1	54.8		ng/L		93	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	59.1	56.8		ng/L		96	70 - 130
Perfluoropentanoic acid (PFPeA)	59.1	55.6		ng/L		94	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	59.1	56.6		ng/L		96	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-246683/22-A

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 246683

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	59.1	59.4		ng/L		101	70 - 130
		LCS	LCS				
		%Recovery	Qualifier	Limits			
13C3 HFPO-DA	84			50 - 200			
13C6 PFDA	99			50 - 200			
13C5 PFHxA	92			50 - 200			
13C4 PFHpA	92			50 - 200			
13C8 PFOA	90			50 - 200			
13C9 PFNA	94			50 - 200			
13C7 PFUnA	99			50 - 200			
13C2 PFDoA	105			50 - 200			
13C4 PFBA	97			50 - 200			
13C5 PFPeA	96			50 - 200			
13C3 PFBS	101			50 - 200			
13C3 PFHxS	115			50 - 200			
13C8 PFOS	114			50 - 200			
13C2-4:2-FTS	112			50 - 200			
13C2-6:2-FTS	104			50 - 200			
13C2-8:2-FTS	105			50 - 200			

Lab Sample ID: MRL 380-246683/21-A

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 246683

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.97	2.01	J	ng/L		102	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	1.97	1.93	J	ng/L		98	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.97	1.73	J	ng/L		88	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.97	2.20	J	ng/L		111	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.97	1.88	J	ng/L		95	50 - 150
Perfluorodecanoic acid (PFDA)	1.97	2.05	J	ng/L		104	50 - 150
Perfluorododecanoic acid (PFDoA)	1.97	2.02	J	ng/L		103	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.97	1.91	J	ng/L		97	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.97	1.90	J	ng/L		96	50 - 150
Perfluorohexanoic acid (PFHxA)	1.97	1.91	J	ng/L		97	50 - 150
Perfluorononanoic acid (PFNA)	1.97	2.03	J	ng/L		103	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.97	2.01	J	ng/L		102	50 - 150
Perfluorooctanoic acid (PFOA)	1.97	2.03	J	ng/L		103	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.97	1.93	J	ng/L		98	50 - 150
Perfluorobutanoic acid (PFBA)	1.97	1.94	J	ng/L		98	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-246683/21-A

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 246683

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	1.97	2.08	J	ng/L		106	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	1.97	2.07	J	ng/L		105	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	1.97	2.32	J	ng/L		117	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	1.97	1.86	J	ng/L		94	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	1.97	2.29	J	ng/L		116	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	1.97	1.94	J	ng/L		98	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	1.97	1.86	J	ng/L		94	50 - 150
Perfluoropentanoic acid (PFPeA)	1.97	2.02	J	ng/L		102	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	1.97	1.97	J	ng/L		100	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	1.97	1.93	J	ng/L		98	50 - 150

Isotope Dilution	%Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	82		50 - 200
13C6 PFDA	96		50 - 200
13C5 PFHxA	92		50 - 200
13C4 PFHpA	92		50 - 200
13C8 PFOA	94		50 - 200
13C9 PFNA	97		50 - 200
13C7 PFUnA	95		50 - 200
13C2 PFDoA	101		50 - 200
13C4 PFBA	98		50 - 200
13C5 PFPeA	100		50 - 200
13C3 PFBS	110		50 - 200
13C3 PFHxS	114		50 - 200
13C8 PFOS	108		50 - 200
13C2-4:2-FTS	116		50 - 200
13C2-6:2-FTS	106		50 - 200
13C2-8:2-FTS	109		50 - 200

Lab Sample ID: 380-230958-A-1-B MS

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 246683

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		119	117		ng/L		99	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		119	116		ng/L		98	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		119	114		ng/L		96	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-230958-A-1-B MS

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 246683

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		119	115		ng/L		97	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		119	114		ng/L		97	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		119	119		ng/L		100	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		119	113		ng/L		96	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		119	110		ng/L		93	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		119	120		ng/L		102	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		119	116		ng/L		98	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		119	123		ng/L		104	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		119	121		ng/L		102	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		119	122		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		119	117		ng/L		99	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		119	109		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		119	125		ng/L		106	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		119	123		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		119	125		ng/L		105	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		119	128		ng/L		108	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		119	121		ng/L		102	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		119	115		ng/L		97	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		119	117		ng/L		99	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		119	117		ng/L		99	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		119	122		ng/L		103	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		119	122		ng/L		103	70 - 130

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C3 HFPO-DA	104		50 - 200
13C6 PFDA	105		50 - 200
13C5 PFHxA	102		50 - 200
13C4 PFHpA	102		50 - 200
13C8 PFOA	98		50 - 200
13C9 PFNA	102		50 - 200
13C7 PFUnA	103		50 - 200
13C2 PFDoA	107		50 - 200
13C4 PFBA	107		50 - 200
13C5 PFPeA	106		50 - 200
13C3 PFBS	110		50 - 200
13C3 PFHxS	103		50 - 200
13C8 PFOS	102		50 - 200

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-230958-A-1-B MS

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 246683

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	105		50 - 200
13C2-6:2-FTS	105		50 - 200
13C2-8:2-FTS	100		50 - 200

Lab Sample ID: 380-230958-A-1-C MSD

Matrix: Water

Analysis Batch: 246855

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 246683

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		118	118		ng/L		100	70 - 130	0	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		118	115		ng/L		97	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		118	118		ng/L		99	70 - 130	3	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		118	124		ng/L		104	70 - 130	7	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		118	114		ng/L		96	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		118	117		ng/L		99	70 - 130	1	30
Perfluorododecanoic acid (PFDoA)	<2.0		118	118		ng/L		99	70 - 130	4	30
Perfluorooheptanoic acid (PFHpA)	<2.0		118	120		ng/L		101	70 - 130	9	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		118	123		ng/L		104	70 - 130	2	30
Perfluorohexanoic acid (PFHxA)	<2.0		118	117		ng/L		99	70 - 130	1	30
Perfluorononanoic acid (PFNA)	<2.0		118	122		ng/L		103	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		118	119		ng/L		101	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	<2.0		118	118		ng/L		100	70 - 130	3	30
Perfluoroundecanoic acid (PFUnA)	<2.0		118	117		ng/L		99	70 - 130	0	30
Perfluorobutanoic acid (PFBA)	<2.0		118	113		ng/L		96	70 - 130	4	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		118	113		ng/L		96	70 - 130	10	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		118	113		ng/L		96	70 - 130	9	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		118	130		ng/L		110	70 - 130	4	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		118	113		ng/L		96	70 - 130	12	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		118	120		ng/L		101	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		118	112		ng/L		95	70 - 130	2	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		118	118		ng/L		100	70 - 130	0	30
Perfluoropentanoic acid (PFPeA)	<2.0		118	115		ng/L		97	70 - 130	2	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		118	120		ng/L		102	70 - 130	2	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		118	128		ng/L		109	70 - 130	5	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	97		50 - 200
13C6 PFDA	106		50 - 200
13C5 PFHxA	103		50 - 200
13C4 PFHpA	99		50 - 200
13C8 PFOA	102		50 - 200
13C9 PFNA	101		50 - 200
13C7 PFUnA	103		50 - 200
13C2 PFDoA	107		50 - 200
13C4 PFBA	105		50 - 200
13C5 PFPeA	108		50 - 200
13C3 PFBS	109		50 - 200
13C3 PFHxS	102		50 - 200
13C8 PFOS	104		50 - 200
13C2-4:2-FTS	108		50 - 200
13C2-6:2-FTS	100		50 - 200
13C2-8:2-FTS	105		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-246690/19-A
Matrix: Water
Analysis Batch: 246857

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 246690

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.57		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.41		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorodecanoic acid (PFDA)	<0.30		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorotetradecanoic acid (PFTA)	<0.53		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Perfluorotridecanoic acid (PFTTrDA)	<0.35		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<0.29		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		08/16/26 16:30	08/17/26 13:17	1
Surrogate	MBL MBL		Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
d5-NEtFOSAA	101		70 - 130			08/16/26 16:30	08/17/26 13:17	1
13C2 PFHxA	103		70 - 130			08/16/26 16:30	08/17/26 13:17	1
13C2 PFDA	102		70 - 130			08/16/26 16:30	08/17/26 13:17	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-246690/19-A

Matrix: Water

Analysis Batch: 246857

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 246690

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	101		70 - 130	08/16/26 16:30	08/17/26 13:17	1

Lab Sample ID: LCS 380-246690/21-A

Matrix: Water

Analysis Batch: 246857

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 246690

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	49.1	45.4		ng/L		92	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	49.1	49.6		ng/L		101	70 - 130
Perfluoroundecanoic acid (PFUnA)	49.1	48.3		ng/L		98	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	49.1	48.2		ng/L		98	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	49.1	48.8		ng/L		99	70 - 130
Perfluorohexanoic acid (PFHxA)	49.1	46.0		ng/L		94	70 - 130
Perfluorododecanoic acid (PFDoA)	49.1	49.5		ng/L		101	70 - 130
Perfluorooctanoic acid (PFOA)	49.1	46.2		ng/L		94	70 - 130
Perfluorodecanoic acid (PFDA)	49.1	47.1		ng/L		96	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	49.1	50.6		ng/L		103	70 - 130
Perfluorobutanesulfonic acid (PFBS)	49.1	47.1		ng/L		96	70 - 130
Perfluoroheptanoic acid (PFHpA)	49.1	49.3		ng/L		100	70 - 130
Perfluorononanoic acid (PFNA)	49.1	48.0		ng/L		98	70 - 130
Perfluorotetradecanoic acid (PFTA)	49.1	44.4		ng/L		90	70 - 130
Perfluorotridecanoic acid (PFTrDA)	49.1	49.4		ng/L		101	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	49.1	48.4		ng/L		98	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	49.1	47.3		ng/L		96	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	49.1	46.7		ng/L		95	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	99		70 - 130
13C2 PFHxA	102		70 - 130
13C2 PFDA	106		70 - 130
13C3-GenX	103		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-246690/20-A

Matrix: Water

Analysis Batch: 246857

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 246690

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	1.96	1.77	J	ng/L		90	50 - 150
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	1.96	1.88	J	ng/L		95	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.96	2.02	J	ng/L		103	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	1.96	1.81	J	ng/L		92	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	1.96	1.90	J	ng/L		97	50 - 150
Perfluorohexanoic acid (PFHxA)	1.96	1.98	J	ng/L		101	50 - 150
Perfluorododecanoic acid (PFDoA)	1.96	2.09	J	ng/L		106	50 - 150
Perfluorooctanoic acid (PFOA)	1.96	1.94	J	ng/L		99	50 - 150
Perfluorodecanoic acid (PFDA)	1.96	2.03	J	ng/L		103	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.96	1.96	J	ng/L		100	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.96	1.98	J	ng/L		101	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.96	2.01	J	ng/L		102	50 - 150
Perfluorononanoic acid (PFNA)	1.96	2.07	J	ng/L		105	50 - 150
Perfluorotetradecanoic acid (PFTA)	1.96	1.82	J	ng/L		93	50 - 150
Perfluorotridecanoic acid (PFTrDA)	1.96	2.02	J	ng/L		103	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	1.96	1.94	J	ng/L		99	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.96	1.83	J	ng/L		93	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.96	1.88	J	ng/L		96	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	103		70 - 130
13C2 PFHxA	97		70 - 130
13C2 PFDA	102		70 - 130
13C3-GenX	95		70 - 130

Lab Sample ID: 380-230654-E-1-A MS

Matrix: Water

Analysis Batch: 246857

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 246690

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		24.6	23.4		ng/L		95	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorooctanesulfonic acid (PFOS)	<2.0		24.6	25.8		ng/L		105	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	25.6		ng/L		104	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	24.0		ng/L		98	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-230654-E-1-A MS

Matrix: Water

Analysis Batch: 246857

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 246690

	Sample	Sample	Spike	MS	MS				%Rec		
Analyte	Result	Qualifier	Added	Result	Qualifier	Unit	D	%Rec	Limits		
N-ethylperfluorooctanesulfonami doacetic acid (NEtFOSAA)	<2.0		24.6	25.0		ng/L		102	70 - 130		
Perfluorohexanoic acid (PFHxA)	<2.0		24.6	25.0		ng/L		101	70 - 130		
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	25.3		ng/L		103	70 - 130		
Perfluorooctanoic acid (PFOA)	<2.0		24.6	25.0		ng/L		102	70 - 130		
Perfluorodecanoic acid (PFDA)	<2.0		24.6	26.3		ng/L		107	70 - 130		
Perfluorohexanesulfonic acid (PFHxS)	<2.0		24.6	26.2		ng/L		105	70 - 130		
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	24.7		ng/L		100	70 - 130		
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	25.9		ng/L		105	70 - 130		
Perfluorononanoic acid (PFNA)	<2.0		24.6	25.7		ng/L		105	70 - 130		
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	23.9		ng/L		97	70 - 130		
Perfluorotridecanoic acid (PFTrDA)	<2.0		24.6	27.2		ng/L		111	70 - 130		
9-Chlorohexadecafluoro-3-oxan onane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		24.6	25.6		ng/L		104	70 - 130		
11-Chloroeicosafluoro-3-oxaund ecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	25.5		ng/L		104	70 - 130		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	24.9		ng/L		101	70 - 130		

Lab Sample ID: 380-230654-F-1-A MSD

Matrix: Water

Analysis Batch: 246857

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 246690

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		24.6	23.8		ng/L		97	70 - 130	2	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		24.6	25.6		ng/L		104	70 - 130	0	30
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	25.5		ng/L		104	70 - 130	0	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	25.4		ng/L		103	70 - 130	6	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		24.6	24.2		ng/L		98	70 - 130	3	30
Perfluorohexanoic acid (PFHxA)	<2.0		24.6	24.8		ng/L		101	70 - 130	1	30
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	26.2		ng/L		106	70 - 130	3	30
Perfluorooctanoic acid (PFOA)	<2.0		24.6	24.7		ng/L		101	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		24.6	25.6		ng/L		104	70 - 130	3	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-230654-F-1-A MSD

Matrix: Water

Analysis Batch: 246857

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 246690

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	<2.0		24.6	26.3		ng/L		105	70 - 130	0	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	25.5		ng/L		104	70 - 130	4	30
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	25.3		ng/L		103	70 - 130	2	30
Perfluorononanoic acid (PFNA)	<2.0		24.6	25.8		ng/L		105	70 - 130	0	30
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	24.8		ng/L		101	70 - 130	4	30
Perfluorotridecanoic acid (PFTTrDA)	<2.0		24.6	27.8		ng/L		113	70 - 130	2	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		24.6	24.9		ng/L		101	70 - 130	3	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	26.0		ng/L		106	70 - 130	2	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	24.3		ng/L		99	70 - 130	3	30
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
d5-NEtFOSAA	99		70 - 130								
13C2 PFHxA	102		70 - 130								
13C2 PFDA	106		70 - 130								
13C3-GenX	98		70 - 130								

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

LCMS

Prep Batch: 246683

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-230960-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	
380-230960-2	FB HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	
MBL 380-246683/20-A	Method Blank	Total/NA	Water	533	
LCS 380-246683/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-246683/21-A	Lab Control Sample	Total/NA	Water	533	
380-230958-A-1-B MS	Matrix Spike	Total/NA	Water	533	
380-230958-A-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Prep Batch: 246690

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-230960-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
380-230960-2	FB HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
MBL 380-246690/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-246690/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-246690/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-230654-E-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-230654-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 246855

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
MBL 380-246683/20-A	Method Blank	Total/NA	Water	533	246683
LCS 380-246683/22-A	Lab Control Sample	Total/NA	Water	533	246683
MRL 380-246683/21-A	Lab Control Sample	Total/NA	Water	533	246683
380-230958-A-1-B MS	Matrix Spike	Total/NA	Water	533	246683
380-230958-A-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	533	246683

Analysis Batch: 246857

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-230960-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	EPA 537.1 V2	246690
380-230960-2	FB HALAWA SHAFT VIEWING POOL	Total/NA	Water	EPA 537.1 V2	246690
MBL 380-246690/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	246690
LCS 380-246690/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	246690
MRL 380-246690/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	246690
380-230654-E-1-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	246690
380-230654-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	246690

Analysis Batch: 247168

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-230960-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	246683
380-230960-2	FB HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	246683

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-1

Date Collected: 08/11/26 09:30

Matrix: Water

Date Received: 08/13/26 09:46

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			246683	N8NE	EA POM	08/16/26 14:20
Total/NA	Analysis	533		1	247168	Y5FM	EA POM	08/18/26 15:48
Total/NA	Prep	537.1 DW			246690	LM3A	EA POM	08/16/26 16:30
Total/NA	Analysis	EPA 537.1 V2		1	246857	Y5FM	EA POM	08/17/26 15:51

Client Sample ID: FB HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-230960-2

Date Collected: 08/11/26 09:30

Matrix: Water

Date Received: 08/13/26 09:46

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			246683	N8NE	EA POM	08/16/26 14:20
Total/NA	Analysis	533		1	247168	Y5FM	EA POM	08/18/26 15:58
Total/NA	Prep	537.1 DW			246690	LM3A	EA POM	08/16/26 16:30
Total/NA	Analysis	EPA 537.1 V2		1	246857	Y5FM	EA POM	08/17/26 16:02

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-230960-1
SDG: PFAS: Halawa Shaft Viewing Pool

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-230960-1	HALAWA SHAFT VIEWING POOL	Water	08/11/26 09:30	08/13/26 09:46	Hawaii
380-230960-2	FB HALAWA SHAFT VIEWING POOL	Water	08/11/26 09:30	08/13/26 09:46	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Client Information Client Contact: Kirk Iwamoto Company: City & County of Honolulu Address: 6330 South Beretania Street; Chemistry Lab City: Honolulu State Zip: HI 96843 Phone: 808-748-5840 (tel) Email: kiwamoto@hbws.org Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill Site:		Sampler: Kai Edison Lab PM: Lopez, Maria Phone: +1 808 748 5840 E-Mail: Maria.Lopez@deleurofinus.com Carrier Tracking No(s): State of Origin:		Coc No: Page: Page 1 of 1 Job #:	
Analysis Requested					
Due Date Requested:		Total Number of Containers:			
TAT Requested (days): RUSH		Special Instructions/Note:			
Compliance Project: A No		Preservation Codes: A - HCL B - NaOH C - Zn Acetate D - Nitric Acid E - NaHSO4 F - MeOH G - Anchor H - Ascorbic Acid I - Ice J - DI Water K - EDTA L - EDA Other:			
PO #: C20525101 exp 05312023		537 1 DW_PREC - 537 1 Full List			
WO #:		533 - All Analytes			
Project #: 38001111		Perform MS/MSD (Yes or No)			
SSOW#:		Field Filtered Sample (Yes or No)			
Sample Date: 11-Aug-2026		Sample Time: 9:30		Matrix (W=water, S=solid, O=soil, ST=Sludge, A=air)	
Sample Identification: HALAWA SHAFT VIEWING POOL		Sample Type: G		Preservation Code:	
FB HALAWA SHAFT VIEWING POOL		Sample Date: 11-Aug-2026		Sample Time: 9:30	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:		Preservation Code:	
Sample Identification:		Sample Type:		Matrix:	
Sample Date:		Sample Time:			

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-230960-1

SDG Number: PFAS: Halawa Shaft Viewing Pool

Login Number: 230960

List Number: 1

Creator: Ngo, Theodore

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	