

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Erwin Kawata  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 310  
Honolulu, Hawaii 96843

Generated 6/16/2026 11:18:46 AM

## JOB DESCRIPTION

RED-HILL  
PFAS: Halawa Shaft Viewing Pool

## JOB NUMBER

380-219303-1

# Eurofins Pomona

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

## Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

## Authorization



Authorized for release by  
Maria Lopez, Project Manager  
[Maria.Lopez@et.eurofinsus.com](mailto:Maria.Lopez@et.eurofinsus.com)  
(626)386-1100

Generated  
6/16/2026 11:18:46 AM



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	11
Surrogate Summary . . . . .	12
Isotope Dilution Summary . . . . .	13
QC Sample Results . . . . .	14
QC Association Summary . . . . .	25
Lab Chronicle . . . . .	26
Certification Summary . . . . .	27
Method Summary . . . . .	28
Sample Summary . . . . .	29
Chain of Custody . . . . .	30
Receipt Checklists . . . . .	31

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Qualifiers

### LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: City & County of Honolulu  
Project: RED-HILL

Job ID: 380-219303-1

**Job ID: 380-219303-1**

**Eurofins Pomona**

## Job Narrative 380-219303-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

### Receipt

The samples were received on 6/11/2026 10:00 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 3.6°C.

### PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-219303-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	3.2		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.9		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.3		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.7		2.0	ng/L	1		EPA 537.1 V2	Total/NA

## Client Sample ID: FB: Halawa Shaft Viewing Pool

Lab Sample ID: 380-219303-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona



# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

**Client Sample ID: Halawa Shaft Viewing Pool**

**Lab Sample ID: 380-219303-1**

Date Collected: 06/09/26 09:35

Matrix: Water

Date Received: 06/11/26 10:00

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.2</b>		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>2.9</b>		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:01	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	95		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C6 PFDA	103		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C5 PFHxA	103		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C4 PFHpA	104		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C8 PFOA	104		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C9 PFNA	110		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C7 PFUnA	108		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C2 PFDoA	119		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C4 PFBA	102		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C5 PFPeA	97		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C3 PFBS	111		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C3 PFHxS	112		50 - 200	06/14/26 13:44	06/15/26 10:01	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

**Client Sample ID: Halawa Shaft Viewing Pool**

**Lab Sample ID: 380-219303-1**

Date Collected: 06/09/26 09:35

Matrix: Water

Date Received: 06/11/26 10:00

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	118		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C2-4:2-FTS	115		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C2-6:2-FTS	106		50 - 200	06/14/26 13:44	06/15/26 10:01	1
13C2-8:2-FTS	105		50 - 200	06/14/26 13:44	06/15/26 10:01	1

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>3.3</b>		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
N-methylperfluorooctanesulfonamide acetic acid (NMeFOSAA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.7</b>		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:33	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>			<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
d5-NEtFOSAA	103		70 - 130			06/12/26 06:41	06/13/26 02:33	1
13C2 PFHxA	108		70 - 130			06/12/26 06:41	06/13/26 02:33	1
13C2 PFDA	111		70 - 130			06/12/26 06:41	06/13/26 02:33	1
13C3-GenX	108		70 - 130			06/12/26 06:41	06/13/26 02:33	1

**Client Sample ID: FB: Halawa Shaft Viewing Pool**

**Lab Sample ID: 380-219303-2**

Date Collected: 06/09/26 09:35

Matrix: Water

Date Received: 06/11/26 10:00

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

**Client Sample ID: FB: Halawa Shaft Viewing Pool**

**Lab Sample ID: 380-219303-2**

Date Collected: 06/09/26 09:35

Matrix: Water

Date Received: 06/11/26 10:00

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		06/14/26 13:44	06/15/26 10:11	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	96		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C6 PFDA	110		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C5 PFHxA	105		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C4 PFHpA	105		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C8 PFOA	107		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C9 PFNA	113		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C7 PFUnA	112		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C2 PFDoA	115		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C4 PFBA	106		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C5 PFPeA	100		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C3 PFBS	105		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C3 PFHxS	105		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C8 PFOS	117		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C2-4:2-FTS	97		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C2-6:2-FTS	103		50 - 200	06/14/26 13:44	06/15/26 10:11	1
13C2-8:2-FTS	103		50 - 200	06/14/26 13:44	06/15/26 10:11	1

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

**Client Sample ID: FB: Halawa Shaft Viewing Pool**

**Lab Sample ID: 380-219303-2**

Date Collected: 06/09/26 09:35

Matrix: Water

Date Received: 06/11/26 10:00

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
N-methylperfluorooctanesulfonamide acetic acid (NMeFOSAA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/12/26 06:41	06/13/26 02:42	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	92		70 - 130	06/12/26 06:41	06/13/26 02:42	1
13C2 PFHxA	95		70 - 130	06/12/26 06:41	06/13/26 02:42	1
13C2 PFDA	101		70 - 130	06/12/26 06:41	06/13/26 02:42	1
13C3-GenX	78		70 - 130	06/12/26 06:41	06/13/26 02:42	1

# Action Limit Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

**Client Sample ID: Halawa Shaft Viewing Pool**

**Lab Sample ID: 380-219303-1**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL		RL	Method	Prep Type
				Limit				
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10		2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.2		ng/L	10		2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10		2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.9		ng/L	4		2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4		2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10		2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.3		ng/L	4		2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4		2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.7		ng/L	10		2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10		2.0	EPA 537.1 V2	Total/NA

**Client Sample ID: FB: Halawa Shaft Viewing Pool**

**Lab Sample ID: 380-219303-2**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL		RL	Method	Prep Type
				Limit				
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10		2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10		2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10		2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4		2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4		2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10		2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4		2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4		2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10		2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10		2.0	EPA 537.1 V2	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-219303-1  
 SDG: PFAS: Halawa Shaft Viewing Pool

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

**Matrix: Water**

**Prep Type: Total/NA**

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-219259-E-1-A MS	Matrix Spike	94	102	108	105
380-219259-F-1-A MSD	Matrix Spike Duplicate	96	111	108	108
380-219303-1	Halawa Shaft Viewing Pool	103	108	111	108
380-219303-2	FB: Halawa Shaft Viewing Pool	92	95	101	78
LCS 380-233412/23-A	Lab Control Sample	90	93	107	91
MBL 380-233412/21-A	Method Blank	110	123	118	114
MRL 380-233412/22-A	Lab Control Sample	95	106	112	86

**Surrogate Legend**

- d5NEFOS = d5-NEtFOSAA
- PFHxA = 13C2 PFHxA
- PFDA = 13C2 PFDA
- GenX = 13C3-GenX



# Isotope Dilution Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

**Matrix: Water**

**Prep Type: Total/NA**

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFD <sub>o</sub> A (50-200)
380-218705-B-1-A MS	Matrix Spike	85	92	91	91	95	102	100	103
380-218705-C-1-A MSD	Matrix Spike Duplicate	70	63	73	75	71	69	71	75
380-219303-1	Halawa Shaft Viewing Pool	95	103	103	104	104	110	108	119
380-219303-2	FB: Halawa Shaft Viewing Pool	96	110	105	105	107	113	112	115
LCS 380-233714/22-A	Lab Control Sample	101	109	102	102	102	115	112	122
MBL 380-233714/20-A	Method Blank	95	105	97	103	114	112	107	120
MRL 380-233714/21-A	Lab Control Sample	91	97	93	96	97	108	107	108

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-218705-B-1-A MS	Matrix Spike	97	91	104	105	114	102	101	103
380-218705-C-1-A MSD	Matrix Spike Duplicate	86	77	98	103	106	95	99	100
380-219303-1	Halawa Shaft Viewing Pool	102	97	111	112	118	115	106	105
380-219303-2	FB: Halawa Shaft Viewing Pool	106	100	105	105	117	97	103	103
LCS 380-233714/22-A	Lab Control Sample	101	92	112	113	119	102	104	109
MBL 380-233714/20-A	Method Blank	109	100	117	119	128	112	118	116
MRL 380-233714/21-A	Lab Control Sample	101	95	113	107	122	109	107	108

**Surrogate Legend**

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFD<sub>o</sub>A = 13C2 PFD<sub>o</sub>A
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

**Lab Sample ID: MBL 380-233714/20-A**  
**Matrix: Water**  
**Analysis Batch: 233869**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 233714**

Analyte	MBL	MBL	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.25		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		06/14/26 13:44	06/15/26 06:53	1

Isotope Dilution	MBL	MBL	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C3 HFPO-DA	95		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C6 PFDA	105		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C5 PFHxA	97		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C4 PFHpA	103		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C8 PFOA	114		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C9 PFNA	112		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C7 PFUnA	107		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C2 PFDoA	120		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C4 PFBA	109		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C5 PFPeA	100		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C3 PFBS	117		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C3 PFHxS	119		50 - 200	06/14/26 13:44	06/15/26 06:53	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MBL 380-233714/20-A**  
**Matrix: Water**  
**Analysis Batch: 233869**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 233714**

Isotope Dilution	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C8 PFOS	128		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C2-4:2-FTS	112		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C2-6:2-FTS	118		50 - 200	06/14/26 13:44	06/15/26 06:53	1
13C2-8:2-FTS	116		50 - 200	06/14/26 13:44	06/15/26 06:53	1

**Lab Sample ID: LCS 380-233714/22-A**  
**Matrix: Water**  
**Analysis Batch: 233869**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 233714**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	121	103		ng/L		85	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	121	110		ng/L		91	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	121	108		ng/L		89	70 - 130
Perfluorobutanesulfonic acid (PFBS)	121	111		ng/L		92	70 - 130
Perfluorodecanoic acid (PFDA)	121	117		ng/L		97	70 - 130
Perfluorododecanoic acid (PFDoA)	121	108		ng/L		89	70 - 130
Perfluoroheptanoic acid (PFHpA)	121	110		ng/L		91	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	121	112		ng/L		93	70 - 130
Perfluorohexanoic acid (PFHxA)	121	113		ng/L		94	70 - 130
Perfluorononanoic acid (PFNA)	121	106		ng/L		88	70 - 130
Perfluorooctanesulfonic acid (PFOS)	121	106		ng/L		88	70 - 130
Perfluorooctanoic acid (PFOA)	121	110		ng/L		91	70 - 130
Perfluoroundecanoic acid (PFUnA)	121	112		ng/L		93	70 - 130
Perfluorobutanoic acid (PFBA)	121	111		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	121	116		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	121	108		ng/L		89	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	121	112		ng/L		93	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	121	113		ng/L		93	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	121	113		ng/L		94	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	121	118		ng/L		98	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	121	117		ng/L		97	70 - 130
Perfluoropentanoic acid (PFPeA)	121	116		ng/L		96	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	121	108		ng/L		89	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: LCS 380-233714/22-A**  
**Matrix: Water**  
**Analysis Batch: 233869**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 233714**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	121	109		ng/L		90	70 - 130
<b>LCS LCS</b>							
<b>Isotope Dilution</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				
13C3 HFPO-DA	101		50 - 200				
13C6 PFDA	109		50 - 200				
13C5 PFHxA	102		50 - 200				
13C4 PFHpA	102		50 - 200				
13C8 PFOA	102		50 - 200				
13C9 PFNA	115		50 - 200				
13C7 PFUnA	112		50 - 200				
13C2 PFDoA	122		50 - 200				
13C4 PFBA	101		50 - 200				
13C5 PFPeA	92		50 - 200				
13C3 PFBS	112		50 - 200				
13C3 PFHxS	113		50 - 200				
13C8 PFOS	119		50 - 200				
13C2-4:2-FTS	102		50 - 200				
13C2-6:2-FTS	104		50 - 200				
13C2-8:2-FTS	109		50 - 200				

**Lab Sample ID: MRL 380-233714/21-A**  
**Matrix: Water**  
**Analysis Batch: 233869**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 233714**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.68	J	ng/L		84	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	1.75	J	ng/L		87	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.96	J	ng/L		98	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.88	J	ng/L		94	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.86	J	ng/L		93	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.08	J	ng/L		103	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.98	J	ng/L		99	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	1.96	J	ng/L		98	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	1.97	J	ng/L		98	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.01	J	ng/L		100	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.00	J	ng/L		100	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.90	J	ng/L		95	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.18	J	ng/L		109	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.88	J	ng/L		94	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	2.17	J	ng/L		108	50 - 150

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MRL 380-233714/21-A**

**Matrix: Water**

**Analysis Batch: 233869**

**Client Sample ID: Lab Control Sample**

**Prep Type: Total/NA**

**Prep Batch: 233714**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	1.85	J	ng/L		92	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	2.05	J	ng/L		102	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	2.03	J	ng/L		101	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	2.07	J	ng/L		103	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.01	1.92	J	ng/L		96	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	1.98	J	ng/L		99	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.90	J	ng/L		95	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	1.99	J	ng/L		99	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	1.82	J	ng/L		91	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	1.95	J	ng/L		97	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	91		50 - 200
13C6 PFDA	97		50 - 200
13C5 PFHxA	93		50 - 200
13C4 PFHpA	96		50 - 200
13C8 PFOA	97		50 - 200
13C9 PFNA	108		50 - 200
13C7 PFUnA	107		50 - 200
13C2 PFDoA	108		50 - 200
13C4 PFBA	101		50 - 200
13C5 PFPeA	95		50 - 200
13C3 PFBS	113		50 - 200
13C3 PFHxS	107		50 - 200
13C8 PFOS	122		50 - 200
13C2-4:2-FTS	109		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	108		50 - 200

**Lab Sample ID: 380-218705-B-1-A MS**

**Matrix: Water**

**Analysis Batch: 233869**

**Client Sample ID: Matrix Spike**

**Prep Type: Total/NA**

**Prep Batch: 233714**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.4	50.1		ng/L		83	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.4	51.3		ng/L		85	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.4	54.0		ng/L		89	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-218705-B-1-A MS**

**Client Sample ID: Matrix Spike**

**Matrix: Water**

**Prep Type: Total/NA**

**Analysis Batch: 233869**

**Prep Batch: 233714**

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Hexafluoropropylene Oxide	<2.0		60.4	56.3		ng/L		93	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		60.4	57.7		ng/L		96	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.4	60.7		ng/L		101	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.4	59.8		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		60.4	57.0		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		60.4	56.8		ng/L		94	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		60.4	59.9		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.4	55.4		ng/L		92	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.4	53.7		ng/L		89	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		60.4	56.9		ng/L		94	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.4	55.9		ng/L		93	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		60.4	57.0		ng/L		94	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.4	58.3		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.4	54.9		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.4	59.6		ng/L		99	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.4	58.1		ng/L		96	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.4	60.1		ng/L		99	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.4	59.2		ng/L		98	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.4	59.1		ng/L		98	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		60.4	56.4		ng/L		93	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.4	55.8		ng/L		92	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.4	57.2		ng/L		95	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	85		50 - 200
13C6 PFDA	92		50 - 200
13C5 PFHxA	91		50 - 200
13C4 PFHpA	91		50 - 200
13C8 PFOA	95		50 - 200
13C9 PFNA	102		50 - 200
13C7 PFUnA	100		50 - 200
13C2 PFDoA	103		50 - 200
13C4 PFBA	97		50 - 200
13C5 PFPeA	91		50 - 200
13C3 PFBS	104		50 - 200
13C3 PFHxS	105		50 - 200
13C8 PFOS	114		50 - 200

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-218705-B-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 233869**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 233714**

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	102		50 - 200
13C2-6:2-FTS	101		50 - 200
13C2-8:2-FTS	103		50 - 200

**Lab Sample ID: 380-218705-C-1-A MSD**  
**Matrix: Water**  
**Analysis Batch: 233869**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 233714**

<b>Analyte</b>	<b>Sample Result</b>	<b>Sample Qualifier</b>	<b>Spike Added</b>	<b>MSD Result</b>	<b>MSD Qualifier</b>	<b>Unit</b>	<b>D</b>	<b>%Rec</b>	<b>%Rec Limits</b>	<b>RPD</b>	<b>RPD Limit</b>
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	50.4		ng/L		84	70 - 130	1	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	50.9		ng/L		84	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	51.1		ng/L		85	70 - 130	5	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.2	55.9		ng/L		93	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		60.2	58.2		ng/L		97	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		60.2	61.6		ng/L		102	70 - 130	2	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	59.8		ng/L		99	70 - 130	0	30
Perfluoroheptanoic acid (PFHpA)	<2.0		60.2	55.0		ng/L		91	70 - 130	4	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		60.2	55.7		ng/L		92	70 - 130	2	30
Perfluorohexanoic acid (PFHxA)	<2.0		60.2	58.5		ng/L		97	70 - 130	2	30
Perfluorononanoic acid (PFNA)	<2.0		60.2	53.7		ng/L		89	70 - 130	3	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.2	55.3		ng/L		92	70 - 130	3	30
Perfluorooctanoic acid (PFOA)	<2.0		60.2	54.4		ng/L		90	70 - 130	5	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	56.3		ng/L		93	70 - 130	1	30
Perfluorobutanoic acid (PFBA)	<2.0		60.2	55.7		ng/L		92	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	56.7		ng/L		94	70 - 130	3	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	56.7		ng/L		94	70 - 130	3	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	55.0		ng/L		91	70 - 130	8	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.2	55.8		ng/L		93	70 - 130	4	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.2	54.0		ng/L		90	70 - 130	11	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	60.2		ng/L		100	70 - 130	2	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	55.3		ng/L		92	70 - 130	7	30
Perfluoropentanoic acid (PFPeA)	<2.0		60.2	55.1		ng/L		91	70 - 130	2	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	55.7		ng/L		92	70 - 130	0	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	53.8		ng/L		89	70 - 130	6	30

Eurofins Pomona

## QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

### Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	70		50 - 200
13C6 PFDA	63		50 - 200
13C5 PFHxA	73		50 - 200
13C4 PFHpA	75		50 - 200
13C8 PFOA	71		50 - 200
13C9 PFNA	69		50 - 200
13C7 PFUnA	71		50 - 200
13C2 PFDoA	75		50 - 200
13C4 PFBA	86		50 - 200
13C5 PFPeA	77		50 - 200
13C3 PFBS	98		50 - 200
13C3 PFHxS	103		50 - 200
13C8 PFOS	106		50 - 200
13C2-4:2-FTS	95		50 - 200
13C2-6:2-FTS	99		50 - 200
13C2-8:2-FTS	100		50 - 200

### Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-233412/21-A  
Matrix: Water  
Analysis Batch: 233618

Client Sample ID: Method Blank  
Prep Type: Total/NA  
Prep Batch: 233412

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
Perfluorotridecanoic acid (PFTTrDA)	<0.36		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		06/12/26 06:41	06/12/26 23:41	1

  

Surrogate	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
d5-NEtFOSAA	110		70 - 130	06/12/26 06:41	06/12/26 23:41	1
13C2 PFHxA	123		70 - 130	06/12/26 06:41	06/12/26 23:41	1
13C2 PFDA	118		70 - 130	06/12/26 06:41	06/12/26 23:41	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

**Lab Sample ID: MBL 380-233412/21-A**  
**Matrix: Water**  
**Analysis Batch: 233618**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 233412**

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	114		70 - 130	06/12/26 06:41	06/12/26 23:41	1

**Lab Sample ID: LCS 380-233412/23-A**  
**Matrix: Water**  
**Analysis Batch: 233618**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 233412**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	50.2	42.9		ng/L		86	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	50.2	51.1		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	50.2	55.7		ng/L		111	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.2	44.0		ng/L		88	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.2	46.0		ng/L		92	70 - 130
Perfluorohexanoic acid (PFHxA)	50.2	49.7		ng/L		99	70 - 130
Perfluorododecanoic acid (PFDoA)	50.2	53.5		ng/L		107	70 - 130
Perfluorooctanoic acid (PFOA)	50.2	51.1		ng/L		102	70 - 130
Perfluorodecanoic acid (PFDA)	50.2	52.6		ng/L		105	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	50.2	51.4		ng/L		102	70 - 130
Perfluorobutanesulfonic acid (PFBS)	50.2	52.7		ng/L		105	70 - 130
Perfluoroheptanoic acid (PFHpA)	50.2	49.6		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	50.2	50.4		ng/L		100	70 - 130
Perfluorotetradecanoic acid (PFTA)	50.2	49.3		ng/L		98	70 - 130
Perfluorotridecanoic acid (PFTrDA)	50.2	53.4		ng/L		106	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	50.2	51.9		ng/L		103	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.2	53.6		ng/L		107	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.2	45.7		ng/L		91	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	90		70 - 130
13C2 PFHxA	93		70 - 130
13C2 PFDA	107		70 - 130
13C3-GenX	91		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

**Lab Sample ID: MRL 380-233412/22-A**  
**Matrix: Water**  
**Analysis Batch: 233618**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 233412**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.85	J	ng/L		92	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	2.20	J	ng/L		110	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	2.44	J	ng/L		121	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.01	1.85	J	ng/L		92	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.01	2.01	J	ng/L		100	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.09	J	ng/L		104	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	2.31	J	ng/L		115	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.18	J	ng/L		109	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.23	J	ng/L		111	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.26	J	ng/L		112	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	2.17	J	ng/L		108	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.13	J	ng/L		106	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.20	J	ng/L		109	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.01	2.12	J	ng/L		106	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.01	2.31	J	ng/L		115	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	2.22	J	ng/L		110	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	2.29	J	ng/L		114	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.77	J	ng/L		88	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	95		70 - 130
13C2 PFHxA	106		70 - 130
13C2 PFDA	112		70 - 130
13C3-GenX	86		70 - 130

**Lab Sample ID: 380-219259-E-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 233618**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 233412**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.2	52.8		ng/L		105	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		50.2	55.0		ng/L		106	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		50.2	54.3		ng/L		108	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.2	46.3		ng/L		92	70 - 130

Eurofins Pomona





# QC Association Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-219303-1  
 SDG: PFAS: Halawa Shaft Viewing Pool

## LCMS

### Prep Batch: 233412

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-219303-1	Halawa Shaft Viewing Pool	Total/NA	Water	537.1 DW	
380-219303-2	FB: Halawa Shaft Viewing Pool	Total/NA	Water	537.1 DW	
MBL 380-233412/21-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-233412/23-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-233412/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-219259-E-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-219259-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

### Analysis Batch: 233618

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-219303-1	Halawa Shaft Viewing Pool	Total/NA	Water	EPA 537.1 V2	233412
380-219303-2	FB: Halawa Shaft Viewing Pool	Total/NA	Water	EPA 537.1 V2	233412
MBL 380-233412/21-A	Method Blank	Total/NA	Water	EPA 537.1 V2	233412
LCS 380-233412/23-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	233412
MRL 380-233412/22-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	233412
380-219259-E-1-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	233412
380-219259-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	233412

### Prep Batch: 233714

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-219303-1	Halawa Shaft Viewing Pool	Total/NA	Water	533	
380-219303-2	FB: Halawa Shaft Viewing Pool	Total/NA	Water	533	
MBL 380-233714/20-A	Method Blank	Total/NA	Water	533	
LCS 380-233714/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-233714/21-A	Lab Control Sample	Total/NA	Water	533	
380-218705-B-1-A MS	Matrix Spike	Total/NA	Water	533	
380-218705-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

### Analysis Batch: 233869

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-219303-1	Halawa Shaft Viewing Pool	Total/NA	Water	533	233714
380-219303-2	FB: Halawa Shaft Viewing Pool	Total/NA	Water	533	233714
MBL 380-233714/20-A	Method Blank	Total/NA	Water	533	233714
LCS 380-233714/22-A	Lab Control Sample	Total/NA	Water	533	233714
MRL 380-233714/21-A	Lab Control Sample	Total/NA	Water	533	233714
380-218705-B-1-A MS	Matrix Spike	Total/NA	Water	533	233714
380-218705-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	233714

# Lab Chronicle

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Client Sample ID: Halawa Shaft Viewing Pool

## Lab Sample ID: 380-219303-1

Date Collected: 06/09/26 09:35

Matrix: Water

Date Received: 06/11/26 10:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			233714	N8NE	EA POM	06/14/26 13:44
Total/NA	Analysis	533		1	233869	SZ9R	EA POM	06/15/26 10:01
Total/NA	Prep	537.1 DW			233412	E9PK	EA POM	06/12/26 06:41
Total/NA	Analysis	EPA 537.1 V2		1	233618	M7ML	EA POM	06/13/26 02:33

## Client Sample ID: FB: Halawa Shaft Viewing Pool

## Lab Sample ID: 380-219303-2

Date Collected: 06/09/26 09:35

Matrix: Water

Date Received: 06/11/26 10:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			233714	N8NE	EA POM	06/14/26 13:44
Total/NA	Analysis	533		1	233869	SZ9R	EA POM	06/15/26 10:11
Total/NA	Prep	537.1 DW			233412	E9PK	EA POM	06/12/26 06:41
Total/NA	Analysis	EPA 537.1 V2		1	233618	M7ML	EA POM	06/13/26 02:42

**Laboratory References:**

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

## Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

\* Accreditation/Certification renewal pending - accreditation/certification considered valid.

# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

**Protocol References:**

EPA = US Environmental Protection Agency

**Laboratory References:**

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-219303-1  
SDG: PFAS: Halawa Shaft Viewing Pool

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-219303-1	Halawa Shaft Viewing Pool	Water	06/09/26 09:35	06/11/26 10:00	Hawaii
380-219303-2	FB: Halawa Shaft Viewing Pool	Water	06/09/26 09:35	06/11/26 10:00	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

**Monrovia, CA (Suite 100)**  
 750 Royal Oaks Drive Suite 100  
 Monrovia, CA 91016  
 Phone (626) 386-1100

# Chain of Custody Record



Environment  
America



COC No: 380-27941-2757.2  
 Page: Page 2 of 2  
 Job #: 380-219303 COC

Carrier Tracking No(s):  
 State of Origin:  
 Lab PM: Arada, Rachelle  
 E-Mail: Rachelle.Arada@et.euronisus.com

Sampler: Jason Rakofsky  
 Phone: +1 808 748 5840  
 PWSID:

Due Date Requested:  
 TAT Requested (days): RUSH  
 Compliance Project: Δ No  
 PO #: C20525101 exp 05312023  
 WO #:  
 Project #: 38001111  
 SSOW#:

Company: City & County of Honolulu  
 Address: 630 South Beretania Street, Chemistry Lab  
 City: Honolulu  
 State, Zip: HI, 96843  
 Phone: 808-748-5840 (tel)  
 Email: kiwamoto@hbws.org  
 Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill  
 Site:

Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (Water, Groundwater, Other)	Preservation Code	Field Filtered Sample (Yes or No)		Perform MS/MSD (Yes or No)		SUBCONTRACT - 626 PAH Phys (EAL) + TICs		80166_GRO_LL (MOD) GRO		80166_DRO_LL_C9 - HNL Ranges C10-C24/C24-C38/C8-C18		526.2_PREC - (MOD) 526plus PLUS TICs		637_1_DW_PREC - 637 1 Full List		633 - All Analytes		Special Instructions/Note:		
						Field Filtered	MS/MSD	Subcontract	80166_GRO_LL	80166_DRO_LL	526.2_PREC	637_1_DW_PREC	633	Field Filtered	MS/MSD	Subcontract	80166_GRO_LL	80166_DRO_LL	526.2_PREC	637_1_DW_PREC	633			
Halawa Shaft Viewing Pool	9-Jun-2026	0935	G	Water																				
Halawa Shaft Viewing Pool Blank	9-Jun-2026		G	Water																				

**Possible Hazard Identification**  
 Non-Hazard  Flammable  Skin Irritant  Poison B  Unknown  Radiological  
 Deliverable Requested I, II, III, IV, Other (specify)

**Empty Kit Relinquished by**  
 Date/Time: 6/9/26 11:00  
 Date/Time: 6/9/26 1000  
 Date/Time:

**Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)**  
 Return To Client  Disposal By Lab  Archive For \_\_\_\_\_ Months  
 Special Instructions/QC Requirements:

Received by	Date/Time	Company
W. W. W.	6/4/26	Company
W. W. W.	6/4/26	Company
W. W. W.	6/4/26	Company

Method of Shipment: FedEx 8729 2130 3160  
 Cooler Temperature(s) °C and Other Remarks: (3) 3.6 + 0.3.6 9/6 (-f. 10.10)



## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-219303-1  
SDG Number: PFAS: Halawa Shaft Viewing Pool

**Login Number: 219303**

**List Number: 1**

**Creator: Avila, Ivan**

**List Source: Eurofins Pomona**

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

