

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 7/31/2026 5:27:04 PM

JOB DESCRIPTION

RED-HILL
PFAS: Halawa Wells P1

JOB NUMBER

380-227787-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1145

Generated
7/31/2026 5:27:04 PM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-227787-1

Job ID: 380-227787-1

Eurofins Pomona

Job Narrative 380-227787-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 7/29/2026 10:40 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.2°C.

PFAS

EPA 537.1 and EPA 533 are two distinct methods for the analysis of PFAS in drinking water. The analyses are conducted on differing instrumentation, with calibrations, extraction solvents and sample preservatives being dissimilar among the two methods. Therefore it is probable and not unexpected to see the methods having slight variations in analytical results: Halawa Wells P1 (331-023-WL065) (380-227787-1). (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	2.4		2.0	ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	3.0		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.3		2.0	ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	4.1		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.6		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanoic acid (PFHxA)	3.3		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	2.1		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.9		2.0	ng/L	1		EPA 537.1 V2	Total/NA

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-1

Date Collected: 07/27/26 10:48

Matrix: Drinking Water

Date Received: 07/29/26 10:40

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9CI-PF3ONS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorohexanesulfonic acid (PFHxS)	2.4		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorohexanoic acid (PFHxA)	3.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorooctanesulfonic acid (PFOS)	2.3		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoropentanoic acid (PFPeA)	4.1		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:03	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	96		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C6 PFDA	104		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C5 PFHxA	100		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C4 PFHpA	103		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C8 PFOA	106		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C9 PFNA	110		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C7 PFUnA	115		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C2 PFDoA	114		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C4 PFBA	108		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C5 PFPeA	95		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C3 PFBS	104		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C3 PFHxS	115		50 - 200	07/30/26 07:53	07/30/26 20:03	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-1

Date Collected: 07/27/26 10:48

Matrix: Drinking Water

Date Received: 07/29/26 10:40

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	113		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C2-4:2-FTS	118		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C2-6:2-FTS	113		50 - 200	07/30/26 07:53	07/30/26 20:03	1
13C2-8:2-FTS	112		50 - 200	07/30/26 07:53	07/30/26 20:03	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorooctanesulfonic acid (PFOS)	2.6		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorohexanoic acid (PFHxA)	3.3		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorooctanoic acid (PFOA)	2.1		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorohexanesulfonic acid (PFHxS)	2.9		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 07:11	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	106		70 - 130	07/30/26 09:35	07/31/26 07:11	1
13C2 PFHxA	112		70 - 130	07/30/26 09:35	07/31/26 07:11	1
13C2 PFDA	113		70 - 130	07/30/26 09:35	07/31/26 07:11	1
13C3-GenX	103		70 - 130	07/30/26 09:35	07/31/26 07:11	1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-2

Date Collected: 07/27/26 10:48

Matrix: Water

Date Received: 07/29/26 10:40

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-2

Date Collected: 07/27/26 10:48

Matrix: Water

Date Received: 07/29/26 10:40

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		07/30/26 07:53	07/30/26 20:12	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	95		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C6 PFDA	100		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C5 PFHxA	100		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C4 PFHpA	105		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C8 PFOA	107		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C9 PFNA	110		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C7 PFUnA	116		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C2 PFDoA	114		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C4 PFBA	106		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C5 PFPeA	101		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C3 PFBS	104		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C3 PFHxS	105		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C8 PFOS	114		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C2-4:2-FTS	110		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C2-6:2-FTS	103		50 - 200	07/30/26 07:53	07/30/26 20:12	1
13C2-8:2-FTS	115		50 - 200	07/30/26 07:53	07/30/26 20:12	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-2

Date Collected: 07/27/26 10:48

Matrix: Water

Date Received: 07/29/26 10:40

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		07/30/26 09:35	07/31/26 08:55	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	93		70 - 130	07/30/26 09:35	07/31/26 08:55	1
13C2 PFHxA	106		70 - 130	07/30/26 09:35	07/31/26 08:55	1
13C2 PFDA	107		70 - 130	07/30/26 09:35	07/31/26 08:55	1
13C3-GenX	93		70 - 130	07/30/26 09:35	07/31/26 08:55	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.4		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.3		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.6		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	2.1		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.9		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-227787-1	HALAWA WELLS P1 (331-023-V	106	112	113	103
380-227787-1 MS	HALAWA WELLS P1 (331-023-WL065)	109	112	110	108
380-227787-1 MSD	HALAWA WELLS P1 (331-023-WL065)	104	108	116	102
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-227787-2	FB: HALAWA WELLS P1 (331-0	93	106	107	93
LCS 380-243415/21-A	Lab Control Sample	98	103	113	93
MBL 380-243415/19-A	Method Blank	95	98	107	85
MRL 380-243415/20-A	Lab Control Sample	101	109	116	94
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-227787-1	HALAWA WELLS P1 (331-023-V)	96	104	100	103	106	110	115	114

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-227787-1	HALAWA WELLS P1 (331-023-V)	108	95	104	115	113	118	113	112

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDaA = 13C2 PFDaA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-227787-2	FB: HALAWA WELLS P1 (331-0)	95	100	100	105	107	110	116	114
380-227793-E-1-A MS	Matrix Spike	109	100	100	107	109	110	119	119
380-227793-F-1-A MSD	Matrix Spike Duplicate	100	107	100	109	109	112	120	119
LCS 380-243406/22-A	Lab Control Sample	96	104	106	102	107	111	122	113
MBL 380-243406/20-A	Method Blank	95	98	104	105	107	110	111	108
MRL 380-243406/21-A	Lab Control Sample	88	94	98	98	100	100	112	107

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-227787-2	FB: HALAWA WELLS P1 (331-0)	106	101	104	105	114	110	103	115
380-227793-E-1-A MS	Matrix Spike	108	102	106	107	112	112	110	113
380-227793-F-1-A MSD	Matrix Spike Duplicate	110	106	107	110	111	105	112	111
LCS 380-243406/22-A	Lab Control Sample	99	97	107	108	113	104	106	107
MBL 380-243406/20-A	Method Blank	95	88	108	111	116	113	105	104
MRL 380-243406/21-A	Lab Control Sample	94	88	104	103	110	112	105	112

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA

Eurofins Pomona

Isotope Dilution Summary

Client: City & County of Honolulu

Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDoA = 13C2 PFDoA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-243406/20-A

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 243406

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluorobutanoic acid (PFBA)	<0.68		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.36		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.47		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.46		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.24		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.45		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoropentanoic acid (PFPeA)	<0.37		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.35		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1
Perfluoropentanesulfonic acid (PFPeS)	<0.38		2.0	ng/L		07/30/26 07:53	07/30/26 17:20	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	95		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C6 PFDA	98		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C5 PFHxA	104		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C4 PFHpA	105		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C8 PFOA	107		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C9 PFNA	110		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C7 PFUnA	111		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C2 PFDoA	108		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C4 PFBA	95		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C5 PFPeA	88		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C3 PFBS	108		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C3 PFHxS	111		50 - 200	07/30/26 07:53	07/30/26 17:20	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-243406/20-A

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 243406

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	116		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C2-4:2-FTS	113		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C2-6:2-FTS	105		50 - 200	07/30/26 07:53	07/30/26 17:20	1
13C2-8:2-FTS	104		50 - 200	07/30/26 07:53	07/30/26 17:20	1

Lab Sample ID: LCS 380-243406/22-A

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 243406

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	59.1	52.2		ng/L		88	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	59.1	51.0		ng/L		86	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	59.1	58.3		ng/L		99	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	59.1	56.8		ng/L		96	70 - 130
Perfluorobutanesulfonic acid (PFBS)	59.1	55.5		ng/L		94	70 - 130
Perfluorodecanoic acid (PFDA)	59.1	60.9		ng/L		103	70 - 130
Perfluorododecanoic acid (PFDaA)	59.1	58.3		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	59.1	60.3		ng/L		102	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	59.1	55.5		ng/L		94	70 - 130
Perfluorohexanoic acid (PFHxA)	59.1	55.6		ng/L		94	70 - 130
Perfluorononanoic acid (PFNA)	59.1	49.5		ng/L		84	70 - 130
Perfluorooctanesulfonic acid (PFOS)	59.1	52.7		ng/L		89	70 - 130
Perfluorooctanoic acid (PFOA)	59.1	54.8		ng/L		93	70 - 130
Perfluoroundecanoic acid (PFUnA)	59.1	54.0		ng/L		91	70 - 130
Perfluorobutanoic acid (PFBA)	59.1	58.3		ng/L		99	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	59.1	56.8		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	59.1	57.6		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	59.1	58.3		ng/L		99	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	59.1	45.7		ng/L		77	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	59.1	52.0		ng/L		88	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	59.1	55.0		ng/L		93	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	59.1	53.6		ng/L		91	70 - 130
Perfluoropentanoic acid (PFPeA)	59.1	57.6		ng/L		98	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	59.1	51.8		ng/L		88	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-243406/22-A

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 243406

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	59.1	55.5		ng/L		94	70 - 130
Isotope Dilution							
	LCS	LCS					
	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	96		50 - 200				
13C6 PFDA	104		50 - 200				
13C5 PFHxA	106		50 - 200				
13C4 PFHpA	102		50 - 200				
13C8 PFOA	107		50 - 200				
13C9 PFNA	111		50 - 200				
13C7 PFUnA	122		50 - 200				
13C2 PFDoA	113		50 - 200				
13C4 PFBA	99		50 - 200				
13C5 PFPeA	97		50 - 200				
13C3 PFBS	107		50 - 200				
13C3 PFHxS	108		50 - 200				
13C8 PFOS	113		50 - 200				
13C2-4:2-FTS	104		50 - 200				
13C2-6:2-FTS	106		50 - 200				
13C2-8:2-FTS	107		50 - 200				

Lab Sample ID: MRL 380-243406/21-A

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 243406

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.97	1.95	J	ng/L		99	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	1.97	1.85	J	ng/L		94	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.97	1.97	J	ng/L		100	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.97	1.95	J	ng/L		99	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.97	1.99	J	ng/L		101	50 - 150
Perfluorodecanoic acid (PFDA)	1.97	2.25	J	ng/L		114	50 - 150
Perfluorododecanoic acid (PFDoA)	1.97	2.12	J	ng/L		107	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.97	2.05	J	ng/L		104	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.97	2.04	J	ng/L		104	50 - 150
Perfluorohexanoic acid (PFHxA)	1.97	2.37	J	ng/L		120	50 - 150
Perfluorononanoic acid (PFNA)	1.97	1.80	J	ng/L		92	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.97	1.91	J	ng/L		97	50 - 150
Perfluorooctanoic acid (PFOA)	1.97	2.07	J	ng/L		105	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.97	2.11	J	ng/L		107	50 - 150
Perfluorobutanoic acid (PFBA)	1.97	2.27	J	ng/L		115	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-243406/21-A

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 243406

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	1.97	1.97	J	ng/L		100	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	1.97	2.04	J	ng/L		104	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	1.97	2.22	J	ng/L		113	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	1.97	1.59	J	ng/L		81	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	1.97	1.96	J	ng/L		99	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	1.97	1.97	J	ng/L		100	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	1.97	1.83	J	ng/L		93	50 - 150
Perfluoropentanoic acid (PFPeA)	1.97	2.21	J	ng/L		112	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	1.97	1.85	J	ng/L		94	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	1.97	1.91	J	ng/L		97	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	88		50 - 200
13C6 PFDA	94		50 - 200
13C5 PFHxA	98		50 - 200
13C4 PFHpA	98		50 - 200
13C8 PFOA	100		50 - 200
13C9 PFNA	100		50 - 200
13C7 PFUnA	112		50 - 200
13C2 PFDoA	107		50 - 200
13C4 PFBA	94		50 - 200
13C5 PFPeA	88		50 - 200
13C3 PFBS	104		50 - 200
13C3 PFHxS	103		50 - 200
13C8 PFOS	110		50 - 200
13C2-4:2-FTS	112		50 - 200
13C2-6:2-FTS	105		50 - 200
13C2-8:2-FTS	112		50 - 200

Lab Sample ID: 380-227793-E-1-A MS

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 243406

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		59.3	52.0		ng/L		88	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		59.3	52.8		ng/L		89	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		59.3	56.8		ng/L		96	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-227793-E-1-A MS

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 243406

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		59.3	57.2		ng/L		97	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	3.4		59.3	62.2		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		59.3	61.9		ng/L		104	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		59.3	55.6		ng/L		94	70 - 130
Perfluoroheptanoic acid (PFHpA)	2.3		59.3	61.0		ng/L		99	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	3.9		59.3	58.6		ng/L		92	70 - 130
Perfluorohexanoic acid (PFHxA)	4.5		59.3	62.9		ng/L		98	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		59.3	50.0		ng/L		84	70 - 130
Perfluorooctanesulfonic acid (PFOS)	5.0		59.3	57.3		ng/L		88	70 - 130
Perfluorooctanoic acid (PFOA)	4.5		59.3	61.9		ng/L		97	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		59.3	60.0		ng/L		101	70 - 130
Perfluorobutanoic acid (PFBA)	2.0		59.3	57.4		ng/L		94	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		59.3	57.5		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		59.3	60.1		ng/L		101	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		59.3	59.6		ng/L		101	70 - 130
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		59.3	47.3		ng/L		80	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		59.3	56.4		ng/L		95	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		59.3	59.1		ng/L		100	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		59.3	58.5		ng/L		99	70 - 130
Perfluoropentanoic acid (PFPeA)	4.6		59.3	61.7		ng/L		96	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		59.3	53.5		ng/L		90	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		59.3	57.0		ng/L		95	70 - 130

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C3 HFPO-DA	109		50 - 200
13C6 PFDA	100		50 - 200
13C5 PFHxA	100		50 - 200
13C4 PFHpA	107		50 - 200
13C8 PFOA	109		50 - 200
13C9 PFNA	110		50 - 200
13C7 PFUnA	119		50 - 200
13C2 PFDoA	119		50 - 200
13C4 PFBA	108		50 - 200
13C5 PFPeA	102		50 - 200
13C3 PFBS	106		50 - 200
13C3 PFHxS	107		50 - 200
13C8 PFOS	112		50 - 200

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-227793-E-1-A MS

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 243406

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	112		50 - 200
13C2-6:2-FTS	110		50 - 200
13C2-8:2-FTS	113		50 - 200

Lab Sample ID: 380-227793-F-1-A MSD

Matrix: Water

Analysis Batch: 243558

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 243406

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		59.2	51.4		ng/L		87	70 - 130	1	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		59.2	51.1		ng/L		86	70 - 130	3	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		59.2	55.9		ng/L		95	70 - 130	2	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		59.2	58.7		ng/L		99	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	3.4		59.2	59.8		ng/L		95	70 - 130	4	30
Perfluorodecanoic acid (PFDA)	<2.0		59.2	59.8		ng/L		101	70 - 130	3	30
Perfluorododecanoic acid (PFDaA)	<2.0		59.2	55.9		ng/L		95	70 - 130	1	30
Perfluoroheptanoic acid (PFHpA)	2.3		59.2	57.6		ng/L		94	70 - 130	6	30
Perfluorohexanesulfonic acid (PFHxS)	3.9		59.2	57.2		ng/L		90	70 - 130	2	30
Perfluorohexanoic acid (PFHxA)	4.5		59.2	61.2		ng/L		96	70 - 130	3	30
Perfluorononanoic acid (PFNA)	<2.0		59.2	49.3		ng/L		83	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	5.0		59.2	56.9		ng/L		88	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	4.5		59.2	59.7		ng/L		93	70 - 130	4	30
Perfluoroundecanoic acid (PFUnA)	<2.0		59.2	55.7		ng/L		94	70 - 130	7	30
Perfluorobutanoic acid (PFBA)	2.0		59.2	56.4		ng/L		92	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		59.2	57.6		ng/L		97	70 - 130	0	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		59.2	55.9		ng/L		95	70 - 130	7	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		59.2	55.8		ng/L		94	70 - 130	7	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		59.2	51.6		ng/L		87	70 - 130	9	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		59.2	59.0		ng/L		100	70 - 130	5	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		59.2	59.4		ng/L		100	70 - 130	0	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		59.2	56.9		ng/L		96	70 - 130	3	30
Perfluoropentanoic acid (PFPeA)	4.6		59.2	59.9		ng/L		93	70 - 130	3	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		59.2	54.1		ng/L		91	70 - 130	1	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		59.2	55.5		ng/L		93	70 - 130	3	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD %Recovery	MSD Qualifier	Limits
13C3 HFPO-DA	100		50 - 200
13C6 PFDA	107		50 - 200
13C5 PFHxA	100		50 - 200
13C4 PFHpA	109		50 - 200
13C8 PFOA	109		50 - 200
13C9 PFNA	112		50 - 200
13C7 PFUnA	120		50 - 200
13C2 PFDoA	119		50 - 200
13C4 PFBA	110		50 - 200
13C5 PFPeA	106		50 - 200
13C3 PFBS	107		50 - 200
13C3 PFHxS	110		50 - 200
13C8 PFOS	111		50 - 200
13C2-4:2-FTS	105		50 - 200
13C2-6:2-FTS	112		50 - 200
13C2-8:2-FTS	111		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-243415/19-A
Matrix: Water
Analysis Batch: 243635

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 243415

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.57		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.41		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorotetradecanoic acid (PFTA)	<0.53		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Perfluorotridecanoic acid (PFTrDA)	<0.35		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		07/30/26 09:35	07/31/26 06:41	1
Surrogate	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	95		70 - 130			07/30/26 09:35	07/31/26 06:41	1
13C2 PFHxA	98		70 - 130			07/30/26 09:35	07/31/26 06:41	1
13C2 PFDA	107		70 - 130			07/30/26 09:35	07/31/26 06:41	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-243415/19-A

Matrix: Water

Analysis Batch: 243635

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 243415

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	85		70 - 130	07/30/26 09:35	07/31/26 06:41	1

Lab Sample ID: LCS 380-243415/21-A

Matrix: Water

Analysis Batch: 243635

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 243415

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	49.1	41.9		ng/L		85	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	49.1	49.9		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	49.1	53.3		ng/L		108	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	49.1	47.1		ng/L		96	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	49.1	45.0		ng/L		92	70 - 130
Perfluorohexanoic acid (PFHxA)	49.1	49.9		ng/L		102	70 - 130
Perfluorododecanoic acid (PFDoA)	49.1	55.0		ng/L		112	70 - 130
Perfluorooctanoic acid (PFOA)	49.1	48.8		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	49.1	50.9		ng/L		104	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	49.1	53.0		ng/L		108	70 - 130
Perfluorobutanesulfonic acid (PFBS)	49.1	48.1		ng/L		98	70 - 130
Perfluoroheptanoic acid (PFHpA)	49.1	49.6		ng/L		101	70 - 130
Perfluorononanoic acid (PFNA)	49.1	52.0		ng/L		106	70 - 130
Perfluorotetradecanoic acid (PFTA)	49.1	55.2		ng/L		112	70 - 130
Perfluorotridecanoic acid (PFTrDA)	49.1	58.1		ng/L		118	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	49.1	50.2		ng/L		102	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	49.1	45.6		ng/L		93	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	49.1	44.4		ng/L		90	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	98		70 - 130
13C2 PFHxA	103		70 - 130
13C2 PFDA	113		70 - 130
13C3-GenX	93		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-243415/20-A

Matrix: Water

Analysis Batch: 243635

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 243415

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.97	1.61	J	ng/L		82	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.97	1.90	J	ng/L		97	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.97	2.24	J	ng/L		114	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	1.97	1.91	J	ng/L		97	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	1.97	1.87	J	ng/L		95	50 - 150
Perfluorohexanoic acid (PFHxA)	1.97	1.85	J	ng/L		94	50 - 150
Perfluorododecanoic acid (PFDoA)	1.97	2.35	J	ng/L		119	50 - 150
Perfluorooctanoic acid (PFOA)	1.97	1.90	J	ng/L		97	50 - 150
Perfluorodecanoic acid (PFDA)	1.97	1.99	J	ng/L		101	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.97	1.95	J	ng/L		99	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.97	1.81	J	ng/L		92	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.97	1.83	J	ng/L		93	50 - 150
Perfluorononanoic acid (PFNA)	1.97	2.02	J	ng/L		103	50 - 150
Perfluorotetradecanoic acid (PFTA)	1.97	2.22	J	ng/L		113	50 - 150
Perfluorotridecanoic acid (PFTrDA)	1.97	2.37	J	ng/L		120	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	1.97	1.92	J	ng/L		97	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.97	1.87	J	ng/L		95	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.97	1.66	J	ng/L		84	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	101		70 - 130
13C2 PFHxA	109		70 - 130
13C2 PFDA	116		70 - 130
13C3-GenX	94		70 - 130

Lab Sample ID: 380-227787-1 MS

Matrix: Drinking Water

Analysis Batch: 243635

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Prep Type: Total/NA

Prep Batch: 243415

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		24.6	25.2		ng/L		102	70 - 130
Perfluorooctanesulfonic acid (PFOS)	2.6		24.6	29.2		ng/L		108	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	28.0		ng/L		114	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	25.4		ng/L		103	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-227787-1 MS

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 243635

Prep Batch: 243415

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		24.6	24.2		ng/L		98	70 - 130
Perfluorohexanoic acid (PFHxA)	3.3		24.6	29.2		ng/L		105	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	28.0		ng/L		114	70 - 130
Perfluorooctanoic acid (PFOA)	2.1		24.6	29.0		ng/L		109	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		24.6	26.7		ng/L		108	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	2.9		24.6	30.4		ng/L		112	70 - 130
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	28.4		ng/L		110	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	27.3		ng/L		106	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		24.6	27.5		ng/L		112	70 - 130
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	27.7		ng/L		113	70 - 130
Perfluorotridecanoic acid (PFTrDA)	<2.0		24.6	29.0		ng/L		118	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		24.6	26.4		ng/L		107	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	25.1		ng/L		102	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	26.1		ng/L		106	70 - 130
Surrogate	MS %Recovery	MS Qualifier	Limits						
d5-NEtFOSAA	109		70 - 130						
13C2 PFHxA	112		70 - 130						
13C2 PFDA	110		70 - 130						
13C3-GenX	108		70 - 130						

Lab Sample ID: 380-227787-1 MSD

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 243635

Prep Batch: 243415

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		24.6	25.5		ng/L		104	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	2.6		24.6	29.0		ng/L		107	70 - 130	0	30
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	28.2		ng/L		115	70 - 130	1	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	24.9		ng/L		101	70 - 130	2	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		24.6	24.6		ng/L		100	70 - 130	2	30
Perfluorohexanoic acid (PFHxA)	3.3		24.6	31.2		ng/L		114	70 - 130	7	30
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	28.9		ng/L		118	70 - 130	3	30
Perfluorooctanoic acid (PFOA)	2.1		24.6	29.1		ng/L		110	70 - 130	0	30
Perfluorodecanoic acid (PFDA)	<2.0		24.6	27.3		ng/L		111	70 - 130	2	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-227787-1 MSD

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 243635

Prep Batch: 243415

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	2.9		24.6	30.4		ng/L		112	70 - 130	0	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	27.4		ng/L		106	70 - 130	3	30
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	28.8		ng/L		112	70 - 130	5	30
Perfluorononanoic acid (PFNA)	<2.0		24.6	27.0		ng/L		110	70 - 130	2	30
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	27.2		ng/L		111	70 - 130	2	30
Perfluorotridecanoic acid (PFTTrDA)	<2.0		24.6	30.0		ng/L		122	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		24.6	27.1		ng/L		110	70 - 130	2	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	25.5		ng/L		104	70 - 130	2	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	26.6		ng/L		108	70 - 130	2	30
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
d5-NEtFOSAA	104		70 - 130								
13C2 PFHxA	108		70 - 130								
13C2 PFDA	116		70 - 130								
13C3-GenX	102		70 - 130								

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

LCMS

Prep Batch: 243406

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-227787-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	533	
380-227787-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	533	
MBL 380-243406/20-A	Method Blank	Total/NA	Water	533	
LCS 380-243406/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-243406/21-A	Lab Control Sample	Total/NA	Water	533	
380-227793-E-1-A MS	Matrix Spike	Total/NA	Water	533	
380-227793-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Prep Batch: 243415

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-227787-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	537.1 DW	
380-227787-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	537.1 DW	
MBL 380-243415/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-243415/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-243415/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-227787-1 MS	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	537.1 DW	
380-227787-1 MSD	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	537.1 DW	

Analysis Batch: 243558

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-227787-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	533	243406
380-227787-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	533	243406
MBL 380-243406/20-A	Method Blank	Total/NA	Water	533	243406
LCS 380-243406/22-A	Lab Control Sample	Total/NA	Water	533	243406
MRL 380-243406/21-A	Lab Control Sample	Total/NA	Water	533	243406
380-227793-E-1-A MS	Matrix Spike	Total/NA	Water	533	243406
380-227793-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	243406

Analysis Batch: 243635

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-227787-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	EPA 537.1 V2	243415
380-227787-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	EPA 537.1 V2	243415
MBL 380-243415/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	243415
LCS 380-243415/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	243415
MRL 380-243415/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	243415
380-227787-1 MS	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	EPA 537.1 V2	243415
380-227787-1 MSD	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	EPA 537.1 V2	243415

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-1

Date Collected: 07/27/26 10:48

Matrix: Drinking Water

Date Received: 07/29/26 10:40

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			243406	XTD8	EA POM	07/30/26 07:53
Total/NA	Analysis	533		1	243558	SZ9R	EA POM	07/30/26 20:03
Total/NA	Prep	537.1 DW			243415	LM3A	EA POM	07/30/26 09:35
Total/NA	Analysis	EPA 537.1 V2		1	243635	SZ9R	EA POM	07/31/26 07:11

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-227787-2

Date Collected: 07/27/26 10:48

Matrix: Water

Date Received: 07/29/26 10:40

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			243406	XTD8	EA POM	07/30/26 07:53
Total/NA	Analysis	533		1	243558	SZ9R	EA POM	07/30/26 20:12
Total/NA	Prep	537.1 DW			243415	LM3A	EA POM	07/30/26 09:35
Total/NA	Analysis	EPA 537.1 V2		1	243635	SZ9R	EA POM	07/31/26 08:55

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-227787-1
SDG: PFAS: Halawa Wells P1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-227787-1	HALAWA WELLS P1 (331-023-WL065)	Drinking Water	07/27/26 10:48	07/29/26 10:40	Hawaii
380-227787-2	FB: HALAWA WELLS P1 (331-023-WL065)	Water	07/27/26 10:48	07/29/26 10:40	Hawaii

Chain of Custody Record

Client Information		Sampler bailey		Lab PM Lopez, Maria		Carrier Tracking No(s)		COC No	
Client Contact kirk iwamoto		Phone +1 808 748 5840		E Mail Maria.Lopez@et.eurofinsus.com		State of Origin		Page Page 1 of 1	
Company City & County of Honolulu		PWSID		Analysis Requested		Job #		Preservation Codes:	
Address 630 South Beretania Street, Chemistry Lab		Due Date Requested		Field Filtered Sample (Yes or No) SUBCONTRACT - 625 PAH Physis LL (EAL) + TICs 3015B_GRO_LL - (MOD) GRO 3015B_DRC_LL_CS - HNL Ranges. C10-C24/C24-C36/C38-C18 25.2_PREC - (MOD) 525plus PLUS TICs 37 ° DW_PREG - 53" Full List All Analysis		Total Number of Containers		A - HCL M Hexane B NaOH N None C - Zn Acetate O - AsNaO2 D Nitric Acid P Na2O4S E NaHSO4 Q Na2SO3 F MeOH R Na2S2O3 G - Amchlor S - H2SO4 H - Ascorbic Acid T TSP Dodecahydrate I Ice U - Acetone J DI Water V MCAA K EDTA W pH 4-5 L EDA Y Trizma Z Other (specify)	
City Honolulu		TAT Requested (days)						Other:	
State Zip HI, 96843		Compliance Project Δ No							
Phone 808-748-5840 (tel)		PO # C20525101 exp 05312023							
Email kiwamoto@hbws.org		WO #							
Project Name RED-HILL/HBWS sites Event Desc RUSH Weekly Red Hill		Project # 38001111							
Site		SSOW#							
Identification		Sample Date		Sample Time		Sample Type (C=Comp, G=grab)		Matrix (W=water, S=solid, O=waste/oil)	
Halawa Wells P1 (331-023-WL065)		27 Jul 2026		1048		C		Water	
FB Halawa Wells P1 (331-023-WL065)		27 Jul-2026		1048					
Do Not Hazard Identification ☐ Non Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☐ Unknown ☐ Radiological		Deliverable Requested I II III IV, Other (specify)		Sample Disposal (A fee may be assessed if samples are not disposed by the 1st 4th) ☐ Return To Client ☐ Disposal By Lab ☐ Archive For _____ Months		Special Instructions/QC Requirements		Special Instructions/No  380-227787 COC	
Empty Kit Relinquished by		Date		Time		Method of Shipment		8750 0967 3959 7/29/26 1040	
Relinquished by		Date/Time		Company		Received by		Date/Time	
Relinquished by		Date/Time		Company		Received by		Date/Time	
Custody Seals Intact Δ Yes Δ No		Custody Seal No		Cooler Temperature(s) °C and Other Remarks (63/18) 2.2 + 0.2 2 gcl - frozen					

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-227787-1

SDG Number: PFAS: Halawa Wells P1

Login Number: 227787

List Number: 1

Creator: Ngo, Theodore

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	