

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 8/27/2026 5:17:39 PM

JOB DESCRIPTION

RED-HILL
PFAS: Aiea Gulch Wells Pump 2

JOB NUMBER

380-231767-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1145

Generated
8/27/2026 5:17:39 PM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	10
Surrogate Summary	11
Isotope Dilution Summary	12
QC Sample Results	14
QC Association Summary	25
Lab Chronicle	26
Certification Summary	27
Method Summary	28
Sample Summary	29
Chain of Custody	30
Receipt Checklists	31



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-231767-1

Job ID: 380-231767-1

Eurofins Pomona

Job Narrative 380-231767-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 8/19/2026 9:30 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.9°C.

PFAS

Method EPA 537.1 V2: Internal standard responses were outside of acceptance limits for the following sample: FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072) (380-231767-2). Insufficient volume for re-extraction. Field reagent blank not required since native sample is ND. 537.1 data excluded for the Field Reagent Blank, analysis of the Field Blank is required only if a field sample contains a method analyte at, or above, the MRL. Sample results showed ND (Non Detect) thus valid for reporting. (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**
PWSID Number: HI0000331

Lab Sample ID: 380-231767-1

☐ No Detections.

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**
PWSID Number: HI0000331

Lab Sample ID: 380-231767-2

☐ No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-231767-1

Date Collected: 08/17/26 11:04

Matrix: Drinking Water

Date Received: 08/19/26 09:30

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 16:29	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	102		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C6 PFDA	113		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C5 PFHxA	107		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C4 PFHpA	107		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C8 PFOA	106		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C9 PFNA	110		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C7 PFUnA	110		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C2 PFDoA	106		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C4 PFBA	104		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C5 PFPeA	108		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C3 PFBS	104		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C3 PFHxS	106		50 - 200	08/20/26 14:08	08/21/26 16:29	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-231767-1

Date Collected: 08/17/26 11:04

Matrix: Drinking Water

Date Received: 08/19/26 09:30

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	110		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C2-4:2-FTS	105		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C2-6:2-FTS	112		50 - 200	08/20/26 14:08	08/21/26 16:29	1
13C2-8:2-FTS	108		50 - 200	08/20/26 14:08	08/21/26 16:29	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		08/20/26 03:50	08/20/26 23:13	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	108		70 - 130	08/20/26 03:50	08/20/26 23:13	1
13C2 PFHxA	110		70 - 130	08/20/26 03:50	08/20/26 23:13	1
13C2 PFDA	110		70 - 130	08/20/26 03:50	08/20/26 23:13	1
13C3-GenX	104		70 - 130	08/20/26 03:50	08/20/26 23:13	1

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-231767-2

Date Collected: 08/17/26 11:04

Matrix: Water

Date Received: 08/19/26 09:30

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-231767-2

Date Collected: 08/17/26 11:04

Matrix: Water

Date Received: 08/19/26 09:30

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		08/20/26 14:08	08/21/26 19:54	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	93		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C6 PFDA	113		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C5 PFHxA	103		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C4 PFHpA	104		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C8 PFOA	107		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C9 PFNA	113		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C7 PFUnA	113		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C2 PFDoA	112		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C4 PFBA	103		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C5 PFPeA	103		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C3 PFBS	98		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C3 PFHxS	103		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C8 PFOS	107		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C2-4:2-FTS	99		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C2-6:2-FTS	107		50 - 200	08/20/26 14:08	08/21/26 19:54	1
13C2-8:2-FTS	113		50 - 200	08/20/26 14:08	08/21/26 19:54	1

Eurofins Pomona

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-231767-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-231767-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)						
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)			
380-231767-1	AIEA GULCH WELLS PUMP 2 (	108	110	110	104			
380-231767-1 MS	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	108	104	103	95			
380-231767-1 MSD	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	103	108	103	101			
Surrogate Legend								
d5NEFOS = d5-NEtFOSAA								
PFHxA = 13C2 PFHxA								
PFDA = 13C2 PFDA								
GenX = 13C3-GenX								

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)						
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)			
LCS 380-247663/22-A	Lab Control Sample	103	100	100	97			
MBL 380-247663/20-A	Method Blank	103	108	115	105			
MRL 380-247663/21-A	Lab Control Sample	106	106	114	108			
Surrogate Legend								
d5NEFOS = d5-NEtFOSAA								
PFHxA = 13C2 PFHxA								
PFDA = 13C2 PFDA								
GenX = 13C3-GenX								

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-231767-1	AIEA GULCH WELLS PUMP 2 (	102	113	107	107	106	110	110	106
380-231767-1 MS	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	109	115	106	106	106	108	111	113
380-231767-1 MSD	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	101	109	100	107	106	108	112	106

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-231767-1	AIEA GULCH WELLS PUMP 2 (	104	108	104	106	110	105	112	108
380-231767-1 MS	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	101	103	100	103	109	99	110	105
380-231767-1 MSD	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	105	110	96	100	106	100	105	103

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDaA = 13C2 PFDaA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-231767-2	FB: AIEA GULCH WELLS PUMI	93	113	103	104	107	113	113	112
LCS 380-247773/22-A	Lab Control Sample	100	112	103	104	101	109	112	112
MBL 380-247773/20-A	Method Blank	97	104	100	101	104	108	108	108
MRL 380-247773/21-A	Lab Control Sample	99	114	105	102	106	109	112	111

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-231767-2	FB: AIEA GULCH WELLS PUMI	103	103	98	103	107	99	107	113
LCS 380-247773/22-A	Lab Control Sample	102	107	102	103	106	98	106	107
MBL 380-247773/20-A	Method Blank	105	104	103	105	104	103	110	96
MRL 380-247773/21-A	Lab Control Sample	102	103	99	105	108	105	105	108

Surrogate Legend

Eurofins Pomona

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDoA = 13C2 PFDoA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-247773/20-A

Matrix: Water

Analysis Batch: 247991

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 247773

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.99		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluorobutanoic acid (PFBA)	<0.68		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.36		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.47		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.46		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.24		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.45		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoropentanoic acid (PFPeA)	<0.37		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.35		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Perfluoropentanesulfonic acid (PFPeS)	<0.38		2.0	ng/L		08/20/26 14:08	08/21/26 15:58	1
Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	97		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C6 PFDA	104		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C5 PFHxA	100		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C4 PFHpA	101		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C8 PFOA	104		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C9 PFNA	108		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C7 PFUnA	108		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C2 PFDoA	108		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C4 PFBA	105		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C5 PFPeA	104		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C3 PFBS	103		50 - 200			08/20/26 14:08	08/21/26 15:58	1
13C3 PFHxS	105		50 - 200			08/20/26 14:08	08/21/26 15:58	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-247773/20-A

Matrix: Water

Analysis Batch: 247991

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 247773

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	104		50 - 200	08/20/26 14:08	08/21/26 15:58	1
13C2-4:2-FTS	103		50 - 200	08/20/26 14:08	08/21/26 15:58	1
13C2-6:2-FTS	110		50 - 200	08/20/26 14:08	08/21/26 15:58	1
13C2-8:2-FTS	96		50 - 200	08/20/26 14:08	08/21/26 15:58	1

Lab Sample ID: LCS 380-247773/22-A

Matrix: Water

Analysis Batch: 247991

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 247773

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	59.1	56.7		ng/L		96	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	59.1	54.6		ng/L		92	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	59.1	57.9		ng/L		98	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	59.1	55.4		ng/L		94	70 - 130
Perfluorobutanesulfonic acid (PFBS)	59.1	59.0		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	59.1	55.5		ng/L		94	70 - 130
Perfluorododecanoic acid (PFDoA)	59.1	54.9		ng/L		93	70 - 130
Perfluoroheptanoic acid (PFHpA)	59.1	60.6		ng/L		103	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	59.1	57.4		ng/L		97	70 - 130
Perfluorohexanoic acid (PFHxA)	59.1	56.2		ng/L		95	70 - 130
Perfluorononanoic acid (PFNA)	59.1	57.2		ng/L		97	70 - 130
Perfluorooctanesulfonic acid (PFOS)	59.1	55.2		ng/L		94	70 - 130
Perfluorooctanoic acid (PFOA)	59.1	61.6		ng/L		104	70 - 130
Perfluoroundecanoic acid (PFUnA)	59.1	58.0		ng/L		98	70 - 130
Perfluorobutanoic acid (PFBA)	59.1	57.8		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	59.1	58.9		ng/L		100	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	59.1	58.6		ng/L		99	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	59.1	54.6		ng/L		92	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	59.1	58.3		ng/L		99	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	59.1	56.1		ng/L		95	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	59.1	57.5		ng/L		97	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	59.1	55.9		ng/L		95	70 - 130
Perfluoropentanoic acid (PFPeA)	59.1	57.4		ng/L		97	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	59.1	54.0		ng/L		91	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-247773/22-A

Matrix: Water

Analysis Batch: 247991

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 247773

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	59.1	55.8		ng/L		94	70 - 130
		LCS	LCS				
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	100		50 - 200				
13C6 PFDA	112		50 - 200				
13C5 PFHxA	103		50 - 200				
13C4 PFHpA	104		50 - 200				
13C8 PFOA	101		50 - 200				
13C9 PFNA	109		50 - 200				
13C7 PFUnA	112		50 - 200				
13C2 PFDoA	112		50 - 200				
13C4 PFBA	102		50 - 200				
13C5 PFPeA	107		50 - 200				
13C3 PFBS	102		50 - 200				
13C3 PFHxS	103		50 - 200				
13C8 PFOS	106		50 - 200				
13C2-4:2-FTS	98		50 - 200				
13C2-6:2-FTS	106		50 - 200				
13C2-8:2-FTS	107		50 - 200				

Lab Sample ID: MRL 380-247773/21-A

Matrix: Water

Analysis Batch: 247991

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 247773

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.97	1.92	J	ng/L		98	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	1.97	1.90	J	ng/L		96	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.97	2.05	J	ng/L		104	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.97	2.01	J	ng/L		102	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.97	2.08	J	ng/L		106	50 - 150
Perfluorodecanoic acid (PFDA)	1.97	1.94	J	ng/L		99	50 - 150
Perfluorododecanoic acid (PFDoA)	1.97	1.98	J	ng/L		100	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.97	2.18	J	ng/L		111	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.97	1.85	J	ng/L		94	50 - 150
Perfluorohexanoic acid (PFHxA)	1.97	1.89	J	ng/L		96	50 - 150
Perfluorononanoic acid (PFNA)	1.97	2.02	J	ng/L		103	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.97	1.93	J	ng/L		98	50 - 150
Perfluorooctanoic acid (PFOA)	1.97	2.10	J	ng/L		106	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.97	1.98	J	ng/L		101	50 - 150
Perfluorobutanoic acid (PFBA)	1.97	1.98	J	ng/L		100	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-247773/21-A

Matrix: Water

Analysis Batch: 247991

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 247773

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	1.97	2.14	J	ng/L		109	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	1.97	1.82	J	ng/L		93	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	1.97	1.94	J	ng/L		99	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	1.97	2.16	J	ng/L		110	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	1.97	1.91	J	ng/L		97	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	1.97	2.09	J	ng/L		106	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	1.97	1.96	J	ng/L		100	50 - 150
Perfluoropentanoic acid (PFPeA)	1.97	2.07	J	ng/L		105	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	1.97	1.80	J	ng/L		91	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	1.97	1.76	J	ng/L		90	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	99		50 - 200
13C6 PFDA	114		50 - 200
13C5 PFHxA	105		50 - 200
13C4 PFHpA	102		50 - 200
13C8 PFOA	106		50 - 200
13C9 PFNA	109		50 - 200
13C7 PFUnA	112		50 - 200
13C2 PFDoA	111		50 - 200
13C4 PFBA	102		50 - 200
13C5 PFPeA	103		50 - 200
13C3 PFBS	99		50 - 200
13C3 PFHxS	105		50 - 200
13C8 PFOS	108		50 - 200
13C2-4:2-FTS	105		50 - 200
13C2-6:2-FTS	105		50 - 200
13C2-8:2-FTS	108		50 - 200

Lab Sample ID: 380-231767-1 MS

Matrix: Drinking Water

Analysis Batch: 247991

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Prep Type: Total/NA

Prep Batch: 247773

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		118	109		ng/L		92	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		118	107		ng/L		91	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		118	115		ng/L		97	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-231767-1 MS

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 247991

Prep Batch: 247773

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		118	107		ng/L		90	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		118	118		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		118	109		ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		118	110		ng/L		93	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		118	115		ng/L		97	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		118	116		ng/L		97	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		118	108		ng/L		91	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		118	114		ng/L		96	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		118	110		ng/L		93	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		118	116		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		118	113		ng/L		95	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		118	117		ng/L		99	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		118	120		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		118	112		ng/L		95	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		118	107		ng/L		91	70 - 130
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		118	111		ng/L		94	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		118	111		ng/L		94	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		118	115		ng/L		97	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		118	111		ng/L		94	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		118	111		ng/L		93	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		118	107		ng/L		90	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		118	111		ng/L		94	70 - 130

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C3 HFPO-DA	109		50 - 200
13C6 PFDA	115		50 - 200
13C5 PFHxA	106		50 - 200
13C4 PFHpA	106		50 - 200
13C8 PFOA	106		50 - 200
13C9 PFNA	108		50 - 200
13C7 PFUnA	111		50 - 200
13C2 PFDoA	113		50 - 200
13C4 PFBA	101		50 - 200
13C5 PFPeA	103		50 - 200
13C3 PFBS	100		50 - 200
13C3 PFHxS	103		50 - 200
13C8 PFOS	109		50 - 200

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-231767-1 MS

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 247991

Prep Batch: 247773

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	99		50 - 200
13C2-6:2-FTS	110		50 - 200
13C2-8:2-FTS	105		50 - 200

Lab Sample ID: 380-231767-1 MSD

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 247991

Prep Batch: 247773

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		118	109		ng/L		93	70 - 130	0	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)	<2.0		118	108		ng/L		91	70 - 130	0	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		118	109		ng/L		92	70 - 130	5	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		118	112		ng/L		95	70 - 130	5	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		118	116		ng/L		98	70 - 130	2	30
Perfluorodecanoic acid (PFDA)	<2.0		118	110		ng/L		93	70 - 130	0	30
Perfluorododecanoic acid (PFDaA)	<2.0		118	117		ng/L		99	70 - 130	6	30
Perfluoroheptanoic acid (PFHpA)	<2.0		118	111		ng/L		94	70 - 130	3	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		118	112		ng/L		95	70 - 130	3	30
Perfluorohexanoic acid (PFHxA)	<2.0		118	116		ng/L		98	70 - 130	7	30
Perfluorononanoic acid (PFNA)	<2.0		118	115		ng/L		97	70 - 130	0	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		118	110		ng/L		93	70 - 130	0	30
Perfluorooctanoic acid (PFOA)	<2.0		118	115		ng/L		98	70 - 130	1	30
Perfluoroundecanoic acid (PFUnA)	<2.0		118	109		ng/L		92	70 - 130	4	30
Perfluorobutanoic acid (PFBA)	<2.0		118	118		ng/L		100	70 - 130	0	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		118	112		ng/L		95	70 - 130	7	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		118	111		ng/L		94	70 - 130	1	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		118	110		ng/L		93	70 - 130	2	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		118	123		ng/L		104	70 - 130	10	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		118	115		ng/L		98	70 - 130	4	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		118	110		ng/L		93	70 - 130	5	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		118	108		ng/L		92	70 - 130	2	30
Perfluoropentanoic acid (PFPeA)	<2.0		118	108		ng/L		91	70 - 130	2	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		118	109		ng/L		92	70 - 130	2	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		118	111		ng/L		94	70 - 130	0	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD %Recovery	MSD Qualifier	Limits
13C3 HFPO-DA	101		50 - 200
13C6 PFDA	109		50 - 200
13C5 PFHxA	100		50 - 200
13C4 PFHpA	107		50 - 200
13C8 PFOA	106		50 - 200
13C9 PFNA	108		50 - 200
13C7 PFUnA	112		50 - 200
13C2 PFDoA	106		50 - 200
13C4 PFBA	105		50 - 200
13C5 PFPeA	110		50 - 200
13C3 PFBS	96		50 - 200
13C3 PFHxS	100		50 - 200
13C8 PFOS	106		50 - 200
13C2-4:2-FTS	100		50 - 200
13C2-6:2-FTS	105		50 - 200
13C2-8:2-FTS	103		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-247663/20-A
Matrix: Water
Analysis Batch: 247827

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 247663

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.57		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.41		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorodecanoic acid (PFDA)	<0.30		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorotetradecanoic acid (PFTA)	<0.53		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Perfluorotridecanoic acid (PFTrDA)	<0.35		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.29		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		08/20/26 03:50	08/20/26 22:43	1
Surrogate	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	103		70 - 130			08/20/26 03:50	08/20/26 22:43	1
13C2 PFHxA	108		70 - 130			08/20/26 03:50	08/20/26 22:43	1
13C2 PFDA	115		70 - 130			08/20/26 03:50	08/20/26 22:43	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-247663/20-A

Matrix: Water

Analysis Batch: 247827

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 247663

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	105		70 - 130	08/20/26 03:50	08/20/26 22:43	1

Lab Sample ID: LCS 380-247663/22-A

Matrix: Water

Analysis Batch: 247827

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 247663

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	24.6	23.6		ng/L		96	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	24.6	25.2		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	24.6	25.6		ng/L		104	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	24.6	24.5		ng/L		100	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	24.6	26.1		ng/L		106	70 - 130
Perfluorohexanoic acid (PFHxA)	24.6	24.5		ng/L		100	70 - 130
Perfluorododecanoic acid (PFDoA)	24.6	25.0		ng/L		102	70 - 130
Perfluorooctanoic acid (PFOA)	24.6	24.8		ng/L		101	70 - 130
Perfluorodecanoic acid (PFDA)	24.6	25.5		ng/L		104	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	24.6	25.0		ng/L		102	70 - 130
Perfluorobutanesulfonic acid (PFBS)	24.6	26.2		ng/L		106	70 - 130
Perfluoroheptanoic acid (PFHpA)	24.6	24.2		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	24.6	25.7		ng/L		105	70 - 130
Perfluorotetradecanoic acid (PFTA)	24.6	18.6		ng/L		76	70 - 130
Perfluorotridecanoic acid (PFTrDA)	24.6	29.5		ng/L		120	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	24.6	25.0		ng/L		102	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	24.6	24.5		ng/L		99	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	24.6	24.2		ng/L		98	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	103		70 - 130
13C2 PFHxA	100		70 - 130
13C2 PFDA	100		70 - 130
13C3-GenX	97		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-247663/21-A

Matrix: Water

Analysis Batch: 247827

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 247663

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.96	2.20	J	ng/L		112	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.96	2.15	J	ng/L		109	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.96	2.29	J	ng/L		116	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	1.96	2.11	J	ng/L		107	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	1.96	1.91	J	ng/L		97	50 - 150
Perfluorohexanoic acid (PFHxA)	1.96	2.30	J	ng/L		117	50 - 150
Perfluorododecanoic acid (PFDoA)	1.96	2.32	J	ng/L		118	50 - 150
Perfluorooctanoic acid (PFOA)	1.96	2.20	J	ng/L		112	50 - 150
Perfluorodecanoic acid (PFDA)	1.96	2.29	J	ng/L		116	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.96	2.14	J	ng/L		109	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.96	2.23	J	ng/L		113	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.96	2.32	J	ng/L		118	50 - 150
Perfluorononanoic acid (PFNA)	1.96	2.34	J	ng/L		119	50 - 150
Perfluorotetradecanoic acid (PFTA)	1.96	1.64	J	ng/L		84	50 - 150
Perfluorotridecanoic acid (PFTrDA)	1.96	2.66	J	ng/L		135	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	1.96	2.14	J	ng/L		109	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.96	2.02	J	ng/L		103	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.96	2.22	J	ng/L		113	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	106		70 - 130
13C2 PFHxA	106		70 - 130
13C2 PFDA	114		70 - 130
13C3-GenX	108		70 - 130

Lab Sample ID: 380-231767-1 MS

Matrix: Drinking Water

Analysis Batch: 247827

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Prep Type: Total/NA

Prep Batch: 247663

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		24.6	21.5		ng/L		87	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		24.6	23.8		ng/L		97	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	24.7		ng/L		100	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	23.3		ng/L		95	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-231767-1 MS

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 247827

Prep Batch: 247663

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		24.6	25.7		ng/L		104	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		24.6	23.7		ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	25.0		ng/L		101	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		24.6	23.2		ng/L		94	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		24.6	25.1		ng/L		102	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		24.6	24.6		ng/L		98	70 - 130
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	25.6		ng/L		104	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	23.9		ng/L		97	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		24.6	24.8		ng/L		101	70 - 130
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	17.8		ng/L		72	70 - 130
Perfluorotridecanoic acid (PFTrDA)	<2.0		24.6	28.2		ng/L		115	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		24.6	24.3		ng/L		99	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	23.8		ng/L		97	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	23.3		ng/L		95	70 - 130
Surrogate	MS %Recovery	MS Qualifier	Limits						
d5-NEtFOSAA	108		70 - 130						
13C2 PFHxA	104		70 - 130						
13C2 PFDA	103		70 - 130						
13C3-GenX	95		70 - 130						

Lab Sample ID: 380-231767-1 MSD

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 247827

Prep Batch: 247663

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		24.6	23.8		ng/L		97	70 - 130	10	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		24.6	23.8		ng/L		97	70 - 130	0	30
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	24.9		ng/L		101	70 - 130	1	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	22.6		ng/L		92	70 - 130	3	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		24.6	25.8		ng/L		105	70 - 130	0	30
Perfluorohexanoic acid (PFHxA)	<2.0		24.6	25.7		ng/L		101	70 - 130	8	30
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	26.3		ng/L		107	70 - 130	5	30
Perfluorooctanoic acid (PFOA)	<2.0		24.6	25.6		ng/L		104	70 - 130	10	30
Perfluorodecanoic acid (PFDA)	<2.0		24.6	26.2		ng/L		107	70 - 130	5	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-231767-1 MSD

Client Sample ID: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 247827

Prep Batch: 247663

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	<2.0		24.6	24.0		ng/L		96	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	25.3		ng/L		103	70 - 130	1	30
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	24.8		ng/L		101	70 - 130	4	30
Perfluorononanoic acid (PFNA)	<2.0		24.6	26.2		ng/L		107	70 - 130	6	30
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	17.8		ng/L		73	70 - 130	0	30
Perfluorotridecanoic acid (PFTTrDA)	<2.0		24.6	30.1		ng/L		122	70 - 130	7	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		24.6	23.3		ng/L		95	70 - 130	4	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	23.5		ng/L		96	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	24.9		ng/L		101	70 - 130	7	30
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
d5-NEtFOSAA	103		70 - 130								
13C2 PFHxA	108		70 - 130								
13C2 PFDA	103		70 - 130								
13C3-GenX	101		70 - 130								

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

LCMS

Prep Batch: 247663

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-231767-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	537.1 DW	
MBL 380-247663/20-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-247663/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-247663/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-231767-1 MS	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	537.1 DW	
380-231767-1 MSD	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	537.1 DW	

Prep Batch: 247773

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-231767-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	
380-231767-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TPC	Total/NA	Water	533	
MBL 380-247773/20-A	Method Blank	Total/NA	Water	533	
LCS 380-247773/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-247773/21-A	Lab Control Sample	Total/NA	Water	533	
380-231767-1 MS	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	
380-231767-1 MSD	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	

Analysis Batch: 247827

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-231767-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	EPA 537.1 V2	247663
MBL 380-247663/20-A	Method Blank	Total/NA	Water	EPA 537.1 V2	247663
LCS 380-247663/22-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	247663
MRL 380-247663/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	247663
380-231767-1 MS	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	EPA 537.1 V2	247663
380-231767-1 MSD	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	EPA 537.1 V2	247663

Analysis Batch: 247991

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-231767-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	247773
380-231767-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TPC	Total/NA	Water	533	247773
MBL 380-247773/20-A	Method Blank	Total/NA	Water	533	247773
LCS 380-247773/22-A	Lab Control Sample	Total/NA	Water	533	247773
MRL 380-247773/21-A	Lab Control Sample	Total/NA	Water	533	247773
380-231767-1 MS	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	247773
380-231767-1 MSD	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	247773

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-231767-1

Date Collected: 08/17/26 11:04

Matrix: Drinking Water

Date Received: 08/19/26 09:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			247773	N8NE	EA POM	08/20/26 14:08
Total/NA	Analysis	533		1	247991	Y5FM	EA POM	08/21/26 16:29
Total/NA	Prep	537.1 DW			247663	G9MN	EA POM	08/20/26 03:50
Total/NA	Analysis	EPA 537.1 V2		1	247827	Y5FM	EA POM	08/20/26 23:13

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-231767-2

Date Collected: 08/17/26 11:04

Matrix: Water

Date Received: 08/19/26 09:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			247773	N8NE	EA POM	08/20/26 14:08
Total/NA	Analysis	533		1	247991	Y5FM	EA POM	08/21/26 19:54

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-231767-1
SDG: PFAS: Aiea Gulch Wells Pump 2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-231767-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Drinking Water	08/17/26 11:04	08/19/26 09:30	HI0000331
380-231767-2	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Water	08/17/26 11:04	08/19/26 09:30	HI0000331

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-231767-1

SDG Number: PFAS: Aiea Gulch Wells Pump 2

Login Number: 231767

List Number: 1

Creator: Del Rosario, Michael

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	