

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 2/12/2026 5:21:14 PM

JOB DESCRIPTION

INTERA - Red-Hill-Incident
Site H1 625.1, 8015B_DRO/_GRO, 8260B

JOB NUMBER

380-195028-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
2/12/2026 5:21:14 PM

Table of Contents

Cover Page 1

Table of Contents 3

Definitions/Glossary 4

Case Narrative 5

Detection Summary 6

Client Sample Results 7

Action Limit Summary 11

Surrogate Summary 12

QC Sample Results 14

QC Association Summary 25

Lab Chronicle 27

Certification Summary 28

Method Summary 29

Sample Summary 30

Chain of Custody 31

Receipt Checklists 33



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Qualifiers

GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

GC VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: INTERA - Red-Hill-Incident

Job ID: 380-195028-1

Job ID: 380-195028-1

Eurofins Pomona

Job Narrative 380-195028-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 1/29/2026 9:52 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 5.4°C.

GC/MS VOA

Method 8260B: Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with analytical batch 570-688946. The laboratory control sample (LCS) was performed in duplicate (LCSD) to provide precision data for this batch. The Trip Blank, BWS2253H1TB (3801950282), provided by the client did not originate from Calscience. This Trip Blank showed a detection of Isopropanol; therefore, its analysis was canceled. Trip Blank analysis is only required when a field sample contains a method analyte at or above the MRL. All field sample results were non-detect, and therefore remain valid for reporting. (XWB4)

BWS2253-H1-TB (380-195028-2)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

GC/MS Semi VOA

Method 625.1: Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with preparation batch 570-689280. The laboratory control sample (LCS) was performed in duplicate (LCSD) to provide precision data for this batch. 625 Sim

Method 625.1 SIM: Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with preparation batch 570-689280. The laboratory control sample (LCS) was performed in duplicate (LCSD) to provide precision data for this batch. 625 Sim

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Gasoline Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Diesel Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Client Sample ID: BWS2253-H1-AQ

Lab Sample ID: 380-195028-1

☐ No Detections.

Client Sample ID: BWS2253-H1-TB

Lab Sample ID: 380-195028-2

☐ No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Client Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Client Sample ID: BWS2253-H1-AQ

Lab Sample ID: 380-195028-1

Date Collected: 01/28/26 09:30

Matrix: Water

Date Received: 01/29/26 09:52

Method: SW846 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<2.0		2.0	ug/L			01/30/26 01:58	1
1,1,1-Trichloroethane	<1.0		1.0	ug/L			01/30/26 01:58	1
1,1,2,2-Tetrachloroethane	<1.0		1.0	ug/L			01/30/26 01:58	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<10		10	ug/L			01/30/26 01:58	1
1,1,2-Trichloroethane	<1.0		1.0	ug/L			01/30/26 01:58	1
1,1-Dichloroethane	<1.0		1.0	ug/L			01/30/26 01:58	1
1,1-Dichloroethene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,1-Dichloropropene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,2,3-Trichlorobenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,2,3-Trichloropropane	<5.0		5.0	ug/L			01/30/26 01:58	1
1,2,4-Trichlorobenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,2,4-Trimethylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,2-Dibromo-3-Chloropropane	<10		10	ug/L			01/30/26 01:58	1
1,2-Dichlorobenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,2-Dichloroethane	<0.50		0.50	ug/L			01/30/26 01:58	1
1,2-Dichloropropane	<1.0		1.0	ug/L			01/30/26 01:58	1
1,3,5-Trimethylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,3-Dichlorobenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
1,3-Dichloropropane	<1.0		1.0	ug/L			01/30/26 01:58	1
1,4-Dichlorobenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
2,2-Dichloropropane	<1.0		1.0	ug/L			01/30/26 01:58	1
2-Butanone	<10		10	ug/L			01/30/26 01:58	1
2-Chlorotoluene	<1.0		1.0	ug/L			01/30/26 01:58	1
2-Hexanone	<10		10	ug/L			01/30/26 01:58	1
4-Chlorotoluene	<1.0		1.0	ug/L			01/30/26 01:58	1
Acetone	<10		10	ug/L			01/30/26 01:58	1
Benzene	<0.50		0.50	ug/L			01/30/26 01:58	1
Bromobenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
Bromochloromethane	<2.0		2.0	ug/L			01/30/26 01:58	1
Bromodichloromethane	<1.0		1.0	ug/L			01/30/26 01:58	1
Bromoform	<5.0		5.0	ug/L			01/30/26 01:58	1
Bromomethane	<10		10	ug/L			01/30/26 01:58	1
Carbon disulfide	<10		10	ug/L			01/30/26 01:58	1
Carbon tetrachloride	<0.50		0.50	ug/L			01/30/26 01:58	1
Chlorobenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
Chloroethane	<5.0		5.0	ug/L			01/30/26 01:58	1
Chloroform	<1.0		1.0	ug/L			01/30/26 01:58	1
Chloromethane	<10		10	ug/L			01/30/26 01:58	1
cis-1,2-Dichloroethene	<1.0		1.0	ug/L			01/30/26 01:58	1
cis-1,3-Dichloropropene	<0.50		0.50	ug/L			01/30/26 01:58	1
Dibromochloromethane	<2.0		2.0	ug/L			01/30/26 01:58	1
Dibromomethane	<1.0		1.0	ug/L			01/30/26 01:58	1
Dichlorodifluoromethane	<5.0		5.0	ug/L			01/30/26 01:58	1
Diethyl ether	<10		10	ug/L			01/30/26 01:58	1
Di-isopropyl ether (DIPE)	<2.0		2.0	ug/L			01/30/26 01:58	1
Ethanol	<100		100	ug/L			01/30/26 01:58	1
Ethylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
Ethylene Dibromide	<1.0		1.0	ug/L			01/30/26 01:58	1
Ethyl-t-butyl ether (ETBE)	<2.0		2.0	ug/L			01/30/26 01:58	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Client Sample ID: BWS2253-H1-AQ

Lab Sample ID: 380-195028-1

Date Collected: 01/28/26 09:30

Matrix: Water

Date Received: 01/29/26 09:52

Method: SW846 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexachloro-1,3-butadiene	<10		10	ug/L			01/30/26 01:58	1
Hexane	<5.0		5.0	ug/L			01/30/26 01:58	1
Isobutyl alcohol	<75		75	ug/L			01/30/26 01:58	1
Isopropanol	<75		75	ug/L			01/30/26 01:58	1
Isopropylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
m,p-Xylene	<2.0		2.0	ug/L			01/30/26 01:58	1
Methylene Chloride	<10		10	ug/L			01/30/26 01:58	1
Methyl-t-Butyl Ether (MTBE)	<1.0		1.0	ug/L			01/30/26 01:58	1
MIBK	<10		10	ug/L			01/30/26 01:58	1
Naphthalene	<10		10	ug/L			01/30/26 01:58	1
n-Butylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
N-Propylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
o-Xylene	<1.0		1.0	ug/L			01/30/26 01:58	1
p-Isopropyltoluene	<1.0		1.0	ug/L			01/30/26 01:58	1
sec-Butylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
Styrene	<1.0		1.0	ug/L			01/30/26 01:58	1
tert-Amyl alcohol	<50		50	ug/L			01/30/26 01:58	1
Tert-amyl-methyl ether (TAME)	<2.0		2.0	ug/L			01/30/26 01:58	1
tert-Butyl alcohol (TBA)	<10		10	ug/L			01/30/26 01:58	1
tert-Butylbenzene	<1.0		1.0	ug/L			01/30/26 01:58	1
Tetrachloroethene	<1.0		1.0	ug/L			01/30/26 01:58	1
Toluene	<1.0		1.0	ug/L			01/30/26 01:58	1
trans-1,2-Dichloroethene	<1.0		1.0	ug/L			01/30/26 01:58	1
trans-1,3-Dichloropropene	<0.50		0.50	ug/L			01/30/26 01:58	1
Trichloroethene	<1.0		1.0	ug/L			01/30/26 01:58	1
Trichlorofluoromethane	<10		10	ug/L			01/30/26 01:58	1
Vinyl acetate	<10		10	ug/L			01/30/26 01:58	1
Vinyl chloride	<0.50		0.50	ug/L			01/30/26 01:58	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		80 - 120		01/30/26 01:58	1
4-Bromofluorobenzene (Surr)	99		80 - 120		01/30/26 01:58	1
Dibromofluoromethane (Surr)	103		80 - 120		01/30/26 01:58	1
Toluene-d8 (Surr)	98		80 - 120		01/30/26 01:58	1

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
2-Methylnaphthalene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Acenaphthene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Acenaphthylene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Anthracene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Benzo[a]anthracene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Benzo[a]pyrene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Benzo[b]fluoranthene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Benzo[g,h,i]perylene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Benzo[k]fluoranthene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Chrysene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Dibenz(a,h)anthracene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Fluoranthene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Client Sample ID: BWS2253-H1-AQ

Lab Sample ID: 380-195028-1

Date Collected: 01/28/26 09:30

Matrix: Water

Date Received: 01/29/26 09:52

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Indeno[1,2,3-cd]pyrene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Naphthalene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Phenanthrene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1
Pyrene	<0.19		0.19	ug/L		01/30/26 12:01	02/04/26 13:46	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	90		28 - 127	01/30/26 12:01	02/04/26 13:46	1
2-Fluorobiphenyl (Surr)	84		31 - 120	01/30/26 12:01	02/04/26 13:46	1
2-Fluorophenol (Surr)	45		17 - 120	01/30/26 12:01	02/04/26 13:46	1
Nitrobenzene-d5 (Surr)	80		27 - 120	01/30/26 12:01	02/04/26 13:46	1
Phenol-d6 (Surr)	29		10 - 120	01/30/26 12:01	02/04/26 13:46	1
p-Terphenyl-d14 (Surr)	77		45 - 120	01/30/26 12:01	02/04/26 13:46	1

Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Butylated Hydroxytoluene	10	T J N	ug/L		6.73	128-37-0	01/30/26 12:01	02/05/26 12:36	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	74		33 - 139	01/30/26 12:01	02/05/26 12:36	1
2-Fluorobiphenyl (Surr)	86		33 - 126	01/30/26 12:01	02/05/26 12:36	1
2-Fluorophenol (Surr)	59		12 - 120	01/30/26 12:01	02/05/26 12:36	1
Nitrobenzene-d5 (Surr)	75		36 - 120	01/30/26 12:01	02/05/26 12:36	1
Phenol-d6 (Surr)	29		10 - 120	01/30/26 12:01	02/05/26 12:36	1
p-Terphenyl-d14 (Surr)	88		47 - 131	01/30/26 12:01	02/05/26 12:36	1

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/06/26 13:43	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	94		38 - 134		02/06/26 13:43	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<26		26	ug/L		02/02/26 09:35	02/06/26 22:12	1
Motor Oil Range Organics [C24-C36]	<26		26	ug/L		02/02/26 09:35	02/06/26 22:12	1
C8-C18	<26		26	ug/L		02/02/26 09:35	02/06/26 22:12	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	114		60 - 130	02/02/26 09:35	02/06/26 22:12	1

Client Sample ID: BWS2253-H1-TB

Lab Sample ID: 380-195028-2

Date Collected: 01/28/26 09:30

Matrix: Water

Date Received: 01/29/26 09:52

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/06/26 18:39	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Client Sample ID: BWS2253-H1-TB
Date Collected: 01/28/26 09:30
Date Received: 01/29/26 09:52

Lab Sample ID: 380-195028-2
Matrix: Water

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	94		38 - 134		02/06/26 18:39	1

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Action Limit Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Client Sample ID: BWS2253-H1-AQ

Lab Sample ID: 380-195028-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	HI Org Limit	RL	Method	Prep Type
1,1,1-Trichloroethane	<1.0		ug/L	200	200.0	1.0	8260B	Total/NA
1,1,2-Trichloroethane	<1.0		ug/L	5	5.000	1.0	8260B	Total/NA
1,1-Dichloroethene	<1.0		ug/L	7	7.000	1.0	8260B	Total/NA
1,2,3-Trichloropropane	<5.0		ug/L		0.6000	5.0	8260B	Total/NA
1,2,4-Trichlorobenzene	<1.0		ug/L	70	70.00	1.0	8260B	Total/NA
1,2-Dibromo-3-Chloropropane	<10		ug/L	0.2		10	8260B	Total/NA
1,2-Dichlorobenzene	<1.0		ug/L	600	600.0	1.0	8260B	Total/NA
1,2-Dichloroethane	<0.50		ug/L	5	5.000	0.50	8260B	Total/NA
1,2-Dichloropropane	<1.0		ug/L	5	5.000	1.0	8260B	Total/NA
1,4-Dichlorobenzene	<1.0		ug/L	75	75.000	1.0	8260B	Total/NA
Benzene	<0.50		ug/L	5	5.000	0.50	8260B	Total/NA
Bromodichloromethane	<1.0		ug/L	80		1.0	8260B	Total/NA
Bromoform	<5.0		ug/L	80		5.0	8260B	Total/NA
Carbon tetrachloride	<0.50		ug/L	5	5.000	0.50	8260B	Total/NA
Chlorobenzene	<1.0		ug/L	100	100.0	1.0	8260B	Total/NA
Chloroform	<1.0		ug/L	80		1.0	8260B	Total/NA
cis-1,2-Dichloroethene	<1.0		ug/L	70	70.00	1.0	8260B	Total/NA
Dibromochloromethane	<2.0		ug/L	80		2.0	8260B	Total/NA
Ethylbenzene	<1.0		ug/L	700	700.0	1.0	8260B	Total/NA
Ethylene Dibromide	<1.0		ug/L	0.05		1.0	8260B	Total/NA
Methylene Chloride	<10		ug/L	5	5.000	10	8260B	Total/NA
Styrene	<1.0		ug/L	100	100.0	1.0	8260B	Total/NA
Tetrachloroethene	<1.0		ug/L	5	5.000	1.0	8260B	Total/NA
Toluene	<1.0		ug/L	1000	1000	1.0	8260B	Total/NA
trans-1,2-Dichloroethene	<1.0		ug/L	100	100.0	1.0	8260B	Total/NA
Trichloroethene	<1.0		ug/L	5	5.000	1.0	8260B	Total/NA
Vinyl chloride	<0.50		ug/L	2	2.000	0.50	8260B	Total/NA
Benzo[a]pyrene	<0.19		ug/L	0.2		0.19	625.1 SIM	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8260B - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		DCA (80-120)	BFB (80-120)	DBFM (80-120)	TOL (80-120)
380-195028-1	BWS2253-H1-AQ	101	99	103	98
LCS 570-688946/1004	Lab Control Sample	99	102	101	99
LCSD 570-688946/6	Lab Control Sample Dup	98	100	101	98
MB 570-688946/9	Method Blank	100	98	102	97

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)
BFB = 4-Bromofluorobenzene (Surr)
DBFM = Dibromofluoromethane (Surr)
TOL = Toluene-d8 (Surr)

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
380-195028-1	BWS2253-H1-AQ	74	86	59	75	29	88
MB 570-689280/1-A	Method Blank	75	78	58	72	30	84

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)
PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-195028-1	BWS2253-H1-AQ	90	84	45	80	29	77
LCS 570-689280/2-A	Lab Control Sample	98	87	63	91	44	94
LCSD 570-689280/3-A	Lab Control Sample Dup	81	76	57	80	39	84
MB 570-689280/1-A	Method Blank	85	73	48	77	32	74

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)
PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Surrogate Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-195028-1	BWS2253-H1-AQ	94
380-195028-2	BWS2253-H1-TB	94
380-195033-C-1 MS	Matrix Spike	97
380-195033-C-1 MSD	Matrix Spike Duplicate	95
LCS 570-692331/7	Lab Control Sample	86
LCSD 570-692331/4	Lab Control Sample Dup	97
MB 570-692331/6	Method Blank	94
MRL 570-692331/5	Lab Control Sample	94

Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-194998-C-1-A MS	Matrix Spike	111
380-194998-D-1-A MSD	Matrix Spike Duplicate	97
380-195023-B-1-A MSD	Matrix Spike Duplicate	95
380-195023-C-1-A MS	Matrix Spike	102
380-195028-1	BWS2253-H1-AQ	114
380-195033-B-1-B MS	Matrix Spike	103
380-195033-B-1-C MSD	Matrix Spike Duplicate	105
LCS 570-689963/2-A	Lab Control Sample	99
LCSD 570-689963/3-A	Lab Control Sample Dup	105
MB 570-689963/1-A	Method Blank	105
MRL 570-689963/4-A	Lab Control Sample	94

Surrogate Legend

OTCSN = n-Octacosane (Surr)

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8260B - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 570-688946/9

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<2.0		2.0	ug/L			01/29/26 21:21	1
1,1,1-Trichloroethane	<1.0		1.0	ug/L			01/29/26 21:21	1
1,1,2,2-Tetrachloroethane	<1.0		1.0	ug/L			01/29/26 21:21	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<10		10	ug/L			01/29/26 21:21	1
1,1,2-Trichloroethane	<1.0		1.0	ug/L			01/29/26 21:21	1
1,1-Dichloroethane	<1.0		1.0	ug/L			01/29/26 21:21	1
1,1-Dichloroethene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,1-Dichloropropene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,2,3-Trichlorobenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,2,3-Trichloropropane	<5.0		5.0	ug/L			01/29/26 21:21	1
1,2,4-Trichlorobenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,2,4-Trimethylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,2-Dibromo-3-Chloropropane	<10		10	ug/L			01/29/26 21:21	1
1,2-Dichlorobenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,2-Dichloroethane	<0.50		0.50	ug/L			01/29/26 21:21	1
1,2-Dichloropropane	<1.0		1.0	ug/L			01/29/26 21:21	1
1,3,5-Trimethylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,3-Dichlorobenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
1,3-Dichloropropane	<1.0		1.0	ug/L			01/29/26 21:21	1
1,4-Dichlorobenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
2,2-Dichloropropane	<1.0		1.0	ug/L			01/29/26 21:21	1
2-Butanone	<10		10	ug/L			01/29/26 21:21	1
2-Chlorotoluene	<1.0		1.0	ug/L			01/29/26 21:21	1
2-Hexanone	<10		10	ug/L			01/29/26 21:21	1
4-Chlorotoluene	<1.0		1.0	ug/L			01/29/26 21:21	1
Acetone	<10		10	ug/L			01/29/26 21:21	1
Benzene	<0.50		0.50	ug/L			01/29/26 21:21	1
Bromobenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
Bromochloromethane	<2.0		2.0	ug/L			01/29/26 21:21	1
Bromodichloromethane	<1.0		1.0	ug/L			01/29/26 21:21	1
Bromoform	<5.0		5.0	ug/L			01/29/26 21:21	1
Bromomethane	<10		10	ug/L			01/29/26 21:21	1
Carbon disulfide	<10		10	ug/L			01/29/26 21:21	1
Carbon tetrachloride	<0.50		0.50	ug/L			01/29/26 21:21	1
Chlorobenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
Chloroethane	<5.0		5.0	ug/L			01/29/26 21:21	1
Chloroform	<1.0		1.0	ug/L			01/29/26 21:21	1
Chloromethane	<10		10	ug/L			01/29/26 21:21	1
cis-1,2-Dichloroethene	<1.0		1.0	ug/L			01/29/26 21:21	1
cis-1,3-Dichloropropene	<0.50		0.50	ug/L			01/29/26 21:21	1
Dibromochloromethane	<2.0		2.0	ug/L			01/29/26 21:21	1
Dibromomethane	<1.0		1.0	ug/L			01/29/26 21:21	1
Dichlorodifluoromethane	<5.0		5.0	ug/L			01/29/26 21:21	1
Diethyl ether	<10		10	ug/L			01/29/26 21:21	1
Di-isopropyl ether (DIPE)	<2.0		2.0	ug/L			01/29/26 21:21	1
Ethanol	<100		100	ug/L			01/29/26 21:21	1
Ethylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
Ethylene Dibromide	<1.0		1.0	ug/L			01/29/26 21:21	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 570-688946/9

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Ethyl-t-butyl ether (ETBE)	<2.0		2.0	ug/L			01/29/26 21:21	1
Hexachloro-1,3-butadiene	<10		10	ug/L			01/29/26 21:21	1
Hexane	<5.0		5.0	ug/L			01/29/26 21:21	1
Isobutyl alcohol	<75		75	ug/L			01/29/26 21:21	1
Isopropanol	<75		75	ug/L			01/29/26 21:21	1
Isopropylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
m,p-Xylene	<2.0		2.0	ug/L			01/29/26 21:21	1
Methylene Chloride	<10		10	ug/L			01/29/26 21:21	1
Methyl-t-Butyl Ether (MTBE)	<1.0		1.0	ug/L			01/29/26 21:21	1
MIBK	<10		10	ug/L			01/29/26 21:21	1
Naphthalene	<10		10	ug/L			01/29/26 21:21	1
n-Butylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
N-Propylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
o-Xylene	<1.0		1.0	ug/L			01/29/26 21:21	1
p-Isopropyltoluene	<1.0		1.0	ug/L			01/29/26 21:21	1
sec-Butylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
Styrene	<1.0		1.0	ug/L			01/29/26 21:21	1
tert-Amyl alcohol	<50		50	ug/L			01/29/26 21:21	1
Tert-amyl-methyl ether (TAME)	<2.0		2.0	ug/L			01/29/26 21:21	1
tert-Butyl alcohol (TBA)	<10		10	ug/L			01/29/26 21:21	1
tert-Butylbenzene	<1.0		1.0	ug/L			01/29/26 21:21	1
Tetrachloroethene	<1.0		1.0	ug/L			01/29/26 21:21	1
Toluene	<1.0		1.0	ug/L			01/29/26 21:21	1
trans-1,2-Dichloroethene	<1.0		1.0	ug/L			01/29/26 21:21	1
trans-1,3-Dichloropropene	<0.50		0.50	ug/L			01/29/26 21:21	1
Trichloroethene	<1.0		1.0	ug/L			01/29/26 21:21	1
Trichlorofluoromethane	<10		10	ug/L			01/29/26 21:21	1
Vinyl acetate	<10		10	ug/L			01/29/26 21:21	1
Vinyl chloride	<0.50		0.50	ug/L			01/29/26 21:21	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	100		80 - 120		01/29/26 21:21	1
4-Bromofluorobenzene (Surr)	98		80 - 120		01/29/26 21:21	1
Dibromofluoromethane (Surr)	102		80 - 120		01/29/26 21:21	1
Toluene-d8 (Surr)	97		80 - 120		01/29/26 21:21	1

Lab Sample ID: LCS 570-688946/1004

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1,1,1,2-Tetrachloroethane	20.0	19.8		ug/L		99	83 - 120
1,1,1-Trichloroethane	20.0	19.8		ug/L		99	78 - 123
1,1,2,2-Tetrachloroethane	20.0	16.6		ug/L		83	74 - 125
1,1,2-Trichloro-1,2,2-trifluoroethane	20.0	20.0		ug/L		100	72 - 130
1,1,2-Trichloroethane	20.0	18.6		ug/L		93	82 - 120
1,1-Dichloroethane	20.0	20.5		ug/L		102	79 - 122
1,1-Dichloroethene	20.0	20.4		ug/L		102	74 - 127

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 570-688946/1004

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1,1-Dichloropropene	20.0	20.4		ug/L		102	75 - 124
1,2,3-Trichlorobenzene	20.0	15.5		ug/L		77	71 - 126
1,2,3-Trichloropropane	20.0	18.6		ug/L		93	73 - 121
1,2,4-Trichlorobenzene	20.0	16.1		ug/L		81	70 - 130
1,2,4-Trimethylbenzene	20.0	18.1		ug/L		90	80 - 125
1,2-Dibromo-3-Chloropropane	20.0	15.1		ug/L		76	64 - 127
1,2-Dichlorobenzene	20.0	18.2		ug/L		91	78 - 122
1,2-Dichloroethane	20.0	19.8		ug/L		99	76 - 123
1,2-Dichloropropane	20.0	20.7		ug/L		104	82 - 120
1,3,5-Trimethylbenzene	20.0	19.7		ug/L		98	79 - 124
1,3-Dichlorobenzene	20.0	17.7		ug/L		88	81 - 120
1,3-Dichloropropane	20.0	19.0		ug/L		95	82 - 120
1,4-Dichlorobenzene	20.0	17.4		ug/L		87	78 - 121
2,2-Dichloropropane	20.0	20.3		ug/L		102	69 - 133
2-Butanone	20.0	22.8		ug/L		114	55 - 142
2-Chlorotoluene	20.0	19.6		ug/L		98	80 - 120
2-Hexanone	20.0	20.7		ug/L		103	54 - 138
4-Chlorotoluene	20.0	17.9		ug/L		89	80 - 121
Acetone	20.0	20.5		ug/L		102	50 - 157
Benzene	20.0	20.1		ug/L		100	80 - 120
Bromobenzene	20.0	19.8		ug/L		99	80 - 120
Bromochloromethane	20.0	19.9		ug/L		99	83 - 123
Bromodichloromethane	20.0	19.9		ug/L		99	83 - 120
Bromoform	20.0	18.3		ug/L		91	75 - 128
Bromomethane	20.0	20.8		ug/L		104	31 - 120
Carbon disulfide	20.0	19.6		ug/L		98	71 - 128
Carbon tetrachloride	20.0	20.5		ug/L		102	72 - 135
Chlorobenzene	20.0	18.9		ug/L		95	84 - 120
Chloroethane	20.0	19.0		ug/L		95	72 - 129
Chloroform	20.0	19.1		ug/L		96	79 - 120
Chloromethane	20.0	23.4		ug/L		117	36 - 148
cis-1,2-Dichloroethene	20.0	20.1		ug/L		101	81 - 120
cis-1,3-Dichloropropene	20.0	19.5		ug/L		98	76 - 125
Dibromochloromethane	20.0	19.4		ug/L		97	82 - 120
Dibromomethane	20.0	19.2		ug/L		96	85 - 120
Dichlorodifluoromethane	20.0	19.1		ug/L		96	52 - 147
Diethyl ether	20.0	20.5		ug/L		103	71 - 127
Di-isopropyl ether (DIPE)	20.0	24.8		ug/L		124	70 - 132
Ethanol	200	267		ug/L		134	48 - 163
Ethylbenzene	20.0	19.4		ug/L		97	81 - 120
Ethylene Dibromide	20.0	19.4		ug/L		97	81 - 120
Ethyl-t-butyl ether (ETBE)	20.0	21.4		ug/L		107	68 - 126
Isopropylbenzene	20.0	19.7		ug/L		99	80 - 122
m,p-Xylene	40.0	39.4		ug/L		99	79 - 122
Methylene Chloride	20.0	18.8		ug/L		94	76 - 120
Methyl-t-Butyl Ether (MTBE)	20.0	19.6		ug/L		98	70 - 124
MIBK	20.0	19.5		ug/L		97	57 - 133
Naphthalene	20.0	13.2		ug/L		66	64 - 124
n-Butylbenzene	20.0	17.6		ug/L		88	78 - 127

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 570-688946/1004

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
N-Propylbenzene	20.0	19.5		ug/L		97	80 - 121
o-Xylene	20.0	19.7		ug/L		98	81 - 120
p-Isopropyltoluene	20.0	18.2		ug/L		91	81 - 124
sec-Butylbenzene	20.0	17.6		ug/L		88	80 - 123
Styrene	20.0	19.8		ug/L		99	82 - 120
tert-Amyl alcohol	100	115		ug/L		115	30 - 157
Tert-amyl-methyl ether (TAME)	20.0	19.3		ug/L		97	71 - 121
tert-Butyl alcohol (TBA)	100	101		ug/L		101	68 - 128
tert-Butylbenzene	20.0	17.5		ug/L		87	80 - 124
Tetrachloroethene	20.0	19.3		ug/L		96	78 - 126
Toluene	20.0	19.7		ug/L		98	82 - 120
trans-1,2-Dichloroethene	20.0	20.9		ug/L		104	79 - 122
trans-1,3-Dichloropropene	20.0	20.0		ug/L		100	73 - 127
Trichloroethene	20.0	19.8		ug/L		99	83 - 120
Trichlorofluoromethane	20.0	21.2		ug/L		106	74 - 135
Vinyl acetate	20.0	18.3		ug/L		91	56 - 135
Vinyl chloride	20.0	22.6		ug/L		113	68 - 131

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	99		80 - 120
4-Bromofluorobenzene (Surr)	102		80 - 120
Dibromofluoromethane (Surr)	101		80 - 120
Toluene-d8 (Surr)	99		80 - 120

Lab Sample ID: LCSD 570-688946/6

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1,1,1,2-Tetrachloroethane	20.0	20.3		ug/L		102	83 - 120	3	20
1,1,1-Trichloroethane	20.0	20.7		ug/L		103	78 - 123	4	20
1,1,2,2-Tetrachloroethane	20.0	17.3		ug/L		86	74 - 125	4	20
1,1,2-Trichloro-1,2,2-trifluoroethane	20.0	21.8		ug/L		109	72 - 130	9	20
1,1,2-Trichloroethane	20.0	19.3		ug/L		97	82 - 120	4	20
1,1-Dichloroethane	20.0	21.5		ug/L		108	79 - 122	5	20
1,1-Dichloroethene	20.0	22.0		ug/L		110	74 - 127	8	20
1,1-Dichloropropene	20.0	21.6		ug/L		108	75 - 124	6	20
1,2,3-Trichlorobenzene	20.0	15.4		ug/L		77	71 - 126	1	20
1,2,3-Trichloropropane	20.0	19.2		ug/L		96	73 - 121	3	20
1,2,4-Trichlorobenzene	20.0	16.5		ug/L		83	70 - 130	3	20
1,2,4-Trimethylbenzene	20.0	19.1		ug/L		96	80 - 125	6	20
1,2-Dibromo-3-Chloropropane	20.0	16.0		ug/L		80	64 - 127	6	20
1,2-Dichlorobenzene	20.0	19.4		ug/L		97	78 - 122	6	20
1,2-Dichloroethane	20.0	20.2		ug/L		101	76 - 123	2	20
1,2-Dichloropropane	20.0	21.3		ug/L		106	82 - 120	3	20
1,3,5-Trimethylbenzene	20.0	20.5		ug/L		103	79 - 124	4	20
1,3-Dichlorobenzene	20.0	18.7		ug/L		93	81 - 120	6	20
1,3-Dichloropropane	20.0	19.6		ug/L		98	82 - 120	3	20

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCSD 570-688946/6

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1,4-Dichlorobenzene	20.0	18.3		ug/L		91	78 - 121	5	20
2,2-Dichloropropane	20.0	21.4		ug/L		107	69 - 133	5	20
2-Butanone	20.0	23.0		ug/L		115	55 - 142	1	21
2-Chlorotoluene	20.0	20.5		ug/L		102	80 - 120	4	20
2-Hexanone	20.0	21.1		ug/L		106	54 - 138	2	21
4-Chlorotoluene	20.0	18.8		ug/L		94	80 - 121	5	20
Acetone	20.0	21.2		ug/L		106	50 - 157	3	26
Benzene	20.0	20.9		ug/L		105	80 - 120	4	20
Bromobenzene	20.0	20.4		ug/L		102	80 - 120	3	20
Bromochloromethane	20.0	20.4		ug/L		102	83 - 123	3	20
Bromodichloromethane	20.0	20.1		ug/L		101	83 - 120	1	20
Bromoform	20.0	19.1		ug/L		96	75 - 128	5	20
Bromomethane	20.0	21.4		ug/L		107	31 - 120	3	22
Carbon disulfide	20.0	21.3		ug/L		106	71 - 128	8	20
Carbon tetrachloride	20.0	22.0		ug/L		110	72 - 135	7	20
Chlorobenzene	20.0	19.8		ug/L		99	84 - 120	5	20
Chloroethane	20.0	19.4		ug/L		97	72 - 129	2	20
Chloroform	20.0	19.7		ug/L		98	79 - 120	3	20
Chloromethane	20.0	23.1		ug/L		116	36 - 148	1	20
cis-1,2-Dichloroethene	20.0	20.4		ug/L		102	81 - 120	2	20
cis-1,3-Dichloropropene	20.0	20.0		ug/L		100	76 - 125	2	20
Dibromochloromethane	20.0	19.8		ug/L		99	82 - 120	2	20
Dibromomethane	20.0	20.1		ug/L		101	85 - 120	5	20
Dichlorodifluoromethane	20.0	18.8		ug/L		94	52 - 147	2	20
Diethyl ether	20.0	20.7		ug/L		104	71 - 127	1	20
Di-isopropyl ether (DIPE)	20.0	25.0		ug/L		125	70 - 132	1	20
Ethanol	200	260		ug/L		130	48 - 163	3	30
Ethylbenzene	20.0	20.2		ug/L		101	81 - 120	4	20
Ethylene Dibromide	20.0	19.7		ug/L		99	81 - 120	1	20
Ethyl-t-butyl ether (ETBE)	20.0	21.7		ug/L		109	68 - 126	1	20
Isopropylbenzene	20.0	20.8		ug/L		104	80 - 122	5	20
m,p-Xylene	40.0	41.4		ug/L		104	79 - 122	5	20
Methylene Chloride	20.0	19.3		ug/L		96	76 - 120	3	20
Methyl-t-Butyl Ether (MTBE)	20.0	19.7		ug/L		98	70 - 124	0	20
MIBK	20.0	19.0		ug/L		95	57 - 133	3	20
Naphthalene	20.0	13.2		ug/L		66	64 - 124	0	20
n-Butylbenzene	20.0	18.7		ug/L		94	78 - 127	6	20
N-Propylbenzene	20.0	20.6		ug/L		103	80 - 121	6	20
o-Xylene	20.0	20.3		ug/L		101	81 - 120	3	20
p-Isopropyltoluene	20.0	19.6		ug/L		98	81 - 124	7	20
sec-Butylbenzene	20.0	19.2		ug/L		96	80 - 123	9	20
Styrene	20.0	20.6		ug/L		103	82 - 120	4	20
tert-Amyl alcohol	100	118		ug/L		118	30 - 157	3	30
Tert-amyl-methyl ether (TAME)	20.0	19.6		ug/L		98	71 - 121	2	20
tert-Butyl alcohol (TBA)	100	101		ug/L		101	68 - 128	0	20
tert-Butylbenzene	20.0	19.0		ug/L		95	80 - 124	8	20
Tetrachloroethene	20.0	21.0		ug/L		105	78 - 126	8	20
Toluene	20.0	20.2		ug/L		101	82 - 120	3	20
trans-1,2-Dichloroethene	20.0	21.3		ug/L		107	79 - 122	2	20

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCSD 570-688946/6

Matrix: Water

Analysis Batch: 688946

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
trans-1,3-Dichloropropene	20.0	19.9		ug/L		100	73 - 127	0	20
Trichloroethene	20.0	21.1		ug/L		106	83 - 120	7	20
Trichlorofluoromethane	20.0	21.4		ug/L		107	74 - 135	1	20
Vinyl acetate	20.0	18.5		ug/L		93	56 - 135	1	24
Vinyl chloride	20.0	22.2		ug/L		111	68 - 131	2	20

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	98		80 - 120
4-Bromofluorobenzene (Surr)	100		80 - 120
Dibromofluoromethane (Surr)	101		80 - 120
Toluene-d8 (Surr)	98		80 - 120

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 570-689280/1-A

Matrix: Water

Analysis Batch: 691696

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 689280

Tentatively Identified Compound	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	01/30/26 12:01	02/05/26 08:34	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	75		33 - 139	01/30/26 12:01	02/05/26 08:34	1
2-Fluorobiphenyl (Surr)	78		33 - 126	01/30/26 12:01	02/05/26 08:34	1
2-Fluorophenol (Surr)	58		12 - 120	01/30/26 12:01	02/05/26 08:34	1
Nitrobenzene-d5 (Surr)	72		36 - 120	01/30/26 12:01	02/05/26 08:34	1
Phenol-d6 (Surr)	30		10 - 120	01/30/26 12:01	02/05/26 08:34	1
p-Terphenyl-d14 (Surr)	84		47 - 131	01/30/26 12:01	02/05/26 08:34	1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Lab Sample ID: MB 570-689280/1-A

Matrix: Water

Analysis Batch: 691132

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 689280

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
2-Methylnaphthalene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Acenaphthene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Acenaphthylene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Anthracene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Benzo[a]anthracene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Benzo[a]pyrene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Benzo[b]fluoranthene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Benzo[g,h,i]perylene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Benzo[k]fluoranthene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Chrysene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Dibenz(a,h)anthracene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: MB 570-689280/1-A

Matrix: Water

Analysis Batch: 691132

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 689280

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluoranthene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Fluorene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Indeno[1,2,3-cd]pyrene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Naphthalene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Phenanthrene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1
Pyrene	<0.20		0.20	ug/L		01/30/26 12:01	02/04/26 07:57	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	85		28 - 127	01/30/26 12:01	02/04/26 07:57	1
2-Fluorobiphenyl (Surr)	73		31 - 120	01/30/26 12:01	02/04/26 07:57	1
2-Fluorophenol (Surr)	48		17 - 120	01/30/26 12:01	02/04/26 07:57	1
Nitrobenzene-d5 (Surr)	77		27 - 120	01/30/26 12:01	02/04/26 07:57	1
Phenol-d6 (Surr)	32		10 - 120	01/30/26 12:01	02/04/26 07:57	1
p-Terphenyl-d14 (Surr)	74		45 - 120	01/30/26 12:01	02/04/26 07:57	1

Lab Sample ID: LCS 570-689280/2-A

Matrix: Water

Analysis Batch: 691132

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 689280

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	20.0	18.7		ug/L		93	47 - 120
2-Methylnaphthalene	20.0	18.0		ug/L		90	43 - 120
Acenaphthene	20.0	18.1		ug/L		91	60 - 132
Acenaphthylene	20.0	18.5		ug/L		93	54 - 126
Anthracene	20.0	18.0		ug/L		90	43 - 120
Benzo[a]anthracene	20.0	18.8		ug/L		94	42 - 133
Benzo[a]pyrene	20.0	19.2		ug/L		96	32 - 148
Benzo[b]fluoranthene	20.0	19.0		ug/L		95	42 - 140
Benzo[g,h,i]perylene	20.0	17.3		ug/L		87	1 - 195
Benzo[k]fluoranthene	20.0	18.2		ug/L		91	25 - 146
Chrysene	20.0	18.2		ug/L		91	44 - 140
Dibenz(a,h)anthracene	20.0	18.2		ug/L		91	1 - 200
Fluoranthene	20.0	18.6		ug/L		93	43 - 121
Fluorene	20.0	18.3		ug/L		92	70 - 120
Indeno[1,2,3-cd]pyrene	20.0	18.2		ug/L		91	1 - 151
Naphthalene	20.0	18.0		ug/L		90	36 - 120
Phenanthrene	20.0	17.9		ug/L		89	65 - 120
Pyrene	20.0	19.5		ug/L		98	70 - 120

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2,4,6-Tribromophenol (Surr)	98		28 - 127
2-Fluorobiphenyl (Surr)	87		31 - 120
2-Fluorophenol (Surr)	63		17 - 120
Nitrobenzene-d5 (Surr)	91		27 - 120
Phenol-d6 (Surr)	44		10 - 120
p-Terphenyl-d14 (Surr)	94		45 - 120

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCSD 570-689280/3-A
Matrix: Water
Analysis Batch: 691132

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 689280

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	20.0	16.7		ug/L		83	47 - 120	11	20
2-Methylnaphthalene	20.0	16.5		ug/L		83	43 - 120	8	20
Acenaphthene	20.0	16.0		ug/L		80	60 - 132	12	29
Acenaphthylene	20.0	16.5		ug/L		83	54 - 126	11	45
Anthracene	20.0	16.3		ug/L		81	43 - 120	10	40
Benzo[a]anthracene	20.0	16.9		ug/L		85	42 - 133	10	32
Benzo[a]pyrene	20.0	17.0		ug/L		85	32 - 148	12	43
Benzo[b]fluoranthene	20.0	16.7		ug/L		84	42 - 140	13	43
Benzo[g,h,i]perylene	20.0	15.7		ug/L		79	1 - 195	9	61
Benzo[k]fluoranthene	20.0	16.8		ug/L		84	25 - 146	8	38
Chrysene	20.0	16.7		ug/L		83	44 - 140	9	53
Dibenz(a,h)anthracene	20.0	16.7		ug/L		83	1 - 200	9	75
Fluoranthene	20.0	16.8		ug/L		84	43 - 121	10	40
Fluorene	20.0	16.4		ug/L		82	70 - 120	11	23
Indeno[1,2,3-cd]pyrene	20.0	16.5		ug/L		83	1 - 151	10	60
Naphthalene	20.0	16.1		ug/L		80	36 - 120	11	39
Phenanthrene	20.0	16.1		ug/L		80	65 - 120	10	24
Pyrene	20.0	17.7		ug/L		89	70 - 120	10	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
2,4,6-Tribromophenol (Surr)	81		28 - 127
2-Fluorobiphenyl (Surr)	76		31 - 120
2-Fluorophenol (Surr)	57		17 - 120
Nitrobenzene-d5 (Surr)	80		27 - 120
Phenol-d6 (Surr)	39		10 - 120
p-Terphenyl-d14 (Surr)	84		45 - 120

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Lab Sample ID: MB 570-692331/6
Matrix: Water
Analysis Batch: 692331

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
GRO (C6-C10)	<10		10	ug/L			02/06/26 11:46	1
Surrogate	MB	MB	Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
4-Bromofluorobenzene (Surr)	94		38 - 134				02/06/26 11:46	1

Lab Sample ID: LCS 570-692331/7
Matrix: Water
Analysis Batch: 692331

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	400	392		ug/L		98	78 - 120

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

Lab Sample ID: LCS 570-692331/7

Matrix: Water

Analysis Batch: 692331

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

	LCS	LCS	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	86		38 - 134

Lab Sample ID: LCSD 570-692331/4

Matrix: Water

Analysis Batch: 692331

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Analysis Data - 02/09/20											
Analyte			Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limits
Gasoline Range Organics (C4-C13)			400	427		ug/L		107	78 - 120	9	10

	LCSD	LCSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	97		38 - 134

Lab Sample ID: MRL 570-692331/5

Matrix: Water

Analysis Batch: 692331

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analysis Data: 06/06/2017											
Analyte			Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits		
Gasoline Range Organics (C4-C13)			10.0	8.27	J	ug/L		83	50 - 150		

	MRL	MRL	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	94		38 - 134

Lab Sample ID: 380-195033-C-1 MS

Matrix: Water

Analysis Batch: 692331

Client Sample ID: Matrix Spike

Prep Type: Total/NA

	Sample	Sample	Spike	MS	MS				%Rec		
Analyte	Result	Qualifier	Added	Result	Qualifier	Unit	D	%Rec	Limits		
Gasoline Range Organics (C4-C13)	<10		400	415		ug/L		104	68 - 122		

	MS	MS	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	97		38 - 134

Lab Sample ID: 380-195033-C-1 MSD

Matrix: Water

Analysis Batch: 692331

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

	Sample	Sample	Spike	MSD	MSD				%Rec		RPD
Analyte	Result	Qualifier	Added	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit
Gasoline Range Organics (C4-C13)	<10		400	408		ug/L		102	68 - 122	2	18

	MSD	MSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	95		38 - 134

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Lab Sample ID: MB 570-689963/1-A

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 689963

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<25		25	ug/L		02/02/26 09:35	02/06/26 14:57	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		02/02/26 09:35	02/06/26 14:57	1
C8-C18	<25		25	ug/L		02/02/26 09:35	02/06/26 14:57	1
Surrogate	MB %Recovery	MB Qualifier	Limits			Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	105		60 - 130			02/02/26 09:35	02/06/26 14:57	1

Lab Sample ID: LCS 570-689963/2-A

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits	
C10-C28	1600	1400		ug/L		88	56 - 127	
Surrogate	LCS %Recovery	LCS Qualifier	Limits					
n-Octacosane (Surr)	99		60 - 130					

Lab Sample ID: LCSD 570-689963/3-A

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	1600	1470		ug/L		92	56 - 127	5	23
Surrogate	LCSD %Recovery	LCSD Qualifier	Limits						
n-Octacosane (Surr)	105		60 - 130						

Lab Sample ID: MRL 570-689963/4-A

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits	
C10-C28	0.0200	<0.020		mg/L		96	50 - 150	
Surrogate	MRL %Recovery	MRL Qualifier	Limits					
n-Octacosane (Surr)	94		60 - 130					

Lab Sample ID: 380-194998-C-1-A MS

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	<26		1640	1720		ug/L		105	70 - 130
Surrogate	MS %Recovery	MS Qualifier	Limits						
n-Octacosane (Surr)	111		60 - 130						

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

Lab Sample ID: 380-194998-D-1-A MSD

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	<26		1650	1530		ug/L		93	70 - 130	12	20
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
n-Octacosane (Surr)	97		60 - 130								

Lab Sample ID: 380-195023-B-1-A MSD

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	<26		1640	1500		ug/L		92	70 - 130	4	20
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
n-Octacosane (Surr)	95		60 - 130								

Lab Sample ID: 380-195023-C-1-A MS

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits		
C10-C28	<26		1610	1570		ug/L		98	70 - 130		
Surrogate	MS %Recovery	MS Qualifier	Limits								
n-Octacosane (Surr)	102		60 - 130								

Lab Sample ID: 380-195033-B-1-B MS

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits		
C10-C28	<26		1670	1630		ug/L		97	70 - 130		
Surrogate	MS %Recovery	MS Qualifier	Limits								
n-Octacosane (Surr)	103		60 - 130								

Lab Sample ID: 380-195033-B-1-C MSD

Matrix: Water

Analysis Batch: 692930

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 689963

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	<26		1650	1620		ug/L		98	70 - 130	0	20
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
n-Octacosane (Surr)	105		60 - 130								

QC Association Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

GC/MS VOA

Analysis Batch: 688946

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195028-1	BWS2253-H1-AQ	Total/NA	Water	8260B	
MB 570-688946/9	Method Blank	Total/NA	Water	8260B	
LCS 570-688946/1004	Lab Control Sample	Total/NA	Water	8260B	
LCSD 570-688946/6	Lab Control Sample Dup	Total/NA	Water	8260B	

GC/MS Semi VOA

Prep Batch: 689280

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195028-1	BWS2253-H1-AQ	Total/NA	Water	625.1	
MB 570-689280/1-A	Method Blank	Total/NA	Water	625.1	
LCS 570-689280/2-A	Lab Control Sample	Total/NA	Water	625.1	
LCSD 570-689280/3-A	Lab Control Sample Dup	Total/NA	Water	625.1	

Analysis Batch: 691132

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195028-1	BWS2253-H1-AQ	Total/NA	Water	625.1 SIM	689280
MB 570-689280/1-A	Method Blank	Total/NA	Water	625.1 SIM	689280
LCS 570-689280/2-A	Lab Control Sample	Total/NA	Water	625.1 SIM	689280
LCSD 570-689280/3-A	Lab Control Sample Dup	Total/NA	Water	625.1 SIM	689280

Analysis Batch: 691696

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195028-1	BWS2253-H1-AQ	Total/NA	Water	625.1	689280
MB 570-689280/1-A	Method Blank	Total/NA	Water	625.1	689280

GC VOA

Analysis Batch: 692331

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195028-1	BWS2253-H1-AQ	Total/NA	Water	8015B GRO LL	
380-195028-2	BWS2253-H1-TB	Total/NA	Water	8015B GRO LL	
MB 570-692331/6	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-692331/7	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-692331/4	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	
MRL 570-692331/5	Lab Control Sample	Total/NA	Water	8015B GRO LL	
380-195033-C-1 MS	Matrix Spike	Total/NA	Water	8015B GRO LL	
380-195033-C-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8015B GRO LL	

GC Semi VOA

Prep Batch: 689963

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195028-1	BWS2253-H1-AQ	Total/NA	Water	3510C	
MB 570-689963/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-689963/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-689963/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-689963/4-A	Lab Control Sample	Total/NA	Water	3510C	
380-194998-C-1-A MS	Matrix Spike	Total/NA	Water	3510C	
380-194998-D-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	3510C	
380-195023-B-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	3510C	
380-195023-C-1-A MS	Matrix Spike	Total/NA	Water	3510C	

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

GC Semi VOA (Continued)

Prep Batch: 689963 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195033-B-1-B MS	Matrix Spike	Total/NA	Water	3510C	
380-195033-B-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	3510C	

Analysis Batch: 692930

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-195028-1	BWS2253-H1-AQ	Total/NA	Water	8015B	689963
MB 570-689963/1-A	Method Blank	Total/NA	Water	8015B	689963
LCS 570-689963/2-A	Lab Control Sample	Total/NA	Water	8015B	689963
LCSD 570-689963/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	689963
MRL 570-689963/4-A	Lab Control Sample	Total/NA	Water	8015B	689963
380-194998-C-1-A MS	Matrix Spike	Total/NA	Water	8015B	689963
380-194998-D-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	8015B	689963
380-195023-B-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	8015B	689963
380-195023-C-1-A MS	Matrix Spike	Total/NA	Water	8015B	689963
380-195033-B-1-B MS	Matrix Spike	Total/NA	Water	8015B	689963
380-195033-B-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	8015B	689963

Lab Chronicle

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Client Sample ID: BWS2253-H1-AQ

Lab Sample ID: 380-195028-1

Date Collected: 01/28/26 09:30

Matrix: Water

Date Received: 01/29/26 09:52

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260B		1	688946	PT	EET CAL 4	01/30/26 01:58
Total/NA	Prep	625.1			689280	OAJ3	EET CAL 4	01/30/26 12:01
Total/NA	Analysis	625.1		1	691696	PQS1	EET CAL 4	02/05/26 12:36
Total/NA	Prep	625.1			689280	OAJ3	EET CAL 4	01/30/26 12:01
Total/NA	Analysis	625.1 SIM		1	691132	PQS1	EET CAL 4	02/04/26 13:46
Total/NA	Analysis	8015B GRO LL		1	692331	A9VE	EET CAL 4	02/06/26 13:43
Total/NA	Prep	3510C			689963	TVD6	EET CAL 4	02/02/26 09:35
Total/NA	Analysis	8015B		1	692930	H6FE	EET CAL 4	02/06/26 22:12

Client Sample ID: BWS2253-H1-TB

Lab Sample ID: 380-195028-2

Date Collected: 01/28/26 09:30

Matrix: Water

Date Received: 01/29/26 09:52

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	692331	A9VE	EET CAL 4	02/06/26 18:39

Laboratory References:

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
A2LA	Dept. of Defense ELAP	7296.01	11-30-26
A2LA	ISO/IEC 17025	7296.01	11-30-26
Alaska (UST)	State	25-005	03-02-26
Arizona	State	AZ0830	11-17-26
California	Los Angeles County Sanitation Districts	9257304	07-31-26
California	SCAQMD LAP	17LA0919	11-30-26
California	State	3082	07-31-26
Kansas	NELAP	E-10420	07-31-26
Nevada	State	CA00111	07-31-26
Oregon	NELAP	4175	02-02-27
USDA	US Federal Programs	525-23-159-97150	06-08-26
Utah	NELAP	CA00111	02-28-26
Washington	State	C916	02-11-26

Method Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Method	Method Description	Protocol	Laboratory
8260B	Volatile Organic Compounds (GC/MS)	SW846	EET CAL 4
625.1	Semivolatile Organic Compounds (GC/MS)	EPA	EET CAL 4
625.1 SIM	Semivolatile Organic Compounds GC/MS (SIM)	EPA	EET CAL 4
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4
625.1	Liquid-Liquid Extraction	40CFR136A	EET CAL 4

Protocol References:

40CFR136A = "Methods for Organic Chemical Analysis of Municipal Industrial Wastewater", 40CFR, Part 136, Appendix A, October 26, 1984 and subsequent revisions.

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Sample Summary

Client: City & County of Honolulu
Project/Site: INTERA - Red-Hill-Incident

Job ID: 380-195028-1
SDG: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-195028-1	BWS2253-H1-AQ	Water	01/28/26 09:30	01/29/26 09:52	Hawaii
380-195028-2	BWS2253-H1-TB	Water	01/28/26 09:30	01/29/26 09:52	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

380-195028 Chain of Custody

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-195028-1

SDG Number: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Login Number: 195028

List Number: 1

Creator: Ngo, Theodore

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-195028-1

SDG Number: Site H1 625.1, 8015B_DRO/_GRO, 8260B

Login Number: 195028

List Number: 2

Creator: Khana, Piyush

List Source: Eurofins Calscience

List Creation: 01/29/26 07:05 PM

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	1.7
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	fgf5
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	