eurofins | Environment Testing

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 10/12/2026 3:39:29 PM

JOB DESCRIPTION

RUSH Weekly Red Hill
PFAS: Ka'amilo Wells P2

JOB NUMBER

380-240706-1

Eurofins Pomona
941 Corporate Center Drive
Pomona CA 91768-2642

See page two for job notes and contact information.

my EOL

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW,Water matrices)

Authorization



Generated
10/12/2026 3:39:29 PM

Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1145

Table of Contents

Cover Page 1
Table of Contents 3
Definitions/Glossary 4
Case Narrative 5
Detection Summary 6
Client Sample Results 7
Action Limit Summary 11
Surrogate Summary 12
Isotope Dilution Summary 13
QC Sample Results 14
QC Association Summary 31
Lab Chronicle 33
Certification Summary 34
Method Summary 35
Sample Summary 36
Chain of Custody 37
Receipt Checklists 38

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Job Narrative
380-240706-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt
The samples were received on 10/7/2026 9:45 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 5.0°C.

PFAS
No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-1

PWSID Number: HI0000331

Analyte	Result	Qualifier	RL		Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid (PFBS)	2.7		2.0		ng/L	1		533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		2.0		ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	3.9		2.0		ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.5		2.0		ng/L	1		533	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		2.0		ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	3.7		2.0		ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.8		2.0		ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanoic acid (PFHxA)	4.0		2.0		ng/L	1		EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		2.0		ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		2.0		ng/L	1		EPA 537.1 V2	Total/NA
Perfluorobutanesulfonic acid (PFBS)	3.0		2.0		ng/L	1		EPA 537.1 V2	Total/NA

Client Sample ID: FB KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-2

PWSID Number: HI0000331

No Detections.

This Detection Summary does not include radiochemical test results.

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: KA'AMILO WELLS P2 (331-600-WL085)
Date Collected: 10/05/26 11:00
Date Received: 10/07/26 09:45

Lab Sample ID: 380-240706-1
Matrix: Water
PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorobutanesulfonic acid (PFBS)	**2.7**		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorohexanesulfonic acid (PFHxS)	**3.6**		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorohexanoic acid (PFHxA)	**3.9**		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorononanoic acid (PFNA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorooctanesulfonic acid (PFOS)	**4.5**		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorooctanoic acid (PFOA)	**3.7**		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoropentanoic acid (PFPeA)	**3.7**		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0		ng/L		10/09/26 07:22	10/10/26 02:30	1

Isotope Dilution	*%Recovery*	*Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
13C3 HFPO-DA	*102*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C6 PFDA	*106*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C5 PFHxA	*103*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C4 PFHpA	*102*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C8 PFOA	*105*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C9 PFNA	*108*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C7 PFUnA	*101*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C2 PFDoA	*100*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C4 PFBA	*106*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C5 PFPeA	*106*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C3 PFBS	*105*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*

Client Sample ID: KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-1

Date Collected: 10/05/26 11:00
Matrix: Water
Date Received: 10/07/26 09:45
PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 PFHxS	*102*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C8 PFOS	*107*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C2-4:2-FTS	*101*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C2-6:2-FTS	*100*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*
13C2-8:2-FTS	*111*		*50 - 200*	*10/09/26 07:22*	*10/10/26 02:30*	*1*

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorooctanesulfonic acid (PFOS)	**4.8**		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorohexanoic acid (PFHxA)	**4.0**		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorooctanoic acid (PFOA)	**3.7**		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorohexanesulfonic acid (PFHxS)	**3.6**		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorobutanesulfonic acid (PFBS)	**3.0**		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorononanoic acid (PFNA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:13	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	*102*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:13*	*1*
13C2 PFHxA	*102*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:13*	*1*
13C2 PFDA	*104*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:13*	*1*
13C3-GenX	*100*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:13*	*1*

Client Sample ID: FB KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-2

Date Collected: 10/05/26 11:00
Matrix: Water
Date Received: 10/07/26 09:45
PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: FB KA'AMILO WELLS P2 (331-600-WL085)
Date Collected: 10/05/26 11:00
Date Received: 10/07/26 09:45

Lab Sample ID: 380-240706-2
Matrix: Water
PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorononanoic acid (PFNA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0		ng/L		10/09/26 16:54	10/10/26 22:46	1

Isotope Dilution	*%Recovery*	*Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
13C3 HFPO-DA	*95*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C6 PFDA	*105*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C5 PFHxA	*99*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C4 PFHpA	*99*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C8 PFOA	*99*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C9 PFNA	*106*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C7 PFUnA	*108*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C2 PFDoA	*106*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C4 PFBA	*103*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C5 PFPeA	*103*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C3 PFBS	*102*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C3 PFHxS	*98*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C8 PFOS	*110*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C2-4:2-FTS	*91*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C2-6:2-FTS	*90*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*
13C2-8:2-FTS	*104*		*50 - 200*	*10/09/26 16:54*	*10/10/26 22:46*	*1*

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: FB KA'AMILO WELLS P2 (331-600-WL085)
Date Collected: 10/05/26 11:00
Date Received: 10/07/26 09:45

Lab Sample ID: 380-240706-2
Matrix: Water
PWSID Number: HI0000331

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorononanoic acid (PFNA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0		ng/L		10/09/26 08:00	10/10/26 05:23	1

Surrogate	*%Recovery*	*Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
d5-NEtFOSAA	*91*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:23*	*1*
13C2 PFHxA	*96*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:23*	*1*
13C2 PFDA	*103*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:23*	*1*
13C3-GenX	*89*		*70 - 130*	*10/09/26 08:00*	*10/10/26 05:23*	*1*

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-1

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	**4.5**		ng/L	**4**	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	**4.8**		ng/L	**4**	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-2

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water **Prep Type: Total/NA**

		Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS	PFHxA	PFDA	GenX
Lab Sample ID	Client Sample ID	(70-130)	(70-130)	(70-130)	(70-130)
380-240618-E-1-A MS	Matrix Spike	91	99	104	97
380-240618-F-1-A MSD	Matrix Spike Duplicate	98	108	106	102
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	102	102	104	100
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	91	96	103	89
LCS 380-257406/22-A	Lab Control Sample	90	99	101	88
MBL 380-257406/20-A	Method Blank	98	104	109	91
MRL 380-257406/21-A	Lab Control Sample	95	105	109	94

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
PFHxA = 13C2 PFHxA
PFDA = 13C2 PFDA
GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water **Prep Type: Total/NA**

		Percent Isotope Dilution Recovery (Acceptance Limits)							
		HFPODA	C6PFDA	13C5PHA	C4PFHA	C8PFOA	C9PFNA	13C7PUA	PFDoA
Lab Sample ID	Client Sample ID	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)
380-240406-B-8-A MS	Matrix Spike	104	107	108	107	106	108	106	107
380-240406-C-8-A MSD	Matrix Spike Duplicate	102	104	103	105	104	105	106	106
380-240694-B-1-A MS	Matrix Spike	90	81	92	86	86	84	84	90
380-240694-C-1-A MSD	Matrix Spike Duplicate	86	70	90	82	78	75	71	75
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	102	106	103	102	105	108	101	100
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	95	105	99	99	99	106	108	106
LCS 380-257407/22-A	Lab Control Sample	103	106	109	110	108	110	105	105
LCS 380-257561/22-A	Lab Control Sample	103	107	99	100	102	105	110	106
MBL 380-257407/20-A	Method Blank	96	102	107	102	105	101	103	101
MBL 380-257561/20-A	Method Blank	82	91	83	84	88	89	94	92
MRL 380-257407/21-A	Lab Control Sample	99	107	109	106	106	109	105	104
MRL 380-257561/21-A	Lab Control Sample	90	99	91	91	94	96	103	101

		Percent Isotope Dilution Recovery (Acceptance Limits)							
		PFBA	PFPeA	C3PFBS	C3PFHS	C8PFOS	42FTS	62FTS	82FTS
Lab Sample ID	Client Sample ID	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)	(50-200)
380-240406-B-8-A MS	Matrix Spike	107	102	106	104	105	111	108	110
380-240406-C-8-A MSD	Matrix Spike Duplicate	105	100	105	104	105	112	104	110
380-240694-B-1-A MS	Matrix Spike	98	94	107	99	117	96	91	109
380-240694-C-1-A MSD	Matrix Spike Duplicate	95	91	101	98	108	92	89	104
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	106	106	105	102	107	101	100	111
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	103	103	102	98	110	91	90	104
LCS 380-257407/22-A	Lab Control Sample	106	105	106	107	108	104	103	111
LCS 380-257561/22-A	Lab Control Sample	104	103	98	98	108	87	89	102
MBL 380-257407/20-A	Method Blank	106	106	103	105	104	101	108	107
MBL 380-257561/20-A	Method Blank	94	93	98	100	106	89	86	106
MRL 380-257407/21-A	Lab Control Sample	107	107	106	105	109	100	102	109
MRL 380-257561/21-A	Lab Control Sample	95	97	99	98	107	90	83	103

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDoA = 13C2 PFDoA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-257407/20-A
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 257407

Analyte	MBL Result	MBL Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorononanoic acid (PFNA)	<0.39		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluorobutanoic acid (PFBA)	<0.68		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.36		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.47		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.46		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.24		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.45		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoropentanoic acid (PFPeA)	<0.37		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.35		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1
Perfluoropentanesulfonic acid (PFPeS)	<0.38		2.0		ng/L		10/09/26 07:22	10/09/26 20:53	1

Isotope Dilution	*MBL %Recovery*	*MBL Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
13C3 HFPO-DA	*96*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C6 PFDA	*102*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C5 PFHxA	*107*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C4 PFHpA	*102*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C8 PFOA	*105*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C9 PFNA	*101*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C7 PFUnA	*103*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C2 PFDoA	*101*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C4 PFBA	*106*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C5 PFPeA	*106*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C3 PFBS	*103*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*
13C3 PFHxS	*105*		*50 - 200*	*10/09/26 07:22*	*10/09/26 20:53*	*1*

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-257407/20-A
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 257407

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	104		50 - 200	10/09/26 07:22	10/09/26 20:53	1
13C2-4:2-FTS	101		50 - 200	10/09/26 07:22	10/09/26 20:53	1
13C2-6:2-FTS	108		50 - 200	10/09/26 07:22	10/09/26 20:53	1
13C2-8:2-FTS	107		50 - 200	10/09/26 07:22	10/09/26 20:53	1

Lab Sample ID: LCS 380-257407/22-A
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257407

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	118	103		ng/L		88	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	118	103		ng/L		87	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	118	108		ng/L		91	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	118	106		ng/L		90	70 - 130
Perfluorobutanesulfonic acid (PFBS)	118	110		ng/L		93	70 - 130
Perfluorodecanoic acid (PFDA)	118	112		ng/L		95	70 - 130
Perfluorododecanoic acid (PFDoA)	118	110		ng/L		93	70 - 130
Perfluoroheptanoic acid (PFHpA)	118	102		ng/L		86	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	118	106		ng/L		90	70 - 130
Perfluorohexanoic acid (PFHxA)	118	105		ng/L		89	70 - 130
Perfluorononanoic acid (PFNA)	118	106		ng/L		90	70 - 130
Perfluorooctanesulfonic acid (PFOS)	118	106		ng/L		90	70 - 130
Perfluorooctanoic acid (PFOA)	118	107		ng/L		91	70 - 130
Perfluoroundecanoic acid (PFUnA)	118	113		ng/L		95	70 - 130
Perfluorobutanoic acid (PFBA)	118	108		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	118	107		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	118	103		ng/L		87	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	118	106		ng/L		89	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	118	102		ng/L		86	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	118	108		ng/L		92	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	118	103		ng/L		87	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	118	105		ng/L		89	70 - 130
Perfluoropentanoic acid (PFPeA)	118	105		ng/L		89	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	118	111		ng/L		94	70 - 130

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-257407/22-A
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257407

Analyte			Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits		
Perfluoropentanesulfonic acid (PFPeS)			118	103		ng/L		87	70 - 130		

Isotope Dilution	*LCS %Recovery*	*LCS Qualifier*	*Limits*
13C3 HFPO-DA	*103*		*50 - 200*
13C6 PFDA	*106*		*50 - 200*
13C5 PFHxA	*109*		*50 - 200*
13C4 PFHpA	*110*		*50 - 200*
13C8 PFOA	*108*		*50 - 200*
13C9 PFNA	*110*		*50 - 200*
13C7 PFUnA	*105*		*50 - 200*
13C2 PFDoA	*105*		*50 - 200*
13C4 PFBA	*106*		*50 - 200*
13C5 PFPeA	*105*		*50 - 200*
13C3 PFBS	*106*		*50 - 200*
13C3 PFHxS	*107*		*50 - 200*
13C8 PFOS	*108*		*50 - 200*
13C2-4:2-FTS	*104*		*50 - 200*
13C2-6:2-FTS	*103*		*50 - 200*
13C2-8:2-FTS	*111*		*50 - 200*

Lab Sample ID: MRL 380-257407/21-A
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257407

Analyte			Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits		
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)			1.97	1.75	J	ng/L		89	50 - 150		
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)			1.97	1.85	J	ng/L		94	50 - 150		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)			1.97	1.85	J	ng/L		94	50 - 150		
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)			1.97	1.94	J	ng/L		98	50 - 150		
Perfluorobutanesulfonic acid (PFBS)			1.97	1.86	J	ng/L		94	50 - 150		
Perfluorodecanoic acid (PFDA)			1.97	1.98	J	ng/L		101	50 - 150		
Perfluorododecanoic acid (PFDoA)			1.97	1.94	J	ng/L		98	50 - 150		
Perfluoroheptanoic acid (PFHpA)			1.97	1.95	J	ng/L		99	50 - 150		
Perfluorohexanesulfonic acid (PFHxS)			1.97	1.93	J	ng/L		98	50 - 150		
Perfluorohexanoic acid (PFHxA)			1.97	1.87	J	ng/L		95	50 - 150		
Perfluorononanoic acid (PFNA)			1.97	1.94	J	ng/L		99	50 - 150		
Perfluorooctanesulfonic acid (PFOS)			1.97	1.88	J	ng/L		96	50 - 150		
Perfluorooctanoic acid (PFOA)			1.97	1.99	J	ng/L		101	50 - 150		
Perfluoroundecanoic acid (PFUnA)			1.97	1.89	J	ng/L		96	50 - 150		
Perfluorobutanoic acid (PFBA)			1.97	1.84	J	ng/L		93	50 - 150		

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-257407/21-A
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257407

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	1.97	2.08	J	ng/L		106	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	1.97	2.10	J	ng/L		107	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	1.97	1.96	J	ng/L		99	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	1.97	1.82	J	ng/L		93	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	1.97	1.87	J	ng/L		95	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	1.97	1.83	J	ng/L		93	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	1.97	1.95	J	ng/L		99	50 - 150
Perfluoropentanoic acid (PFPeA)	1.97	1.74	J	ng/L		88	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	1.97	1.91	J	ng/L		97	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	1.97	1.76	J	ng/L		89	50 - 150

Isotope Dilution	*MRL %Recovery*	*MRL Qualifier*	*Limits*
13C3 HFPO-DA	*99*		*50 - 200*
13C6 PFDA	*107*		*50 - 200*
13C5 PFHxA	*109*		*50 - 200*
13C4 PFHpA	*106*		*50 - 200*
13C8 PFOA	*106*		*50 - 200*
13C9 PFNA	*109*		*50 - 200*
13C7 PFUnA	*105*		*50 - 200*
13C2 PFDoA	*104*		*50 - 200*
13C4 PFBA	*107*		*50 - 200*
13C5 PFPeA	*107*		*50 - 200*
13C3 PFBS	*106*		*50 - 200*
13C3 PFHxS	*105*		*50 - 200*
13C8 PFOS	*109*		*50 - 200*
13C2-4:2-FTS	*100*		*50 - 200*
13C2-6:2-FTS	*102*		*50 - 200*
13C2-8:2-FTS	*109*		*50 - 200*

Lab Sample ID: 380-240406-B-8-A MS
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 257407

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		118	104		ng/L		88	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		118	107		ng/L		91	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		118	106		ng/L		90	70 - 130

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-240406-B-8-A MS
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 257407

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		118	107		ng/L		90	70 - 130
Perfluorobutanesulfonic acid (PFBS)	<2.0		118	104		ng/L		88	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		118	110		ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		118	108		ng/L		91	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		118	106		ng/L		90	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		118	107		ng/L		90	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		118	103		ng/L		87	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		118	105		ng/L		89	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		118	107		ng/L		90	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		118	110		ng/L		93	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		118	112		ng/L		95	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		118	109		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		118	107		ng/L		90	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		118	104		ng/L		88	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		118	108		ng/L		91	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		118	103		ng/L		87	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		118	112		ng/L		95	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		118	98.6		ng/L		83	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		118	108		ng/L		91	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		118	105		ng/L		89	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		118	108		ng/L		92	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		118	104		ng/L		88	70 - 130

Isotope Dilution	*MS %Recovery*	*MS Qualifier*	*Limits*
13C3 HFPO-DA	*104*		*50 - 200*
13C6 PFDA	*107*		*50 - 200*
13C5 PFHxA	*108*		*50 - 200*
13C4 PFHpA	*107*		*50 - 200*
13C8 PFOA	*106*		*50 - 200*
13C9 PFNA	*108*		*50 - 200*
13C7 PFUnA	*106*		*50 - 200*
13C2 PFDoA	*107*		*50 - 200*
13C4 PFBA	*107*		*50 - 200*
13C5 PFPeA	*102*		*50 - 200*
13C3 PFBS	*106*		*50 - 200*
13C3 PFHxS	*104*		*50 - 200*
13C8 PFOS	*105*		*50 - 200*

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-240406-B-8-A MS
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 257407

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	111		50 - 200
13C2-6:2-FTS	108		50 - 200
13C2-8:2-FTS	110		50 - 200

Lab Sample ID: 380-240406-C-8-A MSD
Matrix: Water
Analysis Batch: 257583

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 257407

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		118	109		ng/L		92	70 - 130	5	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		118	108		ng/L		91	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		118	108		ng/L		91	70 - 130	2	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		118	109		ng/L		92	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		118	108		ng/L		91	70 - 130	4	30
Perfluorodecanoic acid (PFDA)	<2.0		118	116		ng/L		98	70 - 130	5	30
Perfluorododecanoic acid (PFDoA)	<2.0		118	110		ng/L		93	70 - 130	2	30
Perfluoroheptanoic acid (PFHpA)	<2.0		118	111		ng/L		94	70 - 130	5	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		118	110		ng/L		93	70 - 130	3	30
Perfluorohexanoic acid (PFHxA)	<2.0		118	109		ng/L		92	70 - 130	6	30
Perfluorononanoic acid (PFNA)	<2.0		118	111		ng/L		94	70 - 130	6	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		118	109		ng/L		92	70 - 130	3	30
Perfluorooctanoic acid (PFOA)	<2.0		118	111		ng/L		94	70 - 130	2	30
Perfluoroundecanoic acid (PFUnA)	<2.0		118	112		ng/L		94	70 - 130	0	30
Perfluorobutanoic acid (PFBA)	<2.0		118	113		ng/L		96	70 - 130	4	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		118	109		ng/L		92	70 - 130	3	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		118	102		ng/L		86	70 - 130	3	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		118	111		ng/L		94	70 - 130	3	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		118	108		ng/L		91	70 - 130	5	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		118	116		ng/L		98	70 - 130	3	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		118	103		ng/L		87	70 - 130	4	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		118	111		ng/L		94	70 - 130	2	30
Perfluoropentanoic acid (PFPeA)	<2.0		118	109		ng/L		92	70 - 130	4	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		118	110		ng/L		93	70 - 130	2	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		118	109		ng/L		92	70 - 130	4	30

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD %Recovery	MSD Qualifier	Limits
13C3 HFPO-DA	102		50 - 200
13C6 PFDA	104		50 - 200
13C5 PFHxA	103		50 - 200
13C4 PFHpA	105		50 - 200
13C8 PFOA	104		50 - 200
13C9 PFNA	105		50 - 200
13C7 PFUnA	106		50 - 200
13C2 PFDoA	106		50 - 200
13C4 PFBA	105		50 - 200
13C5 PFPeA	100		50 - 200
13C3 PFBS	105		50 - 200
13C3 PFHxS	104		50 - 200
13C8 PFOS	105		50 - 200
13C2-4:2-FTS	112		50 - 200
13C2-6:2-FTS	104		50 - 200
13C2-8:2-FTS	110		50 - 200

Lab Sample ID: MBL 380-257561/20-A
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 257561

Analyte	MBL Result	MBL Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorononanoic acid (PFNA)	<0.39		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluorobutanoic acid (PFBA)	<0.68		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.36		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.47		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.46		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.24		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.45		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-257561/20-A
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 257561

Analyte	MBL Result	MBL Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
Perfluoropentanoic acid (PFPeA)	<0.37		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.35		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1
Perfluoropentanesulfonic acid (PFPeS)	<0.38		2.0		ng/L		10/09/26 16:54	10/10/26 17:36	1

Isotope Dilution	*MBL %Recovery*	*MBL Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
13C3 HFPO-DA	*82*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C6 PFDA	*91*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C5 PFHxA	*83*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C4 PFHpA	*84*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C8 PFOA	*88*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C9 PFNA	*89*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C7 PFUnA	*94*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C2 PFDoA	*92*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C4 PFBA	*94*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C5 PFPeA	*93*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C3 PFBS	*98*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C3 PFHxS	*100*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C8 PFOS	*106*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C2-4:2-FTS	*89*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C2-6:2-FTS	*86*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*
13C2-8:2-FTS	*106*		*50 - 200*	*10/09/26 16:54*	*10/10/26 17:36*	*1*

Lab Sample ID: LCS 380-257561/22-A
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257561

Analyte			Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits		
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)			118	105		ng/L		89	70 - 130		
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)			118	111		ng/L		94	70 - 130		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)			118	122		ng/L		103	70 - 130		
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)			118	112		ng/L		95	70 - 130		
Perfluorobutanesulfonic acid (PFBS)			118	114		ng/L		97	70 - 130		
Perfluorodecanoic acid (PFDA)			118	119		ng/L		101	70 - 130		
Perfluorododecanoic acid (PFDoA)			118	112		ng/L		95	70 - 130		
Perfluoroheptanoic acid (PFHpA)			118	117		ng/L		99	70 - 130		
Perfluorohexanesulfonic acid (PFHxS)			118	112		ng/L		95	70 - 130		
Perfluorohexanoic acid (PFHxA)			118	119		ng/L		101	70 - 130		
Perfluorononanoic acid (PFNA)			118	116		ng/L		98	70 - 130		
Perfluorooctanesulfonic acid (PFOS)			118	109		ng/L		93	70 - 130		
Perfluorooctanoic acid (PFOA)			118	115		ng/L		98	70 - 130		

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-257561/22-A
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257561

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoroundecanoic acid (PFUnA)	118	116		ng/L		98	70 - 130
Perfluorobutanoic acid (PFBA)	118	115		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	118	113		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	118	108		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	118	110		ng/L		93	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	118	120		ng/L		102	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	118	118		ng/L		100	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	118	111		ng/L		94	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	118	112		ng/L		95	70 - 130
Perfluoropentanoic acid (PFPeA)	118	114		ng/L		96	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	118	106		ng/L		90	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	118	109		ng/L		93	70 - 130

Isotope Dilution	*LCS %Recovery*	*LCS Qualifier*	*Limits*
13C3 HFPO-DA	*103*		*50 - 200*
13C6 PFDA	*107*		*50 - 200*
13C5 PFHxA	*99*		*50 - 200*
13C4 PFHpA	*100*		*50 - 200*
13C8 PFOA	*102*		*50 - 200*
13C9 PFNA	*105*		*50 - 200*
13C7 PFUnA	*110*		*50 - 200*
13C2 PFDoA	*106*		*50 - 200*
13C4 PFBA	*104*		*50 - 200*
13C5 PFPeA	*103*		*50 - 200*
13C3 PFBS	*98*		*50 - 200*
13C3 PFHxS	*98*		*50 - 200*
13C8 PFOS	*108*		*50 - 200*
13C2-4:2-FTS	*87*		*50 - 200*
13C2-6:2-FTS	*89*		*50 - 200*
13C2-8:2-FTS	*102*		*50 - 200*

Lab Sample ID: MRL 380-257561/21-A
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257561

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.97	2.15	J	ng/L		109	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	1.97	2.12	J	ng/L		108	50 - 150

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-257561/21-A
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257561

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.97	2.01	J	ng/L		102	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.97	2.08	J	ng/L		106	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.97	1.96	J	ng/L		100	50 - 150
Perfluorodecanoic acid (PFDA)	1.97	2.22	J	ng/L		113	50 - 150
Perfluorododecanoic acid (PFDoA)	1.97	2.26	J	ng/L		115	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.97	2.20	J	ng/L		112	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.97	2.04	J	ng/L		103	50 - 150
Perfluorohexanoic acid (PFHxA)	1.97	2.07	J	ng/L		105	50 - 150
Perfluorononanoic acid (PFNA)	1.97	2.11	J	ng/L		107	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.97	2.09	J	ng/L		106	50 - 150
Perfluorooctanoic acid (PFOA)	1.97	1.99	J	ng/L		101	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.97	2.21	J	ng/L		112	50 - 150
Perfluorobutanoic acid (PFBA)	1.97	2.06	J	ng/L		105	50 - 150
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	1.97	2.08	J	ng/L		106	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	1.97	2.12	J	ng/L		108	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	1.97	2.37	J	ng/L		121	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	1.97	2.08	J	ng/L		106	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	1.97	2.29	J	ng/L		116	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	1.97	1.98	J	ng/L		101	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	1.97	1.94	J	ng/L		98	50 - 150
Perfluoropentanoic acid (PFPeA)	1.97	1.92	J	ng/L		98	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	1.97	1.88	J	ng/L		96	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	1.97	1.99	J	ng/L		101	50 - 150

Isotope Dilution	*MRL %Recovery*	*MRL Qualifier*	*Limits*
13C3 HFPO-DA	*90*		*50 - 200*
13C6 PFDA	*99*		*50 - 200*
13C5 PFHxA	*91*		*50 - 200*
13C4 PFHpA	*91*		*50 - 200*
13C8 PFOA	*94*		*50 - 200*
13C9 PFNA	*96*		*50 - 200*
13C7 PFUnA	*103*		*50 - 200*
13C2 PFDoA	*101*		*50 - 200*
13C4 PFBA	*95*		*50 - 200*
13C5 PFPeA	*97*		*50 - 200*
13C3 PFBS	*99*		*50 - 200*
13C3 PFHxS	*98*		*50 - 200*

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-257561/21-A
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257561

Isotope Dilution	*MRL %Recovery*	*MRL Qualifier*	*Limits*
13C8 PFOS	*107*		*50 - 200*
13C2-4:2-FTS	*90*		*50 - 200*
13C2-6:2-FTS	*83*		*50 - 200*
13C2-8:2-FTS	*103*		*50 - 200*

Lab Sample ID: 380-240694-B-1-A MS
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 257561

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		59.1	51.3		ng/L		87	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		59.1	53.9		ng/L		91	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		59.1	58.9		ng/L		100	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		59.1	58.9		ng/L		100	70 - 130
Perfluorobutanesulfonic acid (PFBS)	<2.0		59.1	57.6		ng/L		96	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0	*5-	59.1	58.7		ng/L		99	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0	*5-	59.1	58.7		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		59.1	60.1		ng/L		102	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	9.6		59.1	70.3		ng/L		103	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		59.1	58.7		ng/L		98	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		59.1	59.9		ng/L		101	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		59.1	56.4		ng/L		94	70 - 130
Perfluorooctanoic acid (PFOA)	4.2		59.1	63.8		ng/L		101	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0	*5-	59.1	59.0		ng/L		100	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		59.1	59.4		ng/L		101	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		59.1	58.3		ng/L		99	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		59.1	55.4		ng/L		94	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		59.1	57.0		ng/L		96	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		59.1	59.2		ng/L		100	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		59.1	66.7		ng/L		113	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		59.1	59.4		ng/L		101	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		59.1	57.8		ng/L		98	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		59.1	57.0		ng/L		96	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		59.1	52.7		ng/L		89	70 - 130

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-240694-B-1-A MS
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 257561

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	<2.0		59.1	59.3		ng/L		100	70 - 130

Isotope Dilution	*MS %Recovery*	*MS Qualifier*	*Limits*
13C3 HFPO-DA	*90*		*50 - 200*
13C6 PFDA	*81*		*50 - 200*
13C5 PFHxA	*92*		*50 - 200*
13C4 PFHpA	*86*		*50 - 200*
13C8 PFOA	*86*		*50 - 200*
13C9 PFNA	*84*		*50 - 200*
13C7 PFUnA	*84*		*50 - 200*
13C2 PFDoA	*90*		*50 - 200*
13C4 PFBA	*98*		*50 - 200*
13C5 PFPeA	*94*		*50 - 200*
13C3 PFBS	*107*		*50 - 200*
13C3 PFHxS	*99*		*50 - 200*
13C8 PFOS	*117*		*50 - 200*
13C2-4:2-FTS	*96*		*50 - 200*
13C2-6:2-FTS	*91*		*50 - 200*
13C2-8:2-FTS	*109*		*50 - 200*

Lab Sample ID: 380-240694-C-1-A MSD
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 257561

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		59.1	52.7		ng/L		89	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		59.1	55.8		ng/L		95	70 - 130	3	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		59.1	58.2		ng/L		99	70 - 130	1	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		59.1	59.0		ng/L		100	70 - 130	0	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		59.1	58.6		ng/L		98	70 - 130	2	30
Perfluorodecanoic acid (PFDA)	<2.0	*5-	59.1	59.8		ng/L		101	70 - 130	2	30
Perfluorododecanoic acid (PFDoA)	<2.0	*5-	59.1	59.0		ng/L		100	70 - 130	0	30
Perfluoroheptanoic acid (PFHpA)	<2.0		59.1	60.6		ng/L		103	70 - 130	1	30
Perfluorohexanesulfonic acid (PFHxS)	9.6		59.1	67.0		ng/L		97	70 - 130	5	30
Perfluorohexanoic acid (PFHxA)	<2.0		59.1	56.5		ng/L		94	70 - 130	4	30
Perfluorononanoic acid (PFNA)	<2.0		59.1	59.3		ng/L		100	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		59.1	57.4		ng/L		96	70 - 130	2	30
Perfluorooctanoic acid (PFOA)	4.2		59.1	64.1		ng/L		101	70 - 130	1	30
Perfluoroundecanoic acid (PFUnA)	<2.0	*5-	59.1	59.7		ng/L		101	70 - 130	1	30
Perfluorobutanoic acid (PFBA)	<2.0		59.1	59.3		ng/L		100	70 - 130	0	30

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-240694-C-1-A MSD
Matrix: Water
Analysis Batch: 257665

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 257561

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		59.1	58.6		ng/L		99	70 - 130	1	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		59.1	58.8		ng/L		100	70 - 130	6	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		59.1	57.3		ng/L		97	70 - 130	0	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		59.1	57.5		ng/L		97	70 - 130	3	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		59.1	66.2		ng/L		112	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		59.1	57.6		ng/L		97	70 - 130	3	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		59.1	57.6		ng/L		98	70 - 130	0	30
Perfluoropentanoic acid (PFPeA)	<2.0		59.1	56.0		ng/L		95	70 - 130	2	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		59.1	53.5		ng/L		91	70 - 130	2	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		59.1	55.1		ng/L		93	70 - 130	7	30

Isotope Dilution	*MSD %Recovery*	*MSD Qualifier*	*Limits*
13C3 HFPO-DA	*86*		*50 - 200*
13C6 PFDA	*70*		*50 - 200*
13C5 PFHxA	*90*		*50 - 200*
13C4 PFHpA	*82*		*50 - 200*
13C8 PFOA	*78*		*50 - 200*
13C9 PFNA	*75*		*50 - 200*
13C7 PFUnA	*71*		*50 - 200*
13C2 PFDoA	*75*		*50 - 200*
13C4 PFBA	*95*		*50 - 200*
13C5 PFPeA	*91*		*50 - 200*
13C3 PFBS	*101*		*50 - 200*
13C3 PFHxS	*98*		*50 - 200*
13C8 PFOS	*108*		*50 - 200*
13C2-4:2-FTS	*92*		*50 - 200*
13C2-6:2-FTS	*89*		*50 - 200*
13C2-8:2-FTS	*104*		*50 - 200*

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-257406/20-A
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 257406

Analyte	MBL Result	MBL Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.57		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-257406/20-A
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 257406

Analyte	MBL Result	MBL Qualifier	RL		Unit	D	Prepared	Analyzed	Dil Fac
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.41		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorodecanoic acid (PFDA)	<0.30		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorononanoic acid (PFNA)	<0.39		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorotetradecanoic acid (PFTA)	<0.53		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
Perfluorotridecanoic acid (PFTrDA)	<0.35		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.29		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0		ng/L		10/09/26 08:00	10/10/26 03:18	1

Surrogate	*MBL %Recovery*	*MBL Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
d5-NEtFOSAA	*98*		*70 - 130*	*10/09/26 08:00*	*10/10/26 03:18*	*1*
13C2 PFHxA	*104*		*70 - 130*	*10/09/26 08:00*	*10/10/26 03:18*	*1*
13C2 PFDA	*109*		*70 - 130*	*10/09/26 08:00*	*10/10/26 03:18*	*1*
13C3-GenX	*91*		*70 - 130*	*10/09/26 08:00*	*10/10/26 03:18*	*1*

Lab Sample ID: LCS 380-257406/22-A
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257406

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	49.1	47.6		ng/L		97	70 - 130
Perfluorooctanesulfonic acid (PFOS)	49.1	53.9		ng/L		110	70 - 130
Perfluoroundecanoic acid (PFUnA)	49.1	54.7		ng/L		111	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	49.1	50.4		ng/L		103	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	49.1	49.6		ng/L		101	70 - 130
Perfluorohexanoic acid (PFHxA)	49.1	49.7		ng/L		101	70 - 130
Perfluorododecanoic acid (PFDoA)	49.1	55.9		ng/L		114	70 - 130
Perfluorooctanoic acid (PFOA)	49.1	53.3		ng/L		109	70 - 130
Perfluorodecanoic acid (PFDA)	49.1	54.8		ng/L		111	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	49.1	52.2		ng/L		106	70 - 130
Perfluorobutanesulfonic acid (PFBS)	49.1	57.6		ng/L		117	70 - 130
Perfluoroheptanoic acid (PFHpA)	49.1	50.9		ng/L		104	70 - 130
Perfluorononanoic acid (PFNA)	49.1	54.4		ng/L		111	70 - 130

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: LCS 380-257406/22-A
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257406

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorotetradecanoic acid (PFTA)	49.1	49.3		ng/L		100	70 - 130
Perfluorotridecanoic acid (PFTrDA)	49.1	58.4		ng/L		119	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	49.1	53.6		ng/L		109	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	49.1	49.3		ng/L		100	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	49.1	46.6		ng/L		95	70 - 130

Surrogate	*LCS %Recovery*	*LCS Qualifier*	*Limits*
d5-NEtFOSAA	*90*		*70 - 130*
13C2 PFHxA	*99*		*70 - 130*
13C2 PFDA	*101*		*70 - 130*
13C3-GenX	*88*		*70 - 130*

Lab Sample ID: MRL 380-257406/21-A
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257406

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.96	2.04	J	ng/L		104	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.96	2.32	J	ng/L		118	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.96	2.33	J	ng/L		119	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	1.96	2.02	J	ng/L		103	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	1.96	2.07	J	ng/L		106	50 - 150
Perfluorohexanoic acid (PFHxA)	1.96	2.24	J	ng/L		114	50 - 150
Perfluorododecanoic acid (PFDoA)	1.96	2.39	J	ng/L		122	50 - 150
Perfluorooctanoic acid (PFOA)	1.96	2.33	J	ng/L		118	50 - 150
Perfluorodecanoic acid (PFDA)	1.96	2.42	J	ng/L		123	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.96	2.06	J	ng/L		105	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.96	2.17	J	ng/L		110	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.96	2.19	J	ng/L		112	50 - 150
Perfluorononanoic acid (PFNA)	1.96	2.33	J	ng/L		119	50 - 150
Perfluorotetradecanoic acid (PFTA)	1.96	2.05	J	ng/L		104	50 - 150
Perfluorotridecanoic acid (PFTrDA)	1.96	2.44	J	ng/L		124	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	1.96	2.20	J	ng/L		112	50 - 150

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-257406/21-A
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 257406

Analyte			Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits		
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)			1.96	2.06	J	ng/L		105	50 - 150		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)			1.96	1.98	J	ng/L		101	50 - 150		

Surrogate	*MRL %Recovery*	*MRL Qualifier*	*Limits*
d5-NEtFOSAA	*95*		*70 - 130*
13C2 PFHxA	*105*		*70 - 130*
13C2 PFDA	*109*		*70 - 130*
13C3-GenX	*94*		*70 - 130*

Lab Sample ID: 380-240618-E-1-A MS
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 257406

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits		
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		49.2	52.3		ng/L		106	70 - 130		
Perfluorooctanesulfonic acid (PFOS)	<2.0		49.2	55.9		ng/L		112	70 - 130		
Perfluoroundecanoic acid (PFUnA)	<2.0		49.2	53.9		ng/L		109	70 - 130		
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		49.2	54.2		ng/L		110	70 - 130		
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		49.2	50.7		ng/L		103	70 - 130		
Perfluorohexanoic acid (PFHxA)	2.0		49.2	55.6		ng/L		109	70 - 130		
Perfluorododecanoic acid (PFDoA)	<2.0		49.2	54.7		ng/L		111	70 - 130		
Perfluorooctanoic acid (PFOA)	<2.0		49.2	57.2		ng/L		114	70 - 130		
Perfluorodecanoic acid (PFDA)	<2.0		49.2	56.2		ng/L		114	70 - 130		
Perfluorohexanesulfonic acid (PFHxS)	<2.0		49.2	55.4		ng/L		109	70 - 130		
Perfluorobutanesulfonic acid (PFBS)	<2.0		49.2	56.9		ng/L		113	70 - 130		
Perfluoroheptanoic acid (PFHpA)	<2.0		49.2	55.1		ng/L		111	70 - 130		
Perfluorononanoic acid (PFNA)	<2.0		49.2	54.0		ng/L		110	70 - 130		
Perfluorotetradecanoic acid (PFTA)	<2.0		49.2	50.4		ng/L		102	70 - 130		
Perfluorotridecanoic acid (PFTrDA)	<2.0		49.2	59.3		ng/L		120	70 - 130		
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		49.2	53.7		ng/L		109	70 - 130		
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		49.2	52.0		ng/L		106	70 - 130		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		49.2	52.3		ng/L		106	70 - 130		

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-240618-E-1-A MS
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 257406

Surrogate	MS %Recovery	MS Qualifier	Limits
d5-NEtFOSAA	91		70 - 130
13C2 PFHxA	99		70 - 130
13C2 PFDA	104		70 - 130
13C3-GenX	97		70 - 130

Lab Sample ID: 380-240618-F-1-A MSD
Matrix: Water
Analysis Batch: 257582

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 257406

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		49.2	51.8		ng/L		105	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		49.2	56.2		ng/L		113	70 - 130	1	30
Perfluoroundecanoic acid (PFUnA)	<2.0		49.2	56.3		ng/L		114	70 - 130	4	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		49.2	53.9		ng/L		110	70 - 130	1	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		49.2	52.9		ng/L		108	70 - 130	4	30
Perfluorohexanoic acid (PFHxA)	2.0		49.2	57.6		ng/L		113	70 - 130	3	30
Perfluorododecanoic acid (PFDoA)	<2.0		49.2	57.3		ng/L		116	70 - 130	5	30
Perfluorooctanoic acid (PFOA)	<2.0		49.2	56.7		ng/L		113	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		49.2	56.4		ng/L		115	70 - 130	0	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		49.2	57.1		ng/L		113	70 - 130	3	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		49.2	56.7		ng/L		113	70 - 130	0	30
Perfluoroheptanoic acid (PFHpA)	<2.0		49.2	56.5		ng/L		113	70 - 130	2	30
Perfluorononanoic acid (PFNA)	<2.0		49.2	56.9		ng/L		116	70 - 130	5	30
Perfluorotetradecanoic acid (PFTA)	<2.0		49.2	50.0		ng/L		102	70 - 130	1	30
Perfluorotridecanoic acid (PFTrDA)	<2.0		49.2	61.4		ng/L		125	70 - 130	4	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		49.2	55.6		ng/L		113	70 - 130	3	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		49.2	53.5		ng/L		109	70 - 130	3	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		49.2	52.6		ng/L		107	70 - 130	1	30

Surrogate	MSD %Recovery	MSD Qualifier	Limits
d5-NEtFOSAA	98		70 - 130
13C2 PFHxA	108		70 - 130
13C2 PFDA	106		70 - 130
13C3-GenX	102		70 - 130

LCMS

Prep Batch: 257406

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	537.1 DW	
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	537.1 DW	
MBL 380-257406/20-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-257406/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-257406/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-240618-E-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-240618-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Prep Batch: 257407

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	533	
MBL 380-257407/20-A	Method Blank	Total/NA	Water	533	
LCS 380-257407/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-257407/21-A	Lab Control Sample	Total/NA	Water	533	
380-240406-B-8-A MS	Matrix Spike	Total/NA	Water	533	
380-240406-C-8-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Prep Batch: 257561

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	533	
MBL 380-257561/20-A	Method Blank	Total/NA	Water	533	
LCS 380-257561/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-257561/21-A	Lab Control Sample	Total/NA	Water	533	
380-240694-B-1-A MS	Matrix Spike	Total/NA	Water	533	
380-240694-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 257582

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	EPA 537.1 V2	257406
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	EPA 537.1 V2	257406
MBL 380-257406/20-A	Method Blank	Total/NA	Water	EPA 537.1 V2	257406
LCS 380-257406/22-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	257406
MRL 380-257406/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	257406
380-240618-E-1-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	257406
380-240618-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	257406

Analysis Batch: 257583

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	533	257407
MBL 380-257407/20-A	Method Blank	Total/NA	Water	533	257407
LCS 380-257407/22-A	Lab Control Sample	Total/NA	Water	533	257407
MRL 380-257407/21-A	Lab Control Sample	Total/NA	Water	533	257407
380-240406-B-8-A MS	Matrix Spike	Total/NA	Water	533	257407
380-240406-C-8-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	257407

Analysis Batch: 257665

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	Total/NA	Water	533	257561
MBL 380-257561/20-A	Method Blank	Total/NA	Water	533	257561
LCS 380-257561/22-A	Lab Control Sample	Total/NA	Water	533	257561
MRL 380-257561/21-A	Lab Control Sample	Total/NA	Water	533	257561

LCMS (Continued)

Analysis Batch: 257665 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-240694-B-1-A MS	Matrix Spike	Total/NA	Water	533	257561
380-240694-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	257561

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-1

Date Collected: 10/05/26 11:00
Matrix: Water
Date Received: 10/07/26 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			257407	XTD8	EA POM	10/09/26 07:22
Total/NA	Analysis	533		1	257583	Y5FM	EA POM	10/10/26 02:30
Total/NA	Prep	537.1 DW			257406	E9PK	EA POM	10/09/26 08:00
Total/NA	Analysis	EPA 537.1 V2		1	257582	Y5FM	EA POM	10/10/26 05:13

Client Sample ID: FB KA'AMILO WELLS P2 (331-600-WL085)

Lab Sample ID: 380-240706-2

Date Collected: 10/05/26 11:00
Matrix: Water
Date Received: 10/07/26 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			257561	N8NE	EA POM	10/09/26 16:54
Total/NA	Analysis	533		1	257665	UKDT	EA POM	10/10/26 22:46
Total/NA	Prep	537.1 DW			257406	E9PK	EA POM	10/09/26 08:00
Total/NA	Analysis	EPA 537.1 V2		1	257582	Y5FM	EA POM	10/10/26 05:23

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-27

Method Summary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RUSH Weekly Red Hill

Job ID: 380-240706-1
SDG: PFAS: Ka'amilo Wells P2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-240706-1	KA'AMILO WELLS P2 (331-600-WL085)	Water	10/05/26 11:00	10/07/26 09:45	HI0000331
380-240706-2	FB KA'AMILO WELLS P2 (331-600-WL085)	Water	10/05/26 11:00	10/07/26 09:45	HI0000331

Monrovia, CA (Suite 100)
750 Royal Oaks Drive Suite 100
Monrovia CA 91016
Phone (626) 386-1100

Chain of Custody Record

eurofins Environment Testing America

Client Information	Sampler: Allison Win	Lab PM: Lopez, Maria	Carrier Tracking No(s):	COC No:
Client Contact: Kirk Iwamoto	Phone: +1 808 748 5840	E-Mail: Maria.Lopez@et.eurofinsus.com	State of Origin:	Page: Page 1 of 1
Company: City & County of Honolulu	PWSID:			Job #: 380-240706 COC

Address: 630 South Beretania Street; Chemistry Lab	Due Date Requested	Analysis Requested
City: Honolulu	TAT Requested (days): RUSH	
State, Zip: HI 96843	Compliance Project: Δ No	
Phone: 808-748-5840 (tel)	PO #: C20525101 exp 05312023	
Email: kiwamoto@hbws.org	WO #:	
Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill	Project #: 38001111	
Site:	SSOW#:	

Preservation Codes.

A - HCL	M - Hexane
B - NaOH	N - None
C - Zn Acetate	O - AsNaO2
D - Nitric Acid	P - Na2O4S
E - NaHSO4	Q - Na2SO3
F - MeOH	R - Na2S2O3
G - Amchlor	S - H2SO4
H - Ascorbic Acid	T - TSP Dodecahydrate
I - Ice	U - Acetone
J - DI Water	V - MCAA
K - EDTA	W - pH 4-5
L - EDA	Y - Trizma
	Z - other (specify)
Other:	NH4 Acetate

Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water S=solid, O=waste/oil, BT=Tissue, A=Air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	537 1_DW_PREC - 537 1 Full List	533 - All Analytes	Total Number of containers	Special Instructions/Note:
			Preservation Code:				Y	Z		
KA'AMILO WELLS P2 (331-600-WL085)	10-05-26	11:00	G	Water			3	3		
FB KA'AMILO WELLS P2 (331-600-WL085)	10-05-26	11:00					1	1		

Possible Hazard Identification
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☐ Unknown ☐ Radiological

Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)
☐ Return To Client ☐ Disposal By Lab ☐ Archive For _____ Months

Deliverable Requested: I II III, IV Other (specify)

Special Instructions/QC Requirements

Empty Kit Relinquished by:	Date	Time:	Method of Shipment: 8782 7693 5745

Relinquished by: [redacted]	Date/Time: 10-06-26 12:00	Company: HBWS	Received by: Yuko	Date/Time: 10/7/26 9:45	Company: FedEx
Relinquished by:	Date/Time:	Company	Received by:	Date/Time:	Company
Relinquished by:	Date/Time:	Company	Received by:	Date/Time:	Company

Custody Seals Intact: Δ Yes Δ No	Custody Seal No.	Cooler Temperature(s) °C and Other Remarks: 6.0 - 0.1 = 5.9 Rel

Ver: 01/16/2019

10/12/2026

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-240706-1
SDG Number: PFAS: Ka'amilo Wells P2

Login Number: 240706
List Number: 1
Creator: Abaya, Ashton

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
ClO4 headspace requirement met (>50% for CA, >30% for other states).	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	