

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 6/23/2026 5:46:07 PM

JOB DESCRIPTION

RED-HILL
PFAS: Moanalua Wells

JOB NUMBER

380-220757-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
6/23/2026 5:46:07 PM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-220757-1

Job ID: 380-220757-1

Eurofins Pomona

Job Narrative 380-220757-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 6/19/2026 9:55 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.2°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-1

☐ No Detections.

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-2

☐ No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-1

Date Collected: 06/17/26 11:00

Matrix: Drinking Water

Date Received: 06/19/26 09:55

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9CI-PF3ONS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:31	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	107		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C6 PFDA	111		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C5 PFHxA	104		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C4 PFHpA	114		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C8 PFOA	119		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C9 PFNA	126		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C7 PFUnA	123		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C2 PFDoA	112		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C4 PFBA	119		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C5 PFPeA	108		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C3 PFBS	106		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C3 PFHxS	113		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C8 PFOS	116		50 - 200	06/22/26 15:55	06/23/26 11:31	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-1

Date Collected: 06/17/26 11:00

Matrix: Drinking Water

Date Received: 06/19/26 09:55

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2-4:2-FTS	95		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C2-6:2-FTS	114		50 - 200	06/22/26 15:55	06/23/26 11:31	1
13C2-8:2-FTS	135		50 - 200	06/22/26 15:55	06/23/26 11:31	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:33	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	103		70 - 130			06/22/26 09:45	06/22/26 18:33	1
13C2 PFHxA	98		70 - 130			06/22/26 09:45	06/22/26 18:33	1
13C2 PFDA	95		70 - 130			06/22/26 09:45	06/22/26 18:33	1
13C3-GenX	89		70 - 130			06/22/26 09:45	06/22/26 18:33	1

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-2

Date Collected: 06/17/26 11:00

Matrix: Water

Date Received: 06/19/26 09:55

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-2

Date Collected: 06/17/26 11:00

Matrix: Water

Date Received: 06/19/26 09:55

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		06/22/26 15:55	06/23/26 11:40	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	101		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C6 PFDA	110		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C5 PFHxA	115		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C4 PFHpA	106		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C8 PFOA	116		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C9 PFNA	122		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C7 PFUnA	114		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C2 PFDoA	109		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C4 PFBA	117		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C5 PFPeA	107		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C3 PFBS	109		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C3 PFHxS	112		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C8 PFOS	116		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C2-4:2-FTS	105		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C2-6:2-FTS	119		50 - 200	06/22/26 15:55	06/23/26 11:40	1
13C2-8:2-FTS	136		50 - 200	06/22/26 15:55	06/23/26 11:40	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-2

Date Collected: 06/17/26 11:00

Matrix: Water

Date Received: 06/19/26 09:55

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		06/22/26 09:45	06/22/26 19:59	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	96		70 - 130			06/22/26 09:45	06/22/26 19:59	1
13C2 PFHxA	103		70 - 130			06/22/26 09:45	06/22/26 19:59	1
13C2 PFDA	98		70 - 130			06/22/26 09:45	06/22/26 19:59	1
13C3-GenX	92		70 - 130			06/22/26 09:45	06/22/26 19:59	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-220757-1	MOANALUA WELLS (331-223-T	103	98	95	89
380-220757-1 MS	MOANALUA WELLS (331-223-TP202)	101	100	100	98
380-220757-1 MSD	MOANALUA WELLS (331-223-TP202)	99	104	101	97
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-220757-2	FB: MOANALUA WELLS (331-2	96	103	98	92
LCS 380-235371/21-A	Lab Control Sample	95	93	98	86
MBL 380-235371/19-A	Method Blank	100	101	99	94
MRL 380-235371/20-A	Lab Control Sample	102	105	98	90
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-220757-1	MOANALUA WELLS (331-223-T	107	111	104	114	119	126	123	112

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-220757-1	MOANALUA WELLS (331-223-T	119	108	106	113	116	95	114	135

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDaA = 13C2 PFDaA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-220625-B-1-A MS	Matrix Spike	108	107	107	109	111	119	114	110
380-220625-C-1-A MSD	Matrix Spike Duplicate	103	102	98	99	107	113	112	105
380-220757-2	FB: MOANALUA WELLS (331-223-TP202)	101	110	115	106	116	122	114	109
LCS 380-235471/22-A	Lab Control Sample	101	102	101	98	103	115	114	104
MBL 380-235471/20-A	Method Blank	109	111	115	117	123	135	121	111
MRL 380-235471/21-A	Lab Control Sample	123	109	128	129	138	136	118	106

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-220625-B-1-A MS	Matrix Spike	114	109	104	107	113	88	107	124
380-220625-C-1-A MSD	Matrix Spike Duplicate	107	103	97	105	109	86	107	122
380-220757-2	FB: MOANALUA WELLS (331-223-TP202)	117	107	109	112	116	105	119	136
LCS 380-235471/22-A	Lab Control Sample	105	99	99	103	110	83	103	117
MBL 380-235471/20-A	Method Blank	123	113	116	119	127	109	128	137
MRL 380-235471/21-A	Lab Control Sample	132	125	126	130	128	119	139	134

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA

Eurofins Pomona

Isotope Dilution Summary

Client: City & County of Honolulu

Project/Site: RED-HILL

13C5PHA = 13C5 PFHxA

C4PFHA = 13C4 PFHpA

C8PFOA = 13C8 PFOA

C9PFNA = 13C9 PFNA

13C7PUA = 13C7 PFUnA

PFDoA = 13C2 PFDoA

PFBA = 13C4 PFBA

PFPeA = 13C5 PFPeA

C3PFBS = 13C3 PFBS

C3PFHS = 13C3 PFHxS

C8PFOS = 13C8 PFOS

42FTS = 13C2-4:2-FTS

62FTS = 13C2-6:2-FTS

82FTS = 13C2-8:2-FTS

Job ID: 380-220757-1

SDG: PFAS: Moanalua Wells

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

16

17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-235471/20-A

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 235471

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		06/22/26 15:55	06/23/26 07:48	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	109		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C6 PFDA	111		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C5 PFHxA	115		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C4 PFHpA	117		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C8 PFOA	123		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C9 PFNA	135		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C7 PFUnA	121		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C2 PFDoA	111		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C4 PFBA	123		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C5 PFPeA	113		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C3 PFBS	116		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C3 PFHxS	119		50 - 200	06/22/26 15:55	06/23/26 07:48	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-235471/20-A

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 235471

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	127		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C2-4:2-FTS	109		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C2-6:2-FTS	128		50 - 200	06/22/26 15:55	06/23/26 07:48	1
13C2-8:2-FTS	137		50 - 200	06/22/26 15:55	06/23/26 07:48	1

Lab Sample ID: LCS 380-235471/22-A

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 235471

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	120	97.6		ng/L		81	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	120	104		ng/L		86	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	114		ng/L		95	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	107		ng/L		89	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	103		ng/L		86	70 - 130
Perfluorodecanoic acid (PFDA)	120	114		ng/L		95	70 - 130
Perfluorododecanoic acid (PFDoA)	120	109		ng/L		91	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	109		ng/L		91	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	106		ng/L		88	70 - 130
Perfluorohexanoic acid (PFHxA)	120	105		ng/L		88	70 - 130
Perfluorononanoic acid (PFNA)	120	103		ng/L		85	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	103		ng/L		85	70 - 130
Perfluorooctanoic acid (PFOA)	120	110		ng/L		91	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	104		ng/L		86	70 - 130
Perfluorobutanoic acid (PFBA)	120	104		ng/L		86	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	105		ng/L		87	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	108		ng/L		90	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	104		ng/L		86	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	103		ng/L		85	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	120	112		ng/L		93	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	109		ng/L		91	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	105		ng/L		87	70 - 130
Perfluoropentanoic acid (PFPeA)	120	108		ng/L		90	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	104		ng/L		86	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-235471/22-A

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 235471

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	105		ng/L		87	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	101		50 - 200				
13C6 PFDA	102		50 - 200				
13C5 PFHxA	101		50 - 200				
13C4 PFHpA	98		50 - 200				
13C8 PFOA	103		50 - 200				
13C9 PFNA	115		50 - 200				
13C7 PFUnA	114		50 - 200				
13C2 PFDoA	104		50 - 200				
13C4 PFBA	105		50 - 200				
13C5 PFPeA	99		50 - 200				
13C3 PFBS	99		50 - 200				
13C3 PFHxS	103		50 - 200				
13C8 PFOS	110		50 - 200				
13C2-4:2-FTS	83		50 - 200				
13C2-6:2-FTS	103		50 - 200				
13C2-8:2-FTS	117		50 - 200				

Lab Sample ID: MRL 380-235471/21-A

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 235471

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.57	J	ng/L		78	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	2.01	1.75	J	ng/L		87	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.87	J	ng/L		93	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.94	J	ng/L		96	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.92	J	ng/L		96	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.03	J	ng/L		101	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.94	J	ng/L		97	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	1.91	J	ng/L		95	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	1.85	J	ng/L		92	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	1.91	J	ng/L		95	50 - 150
Perfluorononanoic acid (PFNA)	2.01	1.84	J	ng/L		92	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.77	J	ng/L		88	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	1.92	J	ng/L		95	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.77	J	ng/L		88	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	2.06	J	ng/L		102	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-235471/21-A

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 235471

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	1.86	J	ng/L		92	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	1.99	J	ng/L		99	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	2.10	J	ng/L		104	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	1.84	J	ng/L		92	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	2.01	1.85	J	ng/L		92	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	1.87	J	ng/L		93	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.81	J	ng/L		90	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	1.90	J	ng/L		95	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	1.87	J	ng/L		93	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	1.84	J	ng/L		92	50 - 150

Isotope Dilution	%Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	123		50 - 200
13C6 PFDA	109		50 - 200
13C5 PFHxA	128		50 - 200
13C4 PFHpA	129		50 - 200
13C8 PFOA	138		50 - 200
13C9 PFNA	136		50 - 200
13C7 PFUnA	118		50 - 200
13C2 PFDoA	106		50 - 200
13C4 PFBA	132		50 - 200
13C5 PFPeA	125		50 - 200
13C3 PFBS	126		50 - 200
13C3 PFHxS	130		50 - 200
13C8 PFOS	128		50 - 200
13C2-4:2-FTS	119		50 - 200
13C2-6:2-FTS	139		50 - 200
13C2-8:2-FTS	134		50 - 200

Lab Sample ID: 380-220625-B-1-A MS

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 235471

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		120	96.2		ng/L		80	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		120	102		ng/L		85	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		120	107		ng/L		89	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-220625-B-1-A MS

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 235471

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		120	107		ng/L		89	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		120	105		ng/L		86	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		120	110		ng/L		91	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		120	103		ng/L		85	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		120	104		ng/L		85	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	2.3		120	109		ng/L		89	70 - 130
Perfluorohexanoic acid (PFHxA)	2.0		120	104		ng/L		85	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		120	103		ng/L		86	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		120	105		ng/L		87	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		120	107		ng/L		88	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		120	104		ng/L		86	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		120	108		ng/L		89	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		120	104		ng/L		86	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		120	108		ng/L		89	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		120	109		ng/L		90	70 - 130
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		120	103		ng/L		86	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		120	105		ng/L		87	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		120	106		ng/L		88	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		120	105		ng/L		87	70 - 130
Perfluoropentanoic acid (PFPeA)	2.9		120	110		ng/L		89	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		120	103		ng/L		85	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		120	104		ng/L		87	70 - 130

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C3 HFPO-DA	108		50 - 200
13C6 PFDA	107		50 - 200
13C5 PFHxA	107		50 - 200
13C4 PFHpA	109		50 - 200
13C8 PFOA	111		50 - 200
13C9 PFNA	119		50 - 200
13C7 PFUnA	114		50 - 200
13C2 PFDoA	110		50 - 200
13C4 PFBA	114		50 - 200
13C5 PFPeA	109		50 - 200
13C3 PFBS	104		50 - 200
13C3 PFHxS	107		50 - 200
13C8 PFOS	113		50 - 200

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-220625-B-1-A MS

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 235471

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	88		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	124		50 - 200

Lab Sample ID: 380-220625-C-1-A MSD

Matrix: Water

Analysis Batch: 235609

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 235471

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		120	100		ng/L		83	70 - 130	4	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		120	104		ng/L		86	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		120	112		ng/L		93	70 - 130	5	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		120	105		ng/L		88	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		120	111		ng/L		91	70 - 130	5	30
Perfluorodecanoic acid (PFDA)	<2.0		120	115		ng/L		95	70 - 130	5	30
Perfluorododecanoic acid (PFDaA)	<2.0		120	108		ng/L		90	70 - 130	5	30
Perfluoroheptanoic acid (PFHpA)	<2.0		120	112		ng/L		93	70 - 130	8	30
Perfluorohexanesulfonic acid (PFHxS)	2.3		120	108		ng/L		88	70 - 130	1	30
Perfluorohexanoic acid (PFHxA)	2.0		120	109		ng/L		89	70 - 130	5	30
Perfluorononanoic acid (PFNA)	<2.0		120	107		ng/L		89	70 - 130	3	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		120	106		ng/L		87	70 - 130	0	30
Perfluorooctanoic acid (PFOA)	<2.0		120	111		ng/L		91	70 - 130	3	30
Perfluoroundecanoic acid (PFUnA)	<2.0		120	106		ng/L		88	70 - 130	2	30
Perfluorobutanoic acid (PFBA)	<2.0		120	113		ng/L		93	70 - 130	5	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		120	103		ng/L		85	70 - 130	1	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		120	104		ng/L		86	70 - 130	3	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		120	106		ng/L		88	70 - 130	3	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		120	102		ng/L		85	70 - 130	1	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		120	114		ng/L		95	70 - 130	9	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		120	112		ng/L		93	70 - 130	6	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		120	105		ng/L		87	70 - 130	1	30
Perfluoropentanoic acid (PFPeA)	2.9		120	113		ng/L		91	70 - 130	2	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		120	107		ng/L		88	70 - 130	4	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		120	104		ng/L		86	70 - 130	0	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD %Recovery	MSD Qualifier	Limits
13C3 HFPO-DA	103		50 - 200
13C6 PFDA	102		50 - 200
13C5 PFHxA	98		50 - 200
13C4 PFHpA	99		50 - 200
13C8 PFOA	107		50 - 200
13C9 PFNA	113		50 - 200
13C7 PFUnA	112		50 - 200
13C2 PFDoA	105		50 - 200
13C4 PFBA	107		50 - 200
13C5 PFPeA	103		50 - 200
13C3 PFBS	97		50 - 200
13C3 PFHxS	105		50 - 200
13C8 PFOS	109		50 - 200
13C2-4:2-FTS	86		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	122		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-235371/19-A
Matrix: Water
Analysis Batch: 235511

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 235371

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		06/22/26 09:45	06/22/26 18:05	1
Surrogate	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	100		70 - 130			06/22/26 09:45	06/22/26 18:05	1
13C2 PFHxA	101		70 - 130			06/22/26 09:45	06/22/26 18:05	1
13C2 PFDA	99		70 - 130			06/22/26 09:45	06/22/26 18:05	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-235371/19-A

Matrix: Water

Analysis Batch: 235511

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 235371

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	94		70 - 130	06/22/26 09:45	06/22/26 18:05	1

Lab Sample ID: LCS 380-235371/21-A

Matrix: Water

Analysis Batch: 235511

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 235371

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	25.1	19.8		ng/L		79	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	25.1	23.6		ng/L		94	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.1	24.2		ng/L		96	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.1	22.6		ng/L		90	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.1	24.1		ng/L		96	70 - 130
Perfluorohexanoic acid (PFHxA)	25.1	22.5		ng/L		90	70 - 130
Perfluorododecanoic acid (PFDoA)	25.1	23.7		ng/L		94	70 - 130
Perfluorooctanoic acid (PFOA)	25.1	25.1		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	25.1	23.6		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.1	24.6		ng/L		98	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.1	24.8		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.1	23.1		ng/L		92	70 - 130
Perfluorononanoic acid (PFNA)	25.1	23.9		ng/L		95	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.1	20.8		ng/L		83	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.1	22.8		ng/L		91	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	25.1	24.0		ng/L		96	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.1	24.4		ng/L		97	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.1	23.0		ng/L		92	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	95		70 - 130
13C2 PFHxA	93		70 - 130
13C2 PFDA	98		70 - 130
13C3-GenX	86		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-235371/20-A

Matrix: Water

Analysis Batch: 235511

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 235371

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.64	J	ng/L		82	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.88	J	ng/L		93	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.85	J	ng/L		92	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.01	1.87	J	ng/L		93	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.01	1.99	J	ng/L		99	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	1.83	J	ng/L		91	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.68	J	ng/L		84	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	1.88	J	ng/L		94	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	1.85	J	ng/L		92	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	1.87	J	ng/L		93	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.92	J	ng/L		95	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	1.84	J	ng/L		92	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.06	J	ng/L		102	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.01	1.54	J	ng/L		77	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.01	1.78	J	ng/L		89	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	1.82	J	ng/L		91	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.80	J	ng/L		90	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.78	J	ng/L		89	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	102		70 - 130
13C2 PFHxA	105		70 - 130
13C2 PFDA	98		70 - 130
13C3-GenX	90		70 - 130

Lab Sample ID: 380-220757-1 MS

Matrix: Drinking Water

Analysis Batch: 235511

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 235371

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.2	26.3		ng/L		104	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.2	26.4		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.2	25.5		ng/L		101	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.2	25.4		ng/L		101	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-220757-1 MS

Matrix: Drinking Water

Analysis Batch: 235511

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 235371

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		25.2	25.9		ng/L		103	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		25.2	26.0		ng/L		98	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		25.2	24.4		ng/L		97	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		25.2	28.2		ng/L		108	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		25.2	25.9		ng/L		103	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		25.2	26.5		ng/L		99	70 - 130
Perfluorobutanesulfonic acid (PFBS)	<2.0		25.2	26.1		ng/L		101	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		25.2	26.3		ng/L		102	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		25.2	27.1		ng/L		108	70 - 130
Perfluorotetradecanoic acid (PFTA)	<2.0		25.2	21.8		ng/L		87	70 - 130
Perfluorotridecanoic acid (PFTrDA)	<2.0		25.2	24.7		ng/L		98	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		25.2	24.6		ng/L		98	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		25.2	24.7		ng/L		98	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		25.2	26.1		ng/L		104	70 - 130
Surrogate	MS %Recovery	MS Qualifier	Limits						
d5-NEtFOSAA	101		70 - 130						
13C2 PFHxA	100		70 - 130						
13C2 PFDA	100		70 - 130						
13C3-GenX	98		70 - 130						

Lab Sample ID: 380-220757-1 MSD

Matrix: Drinking Water

Analysis Batch: 235511

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 235371

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	24.0		ng/L		96	70 - 130	9	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.1	27.2		ng/L		102	70 - 130	3	30
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	26.0		ng/L		103	70 - 130	2	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	25.5		ng/L		101	70 - 130	0	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		25.1	25.8		ng/L		103	70 - 130	0	30
Perfluorohexanoic acid (PFHxA)	<2.0		25.1	28.1		ng/L		107	70 - 130	8	30
Perfluorododecanoic acid (PFDoA)	<2.0		25.1	24.4		ng/L		97	70 - 130	0	30
Perfluorooctanoic acid (PFOA)	<2.0		25.1	28.1		ng/L		108	70 - 130	0	30
Perfluorodecanoic acid (PFDA)	<2.0		25.1	26.1		ng/L		104	70 - 130	1	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-220757-1 MSD

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 235511

Prep Batch: 235371

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	<2.0		25.1	27.9		ng/L		105	70 - 130	5	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		25.1	26.9		ng/L		105	70 - 130	3	30
Perfluoroheptanoic acid (PFHpA)	<2.0		25.1	26.3		ng/L		102	70 - 130	0	30
Perfluorononanoic acid (PFNA)	<2.0		25.1	27.1		ng/L		108	70 - 130	0	30
Perfluorotetradecanoic acid (PFTA)	<2.0		25.1	22.5		ng/L		90	70 - 130	3	30
Perfluorotridecanoic acid (PFTTrDA)	<2.0		25.1	25.3		ng/L		101	70 - 130	2	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		25.1	25.8		ng/L		103	70 - 130	5	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		25.1	25.5		ng/L		102	70 - 130	3	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		25.1	26.5		ng/L		106	70 - 130	2	30

Surrogate	MSD %Recovery	MSD Qualifier	Limits
d5-NEtFOSAA	99		70 - 130
13C2 PFHxA	104		70 - 130
13C2 PFDA	101		70 - 130
13C3-GenX	97		70 - 130

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

LCMS

Prep Batch: 235371

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-220757-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	537.1 DW	
380-220757-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	537.1 DW	
MBL 380-235371/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-235371/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-235371/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-220757-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	537.1 DW	
380-220757-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	537.1 DW	

Prep Batch: 235471

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-220757-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	533	
380-220757-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	533	
MBL 380-235471/20-A	Method Blank	Total/NA	Water	533	
LCS 380-235471/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-235471/21-A	Lab Control Sample	Total/NA	Water	533	
380-220625-B-1-A MS	Matrix Spike	Total/NA	Water	533	
380-220625-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 235511

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-220757-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	EPA 537.1 V2	235371
380-220757-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	EPA 537.1 V2	235371
MBL 380-235371/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	235371
LCS 380-235371/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	235371
MRL 380-235371/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	235371
380-220757-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	EPA 537.1 V2	235371
380-220757-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	EPA 537.1 V2	235371

Analysis Batch: 235609

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-220757-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	533	235471
380-220757-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	533	235471
MBL 380-235471/20-A	Method Blank	Total/NA	Water	533	235471
LCS 380-235471/22-A	Lab Control Sample	Total/NA	Water	533	235471
MRL 380-235471/21-A	Lab Control Sample	Total/NA	Water	533	235471
380-220625-B-1-A MS	Matrix Spike	Total/NA	Water	533	235471
380-220625-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	235471

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-1

Date Collected: 06/17/26 11:00

Matrix: Drinking Water

Date Received: 06/19/26 09:55

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			235471	E2HD	EA POM	06/22/26 15:55
Total/NA	Analysis	533		1	235609	SZ9R	EA POM	06/23/26 11:31
Total/NA	Prep	537.1 DW			235371	LM3A	EA POM	06/22/26 09:45
Total/NA	Analysis	EPA 537.1 V2		1	235511	SZ9R	EA POM	06/22/26 18:33

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-220757-2

Date Collected: 06/17/26 11:00

Matrix: Water

Date Received: 06/19/26 09:55

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			235471	E2HD	EA POM	06/22/26 15:55
Total/NA	Analysis	533		1	235609	SZ9R	EA POM	06/23/26 11:40
Total/NA	Prep	537.1 DW			235371	LM3A	EA POM	06/22/26 09:45
Total/NA	Analysis	EPA 537.1 V2		1	235511	SZ9R	EA POM	06/22/26 19:59

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-220757-1
SDG: PFAS: Moanalua Wells

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-220757-1	MOANALUA WELLS (331-223-TP202)	Drinking Water	06/17/26 11:00	06/19/26 09:55	Hawaii
380-220757-2	FB: MOANALUA WELLS (331-223-TP202)	Water	06/17/26 11:00	06/19/26 09:55	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

[illegible]

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-220757-1
SDG Number: PFAS: Moanalua Wells

Login Number: 220757

List Number: 1

Creator: Segura, Ryan

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	