

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 9/28/2026 5:39:58 PM

JOB DESCRIPTION

RED-HILL
PFAS: Halawa Wells P1

JOB NUMBER

380-237883-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1145

Generated
9/28/2026 5:39:58 PM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-237883-1

Job ID: 380-237883-1

Eurofins Pomona

Job Narrative 380-237883-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 9/23/2026 9:06 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 4.2°C.

PFAS

EPA 537.1 and EPA 533 are two distinct methods for the analysis of PFAS in drinking water. The analyses are conducted on differing instrumentation, with calibrations, extraction solvents and sample preservatives being dissimilar among the two methods. Therefore it is probable and not unexpected to see the methods having slight variations in analytical results: Halawa Wells P1 (331-023-WL065) (380-237883-1). (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	2.7		2.0	ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	2.9		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.7		2.0	ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	3.7		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.7		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanoic acid (PFHxA)	3.3		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	2.2		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.2		2.0	ng/L	1		EPA 537.1 V2	Total/NA

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-1

Date Collected: 09/21/26 10:49

Matrix: Drinking Water

Date Received: 09/23/26 09:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorohexanesulfonic acid (PFHxS)	2.7		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorohexanoic acid (PFHxA)	2.9		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorooctanesulfonic acid (PFOS)	2.7		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoropentanoic acid (PFPeA)	3.7		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:04	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	103		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C6 PFDA	110		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C5 PFHxA	108		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C4 PFHpA	106		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C8 PFOA	103		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C9 PFNA	106		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C7 PFUnA	112		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C2 PFDoA	112		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C4 PFBA	109		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C5 PFPeA	111		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C3 PFBS	104		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C3 PFHxS	102		50 - 200	09/24/26 08:09	09/24/26 23:04	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-1

Date Collected: 09/21/26 10:49

Matrix: Drinking Water

Date Received: 09/23/26 09:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	102		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C2-4:2-FTS	104		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C2-6:2-FTS	89		50 - 200	09/24/26 08:09	09/24/26 23:04	1
13C2-8:2-FTS	91		50 - 200	09/24/26 08:09	09/24/26 23:04	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorooctanesulfonic acid (PFOS)	2.7		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorohexanoic acid (PFHxA)	3.3		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorooctanoic acid (PFOA)	2.2		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorohexanesulfonic acid (PFHxS)	3.2		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 11:51	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	103		70 - 130	09/24/26 12:23	09/25/26 11:51	1
13C2 PFHxA	109		70 - 130	09/24/26 12:23	09/25/26 11:51	1
13C2 PFDA	111		70 - 130	09/24/26 12:23	09/25/26 11:51	1
13C3-GenX	112		70 - 130	09/24/26 12:23	09/25/26 11:51	1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-2

Date Collected: 09/21/26 10:49

Matrix: Water

Date Received: 09/23/26 09:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-2

Date Collected: 09/21/26 10:49

Matrix: Water

Date Received: 09/23/26 09:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		09/24/26 08:09	09/24/26 23:17	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	104		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C6 PFDA	111		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C5 PFHxA	105		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C4 PFHpA	108		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C8 PFOA	104		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C9 PFNA	109		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C7 PFUnA	113		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C2 PFDoA	113		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C4 PFBA	110		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C5 PFPeA	109		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C3 PFBS	105		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C3 PFHxS	101		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C8 PFOS	108		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C2-4:2-FTS	102		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C2-6:2-FTS	94		50 - 200			09/24/26 08:09	09/24/26 23:17	1
13C2-8:2-FTS	96		50 - 200			09/24/26 08:09	09/24/26 23:17	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-2

Date Collected: 09/21/26 10:49

Matrix: Water

Date Received: 09/23/26 09:06

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/24/26 12:23	09/25/26 12:00	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	101		70 - 130			09/24/26 12:23	09/25/26 12:00	1
13C2 PFHxA	106		70 - 130			09/24/26 12:23	09/25/26 12:00	1
13C2 PFDA	114		70 - 130			09/24/26 12:23	09/25/26 12:00	1
13C3-GenX	98		70 - 130			09/24/26 12:23	09/25/26 12:00	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.7		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.7		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.7		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	2.2		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.2		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-237883-1	HALAWA WELLS P1 (331-023-V)	103	109	111	112
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-237883-2	FB: HALAWA WELLS P1 (331-0	101	106	114	98
670-98380-B-1-A MS	Matrix Spike	103	106	108	108
670-98380-C-1-A MSD	Matrix Spike Duplicate	102	104	104	108
LCS 380-254478/22-A	Lab Control Sample	97	106	102	100
MBL 380-254478/20-A	Method Blank	96	99	97	99
MRL 380-254478/21-A	Lab Control Sample	99	106	98	104
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-237883-1	HALAWA WELLS P1 (331-023-V	103	110	108	106	103	106	112	112

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-237883-1	HALAWA WELLS P1 (331-023-V	109	111	104	102	102	104	89	91

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDaA = 13C2 PFDaA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-237863-B-1-A MS	Matrix Spike	92	94	97	91	88	91	98	103
380-237863-C-1-A MSD	Matrix Spike Duplicate	84	65	83	79	73	69	68	74
380-237883-2	FB: HALAWA WELLS P1 (331-023-WL065)	104	111	105	108	104	109	113	113
LCS 380-254453/22-A	Lab Control Sample	113	117	115	112	114	113	118	123
MBL 380-254453/20-A	Method Blank	97	107	103	98	99	99	108	105
MRL 380-254453/21-A	Lab Control Sample	104	109	108	105	108	109	113	111

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-237863-B-1-A MS	Matrix Spike	103	99	106	100	109	111	101	98
380-237863-C-1-A MSD	Matrix Spike Duplicate	96	90	100	96	101	108	90	101
380-237883-2	FB: HALAWA WELLS P1 (331-023-WL065)	110	109	105	101	108	102	94	96
LCS 380-254453/22-A	Lab Control Sample	117	118	110	112	114	113	105	104
MBL 380-254453/20-A	Method Blank	104	111	103	105	106	103	88	100
MRL 380-254453/21-A	Lab Control Sample	110	113	104	105	110	100	94	99

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA

Eurofins Pomona

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDoA = 13C2 PFDoA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-254453/20-A

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 254453

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.99		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluorobutanoic acid (PFBA)	<0.68		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.36		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.47		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.46		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.24		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.45		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoropentanoic acid (PFPeA)	<0.37		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.35		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Perfluoropentanesulfonic acid (PFPeS)	<0.38		2.0	ng/L		09/24/26 08:09	09/24/26 20:29	1
Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	97		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C6 PFDA	107		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C5 PFHxA	103		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C4 PFHpA	98		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C8 PFOA	99		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C9 PFNA	99		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C7 PFUnA	108		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C2 PFDoA	105		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C4 PFBA	104		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C5 PFPeA	111		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C3 PFBS	103		50 - 200			09/24/26 08:09	09/24/26 20:29	1
13C3 PFHxS	105		50 - 200			09/24/26 08:09	09/24/26 20:29	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-254453/20-A

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 254453

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	106		50 - 200	09/24/26 08:09	09/24/26 20:29	1
13C2-4:2-FTS	103		50 - 200	09/24/26 08:09	09/24/26 20:29	1
13C2-6:2-FTS	88		50 - 200	09/24/26 08:09	09/24/26 20:29	1
13C2-8:2-FTS	100		50 - 200	09/24/26 08:09	09/24/26 20:29	1

Lab Sample ID: LCS 380-254453/22-A

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 254453

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	59.1	52.7		ng/L		89	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	59.1	51.9		ng/L		88	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	59.1	57.1		ng/L		97	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	59.1	51.7		ng/L		88	70 - 130
Perfluorobutanesulfonic acid (PFBS)	59.1	52.2		ng/L		88	70 - 130
Perfluorodecanoic acid (PFDA)	59.1	54.2		ng/L		92	70 - 130
Perfluorododecanoic acid (PFDaA)	59.1	50.1		ng/L		85	70 - 130
Perfluoroheptanoic acid (PFHpA)	59.1	55.6		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	59.1	51.6		ng/L		87	70 - 130
Perfluorohexanoic acid (PFHxA)	59.1	51.3		ng/L		87	70 - 130
Perfluorononanoic acid (PFNA)	59.1	52.3		ng/L		89	70 - 130
Perfluorooctanesulfonic acid (PFOS)	59.1	52.0		ng/L		88	70 - 130
Perfluorooctanoic acid (PFOA)	59.1	50.4		ng/L		85	70 - 130
Perfluoroundecanoic acid (PFUnA)	59.1	53.9		ng/L		91	70 - 130
Perfluorobutanoic acid (PFBA)	59.1	51.3		ng/L		87	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	59.1	53.9		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	59.1	47.3		ng/L		80	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	59.1	52.7		ng/L		89	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	59.1	50.2		ng/L		85	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	59.1	52.2		ng/L		88	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	59.1	51.1		ng/L		87	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	59.1	50.6		ng/L		86	70 - 130
Perfluoropentanoic acid (PFPeA)	59.1	52.0		ng/L		88	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	59.1	48.3		ng/L		82	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-254453/22-A

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 254453

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	59.1	50.6		ng/L		86	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	113		50 - 200				
13C6 PFDA	117		50 - 200				
13C5 PFHxA	115		50 - 200				
13C4 PFHpA	112		50 - 200				
13C8 PFOA	114		50 - 200				
13C9 PFNA	113		50 - 200				
13C7 PFUnA	118		50 - 200				
13C2 PFDoA	123		50 - 200				
13C4 PFBA	117		50 - 200				
13C5 PFPeA	118		50 - 200				
13C3 PFBS	110		50 - 200				
13C3 PFHxS	112		50 - 200				
13C8 PFOS	114		50 - 200				
13C2-4:2-FTS	113		50 - 200				
13C2-6:2-FTS	105		50 - 200				
13C2-8:2-FTS	104		50 - 200				

Lab Sample ID: MRL 380-254453/21-A

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 254453

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.97	2.02	J	ng/L		102	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	1.97	2.01	J	ng/L		102	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.97	2.07	J	ng/L		105	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.97	1.89	J	ng/L		96	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.97	1.86	J	ng/L		94	50 - 150
Perfluorodecanoic acid (PFDA)	1.97	2.01	J	ng/L		102	50 - 150
Perfluorododecanoic acid (PFDoA)	1.97	1.98	J	ng/L		100	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.97	2.07	J	ng/L		105	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.97	1.91	J	ng/L		97	50 - 150
Perfluorohexanoic acid (PFHxA)	1.97	1.93	J	ng/L		98	50 - 150
Perfluorononanoic acid (PFNA)	1.97	1.97	J	ng/L		100	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.97	1.94	J	ng/L		99	50 - 150
Perfluorooctanoic acid (PFOA)	1.97	1.94	J	ng/L		99	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.97	2.06	J	ng/L		104	50 - 150
Perfluorobutanoic acid (PFBA)	1.97	1.96	J	ng/L		100	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-254453/21-A

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 254453

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	1.97	2.21	J	ng/L		112	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	1.97	1.96	J	ng/L		100	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	1.97	2.08	J	ng/L		106	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	1.97	1.97	J	ng/L		100	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	1.97	1.99	J	ng/L		101	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	1.97	2.00	J	ng/L		101	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	1.97	1.97	J	ng/L		100	50 - 150
Perfluoropentanoic acid (PFPeA)	1.97	1.95	J	ng/L		99	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	1.97	1.79	J	ng/L		91	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	1.97	1.83	J	ng/L		93	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	104		50 - 200
13C6 PFDA	109		50 - 200
13C5 PFHxA	108		50 - 200
13C4 PFHpA	105		50 - 200
13C8 PFOA	108		50 - 200
13C9 PFNA	109		50 - 200
13C7 PFUnA	113		50 - 200
13C2 PFDoA	111		50 - 200
13C4 PFBA	110		50 - 200
13C5 PFPeA	113		50 - 200
13C3 PFBS	104		50 - 200
13C3 PFHxS	105		50 - 200
13C8 PFOS	110		50 - 200
13C2-4:2-FTS	100		50 - 200
13C2-6:2-FTS	94		50 - 200
13C2-8:2-FTS	99		50 - 200

Lab Sample ID: 380-237863-B-1-A MS

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 254453

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		59.3	55.2		ng/L		93	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		59.3	57.5		ng/L		97	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		59.3	57.0		ng/L		96	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-237863-B-1-A MS

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 254453

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		59.3	56.2		ng/L		95	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	6.8		59.3	65.3		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		59.3	58.6		ng/L		99	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		59.3	58.4		ng/L		98	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		59.3	64.3		ng/L		106	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	7.9		59.3	67.2		ng/L		100	70 - 130
Perfluorohexanoic acid (PFHxA)	3.1		59.3	59.9		ng/L		96	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		59.3	61.0		ng/L		100	70 - 130
Perfluorooctanesulfonic acid (PFOS)	11		59.3	66.8		ng/L		94	70 - 130
Perfluorooctanoic acid (PFOA)	5.8		59.3	63.8		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		59.3	58.9		ng/L		99	70 - 130
Perfluorobutanoic acid (PFBA)	8.1		59.3	66.3		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		59.3	60.6		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		59.3	59.5		ng/L		100	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		59.3	58.1		ng/L		98	70 - 130
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		59.3	54.0		ng/L		91	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		59.3	66.6		ng/L		112	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		59.3	52.7		ng/L		89	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		59.3	57.0		ng/L		96	70 - 130
Perfluoropentanoic acid (PFPeA)	3.2		59.3	60.7		ng/L		97	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		59.3	53.4		ng/L		90	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		59.3	60.1		ng/L		100	70 - 130

Isotope Dilution	MS %Recovery	MS Qualifier	MS Limits
13C3 HFPO-DA	92		50 - 200
13C6 PFDA	94		50 - 200
13C5 PFHxA	97		50 - 200
13C4 PFHpA	91		50 - 200
13C8 PFOA	88		50 - 200
13C9 PFNA	91		50 - 200
13C7 PFUnA	98		50 - 200
13C2 PFDoA	103		50 - 200
13C4 PFBA	103		50 - 200
13C5 PFPeA	99		50 - 200
13C3 PFBS	106		50 - 200
13C3 PFHxS	100		50 - 200
13C8 PFOS	109		50 - 200

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-237863-B-1-A MS

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 254453

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	111		50 - 200
13C2-6:2-FTS	101		50 - 200
13C2-8:2-FTS	98		50 - 200

Lab Sample ID: 380-237863-C-1-A MSD

Matrix: Water

Analysis Batch: 254567

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 254453

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		59.3	58.3		ng/L		98	70 - 130	5	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		59.3	59.9		ng/L		101	70 - 130	4	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		59.3	59.3		ng/L		100	70 - 130	4	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		59.3	56.8		ng/L		96	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)	6.8		59.3	67.2		ng/L		102	70 - 130	3	30
Perfluorodecanoic acid (PFDA)	<2.0		59.3	62.2		ng/L		105	70 - 130	6	30
Perfluorododecanoic acid (PFDaA)	<2.0		59.3	59.4		ng/L		100	70 - 130	2	30
Perfluoroheptanoic acid (PFHpA)	<2.0		59.3	61.6		ng/L		102	70 - 130	4	30
Perfluorohexanesulfonic acid (PFHxS)	7.9		59.3	68.1		ng/L		102	70 - 130	1	30
Perfluorohexanoic acid (PFHxA)	3.1		59.3	62.4		ng/L		100	70 - 130	4	30
Perfluorononanoic acid (PFNA)	<2.0		59.3	63.4		ng/L		104	70 - 130	4	30
Perfluorooctanesulfonic acid (PFOS)	11		59.3	70.6		ng/L		100	70 - 130	6	30
Perfluorooctanoic acid (PFOA)	5.8		59.3	65.8		ng/L		101	70 - 130	3	30
Perfluoroundecanoic acid (PFUnA)	<2.0		59.3	60.3		ng/L		102	70 - 130	2	30
Perfluorobutanoic acid (PFBA)	8.1		59.3	66.3		ng/L		98	70 - 130	0	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		59.3	57.3		ng/L		97	70 - 130	6	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		59.3	59.7		ng/L		101	70 - 130	0	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		59.3	63.6		ng/L		107	70 - 130	9	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		59.3	59.1		ng/L		100	70 - 130	9	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		59.3	70.0		ng/L		118	70 - 130	5	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		59.3	55.0		ng/L		93	70 - 130	4	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		59.3	58.7		ng/L		99	70 - 130	3	30
Perfluoropentanoic acid (PFPeA)	3.2		59.3	60.5		ng/L		97	70 - 130	0	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		59.3	56.7		ng/L		96	70 - 130	6	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		59.3	63.2		ng/L		105	70 - 130	5	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD %Recovery	MSD Qualifier	Limits
13C3 HFPO-DA	84		50 - 200
13C6 PFDA	65		50 - 200
13C5 PFHxA	83		50 - 200
13C4 PFHpA	79		50 - 200
13C8 PFOA	73		50 - 200
13C9 PFNA	69		50 - 200
13C7 PFUnA	68		50 - 200
13C2 PFDoA	74		50 - 200
13C4 PFBA	96		50 - 200
13C5 PFPeA	90		50 - 200
13C3 PFBS	100		50 - 200
13C3 PFHxS	96		50 - 200
13C8 PFOS	101		50 - 200
13C2-4:2-FTS	108		50 - 200
13C2-6:2-FTS	90		50 - 200
13C2-8:2-FTS	101		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-254478/20-A
Matrix: Water
Analysis Batch: 254713

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 254478

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<0.98		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorooctanesulfonic acid (PFOS)	<0.42		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluoroundecanoic acid (PFUnA)	<0.41		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.57		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.41		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorohexanoic acid (PFHxA)	<0.45		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorododecanoic acid (PFDoA)	<0.53		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorooctanoic acid (PFOA)	<0.37		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorodecanoic acid (PFDA)	<0.30		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorohexanesulfonic acid (PFHxS)	<0.31		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorobutanesulfonic acid (PFBS)	<0.36		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluoroheptanoic acid (PFHpA)	<0.38		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorononanoic acid (PFNA)	<0.39		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorotetradecanoic acid (PFTA)	<0.53		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Perfluorotridecanoic acid (PFTrDA)	<0.35		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.29		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.59		2.0	ng/L		09/24/26 12:23	09/25/26 10:05	1
Surrogate	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	96		70 - 130			09/24/26 12:23	09/25/26 10:05	1
13C2 PFHxA	99		70 - 130			09/24/26 12:23	09/25/26 10:05	1
13C2 PFDA	97		70 - 130			09/24/26 12:23	09/25/26 10:05	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-254478/20-A

Matrix: Water

Analysis Batch: 254713

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 254478

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	99		70 - 130	09/24/26 12:23	09/25/26 10:05	1

Lab Sample ID: LCS 380-254478/22-A

Matrix: Water

Analysis Batch: 254713

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 254478

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	49.0	51.5		ng/L		105	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	49.0	49.8		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	49.0	47.4		ng/L		97	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	49.0	47.3		ng/L		96	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	49.0	47.5		ng/L		97	70 - 130
Perfluorohexanoic acid (PFHxA)	49.0	50.4		ng/L		103	70 - 130
Perfluorododecanoic acid (PFDoA)	49.0	48.7		ng/L		99	70 - 130
Perfluorooctanoic acid (PFOA)	49.0	50.3		ng/L		103	70 - 130
Perfluorodecanoic acid (PFDA)	49.0	47.6		ng/L		97	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	49.0	51.1		ng/L		104	70 - 130
Perfluorobutanesulfonic acid (PFBS)	49.0	50.6		ng/L		103	70 - 130
Perfluoroheptanoic acid (PFHpA)	49.0	49.5		ng/L		101	70 - 130
Perfluorononanoic acid (PFNA)	49.0	51.1		ng/L		104	70 - 130
Perfluorotetradecanoic acid (PFTA)	49.0	47.6		ng/L		97	70 - 130
Perfluorotridecanoic acid (PFTrDA)	49.0	48.1		ng/L		98	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	49.0	49.5		ng/L		101	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	49.0	47.7		ng/L		97	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	49.0	50.0		ng/L		102	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	97		70 - 130
13C2 PFHxA	106		70 - 130
13C2 PFDA	102		70 - 130
13C3-GenX	100		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-254478/21-A

Matrix: Water

Analysis Batch: 254713

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 254478

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	1.96	1.97	J	ng/L		100	50 - 150
Perfluorooctanesulfonic acid (PFOS)	1.96	1.96	J	ng/L		100	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.96	2.04	J	ng/L		104	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	1.96	1.81	J	ng/L		92	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	1.96	1.98	J	ng/L		101	50 - 150
Perfluorohexanoic acid (PFHxA)	1.96	1.95	J	ng/L		99	50 - 150
Perfluorododecanoic acid (PFDoA)	1.96	2.14	J	ng/L		109	50 - 150
Perfluorooctanoic acid (PFOA)	1.96	2.00	J	ng/L		102	50 - 150
Perfluorodecanoic acid (PFDA)	1.96	2.01	J	ng/L		102	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.96	1.95	J	ng/L		99	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.96	1.89	J	ng/L		96	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.96	1.93	J	ng/L		98	50 - 150
Perfluorononanoic acid (PFNA)	1.96	1.98	J	ng/L		101	50 - 150
Perfluorotetradecanoic acid (PFTA)	1.96	1.76	J	ng/L		90	50 - 150
Perfluorotridecanoic acid (PFTrDA)	1.96	1.90	J	ng/L		97	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	1.96	1.85	J	ng/L		94	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	1.96	1.82	J	ng/L		93	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	1.96	1.91	J	ng/L		97	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	99		70 - 130
13C2 PFHxA	106		70 - 130
13C2 PFDA	98		70 - 130
13C3-GenX	104		70 - 130

Lab Sample ID: 670-98380-B-1-A MS

Matrix: Water

Analysis Batch: 254713

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 254478

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		24.6	26.9		ng/L		110	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		24.6	25.7		ng/L		105	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	25.4		ng/L		103	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	24.8		ng/L		101	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 670-98380-B-1-A MS

Matrix: Water

Analysis Batch: 254713

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 254478

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		24.6	25.3		ng/L		103	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		24.6	27.4		ng/L		111	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	27.0		ng/L		110	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		24.6	26.9		ng/L		110	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		24.6	26.5		ng/L		108	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		24.6	26.1		ng/L		106	70 - 130
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	26.5		ng/L		108	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	26.9		ng/L		109	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		24.6	28.0		ng/L		114	70 - 130
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	26.8		ng/L		109	70 - 130
Perfluorotridecanoic acid (PFTrDA)	<2.0		24.6	27.6		ng/L		113	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		24.6	26.2		ng/L		107	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	24.9		ng/L		102	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	26.8		ng/L		109	70 - 130
Surrogate	MS %Recovery	MS Qualifier	Limits						
d5-NEtFOSAA	103		70 - 130						
13C2 PFHxA	106		70 - 130						
13C2 PFDA	108		70 - 130						
13C3-GenX	108		70 - 130						

Lab Sample ID: 670-98380-C-1-A MSD

Matrix: Water

Analysis Batch: 254713

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 254478

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		24.6	24.5		ng/L		100	70 - 130	9	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		24.6	25.5		ng/L		104	70 - 130	1	30
Perfluoroundecanoic acid (PFUnA)	<2.0		24.6	25.4		ng/L		103	70 - 130	0	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		24.6	25.1		ng/L		102	70 - 130	1	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		24.6	24.6		ng/L		100	70 - 130	3	30
Perfluorohexanoic acid (PFHxA)	<2.0		24.6	26.4		ng/L		107	70 - 130	4	30
Perfluorododecanoic acid (PFDoA)	<2.0		24.6	26.7		ng/L		108	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	<2.0		24.6	27.3		ng/L		111	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		24.6	25.3		ng/L		103	70 - 130	4	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 670-98380-C-1-A MSD

Matrix: Water

Analysis Batch: 254713

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 254478

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	<2.0		24.6	26.0		ng/L		106	70 - 130	0	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		24.6	26.1		ng/L		106	70 - 130	1	30
Perfluoroheptanoic acid (PFHpA)	<2.0		24.6	26.4		ng/L		107	70 - 130	2	30
Perfluorononanoic acid (PFNA)	<2.0		24.6	27.2		ng/L		111	70 - 130	3	30
Perfluorotetradecanoic acid (PFTA)	<2.0		24.6	25.9		ng/L		105	70 - 130	4	30
Perfluorotridecanoic acid (PFTTrDA)	<2.0		24.6	26.3		ng/L		107	70 - 130	5	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		24.6	25.8		ng/L		105	70 - 130	2	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		24.6	25.0		ng/L		102	70 - 130	0	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		24.6	25.6		ng/L		104	70 - 130	5	30
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
d5-NEtFOSAA	102		70 - 130								
13C2 PFHxA	104		70 - 130								
13C2 PFDA	104		70 - 130								
13C3-GenX	108		70 - 130								

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

LCMS

Prep Batch: 254453

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-237883-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	533	
380-237883-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	533	
MBL 380-254453/20-A	Method Blank	Total/NA	Water	533	
LCS 380-254453/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-254453/21-A	Lab Control Sample	Total/NA	Water	533	
380-237863-B-1-A MS	Matrix Spike	Total/NA	Water	533	
380-237863-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Prep Batch: 254478

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-237883-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	537.1 DW	
380-237883-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	537.1 DW	
MBL 380-254478/20-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-254478/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-254478/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
670-98380-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
670-98380-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 254567

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-237883-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	533	254453
380-237883-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	533	254453
MBL 380-254453/20-A	Method Blank	Total/NA	Water	533	254453
LCS 380-254453/22-A	Lab Control Sample	Total/NA	Water	533	254453
MRL 380-254453/21-A	Lab Control Sample	Total/NA	Water	533	254453
380-237863-B-1-A MS	Matrix Spike	Total/NA	Water	533	254453
380-237863-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	254453

Analysis Batch: 254713

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-237883-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	EPA 537.1 V2	254478
380-237883-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	EPA 537.1 V2	254478
MBL 380-254478/20-A	Method Blank	Total/NA	Water	EPA 537.1 V2	254478
LCS 380-254478/22-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	254478
MRL 380-254478/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	254478
670-98380-B-1-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	254478
670-98380-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	254478

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-1

Date Collected: 09/21/26 10:49

Matrix: Drinking Water

Date Received: 09/23/26 09:06

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			254453	XTD8	EA POM	09/24/26 08:09
Total/NA	Analysis	533		1	254567	Y5FM	EA POM	09/24/26 23:04
Total/NA	Prep	537.1 DW			254478	LM3A	EA POM	09/24/26 12:23
Total/NA	Analysis	EPA 537.1 V2		1	254713	UKDT	EA POM	09/25/26 11:51

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-237883-2

Date Collected: 09/21/26 10:49

Matrix: Water

Date Received: 09/23/26 09:06

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			254453	XTD8	EA POM	09/24/26 08:09
Total/NA	Analysis	533		1	254567	Y5FM	EA POM	09/24/26 23:17
Total/NA	Prep	537.1 DW			254478	LM3A	EA POM	09/24/26 12:23
Total/NA	Analysis	EPA 537.1 V2		1	254713	UKDT	EA POM	09/25/26 12:00

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-27

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-237883-1
SDG: PFAS: Halawa Wells P1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-237883-1	HALAWA WELLS P1 (331-023-WL065)	Drinking Water	09/21/26 10:49	09/23/26 09:06	Hawaii
380-237883-2	FB: HALAWA WELLS P1 (331-023-WL065)	Water	09/21/26 10:49	09/23/26 09:06	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Monrovia, CA (Suite 100)

750 Royal Oaks Drive Suite 100
Monrovia, CA 91016
Phone (626) 396-1100

Chain of Custody Record



Enviro
Anne



380-237883 COC

Client Information Client Contact: Kirk Iwamoto Phone: +1 808 748 5840 City & County of Honolulu Address: 630 South Beretania Street, Chemistry Lab City: Honolulu State, Zip: HI, 96843 Phone: 808-748-5840 (tel) Email: kiwamoto@hbws.org Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill Site:		Lab PM: Lopez, Maria E-Mail: Maria.Lopez@et.eurofins.com Carrier Tracking No(s): State of Origin: Page: Page 1 of 1 Job #:	
Due Date Requested: TAT Requested (days): Compliance Project: ☐ No PO #: C20525101 exp 05312023 WO #: 38001111 Project #: 38001111 SSOW#:		Analysis Requested Field Filled Sample (Yes or No) ☒ Perform MS/MSD (Yes or No) ☒ SUBCONTRACT - 625 PAH Physals LL (EAL) + TICs 80168_GRO_LL - (MOD) GRO 80168_DRO_LL_G8 - HNL Ranges: C10-C24/C24-C38/C8-C18 625.2 PREC - (MOD) 525plus PLUS TICs 637.1 DW_PREC - 637.1 Full List 633 - All Analytes Total Number of Containers:	
Sample Identification Halawa Wells P1 (331-023-WL065) FB: Halawa Wells P1 (331-023-WL065)		Preservation Codes: A - HCL B - NaOH C - Zn Acetate D - Nitric Acid E - NaHSO4 F - MeOH G - Amchlor H - Ascorbic Acid I - Ice J - DI Water K - EDTA L - EDA Other: M - Hexane N - None O - AsNaO2 P - Na2OAS Q - Na2SO3 R - Na2S2O3 S - H2SO4 T - TSP Dodecahydrate U - Acetone V - MCAA W - pH 4.5 Y - Trizma Z - other (specify)	
Sample Date: 21-Sep-2026 Sample Time: 1049 Sample Type (C=Comp, G=grab): G Matrix (Water, Sediment, Organics, etc.): Water		Special Instructions/Note: Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) ☐ Return To Client ☐ Disposal By Lab ☐ Archive For Months	
Possible Hazard Identification ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☐ Unknown ☐ Radiological Deliverable Requested: I, II, III, IV, Other (specify)		Special Instructions/QC Requirements:	
Empty Kit Relinquished by:		Date: 22-Sep-2026 1400 Company: HBWS	
Relinquished by:		Date/Time: 09/23/26 09:06 Company:	
Custody Seals Intact: ☐ Yes ☐ No		Cooler Temperature(s) °C and Other Remarks: 4.4-0.2-4.2 DEL-POWER	

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-237883-1

SDG Number: PFAS: Halawa Wells P1

Login Number: 237883

List Number: 1

Creator: Tran, Kristine

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	