

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 9/10/2026 11:30:19 AM

JOB DESCRIPTION

RED-HILL
Weekly: Ka'amilo Wells P1

JOB NUMBER

380-233066-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1145

Generated
9/10/2026 11:30:19 AM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
QC Sample Results	14
QC Association Summary	31
Lab Chronicle	33
Certification Summary	34
Method Summary	36
Sample Summary	37
Chain of Custody	38
Receipt Checklists	45



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Qualifiers

GC/MS Semi VOA

Qualifier	Qualifier Description
*1	LCS/LCSD RPD exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

GC Semi VOA

Qualifier	Qualifier Description
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased
F1	MS and/or MSD recovery exceeds control limits.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-233066-1

Job ID: 380-233066-1

Eurofins Pomona

Job Narrative 380-233066-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 8/26/2026 10:12 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.8°C.

GC/MS Semi VOA

Method 625.1 SIM: The RPD of the laboratory control sample (LCS) and laboratory control sample duplicate (LCSD) for preparation batch 570-791672 and analytical batch 570-794330 recovered outside control limits for the following analytes: 1-Methylnaphthalene and 2-Methylnaphthalene.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Gasoline Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Diesel Range Organics

Method 8015B: The matrix spike / matrix spike duplicate (MS/MSD) recoveries for preparation batch 570-791434 and analytical batch 570-796133 were outside control limits for one or more analytes. See QC Sample Results for detail. Sample matrix interference and/or non-homogeneity are suspected because the associated laboratory control sample (LCS) recovery is within acceptance limits.

Method 8015B: The method reporting limit check (MRL) for preparation batch 570-791434 and analytical batch 570-796453 recovered outside control limits for the following analytes: C10-C28. These analytes were biased high in the MRL and were not detected in the associated samples; therefore, the data have been reported.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Client Sample ID: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Dieldrin	0.081		0.0098	ug/L	1		525.2	Total/NA
Heptachlor epoxide (isomer B)	0.012		0.0098	ug/L	1		525.2	Total/NA

Client Sample ID: TB: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Client Sample ID: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-1

Date Collected: 08/24/26 12:48

Matrix: Water

Date Received: 08/26/26 10:12

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
2,4'-DDD	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
2,4'-DDE	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
2,4'-DDT	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
2-Methylnaphthalene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
4,4'-DDD	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
4,4'-DDE	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
4,4'-DDT	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Acenaphthene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Acenaphthylene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Acetochlor	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Alachlor	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
alpha-BHC	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
alpha-Chlordane	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Anthracene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 19:55	1
Atrazine	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Benz(a)anthracene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Benzo[a]pyrene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 19:55	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 19:55	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 19:55	1
beta-BHC	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		08/28/26 10:18	08/31/26 19:55	1
Bromacil	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Butachlor	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Butylbenzylphthalate	<0.49		0.49	ug/L		08/28/26 10:18	08/31/26 19:55	1
Chlorobenzilate	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Chloroneb	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Chlorpyrifos	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Chrysene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 19:55	1
delta-BHC	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		08/28/26 10:18	08/31/26 19:55	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Dieldrin	0.081		0.0098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Diethylphthalate	<0.49		0.49	ug/L		08/28/26 10:18	08/31/26 19:55	1
Dimethylphthalate	<0.49		0.49	ug/L		08/28/26 10:18	08/31/26 19:55	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		08/28/26 10:18	08/31/26 19:55	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Endosulfan sulfate	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Endrin	<0.0098		0.0098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Endrin aldehyde	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
EPTC	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Fluoranthene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Client Sample ID: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-1

Date Collected: 08/24/26 12:48

Matrix: Water

Date Received: 08/26/26 10:12

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
gamma-Chlordane	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Heptachlor	<0.0098		0.0098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Heptachlor epoxide (isomer B)	0.012		0.0098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Hexachlorobenzene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Isophorone	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Lindane	<0.0098		0.0098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Malathion	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Methoxychlor	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Metolachlor	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Molinate	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Naphthalene	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Parathion	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Phenanthrene	<0.039		0.039	ug/L		08/28/26 10:18	08/31/26 19:55	1
Propachlor	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Pyrene	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Simazine	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Terbacil	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Terbutylazine	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Thiobencarb	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		08/28/26 10:18	08/31/26 19:55	1
trans-Nonachlor	<0.049		0.049	ug/L		08/28/26 10:18	08/31/26 19:55	1
Trifluralin	<0.098		0.098	ug/L		08/28/26 10:18	08/31/26 19:55	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	08/28/26 10:18	08/31/26 19:55	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	95		70 - 130	08/28/26 10:18	08/31/26 19:55	1
Perylene-d12	92		70 - 130	08/28/26 10:18	08/31/26 19:55	1
Triphenylphosphate	102		70 - 130	08/28/26 10:18	08/31/26 19:55	1

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19	*1	0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
2-Methylnaphthalene	<0.19	*1	0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Acenaphthene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Acenaphthylene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Anthracene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Benzo[a]anthracene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Benzo[a]pyrene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Benzo[b]fluoranthene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Benzo[g,h,i]perylene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Benzo[k]fluoranthene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Chrysene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Dibenz(a,h)anthracene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Fluoranthene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Client Sample ID: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-1

Date Collected: 08/24/26 12:48

Matrix: Water

Date Received: 08/26/26 10:12

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Indeno[1,2,3-cd]pyrene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Naphthalene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Phenanthrene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1
Pyrene	<0.19		0.19	ug/L		08/28/26 06:39	09/02/26 18:04	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	94		28 - 127	08/28/26 06:39	09/02/26 18:04	1
2-Fluorobiphenyl (Surr)	85		31 - 120	08/28/26 06:39	09/02/26 18:04	1
2-Fluorophenol (Surr)	55		17 - 120	08/28/26 06:39	09/02/26 18:04	1
Nitrobenzene-d5 (Surr)	87		27 - 120	08/28/26 06:39	09/02/26 18:04	1
Phenol-d6 (Surr)	34		10 - 120	08/28/26 06:39	09/02/26 18:04	1
p-Terphenyl-d14 (Surr)	83		45 - 120	08/28/26 06:39	09/02/26 18:04	1

Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	08/28/26 06:39	09/09/26 19:41	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	72		33 - 139	08/28/26 06:39	09/09/26 19:41	1
2-Fluorobiphenyl (Surr)	76		33 - 126	08/28/26 06:39	09/09/26 19:41	1
2-Fluorophenol (Surr)	44		12 - 120	08/28/26 06:39	09/09/26 19:41	1
Nitrobenzene-d5 (Surr)	77		36 - 120	08/28/26 06:39	09/09/26 19:41	1
Phenol-d6 (Surr)	23		10 - 120	08/28/26 06:39	09/09/26 19:41	1
p-Terphenyl-d14 (Surr)	79		47 - 131	08/28/26 06:39	09/09/26 19:41	1

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			09/04/26 19:37	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	93		38 - 134		09/04/26 19:37	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<26		26	ug/L		08/27/26 15:57	09/06/26 05:00	1
Motor Oil Range Organics [C24-C36]	<26		26	ug/L		08/27/26 15:57	09/06/26 05:00	1
C8-C18	<26		26	ug/L		08/27/26 15:57	09/06/26 05:00	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	87		60 - 130	08/27/26 15:57	09/06/26 05:00	1

Client Sample ID: TB: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-2

Date Collected: 08/24/26 12:48

Matrix: Water

Date Received: 08/26/26 10:12

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			09/04/26 14:16	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Client Sample ID: TB: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-2

Date Collected: 08/24/26 12:48

Matrix: Water

Date Received: 08/26/26 10:12

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
4-Bromofluorobenzene (Surr)	94		38 - 134		09/04/26 14:16	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Client Sample ID: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	<0.049		ug/L	2	0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.020		ug/L	0.2	0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.59		ug/L	6	0.59	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.59		ug/L	400	0.59	525.2	Total/NA
Endrin	<0.0098		ug/L	2	0.0098	525.2	Total/NA
Heptachlor	<0.0098		ug/L	0.4	0.0098	525.2	Total/NA
Heptachlor epoxide (isomer B)	0.012		ug/L	0.2	0.0098	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50	0.049	525.2	Total/NA
Lindane	<0.0098		ug/L	0.2	0.0098	525.2	Total/NA
Methoxychlor	<0.049		ug/L	40	0.049	525.2	Total/NA
Simazine	<0.049		ug/L	4	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.19		ug/L	0.2	0.19	625.1 SIM	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	95	92	102
380-233443-B-3-A MS	Matrix Spike	97	98	102
380-233458-V-1-A DU	Duplicate	98	90	99
LCS 380-249512/23-A	Lab Control Sample	96	96	100
LCSD 380-249512/24-A	Lab Control Sample Dup	97	97	104
MB 380-249512/21-A	Method Blank	96	92	100
MRL 380-249512/22-A	Lab Control Sample	95	90	98
Surrogate Legend				
2NMX = 2-Nitro-m-xylene				
PRY = Perylene-d12				
TPP = Triphenylphosphate				

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	72	76	44	77	23	79
MB 570-791672/1-A	Method Blank	67	69	49	71	24	90
Surrogate Legend							
TBP = 2,4,6-Tribromophenol (Surr)							
FBP = 2-Fluorobiphenyl (Surr)							
2FP = 2-Fluorophenol (Surr)							
NBZ = Nitrobenzene-d5 (Surr)							
PHL6 = Phenol-d6 (Surr)							
TPHd14 = p-Terphenyl-d14 (Surr)							

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-233060-A-1-A MS	Matrix Spike	77	74	49	72	31	79
380-233060-A-1-B MSD	Matrix Spike Duplicate	82	81	58	80	37	85
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	94	85	55	87	34	83
LCS 570-791672/2-A	Lab Control Sample	89	84	65	85	41	89
LCSD 570-791672/3-A	Lab Control Sample Dup	84	86	61	64	39	86
MB 570-791672/1-A	Method Blank	86	73	50	75	31	85
Surrogate Legend							
TBP = 2,4,6-Tribromophenol (Surr)							
FBP = 2-Fluorobiphenyl (Surr)							
2FP = 2-Fluorophenol (Surr)							
NBZ = Nitrobenzene-d5 (Surr)							
PHL6 = Phenol-d6 (Surr)							
TPHd14 = p-Terphenyl-d14 (Surr)							

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)	
Lab Sample ID	Client Sample ID	BFB1 (38-134)	
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	93	
380-233066-2	TB: Ka'amilo Wells P1 (331-031-WL008)	94	
380-233271-B-1 MS	Matrix Spike	93	
380-233271-E-1 MSD	Matrix Spike Duplicate	92	
LCS 570-795451/3	Lab Control Sample	83	
LCSD 570 795451/4	Lab Control Sample Dup	88	
MB 570-795451/6	Method Blank	86	
MRL 570-795451/5	Lab Control Sample	89	
Surrogate Legend			
BFB = 4-Bromofluorobenzene (Surr)			

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)	
Lab Sample ID	Client Sample ID	OTCSN1 (60-130)	
380-233060-C-1-C MS	Matrix Spike	82	
380-233060-C-1-D MSD	Matrix Spike Duplicate	92	
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	87	
LCS 570-791434/2-A	Lab Control Sample	93	
LCSD 570-791434/3-A	Lab Control Sample Dup	91	
MB 570-791434/1-A	Method Blank	83	
MRL 570-791434/4-A - RA	Lab Control Sample	105	
Surrogate Legend			
OTCSN = n-Octacosane (Surr)			

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 380-249512/21-A
Matrix: Water
Analysis Batch: 249987

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 249512

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
2,4'-DDD	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
2,4'-DDE	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
2,4'-DDT	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
2,4-Dinitrotoluene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
2,6-Dinitrotoluene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
2-Methylnaphthalene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
4,4'-DDD	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
4,4'-DDE	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
4,4'-DDT	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Acenaphthene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Acenaphthylene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Acetochlor	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Alachlor	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
alpha-BHC	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
alpha-Chlordane	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Anthracene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 14:47	1
Atrazine	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Benz(a)anthracene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Benzo[a]pyrene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 14:47	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 14:47	1
Benzo[g,h,i]perylene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 14:47	1
beta-BHC	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		08/28/26 10:18	08/31/26 14:47	1
Bromacil	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Butachlor	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Butylbenzylphthalate	<0.50		0.50	ug/L		08/28/26 10:18	08/31/26 14:47	1
Chlorobenzilate	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Chloroneb	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Chlorothalonil (Draconil, Bravo)	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Chlorpyrifos	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Chrysene	<0.020		0.020	ug/L		08/28/26 10:18	08/31/26 14:47	1
delta-BHC	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		08/28/26 10:18	08/31/26 14:47	1
Dibenz(a,h)anthracene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Diclorvos (DDVP)	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Dieldrin	<0.0099		0.0099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Diethylphthalate	<0.50		0.50	ug/L		08/28/26 10:18	08/31/26 14:47	1
Dimethylphthalate	<0.50		0.50	ug/L		08/28/26 10:18	08/31/26 14:47	1
Di-n-butyl phthalate	<0.99		0.99	ug/L		08/28/26 10:18	08/31/26 14:47	1
Di-n-octyl phthalate	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Endosulfan I (Alpha)	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Endosulfan II (Beta)	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Endosulfan sulfate	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Endrin	<0.0099		0.0099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Endrin aldehyde	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
EPTC	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 380-249512/21-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 249512

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluoranthene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Fluorene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
gamma-Chlordane	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Heptachlor	<0.0099		0.0099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Heptachlor epoxide (isomer B)	<0.0099		0.0099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Hexachlorobenzene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Hexachlorocyclopentadiene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Indeno[1,2,3-cd]pyrene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Isophorone	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Lindane	<0.0099		0.0099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Malathion	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Methoxychlor	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Metolachlor	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Molinate	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Naphthalene	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Parathion	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Pendimethalin (Penoxaline)	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Phenanthrene	<0.040		0.040	ug/L		08/28/26 10:18	08/31/26 14:47	1
Propachlor	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Pyrene	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Simazine	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Terbacil	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Terbutylazine	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Thiobencarb	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		08/28/26 10:18	08/31/26 14:47	1
trans-Nonachlor	<0.050		0.050	ug/L		08/28/26 10:18	08/31/26 14:47	1
Trifluralin	<0.099		0.099	ug/L		08/28/26 10:18	08/31/26 14:47	1

Tentatively Identified Compound	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Cyclopentene, 1,2,3,3,4-pentamethyl-	1.09	T J N	ug/L		2.58	197390-29-7	08/28/26 10:18	08/31/26 14:47	1
Undecane	7.10	T J N	ug/L		3.18	1120-21-4	08/28/26 10:18	08/31/26 14:47	1
9-Octadecenamide, (Z)-	4.49	T J N	ug/L		7.99	301-02-0	08/28/26 10:18	08/31/26 14:47	1
13-Docosenamide, (Z)-	1.39	T J N	ug/L		10.55	112-84-5	08/28/26 10:18	08/31/26 14:47	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	96		70 - 130	08/28/26 10:18	08/31/26 14:47	1
Perylene-d12	92		70 - 130	08/28/26 10:18	08/31/26 14:47	1
Triphenylphosphate	100		70 - 130	08/28/26 10:18	08/31/26 14:47	1

Lab Sample ID: LCS 380-249512/23-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
1-Methylnaphthalene	1.98	2.09		ug/L		105	70 - 130
2,4'-DDD	1.98	1.98		ug/L		100	70 - 130
2,4'-DDE	1.98	2.00		ug/L		101	70 - 130
2,4'-DDT	1.98	2.05		ug/L		103	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-249512/23-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2,4-Dinitrotoluene	1.98	2.08		ug/L		105	70 - 130
2,6-Dinitrotoluene	1.98	2.03		ug/L		102	70 - 130
2-Methylnaphthalene	1.98	2.07		ug/L		105	70 - 130
4,4'-DDD	1.98	2.06		ug/L		104	70 - 130
4,4'-DDE	1.98	2.03		ug/L		103	70 - 130
4,4'-DDT	1.98	2.04		ug/L		103	70 - 130
Acenaphthene	1.98	2.04		ug/L		103	70 - 130
Acenaphthylene	1.98	1.69		ug/L		85	70 - 130
Acetochlor	1.98	2.03		ug/L		102	70 - 130
Alachlor	1.98	2.05		ug/L		104	70 - 130
alpha-BHC	1.98	2.06		ug/L		104	70 - 130
alpha-Chlordane	1.98	2.01		ug/L		102	70 - 130
Anthracene	1.98	1.90		ug/L		96	70 - 130
Atrazine	1.98	2.12		ug/L		107	70 - 130
Benz(a)anthracene	1.98	2.05		ug/L		103	70 - 130
Benzo[a]pyrene	1.98	1.99		ug/L		101	70 - 130
Benzo[b]fluoranthene	1.98	2.03		ug/L		102	70 - 130
Benzo[g,h,i]perylene	1.98	1.98		ug/L		100	70 - 130
Benzo[k]fluoranthene	1.98	2.12		ug/L		107	70 - 130
beta-BHC	1.98	2.04		ug/L		103	70 - 130
Bis(2-ethylhexyl) phthalate	1.98	2.06		ug/L		104	70 - 130
Bromacil	1.98	1.89		ug/L		96	70 - 130
Butachlor	1.98	2.08		ug/L		105	70 - 130
Butylbenzylphthalate	1.98	2.29		ug/L		115	70 - 130
Chlorobenzilate	1.98	1.94		ug/L		98	70 - 130
Chloroneb	1.98	2.09		ug/L		106	70 - 130
Chlorothalonil (Draconil, Bravo)	1.98	2.12		ug/L		107	70 - 130
Chlorpyrifos	1.98	2.01		ug/L		101	70 - 130
Chrysene	1.98	2.07		ug/L		104	70 - 130
delta-BHC	1.98	2.08		ug/L		105	70 - 130
Di(2-ethylhexyl)adipate	1.98	2.17		ug/L		110	70 - 130
Dibenz(a,h)anthracene	1.98	1.92		ug/L		97	70 - 130
Diclorvos (DDVP)	1.98	1.99		ug/L		100	70 - 130
Dieldrin	1.98	2.00		ug/L		101	70 - 130
Diethylphthalate	1.98	2.16		ug/L		109	70 - 130
Dimethylphthalate	1.98	2.12		ug/L		107	70 - 130
Di-n-butyl phthalate	3.96	4.28		ug/L		108	70 - 130
Di-n-octyl phthalate	1.98	1.97		ug/L		99	70 - 130
Endosulfan I (Alpha)	1.98	1.98		ug/L		100	70 - 130
Endosulfan II (Beta)	1.98	2.01		ug/L		101	70 - 130
Endosulfan sulfate	1.98	2.04		ug/L		103	70 - 130
Endrin	1.98	2.03		ug/L		102	70 - 130
Endrin aldehyde	1.98	1.97		ug/L		100	60 - 130
EPTC	1.98	2.03		ug/L		102	70 - 130
Fluoranthene	1.98	2.14		ug/L		108	70 - 130
Fluorene	1.98	2.16		ug/L		109	70 - 130
gamma-Chlordane	1.98	1.98		ug/L		100	70 - 130
Heptachlor	1.98	1.95		ug/L		99	70 - 130
Heptachlor epoxide (isomer B)	1.98	2.02		ug/L		102	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-249512/23-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexachlorobenzene	1.98	1.99		ug/L		101	70 - 130
Hexachlorocyclopentadiene	1.98	1.97		ug/L		100	70 - 130
Indeno[1,2,3-cd]pyrene	1.98	1.97		ug/L		99	70 - 130
Isophorone	1.98	1.98		ug/L		100	70 - 130
Lindane	1.98	2.07		ug/L		105	70 - 130
Malathion	1.98	2.07		ug/L		105	70 - 130
Methoxychlor	1.98	2.12		ug/L		107	70 - 130
Metolachlor	1.98	2.13		ug/L		108	70 - 130
Molinate	1.98	2.08		ug/L		105	70 - 130
Naphthalene	1.98	1.98		ug/L		100	70 - 130
Parathion	1.98	2.11		ug/L		107	70 - 130
Pendimethalin (Penoxaline)	1.98	2.06		ug/L		104	70 - 130
Phenanthrene	1.98	2.09		ug/L		106	70 - 130
Propachlor	1.98	2.10		ug/L		106	70 - 130
Pyrene	1.98	2.12		ug/L		107	70 - 130
Simazine	1.98	2.06		ug/L		104	70 - 130
Terbacil	1.98	1.99		ug/L		101	70 - 130
Terbutylazine	1.98	2.11		ug/L		106	70 - 130
Thiobencarb	1.98	2.15		ug/L		108	70 - 130
trans-Nonachlor	1.98	1.98		ug/L		100	70 - 130
Trifluralin	1.98	1.95		ug/L		99	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2-Nitro-m-xylene	96		70 - 130
Perylene-d12	96		70 - 130
Triphenylphosphate	100		70 - 130

Lab Sample ID: LCSD 380-249512/24-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	1.98	2.09		ug/L		105	70 - 130	0	20
2,4'-DDD	1.98	2.03		ug/L		102	70 - 130	3	20
2,4'-DDE	1.98	2.04		ug/L		103	70 - 130	2	20
2,4'-DDT	1.98	2.11		ug/L		106	70 - 130	3	20
2,4-Dinitrotoluene	1.98	2.18		ug/L		110	70 - 130	5	20
2,6-Dinitrotoluene	1.98	2.15		ug/L		108	70 - 130	6	20
2-Methylnaphthalene	1.98	2.06		ug/L		104	70 - 130	1	20
4,4'-DDD	1.98	2.11		ug/L		107	70 - 130	3	20
4,4'-DDE	1.98	2.06		ug/L		104	70 - 130	1	20
4,4'-DDT	1.98	2.12		ug/L		107	70 - 130	4	20
Acenaphthene	1.98	2.07		ug/L		104	70 - 130	1	20
Acenaphthylene	1.98	1.85		ug/L		93	70 - 130	9	20
Acetochlor	1.98	2.06		ug/L		104	70 - 130	2	20
Alachlor	1.98	2.08		ug/L		105	70 - 130	1	20
alpha-BHC	1.98	2.09		ug/L		105	70 - 130	2	20
alpha-Chlordane	1.98	2.02		ug/L		102	70 - 130	1	20

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCSD 380-249512/24-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Anthracene	1.98	1.83		ug/L		92	70 - 130	4	20
Atrazine	1.98	2.17		ug/L		109	70 - 130	2	20
Benz(a)anthracene	1.98	2.10		ug/L		106	70 - 130	2	20
Benzo[a]pyrene	1.98	2.02		ug/L		102	70 - 130	1	20
Benzo[b]fluoranthene	1.98	2.05		ug/L		103	70 - 130	1	20
Benzo[g,h,i]perylene	1.98	1.98		ug/L		100	70 - 130	0	20
Benzo[k]fluoranthene	1.98	2.13		ug/L		107	70 - 130	0	20
beta-BHC	1.98	2.12		ug/L		107	70 - 130	4	20
Bis(2-ethylhexyl) phthalate	1.98	2.19		ug/L		110	70 - 130	6	20
Bromacil	1.98	2.04		ug/L		103	70 - 130	8	20
Butachlor	1.98	2.11		ug/L		106	70 - 130	1	20
Butylbenzylphthalate	1.98	2.23		ug/L		112	70 - 130	2	20
Chlorobenzilate	1.98	2.00		ug/L		101	70 - 130	3	20
Chloroneb	1.98	2.10		ug/L		106	70 - 130	0	20
Chlorothalonil (Draconil, Bravo)	1.98	2.16		ug/L		109	70 - 130	2	20
Chlorpyrifos	1.98	2.06		ug/L		104	70 - 130	2	20
Chrysene	1.98	2.06		ug/L		104	70 - 130	1	20
delta-BHC	1.98	2.13		ug/L		107	70 - 130	3	20
Di(2-ethylhexyl)adipate	1.98	2.36		ug/L		119	70 - 130	8	20
Dibenz(a,h)anthracene	1.98	2.03		ug/L		102	70 - 130	6	20
Diclorvos (DDVP)	1.98	2.03		ug/L		102	70 - 130	2	20
Dieldrin	1.98	2.03		ug/L		102	70 - 130	2	20
Diethylphthalate	1.98	2.21		ug/L		111	70 - 130	2	20
Dimethylphthalate	1.98	2.15		ug/L		109	70 - 130	1	20
Di-n-butyl phthalate	3.97	4.35		ug/L		110	70 - 130	1	20
Di-n-octyl phthalate	1.98	2.11		ug/L		106	70 - 130	7	20
Endosulfan I (Alpha)	1.98	2.02		ug/L		102	70 - 130	2	20
Endosulfan II (Beta)	1.98	2.05		ug/L		103	70 - 130	2	20
Endosulfan sulfate	1.98	2.06		ug/L		104	70 - 130	1	20
Endrin	1.98	2.07		ug/L		105	70 - 130	2	20
Endrin aldehyde	1.98	2.03		ug/L		102	60 - 130	3	20
EPTC	1.98	2.07		ug/L		104	70 - 130	2	20
Fluoranthene	1.98	2.14		ug/L		108	70 - 130	0	20
Fluorene	1.98	2.15		ug/L		108	70 - 130	1	20
gamma-Chlordane	1.98	2.03		ug/L		102	70 - 130	3	20
Heptachlor	1.98	1.99		ug/L		100	70 - 130	2	20
Heptachlor epoxide (isomer B)	1.98	2.04		ug/L		103	70 - 130	1	20
Hexachlorobenzene	1.98	1.98		ug/L		100	70 - 130	1	20
Hexachlorocyclopentadiene	1.98	1.93		ug/L		97	70 - 130	2	20
Indeno[1,2,3-cd]pyrene	1.98	2.00		ug/L		101	70 - 130	1	20
Isophorone	1.98	2.00		ug/L		101	70 - 130	1	20
Lindane	1.98	2.14		ug/L		108	70 - 130	3	20
Malathion	1.98	2.13		ug/L		107	70 - 130	2	20
Methoxychlor	1.98	2.16		ug/L		109	70 - 130	1	20
Metolachlor	1.98	2.17		ug/L		109	70 - 130	2	20
Molinate	1.98	2.12		ug/L		107	70 - 130	2	20
Naphthalene	1.98	1.98		ug/L		100	70 - 130	0	20
Parathion	1.98	2.16		ug/L		109	70 - 130	2	20
Pendimethalin (Penoxaline)	1.98	2.12		ug/L		107	70 - 130	3	20

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCSD 380-249512/24-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Phenanthrene	1.98	2.01		ug/L		101	70 - 130	4	20
Propachlor	1.98	2.16		ug/L		109	70 - 130	3	20
Pyrene	1.98	2.11		ug/L		107	70 - 130	0	20
Simazine	1.98	2.14		ug/L		108	70 - 130	4	20
Terbacil	1.98	2.12		ug/L		107	70 - 130	6	20
Terbutylazine	1.98	2.19		ug/L		110	70 - 130	4	20
Thiobencarb	1.98	2.18		ug/L		110	70 - 130	2	20
trans-Nonachlor	1.98	2.02		ug/L		102	70 - 130	2	20
Trifluralin	1.98	2.03		ug/L		102	70 - 130	4	20

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
2-Nitro-m-xylene	97		70 - 130
Perylene-d12	97		70 - 130
Triphenylphosphate	104		70 - 130

Lab Sample ID: MRL 380-249512/22-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.0995	0.0879	J	ug/L		88	50 - 150
2,4'-DDD	0.0995	0.0863	J	ug/L		87	50 - 150
2,4'-DDE	0.0995	0.0987	J	ug/L		99	50 - 150
2,4'-DDT	0.0995	0.101		ug/L		102	50 - 150
2,4-Dinitrotoluene	0.0995	0.102		ug/L		103	50 - 150
2,6-Dinitrotoluene	0.0995	0.106		ug/L		107	50 - 150
2-Methylnaphthalene	0.0995	0.0891	J	ug/L		90	50 - 150
4,4'-DDD	0.0995	0.105		ug/L		105	50 - 150
4,4'-DDE	0.0995	0.103		ug/L		103	50 - 150
4,4'-DDT	0.0995	0.110		ug/L		110	50 - 150
Acenaphthene	0.0995	0.0935	J	ug/L		94	50 - 150
Acenaphthylene	0.0995	0.0742	J	ug/L		75	50 - 150
Acetochlor	0.0995	0.102		ug/L		102	50 - 150
Alachlor	0.0497	0.0558		ug/L		112	50 - 150
alpha-BHC	0.0995	0.0968	J	ug/L		97	50 - 150
alpha-Chlordane	0.0249	<0.029		ug/L		114	50 - 150
Anthracene	0.0199	0.0191	J	ug/L		96	50 - 150
Atrazine	0.0497	0.0583		ug/L		117	50 - 150
Benz(a)anthracene	0.0497	0.0496	J	ug/L		100	50 - 150
Benzo[a]pyrene	0.0199	0.0180	J	ug/L		91	50 - 150
Benzo[b]fluoranthene	0.0199	0.0195	J	ug/L		98	50 - 150
Benzo[g,h,i]perylene	0.0497	0.0439	J	ug/L		88	50 - 150
Benzo[k]fluoranthene	0.0199	0.0212		ug/L		107	50 - 150
beta-BHC	0.0995	0.108		ug/L		108	50 - 150
Bis(2-ethylhexyl) phthalate	0.597	0.680		ug/L		114	50 - 150
Bromacil	0.0995	0.0981	J	ug/L		99	50 - 150
Butachlor	0.0497	0.0517		ug/L		104	50 - 150
Butylbenzylphthalate	0.497	0.601		ug/L		121	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-249512/22-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Chlorobenzilate	0.0995	0.0920	J	ug/L		92	50 - 150
Chloroneb	0.0995	0.107		ug/L		108	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0995	0.105		ug/L		106	50 - 150
Chlorpyrifos	0.0497	0.0482	J	ug/L		97	50 - 150
Chrysene	0.0199	0.0236		ug/L		119	50 - 150
delta-BHC	0.0995	0.101		ug/L		101	50 - 150
Di(2-ethylhexyl)adipate	0.597	0.707		ug/L		118	50 - 150
Dibenz(a,h)anthracene	0.0497	0.0450	J	ug/L		90	50 - 150
Diclorvos (DDVP)	0.0497	0.0501		ug/L		101	50 - 150
Dieldrin	0.00995	0.0110		ug/L		111	50 - 150
Diethylphthalate	0.497	0.584		ug/L		117	50 - 150
Dimethylphthalate	0.497	0.552		ug/L		111	50 - 150
Di-n-butyl phthalate	0.497	0.600	J	ug/L		121	49 - 243
Di-n-octyl phthalate	0.0995	0.0895	J	ug/L		90	50 - 150
Endosulfan I (Alpha)	0.0995	0.0955	J	ug/L		96	50 - 150
Endosulfan II (Beta)	0.0995	0.101		ug/L		101	50 - 150
Endosulfan sulfate	0.0995	0.112		ug/L		112	50 - 150
Endrin	0.00995	0.0101		ug/L		102	50 - 150
Endrin aldehyde	0.0995	0.104		ug/L		105	50 - 150
EPTC	0.0995	0.0903	J	ug/L		91	50 - 150
Fluoranthene	0.0995	0.0996		ug/L		100	50 - 150
Fluorene	0.0497	0.0508		ug/L		102	50 - 150
gamma-Chlordane	0.0249	0.0262	J	ug/L		105	50 - 150
Heptachlor	0.00995	0.0106		ug/L		107	50 - 150
Heptachlor epoxide (isomer B)	0.00995	0.0115		ug/L		116	50 - 150
Hexachlorobenzene	0.0497	0.0474	J	ug/L		95	50 - 150
Hexachlorocyclopentadiene	0.0497	0.0392	J	ug/L		79	50 - 150
Indeno[1,2,3-cd]pyrene	0.0497	0.0466	J	ug/L		94	50 - 150
Isophorone	0.0995	0.0917	J	ug/L		92	50 - 150
Lindane	0.00995	0.0127		ug/L		128	50 - 150
Malathion	0.0995	0.0984	J	ug/L		99	50 - 150
Methoxychlor	0.0497	0.0522		ug/L		105	50 - 150
Metolachlor	0.0497	0.0510		ug/L		103	50 - 150
Molinate	0.0995	0.110		ug/L		111	50 - 150
Naphthalene	0.0995	0.0882	J	ug/L		89	50 - 150
Parathion	0.0995	0.0882	J	ug/L		89	50 - 150
Pendimethalin (Penoxaline)	0.0995	0.0937	J	ug/L		94	50 - 150
Phenanthrene	0.0398	0.0394	J	ug/L		99	50 - 150
Propachlor	0.0497	0.0524		ug/L		105	50 - 150
Pyrene	0.0497	0.0518		ug/L		104	50 - 150
Simazine	0.0497	0.0503		ug/L		101	50 - 150
Terbacil	0.0995	0.101		ug/L		102	50 - 150
Terbuthylazine	0.0995	0.108		ug/L		109	50 - 150
Thiobencarb	0.0995	0.107		ug/L		108	50 - 150
trans-Nonachlor	0.0249	0.0264	J	ug/L		106	50 - 150
Trifluralin	0.0995	0.0984	J	ug/L		99	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-249512/22-A

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 249512

Surrogate	MRL %Recovery	MRL Qualifier	Limits
2-Nitro-m-xylene	95		70 - 130
Perylene-d12	90		70 - 130
Triphenylphosphate	98		70 - 130

Lab Sample ID: 380-233443-B-3-A MS

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.096		1.92	2.03		ug/L		106	70 - 130
2,4'-DDD	<0.096		1.92	1.91		ug/L		99	70 - 130
2,4'-DDE	<0.096		1.92	1.91		ug/L		100	70 - 130
2,4'-DDT	<0.096		1.92	1.98		ug/L		103	70 - 130
2,4-Dinitrotoluene	<0.096		1.92	2.12		ug/L		110	70 - 130
2,6-Dinitrotoluene	<0.096		1.92	2.11		ug/L		110	70 - 130
2-Methylnaphthalene	<0.096		1.92	2.01		ug/L		104	70 - 130
4,4'-DDD	<0.096		1.92	2.01		ug/L		105	70 - 130
4,4'-DDE	<0.096		1.92	1.92		ug/L		100	70 - 130
4,4'-DDT	<0.096		1.92	1.98		ug/L		103	70 - 130
Acenaphthene	<0.096		1.92	1.97		ug/L		102	70 - 130
Acenaphthylene	<0.096		1.92	1.85		ug/L		96	70 - 130
Acetochlor	<0.096		1.92	1.97		ug/L		102	70 - 130
Alachlor	<0.048		1.92	2.00		ug/L		104	70 - 130
alpha-BHC	<0.096		1.92	1.97		ug/L		103	70 - 130
alpha-Chlordane	<0.048		1.92	1.93		ug/L		101	70 - 130
Anthracene	<0.019		1.92	1.48		ug/L		77	70 - 130
Atrazine	<0.048		1.92	1.95		ug/L		102	70 - 130
Benz(a)anthracene	<0.048		1.92	1.96		ug/L		102	70 - 130
Benzo[a]pyrene	<0.019		1.92	1.88		ug/L		98	70 - 130
Benzo[b]fluoranthene	<0.019		1.92	2.08		ug/L		108	70 - 130
Benzo[g,h,i]perylene	<0.048		1.92	1.98		ug/L		103	70 - 130
Benzo[k]fluoranthene	<0.019		1.92	2.02		ug/L		105	70 - 130
beta-BHC	<0.096		1.92	1.98		ug/L		103	70 - 130
Bis(2-ethylhexyl) phthalate	<0.58		1.92	2.14		ug/L		111	70 - 130
Bromacil	<0.096		1.92	2.08		ug/L		109	70 - 130
Butachlor	<0.048		1.92	2.04		ug/L		106	70 - 130
Butylbenzylphthalate	<0.48		1.92	2.22		ug/L		115	70 - 130
Chlorobenzilate	<0.096		1.92	2.02		ug/L		105	70 - 130
Chloroneb	<0.096		1.92	2.00		ug/L		104	70 - 130
Chlorothalonil (Draconil, Bravo)	<0.096		1.92	2.08		ug/L		108	70 - 130
Chlorpyrifos	<0.048		1.92	1.93		ug/L		100	70 - 130
Chrysene	<0.019		1.92	2.02		ug/L		105	70 - 130
delta-BHC	<0.096		1.92	1.99		ug/L		103	70 - 130
Di(2-ethylhexyl)adipate	<0.58		1.92	2.17		ug/L		113	70 - 130
Dibenz(a,h)anthracene	<0.048		1.92	2.00		ug/L		104	70 - 130
Diclorvos (DDVP)	<0.048		1.92	1.98		ug/L		103	70 - 130
Dieldrin	<0.0096		1.92	1.95		ug/L		102	70 - 130
Diethylphthalate	<0.48		1.92	2.11		ug/L		110	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-233443-B-3-A MS

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Dimethylphthalate	<0.48		1.92	2.10		ug/L		109	70 - 130
Di-n-butyl phthalate	<0.96		3.84	4.18		ug/L		104	70 - 130
Di-n-octyl phthalate	<0.096		1.92	2.02		ug/L		105	70 - 130
Endosulfan I (Alpha)	<0.096		1.92	1.94		ug/L		101	70 - 130
Endosulfan II (Beta)	<0.096		1.92	1.98		ug/L		103	70 - 130
Endosulfan sulfate	<0.096		1.92	1.96		ug/L		102	70 - 130
Endrin	<0.0096		1.92	1.97		ug/L		102	70 - 130
Endrin aldehyde	<0.096		1.92	1.50		ug/L		78	60 - 130
EPTC	<0.096		1.92	2.02		ug/L		105	70 - 130
Fluoranthene	<0.096		1.92	2.07		ug/L		108	70 - 130
Fluorene	<0.048		1.92	2.08		ug/L		108	70 - 130
gamma-Chlordane	<0.048		1.92	1.94		ug/L		101	70 - 130
Heptachlor	<0.0096		1.92	1.94		ug/L		101	70 - 130
Heptachlor epoxide (isomer B)	<0.0096		1.92	1.97		ug/L		103	70 - 130
Hexachlorobenzene	<0.048		1.92	1.91		ug/L		99	70 - 130
Hexachlorocyclopentadiene	<0.048		1.92	1.93		ug/L		101	70 - 130
Indeno[1,2,3-cd]pyrene	<0.048		1.92	2.01		ug/L		104	70 - 130
Isophorone	<0.096		1.92	1.92		ug/L		100	70 - 130
Lindane	<0.0096		1.92	2.01		ug/L		104	70 - 130
Malathion	<0.096		1.92	2.00		ug/L		104	70 - 130
Methoxychlor	<0.048		1.92	2.11		ug/L		110	70 - 130
Metolachlor	<0.048		1.92	2.08		ug/L		108	70 - 130
Molinate	<0.096		1.92	2.05		ug/L		107	70 - 130
Naphthalene	<0.096		1.92	1.92		ug/L		100	70 - 130
Parathion	<0.096		1.92	2.09		ug/L		109	70 - 130
Pendimethalin (Penoxaline)	<0.096		1.92	2.07		ug/L		108	70 - 130
Phenanthrene	<0.038		1.92	1.98		ug/L		103	70 - 130
Propachlor	<0.048		1.92	2.10		ug/L		109	70 - 130
Pyrene	<0.048		1.92	2.02		ug/L		105	70 - 130
Simazine	<0.048		1.92	1.96		ug/L		102	70 - 130
Terbacil	<0.096		1.92	2.05		ug/L		107	70 - 130
Terbutylazine	<0.096		1.92	1.97		ug/L		103	70 - 130
Thiobencarb	<0.096		1.92	2.10		ug/L		109	70 - 130
trans-Nonachlor	<0.048		1.92	1.95		ug/L		101	70 - 130
Trifluralin	<0.096		1.92	1.97		ug/L		103	70 - 130

Surrogate	MS %Recovery	MS Qualifier	Limits
2-Nitro-m-xylene	97		70 - 130
Perylene-d12	98		70 - 130
Triphenylphosphate	102		70 - 130

Lab Sample ID: 380-233458-V-1-A DU

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	RPD Limit
1-Methylnaphthalene	<0.096		<0.096		ug/L		NC	20
2,4'-DDD	<0.096		<0.096		ug/L		NC	20

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-233458-V-1-A DU

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
2,4'-DDE	<0.096		<0.096		ug/L		NC	20
2,4'-DDT	<0.096		<0.096		ug/L		NC	20
2,4-Dinitrotoluene	<0.096		<0.096		ug/L		NC	20
2,6-Dinitrotoluene	<0.096		<0.096		ug/L		NC	20
2-Methylnaphthalene	<0.096		<0.096		ug/L		NC	20
4,4'-DDD	<0.096		<0.096		ug/L		NC	20
4,4'-DDE	<0.096		<0.096		ug/L		NC	20
4,4'-DDT	<0.096		<0.096		ug/L		NC	20
Acenaphthene	<0.096		<0.096		ug/L		NC	20
Acenaphthylene	<0.096		<0.096		ug/L		NC	20
Acetochlor	<0.096		<0.096		ug/L		NC	20
Alachlor	<0.048		<0.048		ug/L		NC	20
alpha-BHC	<0.096		<0.096		ug/L		NC	20
alpha-Chlordane	<0.048		<0.048		ug/L		NC	20
Anthracene	<0.019		<0.019		ug/L		NC	20
Atrazine	<0.048		<0.048		ug/L		NC	20
Benz(a)anthracene	<0.048		<0.048		ug/L		NC	20
Benzo[a]pyrene	<0.019		<0.019		ug/L		NC	20
Benzo[b]fluoranthene	<0.019		<0.019		ug/L		NC	20
Benzo[g,h,i]perylene	<0.048		<0.048		ug/L		NC	20
Benzo[k]fluoranthene	<0.019		<0.019		ug/L		NC	20
beta-BHC	<0.096		<0.096		ug/L		NC	20
Bis(2-ethylhexyl) phthalate	<0.58		<0.57		ug/L		NC	20
Bromacil	<0.096		<0.096		ug/L		NC	20
Butachlor	<0.048		<0.048		ug/L		NC	20
Butylbenzylphthalate	<0.48		<0.48		ug/L		NC	20
Chlorobenzilate	<0.096		<0.096		ug/L		NC	20
Chloroneb	<0.096		<0.096		ug/L		NC	20
Chlorothalonil (Draconil, Bravo)	<0.096		<0.096		ug/L		NC	20
Chlorpyrifos	<0.048		<0.048		ug/L		NC	20
Chrysene	<0.019		<0.019		ug/L		NC	20
delta-BHC	<0.096		<0.096		ug/L		NC	20
Di(2-ethylhexyl)adipate	<0.58		<0.57		ug/L		NC	20
Dibenz(a,h)anthracene	<0.048		<0.048		ug/L		NC	20
Diclorvos (DDVP)	<0.048		<0.048		ug/L		NC	20
Dieldrin	<0.0096		<0.0096		ug/L		NC	20
Diethylphthalate	<0.48		<0.48		ug/L		NC	20
Dimethylphthalate	<0.48		<0.48		ug/L		NC	20
Di-n-butyl phthalate	<0.96		<0.96		ug/L		NC	20
Di-n-octyl phthalate	<0.096		<0.096		ug/L		NC	20
Endosulfan I (Alpha)	<0.096		<0.096		ug/L		NC	20
Endosulfan II (Beta)	<0.096		<0.096		ug/L		NC	20
Endosulfan sulfate	<0.096		<0.096		ug/L		NC	20
Endrin	<0.0096		<0.0096		ug/L		NC	20
Endrin aldehyde	<0.096		<0.096		ug/L		NC	20
EPTC	<0.096		<0.096		ug/L		NC	20
Fluoranthene	<0.096		<0.096		ug/L		NC	20
Fluorene	<0.048		<0.048		ug/L		NC	20
gamma-Chlordane	<0.048		<0.048		ug/L		NC	20

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-233458-V-1-A DU

Matrix: Water

Analysis Batch: 249987

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 249512

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
Heptachlor	<0.0096		<0.0096		ug/L		NC	20
Heptachlor epoxide (isomer B)	<0.0096		<0.0096		ug/L		NC	20
Hexachlorobenzene	<0.048		<0.048		ug/L		NC	20
Hexachlorocyclopentadiene	<0.048		<0.048		ug/L		NC	20
Indeno[1,2,3-cd]pyrene	<0.048		<0.048		ug/L		NC	20
Isophorone	<0.096		<0.096		ug/L		NC	20
Lindane	<0.0096		<0.0096		ug/L		NC	20
Malathion	<0.096		<0.096		ug/L		NC	20
Methoxychlor	<0.048		<0.048		ug/L		NC	20
Metolachlor	<0.048		<0.048		ug/L		NC	20
Molinate	<0.096		<0.096		ug/L		NC	20
Naphthalene	<0.096		<0.096		ug/L		NC	20
Parathion	<0.096		<0.096		ug/L		NC	20
Pendimethalin (Penoxaline)	<0.096		<0.096		ug/L		NC	20
Phenanthrene	<0.038		<0.038		ug/L		NC	20
Propachlor	<0.048		<0.048		ug/L		NC	20
Pyrene	<0.048		<0.048		ug/L		NC	20
Simazine	<0.048		<0.048		ug/L		NC	20
Terbacil	<0.096		<0.096		ug/L		NC	20
Terbuthylazine	<0.096		<0.096		ug/L		NC	20
Thiobencarb	<0.096		<0.096		ug/L		NC	20
Total Permethrin (mixed isomers)	<0.19		<0.19		ug/L		NC	20
trans-Nonachlor	<0.048		<0.048		ug/L		NC	20
Trifluralin	<0.096		<0.096		ug/L		NC	20

Surrogate	%Recovery	Qualifier	Limits
2-Nitro-m-xylene	98		70 - 130
Perylene-d12	90		70 - 130
Triphenylphosphate	99		70 - 130

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 570-791672/1-A

Matrix: Water

Analysis Batch: 797504

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 791672

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	08/28/26 06:39	09/09/26 17:05	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	67		33 - 139	08/28/26 06:39	09/09/26 17:05	1
2-Fluorobiphenyl (Surr)	69		33 - 126	08/28/26 06:39	09/09/26 17:05	1
2-Fluorophenol (Surr)	49		12 - 120	08/28/26 06:39	09/09/26 17:05	1
Nitrobenzene-d5 (Surr)	71		36 - 120	08/28/26 06:39	09/09/26 17:05	1
Phenol-d6 (Surr)	24		10 - 120	08/28/26 06:39	09/09/26 17:05	1
p-Terphenyl-d14 (Surr)	90		47 - 131	08/28/26 06:39	09/09/26 17:05	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Lab Sample ID: MB 570-791672/1-A

Matrix: Water

Analysis Batch: 794330

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 791672

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
2-Methylnaphthalene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Acenaphthene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Acenaphthylene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Anthracene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Benzo[a]anthracene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Benzo[a]pyrene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Benzo[b]fluoranthene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Benzo[g,h,i]perylene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Benzo[k]fluoranthene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Chrysene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Dibenz(a,h)anthracene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Fluoranthene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Fluorene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Indeno[1,2,3-cd]pyrene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Naphthalene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Phenanthrene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1
Pyrene	<0.20		0.20	ug/L		08/28/26 06:39	09/02/26 13:16	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	86		28 - 127	08/28/26 06:39	09/02/26 13:16	1
2-Fluorobiphenyl (Surr)	73		31 - 120	08/28/26 06:39	09/02/26 13:16	1
2-Fluorophenol (Surr)	50		17 - 120	08/28/26 06:39	09/02/26 13:16	1
Nitrobenzene-d5 (Surr)	75		27 - 120	08/28/26 06:39	09/02/26 13:16	1
Phenol-d6 (Surr)	31		10 - 120	08/28/26 06:39	09/02/26 13:16	1
p-Terphenyl-d14 (Surr)	85		45 - 120	08/28/26 06:39	09/02/26 13:16	1

Lab Sample ID: LCS 570-791672/2-A

Matrix: Water

Analysis Batch: 794330

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 791672

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	20.0	16.8		ug/L		84	47 - 120
2-Methylnaphthalene	20.0	16.4		ug/L		82	43 - 120
Acenaphthene	20.0	16.9		ug/L		85	60 - 132
Acenaphthylene	20.0	15.6		ug/L		78	54 - 126
Anthracene	20.0	17.5		ug/L		87	43 - 120
Benzo[a]anthracene	20.0	18.6		ug/L		93	42 - 133
Benzo[a]pyrene	20.0	18.6		ug/L		93	32 - 148
Benzo[b]fluoranthene	20.0	19.0		ug/L		95	42 - 140
Benzo[g,h,i]perylene	20.0	25.0		ug/L		125	1 - 195
Benzo[k]fluoranthene	20.0	18.2		ug/L		91	25 - 146
Chrysene	20.0	18.9		ug/L		95	44 - 140
Dibenz(a,h)anthracene	20.0	26.9		ug/L		134	1 - 200
Fluoranthene	20.0	18.5		ug/L		92	43 - 121
Fluorene	20.0	16.9		ug/L		84	70 - 120
Indeno[1,2,3-cd]pyrene	20.0	25.4		ug/L		127	1 - 151
Naphthalene	20.0	17.3		ug/L		86	36 - 120

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCS 570-791672/2-A

Matrix: Water

Analysis Batch: 794330

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 791672

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Phenanthrene	20.0	18.2		ug/L		91	65 - 120
Pyrene	20.0	18.7		ug/L		93	70 - 120

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2,4,6-Tribromophenol (Surr)	89		28 - 127
2-Fluorobiphenyl (Surr)	84		31 - 120
2-Fluorophenol (Surr)	65		17 - 120
Nitrobenzene-d5 (Surr)	85		27 - 120
Phenol-d6 (Surr)	41		10 - 120
p-Terphenyl-d14 (Surr)	89		45 - 120

Lab Sample ID: LCSD 570-791672/3-A

Matrix: Water

Analysis Batch: 794330

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 791672

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	20.0	12.5	*1	ug/L		62	47 - 120	30	20
2-Methylnaphthalene	20.0	12.1	*1	ug/L		60	43 - 120	30	20
Acenaphthene	20.0	16.4		ug/L		82	60 - 132	3	29
Acenaphthylene	20.0	16.6		ug/L		83	54 - 126	6	45
Anthracene	20.0	16.5		ug/L		82	43 - 120	6	40
Benzo[a]anthracene	20.0	17.7		ug/L		88	42 - 133	5	32
Benzo[a]pyrene	20.0	18.6		ug/L		93	32 - 148	0	43
Benzo[b]fluoranthene	20.0	17.6		ug/L		88	42 - 140	7	43
Benzo[g,h,i]perylene	20.0	17.7		ug/L		89	1 - 195	34	61
Benzo[k]fluoranthene	20.0	17.1		ug/L		85	25 - 146	7	38
Chrysene	20.0	17.7		ug/L		88	44 - 140	7	53
Dibenz(a,h)anthracene	20.0	18.3		ug/L		92	1 - 200	38	75
Fluoranthene	20.0	17.3		ug/L		87	43 - 121	7	40
Fluorene	20.0	16.9		ug/L		84	70 - 120	0	23
Indeno[1,2,3-cd]pyrene	20.0	17.7		ug/L		89	1 - 151	35	60
Naphthalene	20.0	11.9		ug/L		59	36 - 120	37	39
Phenanthrene	20.0	17.0		ug/L		85	65 - 120	7	24
Pyrene	20.0	18.2		ug/L		91	70 - 120	3	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
2,4,6-Tribromophenol (Surr)	84		28 - 127
2-Fluorobiphenyl (Surr)	86		31 - 120
2-Fluorophenol (Surr)	61		17 - 120
Nitrobenzene-d5 (Surr)	64		27 - 120
Phenol-d6 (Surr)	39		10 - 120
p-Terphenyl-d14 (Surr)	86		45 - 120

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: 380-233060-A-1-A MS

Matrix: Water

Analysis Batch: 794330

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 791672

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.19	*1	19.4	14.5		ug/L		75	36 - 120
2-Methylnaphthalene	<0.19	*1	19.4	14.2		ug/L		73	32 - 124
Acenaphthene	<0.19		19.4	14.3		ug/L		74	47 - 145
Acenaphthylene	<0.19		19.4	13.8		ug/L		71	33 - 145
Anthracene	<0.19		19.4	14.5		ug/L		75	27 - 133
Benzo[a]anthracene	<0.19		19.4	16.0		ug/L		82	33 - 143
Benzo[a]pyrene	<0.19		19.4	16.7		ug/L		86	17 - 163
Benzo[b]fluoranthene	<0.19		19.4	15.8		ug/L		81	24 - 159
Benzo[g,h,i]perylene	<0.19		19.4	15.8		ug/L		82	1 - 219
Benzo[k]fluoranthene	<0.19		19.4	15.5		ug/L		80	11 - 162
Chrysene	<0.19		19.4	15.8		ug/L		82	17 - 168
Dibenz(a,h)anthracene	<0.19		19.4	16.4		ug/L		85	1 - 227
Fluoranthene	<0.19		19.4	15.6		ug/L		80	26 - 137
Fluorene	<0.19		19.4	14.5		ug/L		75	59 - 121
Indeno[1,2,3-cd]pyrene	<0.19		19.4	16.1		ug/L		83	1 - 171
Naphthalene	<0.19		19.4	14.1		ug/L		73	21 - 133
Phenanthrene	<0.19		19.4	14.9		ug/L		77	54 - 120
Pyrene	<0.19		19.4	16.3		ug/L		84	52 - 120

Surrogate	MS %Recovery	MS Qualifier	Limits
2,4,6-Tribromophenol (Surr)	77		28 - 127
2-Fluorobiphenyl (Surr)	74		31 - 120
2-Fluorophenol (Surr)	49		17 - 120
Nitrobenzene-d5 (Surr)	72		27 - 120
Phenol-d6 (Surr)	31		10 - 120
p-Terphenyl-d14 (Surr)	79		45 - 120

Lab Sample ID: 380-233060-A-1-B MSD

Matrix: Water

Analysis Batch: 794330

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 791672

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	<0.19	*1	19.3	16.1		ug/L		84	36 - 120	11	30
2-Methylnaphthalene	<0.19	*1	19.3	15.9		ug/L		82	32 - 124	11	30
Acenaphthene	<0.19		19.3	15.4		ug/L		80	47 - 145	8	48
Acenaphthylene	<0.19		19.3	15.2		ug/L		79	33 - 145	9	74
Anthracene	<0.19		19.3	15.5		ug/L		80	27 - 133	7	66
Benzo[a]anthracene	<0.19		19.3	16.9		ug/L		88	33 - 143	6	53
Benzo[a]pyrene	<0.19		19.3	18.1		ug/L		94	17 - 163	8	72
Benzo[b]fluoranthene	<0.19		19.3	17.0		ug/L		88	24 - 159	7	71
Benzo[g,h,i]perylene	<0.19		19.3	17.2		ug/L		89	1 - 219	8	97
Benzo[k]fluoranthene	<0.19		19.3	16.5		ug/L		86	11 - 162	6	63
Chrysene	<0.19		19.3	16.7		ug/L		87	17 - 168	6	87
Dibenz(a,h)anthracene	<0.19		19.3	17.7		ug/L		92	1 - 227	7	126
Fluoranthene	<0.19		19.3	16.7		ug/L		87	26 - 137	7	66
Fluorene	<0.19		19.3	15.9		ug/L		83	59 - 121	9	38
Indeno[1,2,3-cd]pyrene	<0.19		19.3	17.5		ug/L		91	1 - 171	8	99
Naphthalene	<0.19		19.3	16.0		ug/L		83	21 - 133	13	65

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: 380-233060-A-1-B MSD

Matrix: Water

Analysis Batch: 794330

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 791672

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Phenanthrene	<0.19		19.3	16.0		ug/L		83	54 - 120	7	39
Pyrene	<0.19		19.3	17.1		ug/L		89	52 - 120	5	49

Surrogate	MSD %Recovery	MSD Qualifier	MSD Limits
2,4,6-Tribromophenol (Surr)	82		28 - 127
2-Fluorobiphenyl (Surr)	81		31 - 120
2-Fluorophenol (Surr)	58		17 - 120
Nitrobenzene-d5 (Surr)	80		27 - 120
Phenol-d6 (Surr)	37		10 - 120
p-Terphenyl-d14 (Surr)	85		45 - 120

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Lab Sample ID: MB 570-795451/6

Matrix: Water

Analysis Batch: 795451

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			09/04/26 10:59	1

Surrogate	MB %Recovery	MB Qualifier	MB Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	86		38 - 134		09/04/26 10:59	1

Lab Sample ID: LCS 570-795451/3

Matrix: Water

Analysis Batch: 795451

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	400	373		ug/L		93	78 - 120

Surrogate	LCS %Recovery	LCS Qualifier	LCS Limits
4-Bromofluorobenzene (Surr)	83		38 - 134

Lab Sample ID: LCSD 570-795451/4

Matrix: Water

Analysis Batch: 795451

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	400	366		ug/L		91	78 - 120	2	10

Surrogate	LCSD %Recovery	LCSD Qualifier	LCSD Limits
4-Bromofluorobenzene (Surr)	88		38 - 134

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

Lab Sample ID: MRL 570-795451/5

Matrix: Water

Analysis Batch: 795451

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

			Spike	MRL	MRL				%Rec		
Analyte			Added	Result	Qualifier	Unit	D	%Rec	Limits		
Gasoline Range Organics (C4-C13)			10.0	13.5		ug/L		135	50 - 150		
Surrogate	MRL	MRL									
	%Recovery	Qualifier	Limits								
4-Bromofluorobenzene (Surr)	89		38 - 134								

Lab Sample ID: 380-233271-B-1 MS

Matrix: Water

Analysis Batch: 795451

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	<10		400	421		ug/L		105	68 - 122
Surrogate	MS %Recovery	MS Qualifier	Limits						
4-Bromofluorobenzene (Surr)	93		38 - 134						

Lab Sample ID: 380-233271-E-1 MSD

Matrix: Water

Analysis Batch: 795451

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

	Sample	Sample	Spike	MSD	MSD				%Rec		RPD
Analyte	Result	Qualifier	Added	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit
Gasoline Range Organics (C4-C13)	<10		400	374		ug/L		94	68 - 122	12	18
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
4-Bromofluorobenzene (Surr)	92		38 - 134								

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Lab Sample ID: MB 570-791434/1-A

Matrix: Water

Analysis Batch: 796133

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 791434

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Diesel Range Organics (C10-C24)	<25		25	ug/L		08/27/26 15:57	09/06/26 01:02	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		08/27/26 15:57	09/06/26 01:02	1
C8-C18	<25		25	ug/L		08/27/26 15:57	09/06/26 01:02	1
Surrogate	%Recovery	MB	MB			Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	83			Limits		08/27/26 15:57	09/06/26 01:02	1

Lab Sample ID: LCS 570-791434/2-A

Matrix: Water

Analysis Batch: 796133

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 791434

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	1600	1110	^3+	ug/L		69	56 - 127

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

Lab Sample ID: LCS 570-791434/2-A

Matrix: Water

Analysis Batch: 796133

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 791434

	LCS	LCS	
Surrogate	%Recovery	Qualifier	Limits
n-Octacosane (Surr)	93		60 - 130

Lab Sample ID: LCSD 570-791434/3-A

Matrix: Water

Analysis Batch: 796133

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 791434

	Spike	LCSD	LCSD						%Rec		RPD	
Analyte	Added	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit			
C10-C28	1600	1150	^3+	ug/L		72	56 - 127	3	23			

	LCSD	LCSD	
Surrogate	%Recovery	Qualifier	Limits
n-Octacosane (Surr)	91		60 - 130

Lab Sample ID: 380-233060-C-1-C MS

Matrix: Water

Analysis Batch: 796133

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 791434

	Sample	Sample	Spike	MS	MS				%Rec			
Analyte	Result	Qualifier	Added	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit	
C10-C28	<26	^3+	1630	1020	F1	ug/L		63	70 - 130			

	MS	MS	
Surrogate	%Recovery	Qualifier	Limits
n-Octacosane (Surr)	82		60 - 130

Lab Sample ID: 380-233060-C-1-D MSD

Matrix: Water

Analysis Batch: 796133

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 791434

	Sample	Sample	Spike	MSD	MSD				%Rec		RPD	
Analyte	Result	Qualifier	Added	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit	
C10-C28	<26	^3+	1670	1160	F1	ug/L		69	70 - 130	12	20	

	MSD	MSD	
Surrogate	%Recovery	Qualifier	Limits
n-Octacosane (Surr)	92		60 - 130

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level - RA

Lab Sample ID: MRL 570-791434/4-A

Matrix: Water

Analysis Batch: 796453

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 791434

	Spike	MRL	MRL						%Rec			
Analyte	Added	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit			
C10-C28 - RA	0.0200	0.0337	^3+	mg/L		169	50 - 150					

	MRL	MRL	
Surrogate	%Recovery	Qualifier	Limits
n-Octacosane (Surr) - RA	105		60 - 130

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

GC/MS Semi VOA

Prep Batch: 249512

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	525.2	
MB 380-249512/21-A	Method Blank	Total/NA	Water	525.2	
LCS 380-249512/23-A	Lab Control Sample	Total/NA	Water	525.2	
LCSD 380-249512/24-A	Lab Control Sample Dup	Total/NA	Water	525.2	
MRL 380-249512/22-A	Lab Control Sample	Total/NA	Water	525.2	
380-233443-B-3-A MS	Matrix Spike	Total/NA	Water	525.2	
380-233458-V-1-A DU	Duplicate	Total/NA	Water	525.2	

Analysis Batch: 249987

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	525.2	249512
MB 380-249512/21-A	Method Blank	Total/NA	Water	525.2	249512
LCS 380-249512/23-A	Lab Control Sample	Total/NA	Water	525.2	249512
LCSD 380-249512/24-A	Lab Control Sample Dup	Total/NA	Water	525.2	249512
MRL 380-249512/22-A	Lab Control Sample	Total/NA	Water	525.2	249512
380-233443-B-3-A MS	Matrix Spike	Total/NA	Water	525.2	249512
380-233458-V-1-A DU	Duplicate	Total/NA	Water	525.2	249512

Prep Batch: 791672

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	625.1	
MB 570-791672/1-A	Method Blank	Total/NA	Water	625.1	
LCS 570-791672/2-A	Lab Control Sample	Total/NA	Water	625.1	
LCSD 570-791672/3-A	Lab Control Sample Dup	Total/NA	Water	625.1	
380-233060-A-1-A MS	Matrix Spike	Total/NA	Water	625.1	
380-233060-A-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	625.1	

Analysis Batch: 794330

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	625.1 SIM	791672
MB 570-791672/1-A	Method Blank	Total/NA	Water	625.1 SIM	791672
LCS 570-791672/2-A	Lab Control Sample	Total/NA	Water	625.1 SIM	791672
LCSD 570-791672/3-A	Lab Control Sample Dup	Total/NA	Water	625.1 SIM	791672
380-233060-A-1-A MS	Matrix Spike	Total/NA	Water	625.1 SIM	791672
380-233060-A-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	625.1 SIM	791672

Analysis Batch: 797504

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	625.1	791672
MB 570-791672/1-A	Method Blank	Total/NA	Water	625.1	791672

GC VOA

Analysis Batch: 795451

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	8015B GRO LL	
380-233066-2	TB: Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	8015B GRO LL	
MB 570-795451/6	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-795451/3	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-795451/4	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	
MRL 570-795451/5	Lab Control Sample	Total/NA	Water	8015B GRO LL	

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

GC VOA (Continued)

Analysis Batch: 795451 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233271-B-1 MS	Matrix Spike	Total/NA	Water	8015B GRO LL	
380-233271-E-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8015B GRO LL	

GC Semi VOA

Prep Batch: 791434

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	3510C	
MB 570-791434/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-791434/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-791434/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-791434/4-A - RA	Lab Control Sample	Total/NA	Water	3510C	
380-233060-C-1-C MS	Matrix Spike	Total/NA	Water	3510C	
380-233060-C-1-D MSD	Matrix Spike Duplicate	Total/NA	Water	3510C	

Analysis Batch: 796133

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Total/NA	Water	8015B	791434
MB 570-791434/1-A	Method Blank	Total/NA	Water	8015B	791434
LCS 570-791434/2-A	Lab Control Sample	Total/NA	Water	8015B	791434
LCSD 570-791434/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	791434
380-233060-C-1-C MS	Matrix Spike	Total/NA	Water	8015B	791434
380-233060-C-1-D MSD	Matrix Spike Duplicate	Total/NA	Water	8015B	791434

Analysis Batch: 796453

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
MRL 570-791434/4-A - RA	Lab Control Sample	Total/NA	Water	8015B	791434

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Client Sample ID: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-1

Date Collected: 08/24/26 12:48

Matrix: Water

Date Received: 08/26/26 10:12

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			249512	L9UA	EA POM	08/28/26 10:18
Total/NA	Analysis	525.2		1	249987	Q8LA	EA POM	08/31/26 19:55
Total/NA	Prep	625.1			791672	H1SH	EET CAL 4	08/28/26 06:39
Total/NA	Analysis	625.1		1	797504	S4EA	EET CAL 4	09/09/26 19:41
Total/NA	Prep	625.1			791672	H1SH	EET CAL 4	08/28/26 06:39
Total/NA	Analysis	625.1 SIM		1	794330	PQS1	EET CAL 4	09/02/26 18:04
Total/NA	Analysis	8015B GRO LL		1	795451	7CF2	EET CAL 4	09/04/26 19:37
Total/NA	Prep	3510C			791434	ATRL	EET CAL 4	08/27/26 15:57
Total/NA	Analysis	8015B		1	796133	H6FE	EET CAL 4	09/06/26 05:00

Client Sample ID: TB: Ka'amilo Wells P1 (331-031-WL008)

Lab Sample ID: 380-233066-2

Date Collected: 08/24/26 12:48

Matrix: Water

Date Received: 08/26/26 10:12

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	795451	7CF2	EET CAL 4	09/04/26 14:16

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Laboratory: Eurofins Pomona

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	09-03-26
The following analytes are included in this report, but the laboratory is not certified by Hawaii State CA00006. This list may include analytes for which the agency does not offer certification :			
Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Water	1-Methylnaphthalene
525.2	525.2	Water	2,4'-DDD
525.2	525.2	Water	2,4'-DDE
525.2	525.2	Water	2,4'-DDT
525.2	525.2	Water	2,4-Dinitrotoluene
525.2	525.2	Water	2,6-Dinitrotoluene
525.2	525.2	Water	2-Methylnaphthalene
525.2	525.2	Water	4,4'-DDD
525.2	525.2	Water	4,4'-DDE
525.2	525.2	Water	4,4'-DDT
525.2	525.2	Water	Acetochlor
525.2	525.2	Water	alpha-BHC
525.2	525.2	Water	alpha-Chlordane
525.2	525.2	Water	beta-BHC
525.2	525.2	Water	Chlorobenzilate
525.2	525.2	Water	Chloroneb
525.2	525.2	Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Water	Chlorpyrifos
525.2	525.2	Water	delta-BHC
525.2	525.2	Water	Diclorvos (DDVP)
525.2	525.2	Water	Endosulfan I (Alpha)
525.2	525.2	Water	Endosulfan II (Beta)
525.2	525.2	Water	Endosulfan sulfate
525.2	525.2	Water	Endrin aldehyde
525.2	525.2	Water	EPTC
525.2	525.2	Water	gamma-Chlordane
525.2	525.2	Water	Isophorone
525.2	525.2	Water	Malathion
525.2	525.2	Water	Parathion
525.2	525.2	Water	Pendimethalin (Penoxaline)
525.2	525.2	Water	Terbacil
525.2	525.2	Water	Terbutylazine
525.2	525.2	Water	Total Permethrin (mixed isomers)
525.2	525.2	Water	trans-Nonachlor

Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
A2LA	Dept. of Defense ELAP	7296.01	11-30-26
A2LA	ISO/IEC 17025	7296.01	11-30-26
Alaska (UST)	State	25-005	03-02-27
Arizona	State	AZ0830	11-16-26
California	Los Angeles County Sanitation Districts	9257304	07-31-28
California	SCAQMD LAP	17LA0919	11-30-26

Eurofins Pomona

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Laboratory: Eurofins Calscience (Continued)

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	3082	07-31-28
Kansas	NELAP	E-10420	07-31-27
Nevada	State	CA00111	07-31-27
Oregon	NELAP	4175	02-02-27
Texas	NELAP	T104704499	07-31-27
USDA	US Federal Programs	525-23-159-97150	09-30-26
Utah	NELAP	CA00111	02-28-27
Washington	State	C916	10-12-26

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA POM
625.1	Semivolatile Organic Compounds (GC/MS)	EPA	EET CAL 4
625.1 SIM	Semivolatile Organic Compounds GC/MS (SIM)	EPA	EET CAL 4
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4
525.2	Extraction of Semivolatile Compounds	EPA	EA POM
625.1	Liquid-Liquid Extraction	40CFR136A	EET CAL 4

Protocol References:

40CFR136A = "Methods for Organic Chemical Analysis of Municipal Industrial Wastewater", 40CFR, Part 136, Appendix A, October 26, 1984 and subsequent revisions.

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-233066-1
SDG: Weekly: Ka'amilo Wells P1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-233066-1	Ka'amilo Wells P1 (331-031-WL008)	Water	08/24/26 12:48	08/26/26 10:12	Hawaii
380-233066-2	TB: Ka'amilo Wells P1 (331-031-WL008)	Water	08/24/26 12:48	08/26/26 10:12	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

941 Corporate Center Drive
Pomona, CA 91768-2642
Phone (626) 386-1100

Chain of Custody Record



380-233066 COC

[illegible]

2841 Dow Avenue Suite 100
Tustin CA 92780
Phone 714-895-5494

[illegible]

Joseph Sanchez

From: Joseph Sanchez
Sent: Thursday, August 27, 2026 6:54 AM
To: Calscience - Sample Receiving
Cc: Theodore Ngo, Michael Del Rosario, Ryan Segura, Kristine Tran
Subject: RE: HONOLULU 525 BOTTLES NEEDED

We also need volume for job #233066 which is in location 74B

Joseph Sanchez

Eurofins Drinking Water and Wastewater West, LLC
941 Corporate Center Drive
Pomona, CA 91768

Mobile: +1 909 247 7307

Email: Joseph.Sanchez@et.eurofinsus.com

Website: www.EurofinsUS.com/Eaton

The information transmitted is intended only for the person or entity to which it is addressed and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon this information by persons or entities other than the intended recipient is prohibited. If you receive this in error, please contact the sender and delete the material from any computer. Email transmission cannot be guaranteed to be secure or error free as information could be intercepted, corrupted, lost, destroyed, arrive late or incomplete. The sender therefore is in no way liable for any errors or omissions in the content of this message which may arise as a result of email transmission. If verification is required, please request a hard copy. We take reasonable precautions to ensure our emails are free from viruses. You need, however, to verify that this email and any attachments are free of viruses, as we can take no responsibility for any computer viruses which might be transferred by way of this email. We may monitor all email communication through our networks. If you contact us by email.

From: Joseph Sanchez
Sent: Thursday, August 27, 2026 6:42 AM
To: Calscience - Sample Receiving <Calscience-SampleReceiving@et.eurofinsus.com>
Cc: Theodore Ngo <Theodore.Ngo@et.eurofinsus.com>; Michael Del Rosario <Michael.DelRosario@et.eurofinsus.com>; Ryan Segura <Ryan.Segura@et.eurofinsus.com>; Kristine Tran <Kristine.Tran@et.eurofinsus.com>
Subject: HONOLULU 525 BOTTLES NEEDED

Our courier will be by today to pick up two 525 bottles that were sent by mistake. Job is #233050 location 71D

Joseph Sanchez

Eurofins Drinking Water and Wastewater West, LLC
941 Corporate Center Drive
Pomona, CA 91768

Mobile: +1 909 247 7307

Email: Joseph.Sanchez@et.eurofinsus.com

Website: www.EurofinsUS.com/Eaton

The information transmitted is intended only for the person or entity to which it is addressed and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon this information by persons or entities other than the intended recipient is prohibited. If you receive this in error, please contact the sender and delete the material from any computer. Email transmission cannot be guaranteed to be secure or error free as information could be intercepted, corrupted, lost, destroyed, arrive late or incomplete. The sender therefore is in no way liable for any errors or omissions in the content of this message which may arise as a result of email transmission. If verification is required, please request a hard copy. We take reasonable precautions to ensure our emails are free from viruses. You need, however, to verify that this email and any attachments are free of viruses, as we can take no responsibility for any computer viruses which might be transferred by way of this email. We may monitor all email communication through our networks. If you contact us by email.

Joseph Sanchez

From: Joseph Sanchez
Sent: Thursday, August 27, 2026 6:54 AM
To: Calscience - Sample Receiving
Cc: Theodore Ngo, Michael Del Rosario, Ryan Segura, Kristine Tran
Subject: RE: HONOLULU 525 BOTTLES NEEDED

We also need volume for job #233066 which is in location 74B

Joseph Sanchez

Eurofins Drinking Water and Wastewater West, LLC
941 Corporate Center Drive
Pomona, CA 91768

Mobile: +1 909 247 7307

Email: Joseph.Sanchez@et.eurofinsus.com
Website: www.EurofinsUS.com/Eaton

The information transmitted is intended only for the person or entity to which it is addressed and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon this information by persons or entities other than the intended recipient is prohibited. If you receive this in error, please contact the sender and delete the material from any computer. Email transmission cannot be guaranteed to be secure or error free as information could be intercepted, corrupted, lost, destroyed, arrive late or incomplete. The sender therefore is in no way liable for any errors or omissions in the content of this message which may arise as a result of email transmission. If verification is required, please request a hard copy. We take reasonable precautions to ensure our emails are free from viruses. You need, however, to verify that this email and any attachments are free of viruses, as we can take no responsibility for any computer viruses which might be transferred by way of this email. We may monitor all email communication through our networks. If you contact us by email.

From: Joseph Sanchez
Sent: Thursday, August 27, 2026 6:42 AM
To: Calscience - Sample Receiving <Calscience-SampleReceiving@et.eurofinsus.com>
Cc: Theodore Ngo <Theodore.Ngo@et.eurofinsus.com>; Michael Del Rosario <Michael.DelRosario@et.eurofinsus.com>; Ryan Segura <Ryan.Segura@et.eurofinsus.com>; Kristine Tran <Kristine.Tran@et.eurofinsus.com>
Subject: HONOLULU 525 BOTTLES NEEDED

Our courier will be by today to pick up two 525 bottles that were sent by mistake. Job is #233050 location 71D

Joseph Sanchez

Eurofins Drinking Water and Wastewater West, LLC
941 Corporate Center Drive
Pomona, CA 91768

Mobile: +1 909 247 7307

Email: Joseph.Sanchez@et.eurofinsus.com
Website: www.EurofinsUS.com/Eaton

The information transmitted is intended only for the person or entity to which it is addressed and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon this information by persons or entities other than the intended recipient is prohibited. If you receive this in error, please contact the sender and delete the material from any computer. Email transmission cannot be guaranteed to be secure or error free as information could be intercepted, corrupted, lost, destroyed, arrive late or incomplete. The sender therefore is in no way liable for any errors or omissions in the content of this message which may arise as a result of email transmission. If verification is required, please request a hard copy. We take reasonable precautions to ensure our emails are free from viruses. You need, however, to verify that this email and any attachments are free of viruses, as we can take no responsibility for any computer viruses which might be transferred by way of this email. We may monitor all email communication through our networks. If you contact us by email.

Chain of Custody Record

[illegible]

9/10/2026

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-233066-1
SDG Number: Weekly: Ka'amilo Wells P1

Login Number: 233066

List Number: 1

Creator: Edrosa, Rey

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-233066-1
SDG Number: Weekly: Ka'amilo Wells P1

Login Number: 233066

List Number: 2

Creator: Khana, Piyush

List Source: Eurofins Calscience

List Creation: 08/26/26 07:26 PM

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	2.7
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	fgf5
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	