

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Erwin Kawata  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 310  
Honolulu, Hawaii 96843

Generated 6/10/2026 3:14:30 PM

## JOB DESCRIPTION

RED-HILL  
Weekly: Halawa Wells P1 (MS/MSD)

## JOB NUMBER

380-216586-1

# Eurofins Pomona

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

## Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

## Authorization



Authorized for release by  
Maria Lopez, Project Manager  
[Maria.Lopez@et.eurofinsus.com](mailto:Maria.Lopez@et.eurofinsus.com)  
(626)386-1100

Generated  
6/10/2026 3:14:30 PM



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	11
Surrogate Summary . . . . .	12
QC Sample Results . . . . .	15
QC Association Summary . . . . .	30
Lab Chronicle . . . . .	32
Certification Summary . . . . .	33
Method Summary . . . . .	35
Sample Summary . . . . .	36
Chain of Custody . . . . .	37
Receipt Checklists . . . . .	47

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Qualifiers

### GC/MS Semi VOA

Qualifier	Qualifier Description
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased
F1	MS and/or MSD recovery exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

### GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
T	Result is a tentatively identified compound (TIC) and an estimated value.

### GC Semi VOA

Qualifier	Qualifier Description
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: City & County of Honolulu  
Project: RED-HILL

Job ID: 380-216586-1

**Job ID: 380-216586-1**

**Eurofins Pomona**

## Job Narrative 380-216586-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

### Receipt

The samples were received on 5/28/2026 10:10 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 4.6°C and 5.8°C.

### Receipt Exceptions

#### GC/MS Semi VOA

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

#### Gasoline Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

#### Diesel Range Organics

Method 8015B: The method reporting limit check (MRL) for preparation batch 570-747016 and analytical batch 570-750788 recovered outside control limits for the following analytes: C10-C28. These analytes were biased high in the MRL and were not detected in the associated samples; therefore, the data have been reported.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**PWSID Number: HI0000331**

**Lab Sample ID: 380-216586-1**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Dieldrin	0.038		0.0099	ug/L	1		525.2	Total/NA
Heptachlor epoxide (isomer B)	0.019	^3+	0.0099	ug/L	1		525.2	Total/NA

**Client Sample ID: TB: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-2**

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona



# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-1**

Date Collected: 05/26/26 10:22

Matrix: Drinking Water

Date Received: 05/28/26 10:10

PWSID Number: HI0000331

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
2,4'-DDD	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
2,4'-DDE	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
2,4'-DDT	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
2,4-Dinitrotoluene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
2,6-Dinitrotoluene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
2-Methylnaphthalene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
4,4'-DDD	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
4,4'-DDE	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
4,4'-DDT	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Acenaphthene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Acenaphthylene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Acetochlor	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Alachlor	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
alpha-BHC	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
alpha-Chlordane	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Anthracene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 18:54	1
Atrazine	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Benz(a)anthracene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Benzo[a]pyrene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 18:54	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 18:54	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 18:54	1
beta-BHC	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		06/02/26 15:52	06/03/26 18:54	1
Bromacil	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Butachlor	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Butylbenzylphthalate	<0.49		0.49	ug/L		06/02/26 15:52	06/03/26 18:54	1
Chlorobenzilate	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Chloroneb	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Chlorothalonil (Draconil, Bravo)	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Chlorpyrifos	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Chrysene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 18:54	1
delta-BHC	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		06/02/26 15:52	06/03/26 18:54	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
<b>Dieldrin</b>	<b>0.038</b>		0.0099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Diethylphthalate	<0.49		0.49	ug/L		06/02/26 15:52	06/03/26 18:54	1
Dimethylphthalate	<0.49		0.49	ug/L		06/02/26 15:52	06/03/26 18:54	1
Di-n-butyl phthalate	<0.99		0.99	ug/L		06/02/26 15:52	06/03/26 18:54	1
Di-n-octyl phthalate	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Endosulfan I (Alpha)	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Endosulfan II (Beta)	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Endosulfan sulfate	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Endrin	<0.0099		0.0099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Endrin aldehyde	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
EPTC	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Fluoranthene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-1**

Date Collected: 05/26/26 10:22

Matrix: Drinking Water

Date Received: 05/28/26 10:10

PWSID Number: HI0000331

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
gamma-Chlordane	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Heptachlor	<0.0099		0.0099	ug/L		06/02/26 15:52	06/03/26 18:54	1
<b>Heptachlor epoxide (isomer B)</b>	<b>0.019</b>	<b>^3+</b>	0.0099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Hexachlorobenzene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Isophorone	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Lindane	<0.0099		0.0099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Malathion	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Methoxychlor	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Metolachlor	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Molinate	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Naphthalene	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Parathion	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Pendimethalin (Penoxaline)	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Phenanthrene	<0.040		0.040	ug/L		06/02/26 15:52	06/03/26 18:54	1
Propachlor	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Pyrene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Simazine	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Terbacil	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Terbutylazine	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Thiobencarb	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		06/02/26 15:52	06/03/26 18:54	1
trans-Nonachlor	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 18:54	1
Trifluralin	<0.099		0.099	ug/L		06/02/26 15:52	06/03/26 18:54	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Unknown	0.53	T J	ug/L		2.97	N/A	06/02/26 15:52	06/03/26 18:54	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	99		70 - 130	06/02/26 15:52	06/03/26 18:54	1
Perylene-d12	96		70 - 130	06/02/26 15:52	06/03/26 18:54	1
Triphenylphosphate	115		70 - 130	06/02/26 15:52	06/03/26 18:54	1

**Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
2-Methylnaphthalene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Acenaphthene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Acenaphthylene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Anthracene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Benzo[a]anthracene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Benzo[a]pyrene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Benzo[b]fluoranthene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Benzo[g,h,i]perylene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Benzo[k]fluoranthene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Chrysene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Dibenz(a,h)anthracene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Fluoranthene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-1**

Date Collected: 05/26/26 10:22

Matrix: Drinking Water

Date Received: 05/28/26 10:10

PWSID Number: HI0000331

**Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Indeno[1,2,3-cd]pyrene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Naphthalene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Phenanthrene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1
Pyrene	<0.19		0.19	ug/L		05/31/26 08:21	06/03/26 11:54	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	83		28 - 127	05/31/26 08:21	06/03/26 11:54	1
2-Fluorobiphenyl (Surr)	84		31 - 120	05/31/26 08:21	06/03/26 11:54	1
2-Fluorophenol (Surr)	45		17 - 120	05/31/26 08:21	06/03/26 11:54	1
Nitrobenzene-d5 (Surr)	79		27 - 120	05/31/26 08:21	06/03/26 11:54	1
Phenol-d6 (Surr)	29		10 - 120	05/31/26 08:21	06/03/26 11:54	1
p-Terphenyl-d14 (Surr)	68		45 - 120	05/31/26 08:21	06/03/26 11:54	1

**Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)**

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	05/31/26 08:21	06/04/26 12:27	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	104		33 - 139	05/31/26 08:21	06/04/26 12:27	1
2-Fluorobiphenyl (Surr)	88		33 - 126	05/31/26 08:21	06/04/26 12:27	1
2-Fluorophenol (Surr)	53		12 - 120	05/31/26 08:21	06/04/26 12:27	1
Nitrobenzene-d5 (Surr)	113		36 - 120	05/31/26 08:21	06/04/26 12:27	1
Phenol-d6 (Surr)	33		10 - 120	05/31/26 08:21	06/04/26 12:27	1
p-Terphenyl-d14 (Surr)	82		47 - 131	05/31/26 08:21	06/04/26 12:27	1

**Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			06/06/26 14:47	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	104		38 - 134		06/06/26 14:47	1

**Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<25		25	ug/L		06/01/26 09:04	06/08/26 13:54	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		06/01/26 09:04	06/08/26 13:54	1
C8-C18	<25		25	ug/L		06/01/26 09:04	06/08/26 13:54	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	104		60 - 130	06/01/26 09:04	06/08/26 13:54	1

**Client Sample ID: TB: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-2**

Date Collected: 05/26/26 10:22

Matrix: Water

Date Received: 05/28/26 10:10

**Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			06/06/26 14:23	1

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

**Client Sample ID: TB: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-2**

**Date Collected: 05/26/26 10:22**

**Matrix: Water**

**Date Received: 05/28/26 10:10**

<u>Surrogate</u>	<u>%Recovery</u>	<u>Qualifier</u>	<u>Limits</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
4-Bromofluorobenzene (Surr)	97		38 - 134		06/06/26 14:23	1

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

# Action Limit Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-1**

**PWSID Number: HI0000331**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	<0.049		ug/L	2	0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.020		ug/L	0.2	0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.59		ug/L	6	0.59	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.59		ug/L	400	0.59	525.2	Total/NA
Endrin	<0.0099		ug/L	2	0.0099	525.2	Total/NA
Heptachlor	<0.0099		ug/L	0.4	0.0099	525.2	Total/NA
Heptachlor epoxide (isomer B)	0.019	^3+	ug/L	0.2	0.0099	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50	0.049	525.2	Total/NA
Lindane	<0.0099		ug/L	0.2	0.0099	525.2	Total/NA
Methoxychlor	<0.049	^3+	ug/L	40	0.049	525.2	Total/NA
Simazine	<0.049	^3+	ug/L	4	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.19		ug/L	0.2	0.19	625.1 SIM	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-216586-1	HALAWA WELLS P1 (331-023-V)	99	96	115

**Surrogate Legend**  
 2NMX = 2-Nitro-m-xylene  
 PRY = Perylene-d12  
 TPP = Triphenylphosphate

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-216945-B-4-A MS	Matrix Spike	100	103	111
380-216945-B-4-B MSD	Matrix Spike Duplicate	98	101	115
LCS 380-231139/23-A	Lab Control Sample	94	98	102
MB 380-231139/21-A	Method Blank	95	94	110
MRL 380-231139/22-A	Lab Control Sample	97	97	109

**Surrogate Legend**  
 2NMX = 2-Nitro-m-xylene  
 PRY = Perylene-d12  
 TPP = Triphenylphosphate

## Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
380-216586-1	HALAWA WELLS P1 (331-023-V)	104	88	53	113	33	82

**Surrogate Legend**  
 TBP = 2,4,6-Tribromophenol (Surr)  
 FBP = 2-Fluorobiphenyl (Surr)  
 2FP = 2-Fluorophenol (Surr)  
 NBZ = Nitrobenzene-d5 (Surr)  
 PHL6 = Phenol-d6 (Surr)  
 TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
MB 570-746664/1-A	Method Blank	106	89	60	108	37	92

**Surrogate Legend**  
 TBP = 2,4,6-Tribromophenol (Surr)  
 FBP = 2-Fluorobiphenyl (Surr)  
 2FP = 2-Fluorophenol (Surr)  
 NBZ = Nitrobenzene-d5 (Surr)

# Surrogate Summary

Client: City & County of Honolulu

Job ID: 380-216586-1

Project/Site: RED-HILL

SDG: Weekly: Halawa Wells P1 (MS/MSD)

PHL6 = Phenol-d6 (Surr)

TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Drinking Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-216586-1	HALAWA WELLS P1 (331-023-v)	83	84	45	79	29	68
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	80	79	54	69	37	78
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	77	83	57	69	39	81

#### Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)

FBP = 2-Fluorobiphenyl (Surr)

2FP = 2-Fluorophenol (Surr)

NBZ = Nitrobenzene-d5 (Surr)

PHL6 = Phenol-d6 (Surr)

TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
LCS 570-746664/2-A	Lab Control Sample	73	75	50	64	34	70
LCSD 570-746664/3-A	Lab Control Sample Dup	69	72	47	61	33	74
MB 570-746664/1-A	Method Blank	85	83	49	80	31	78

#### Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)

FBP = 2-Fluorobiphenyl (Surr)

2FP = 2-Fluorophenol (Surr)

NBZ = Nitrobenzene-d5 (Surr)

PHL6 = Phenol-d6 (Surr)

TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Drinking Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1
		(38-134)
380-216586-1	HALAWA WELLS P1 (331-023-v)	104
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	100
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	89

#### Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

# Surrogate Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-216586-2	TB: HALAWA WELLS P1 (331-0	97
LCS 570-750190/3	Lab Control Sample	98
LCSD 570-750190/4	Lab Control Sample Dup	98
MB 570-750190/6	Method Blank	102
MRL 570-750190/5	Lab Control Sample	99

#### Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Drinking Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-216586-1	HALAWA WELLS P1 (331-023-v	104
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	110
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	102

#### Surrogate Legend

OTCSN = n-Octacosane (Surr)

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
LCS 570-747016/2-A	Lab Control Sample	108
LCSD 570-747016/3-A	Lab Control Sample Dup	109
MB 570-747016/1-A	Method Blank	103
MRL 570-747016/4-A	Lab Control Sample	98

#### Surrogate Legend

OTCSN = n-Octacosane (Surr)

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 380-231139/21-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
1-Methylnaphthalene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
2,4'-DDD	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
2,4'-DDE	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
2,4'-DDT	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
2-Methylnaphthalene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
4,4'-DDD	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
4,4'-DDE	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
4,4'-DDT	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Acenaphthene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Acenaphthylene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Acetochlor	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Alachlor	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
alpha-BHC	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
alpha-Chlordane	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Anthracene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 14:34	1
Atrazine	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Benz(a)anthracene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Benzo[a]pyrene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 14:34	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 14:34	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 14:34	1
beta-BHC	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		06/02/26 15:52	06/03/26 14:34	1
Bromacil	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Butachlor	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Butylbenzylphthalate	<0.49		0.49	ug/L		06/02/26 15:52	06/03/26 14:34	1
Chlorobenzilate	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Chloroneb	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Chlorpyrifos	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Chrysene	<0.020		0.020	ug/L		06/02/26 15:52	06/03/26 14:34	1
delta-BHC	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		06/02/26 15:52	06/03/26 14:34	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Dieldrin	<0.0098		0.0098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Diethylphthalate	<0.49		0.49	ug/L		06/02/26 15:52	06/03/26 14:34	1
Dimethylphthalate	<0.49		0.49	ug/L		06/02/26 15:52	06/03/26 14:34	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		06/02/26 15:52	06/03/26 14:34	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Endosulfan sulfate	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Endrin	<0.0098		0.0098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Endrin aldehyde	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
EPTC	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MB 380-231139/21-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluoranthene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Fluorene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
gamma-Chlordane	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Heptachlor	<0.0098		0.0098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Heptachlor epoxide (isomer B)	<0.0098	^3+	0.0098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Hexachlorobenzene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Isophorone	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Lindane	<0.0098		0.0098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Malathion	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Methoxychlor	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Metolachlor	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Molinate	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Naphthalene	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Parathion	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Phenanthrene	<0.039		0.039	ug/L		06/02/26 15:52	06/03/26 14:34	1
Propachlor	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Pyrene	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Simazine	<0.049	^3+	0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Terbacil	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Terbutylazine	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Thiobencarb	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		06/02/26 15:52	06/03/26 14:34	1
trans-Nonachlor	<0.049		0.049	ug/L		06/02/26 15:52	06/03/26 14:34	1
Trifluralin	<0.098		0.098	ug/L		06/02/26 15:52	06/03/26 14:34	1

Tentatively Identified Compound	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	06/02/26 15:52	06/03/26 14:34	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	95		70 - 130	06/02/26 15:52	06/03/26 14:34	1
Perylene-d12	94		70 - 130	06/02/26 15:52	06/03/26 14:34	1
Triphenylphosphate	110		70 - 130	06/02/26 15:52	06/03/26 14:34	1

**Lab Sample ID: LCS 380-231139/23-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
1-Methylnaphthalene	1.95	1.86		ug/L		95	70 - 130
2,4'-DDD	1.95	1.85		ug/L		95	70 - 130
2,4'-DDE	1.95	1.92		ug/L		98	70 - 130
2,4'-DDT	1.95	1.74		ug/L		89	70 - 130
2,4-Dinitrotoluene	1.95	1.96		ug/L		100	70 - 130
2,6-Dinitrotoluene	1.95	2.03		ug/L		104	70 - 130
2-Methylnaphthalene	1.95	1.82		ug/L		93	70 - 130

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCS 380-231139/23-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
4,4'-DDD	1.95	1.80		ug/L		92	70 - 130
4,4'-DDE	1.95	1.88		ug/L		96	70 - 130
4,4'-DDT	1.95	1.80		ug/L		92	70 - 130
Acenaphthene	1.95	1.97		ug/L		101	70 - 130
Acenaphthylene	1.95	1.75		ug/L		90	70 - 130
Acetochlor	1.95	2.07		ug/L		106	70 - 130
Alachlor	1.95	2.09		ug/L		107	70 - 130
alpha-BHC	1.95	2.02		ug/L		103	70 - 130
alpha-Chlordane	1.95	1.74		ug/L		89	70 - 130
Anthracene	1.95	1.80		ug/L		92	70 - 130
Atrazine	1.95	1.98		ug/L		101	70 - 130
Benz(a)anthracene	1.95	1.93		ug/L		99	70 - 130
Benzo[a]pyrene	1.95	1.96		ug/L		100	70 - 130
Benzo[b]fluoranthene	1.95	2.00		ug/L		102	70 - 130
Benzo[g,h,i]perylene	1.95	2.09		ug/L		107	70 - 130
Benzo[k]fluoranthene	1.95	2.00		ug/L		102	70 - 130
beta-BHC	1.95	2.03		ug/L		104	70 - 130
Bis(2-ethylhexyl) phthalate	1.95	2.14		ug/L		109	70 - 130
Bromacil	1.95	2.01		ug/L		103	70 - 130
Butachlor	1.95	2.22		ug/L		113	70 - 130
Butylbenzylphthalate	1.95	2.30		ug/L		118	70 - 130
Chlorobenzilate	1.95	2.02		ug/L		103	70 - 130
Chloroneb	1.95	1.99		ug/L		102	70 - 130
Chlorothalonil (Draconil, Bravo)	1.95	1.94		ug/L		99	70 - 130
Chlorpyrifos	1.95	1.80		ug/L		92	70 - 130
Chrysene	1.95	1.84		ug/L		94	70 - 130
delta-BHC	1.95	2.00		ug/L		103	70 - 130
Di(2-ethylhexyl)adipate	1.95	2.33		ug/L		119	70 - 130
Dibenz(a,h)anthracene	1.95	2.07		ug/L		106	70 - 130
Diclorvos (DDVP)	1.95	2.10		ug/L		107	70 - 130
Dieldrin	1.95	2.07		ug/L		106	70 - 130
Diethylphthalate	1.95	2.08		ug/L		106	70 - 130
Dimethylphthalate	1.95	2.02		ug/L		103	70 - 130
Di-n-butyl phthalate	3.91	4.48		ug/L		115	70 - 130
Di-n-octyl phthalate	1.95	2.09		ug/L		107	70 - 130
Endosulfan I (Alpha)	1.95	1.87		ug/L		96	70 - 130
Endosulfan II (Beta)	1.95	1.89		ug/L		97	70 - 130
Endosulfan sulfate	1.95	2.04		ug/L		104	70 - 130
Endrin	1.95	2.19		ug/L		112	70 - 130
Endrin aldehyde	1.95	1.95		ug/L		100	60 - 130
EPTC	1.95	1.90		ug/L		97	70 - 130
Fluoranthene	1.95	1.81		ug/L		93	70 - 130
Fluorene	1.95	1.91		ug/L		98	70 - 130
gamma-Chlordane	1.95	1.80		ug/L		92	70 - 130
Heptachlor	1.95	2.10		ug/L		108	70 - 130
Heptachlor epoxide (isomer B)	1.95	2.02		ug/L		103	70 - 130
Hexachlorobenzene	1.95	1.80		ug/L		92	70 - 130
Hexachlorocyclopentadiene	1.95	1.76		ug/L		90	70 - 130
Indeno[1,2,3-cd]pyrene	1.95	2.20		ug/L		113	70 - 130

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCS 380-231139/23-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Isophorone	1.95	1.83		ug/L		94	70 - 130
Lindane	1.95	2.13		ug/L		109	70 - 130
Malathion	1.95	1.96		ug/L		101	70 - 130
Methoxychlor	1.95	1.97		ug/L		101	70 - 130
Metolachlor	1.95	2.08		ug/L		106	70 - 130
Molinate	1.95	1.95		ug/L		100	70 - 130
Naphthalene	1.95	1.82		ug/L		93	70 - 130
Parathion	1.95	2.15		ug/L		110	70 - 130
Pendimethalin (Penoxaline)	1.95	1.84		ug/L		94	70 - 130
Phenanthrene	1.95	1.85		ug/L		94	70 - 130
Propachlor	1.95	2.09		ug/L		107	70 - 130
Pyrene	1.95	1.86		ug/L		95	70 - 130
Simazine	1.95	1.94		ug/L		99	70 - 130
Terbacil	1.95	2.24		ug/L		115	70 - 130
Terbutylazine	1.95	2.07		ug/L		106	70 - 130
Thiobencarb	1.95	1.94		ug/L		99	70 - 130
trans-Nonachlor	1.95	1.75		ug/L		90	70 - 130
Trifluralin	1.95	2.13		ug/L		109	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2-Nitro-m-xylene	94		70 - 130
Perylene-d12	98		70 - 130
Triphenylphosphate	102		70 - 130

**Lab Sample ID: MRL 380-231139/22-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.0978	0.0984		ug/L		101	50 - 150
2,4'-DDD	0.0978	0.0947	J	ug/L		97	50 - 150
2,4'-DDE	0.0978	0.102		ug/L		104	50 - 150
2,4'-DDT	0.0978	0.121		ug/L		124	50 - 150
2,4-Dinitrotoluene	0.0978	0.107		ug/L		110	50 - 150
2,6-Dinitrotoluene	0.0978	0.0966	J	ug/L		99	50 - 150
2-Methylnaphthalene	0.0978	0.103		ug/L		105	50 - 150
4,4'-DDD	0.0978	0.115		ug/L		118	50 - 150
4,4'-DDE	0.0978	0.120		ug/L		123	50 - 150
4,4'-DDT	0.0978	0.121		ug/L		124	50 - 150
Acenaphthene	0.0978	0.0991		ug/L		101	50 - 150
Acenaphthylene	0.0978	0.0740	J	ug/L		76	50 - 150
Acetochlor	0.0978	0.125		ug/L		128	50 - 150
Alachlor	0.0489	0.0728		ug/L		149	50 - 150
alpha-BHC	0.0978	0.122		ug/L		124	50 - 150
alpha-Chlordane	0.0244	0.0314	J	ug/L		128	50 - 150
Anthracene	0.0196	0.0245		ug/L		125	50 - 150
Atrazine	0.0489	0.0594		ug/L		121	50 - 150
Benz(a)anthracene	0.0489	0.0558		ug/L		114	50 - 150

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MRL 380-231139/22-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Spike Added	MRL	MRL	Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
Benzo[a]pyrene	0.0196	0.0239		ug/L		122	50 - 150
Benzo[b]fluoranthene	0.0196	0.0243		ug/L		124	50 - 150
Benzo[g,h,i]perylene	0.0489	0.0659		ug/L		135	50 - 150
Benzo[k]fluoranthene	0.0196	0.0227		ug/L		116	50 - 150
beta-BHC	0.0978	0.129		ug/L		132	50 - 150
Bis(2-ethylhexyl) phthalate	0.587	0.689		ug/L		117	50 - 150
Bromacil	0.0978	0.139		ug/L		142	50 - 150
Butachlor	0.0489	0.0765	^3+	ug/L		156	50 - 150
Butylbenzylphthalate	0.489	0.611		ug/L		125	50 - 150
Chlorobenzilate	0.0978	0.132		ug/L		135	50 - 150
Chloroneb	0.0978	0.108		ug/L		110	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0978	0.110		ug/L		112	50 - 150
Chlorpyrifos	0.0489	0.0782	^3+	ug/L		160	50 - 150
Chrysene	0.0196	0.0223		ug/L		114	50 - 150
delta-BHC	0.0978	0.116		ug/L		118	50 - 150
Di(2-ethylhexyl)adipate	0.587	0.778		ug/L		133	50 - 150
Dibenz(a,h)anthracene	0.0489	0.0556		ug/L		114	50 - 150
Diclorvos (DDVP)	0.0489	0.0681		ug/L		139	50 - 150
Dieldrin	0.00978	0.0122		ug/L		125	50 - 150
Diethylphthalate	0.489	0.561		ug/L		115	50 - 150
Dimethylphthalate	0.489	0.542		ug/L		111	50 - 150
Di-n-butyl phthalate	0.489	0.657	J	ug/L		134	49 - 243
Di-n-octyl phthalate	0.0978	0.124		ug/L		127	50 - 150
Endosulfan I (Alpha)	0.0978	0.110		ug/L		113	50 - 150
Endosulfan II (Beta)	0.0978	0.143		ug/L		147	50 - 150
Endosulfan sulfate	0.0978	0.117		ug/L		120	50 - 150
Endrin	0.00978	0.0110		ug/L		112	50 - 150
Endrin aldehyde	0.0978	0.130		ug/L		133	50 - 150
EPTC	0.0978	0.106		ug/L		109	50 - 150
Fluoranthene	0.0978	0.117		ug/L		119	50 - 150
Fluorene	0.0489	0.0548		ug/L		112	50 - 150
gamma-Chlordane	0.0244	0.0361	J	ug/L		148	50 - 150
Heptachlor	0.00978	0.0135		ug/L		138	50 - 150
Heptachlor epoxide (isomer B)	0.00978	0.0154	^3+	ug/L		158	50 - 150
Hexachlorobenzene	0.0489	0.0682		ug/L		139	50 - 150
Hexachlorocyclopentadiene	0.0489	0.0471	J	ug/L		96	50 - 150
Indeno[1,2,3-cd]pyrene	0.0489	0.0619		ug/L		127	50 - 150
Isophorone	0.0978	0.106		ug/L		108	50 - 150
Lindane	0.00978	0.0138		ug/L		142	50 - 150
Malathion	0.0978	0.119		ug/L		122	50 - 150
Methoxychlor	0.0489	0.0766	^3+	ug/L		157	50 - 150
Metolachlor	0.0489	0.0631		ug/L		129	50 - 150
Molinate	0.0978	0.120		ug/L		123	50 - 150
Naphthalene	0.0978	0.102		ug/L		104	50 - 150
Parathion	0.0978	0.112		ug/L		115	50 - 150
Pendimethalin (Penoxaline)	0.0978	0.107		ug/L		110	50 - 150
Phenanthrene	0.0391	0.0453		ug/L		116	50 - 150
Propachlor	0.0489	0.0795	^3+	ug/L		163	50 - 150
Pyrene	0.0489	0.0600		ug/L		123	50 - 150

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MRL 380-231139/22-A**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Simazine	0.0489	0.0768	^3+	ug/L		157	50 - 150
Terbacil	0.0978	0.136		ug/L		139	50 - 150
Terbutylazine	0.0978	0.116		ug/L		119	50 - 150
Thiobencarb	0.0978	0.117		ug/L		120	50 - 150
trans-Nonachlor	0.0244	0.0316	J	ug/L		129	50 - 150
Trifluralin	0.0978	0.121		ug/L		124	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
2-Nitro-m-xylene	97		70 - 130
Perylene-d12	97		70 - 130
Triphenylphosphate	109		70 - 130

**Lab Sample ID: 380-216945-B-4-A MS**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.097		1.95	1.92		ug/L		98	70 - 130
2,4'-DDD	<0.097		1.95	1.98		ug/L		102	70 - 130
2,4'-DDE	<0.097	F1	1.95	0.873	F1	ug/L		45	70 - 130
2,4'-DDT	<0.097		1.95	1.97		ug/L		101	70 - 130
2,4-Dinitrotoluene	<0.097		1.95	2.28		ug/L		117	70 - 130
2,6-Dinitrotoluene	<0.097		1.95	2.33		ug/L		119	70 - 130
2-Methylnaphthalene	<0.097		1.95	1.90		ug/L		97	70 - 130
4,4'-DDD	<0.097		1.95	1.95		ug/L		100	70 - 130
4,4'-DDE	<0.097		1.95	1.67		ug/L		86	70 - 130
4,4'-DDT	<0.097		1.95	1.96		ug/L		100	70 - 130
Acenaphthene	<0.097		1.95	1.94		ug/L		100	70 - 130
Acenaphthylene	<0.097		1.95	1.95		ug/L		100	70 - 130
Acetochlor	<0.097		1.95	2.18		ug/L		112	70 - 130
Alachlor	<0.049		1.95	2.17		ug/L		112	70 - 130
alpha-BHC	<0.097		1.95	1.95		ug/L		100	70 - 130
alpha-Chlordane	<0.049		1.95	1.91		ug/L		98	70 - 130
Anthracene	<0.019		1.95	1.87		ug/L		96	70 - 130
Atrazine	<0.049		1.95	2.02		ug/L		103	70 - 130
Benz(a)anthracene	<0.049		1.95	2.20		ug/L		113	70 - 130
Benzo[a]pyrene	<0.019		1.95	2.48		ug/L		127	70 - 130
Benzo[b]fluoranthene	<0.019		1.95	2.36		ug/L		121	70 - 130
Benzo[g,h,i]perylene	<0.049		1.95	2.45		ug/L		126	70 - 130
Benzo[k]fluoranthene	<0.019		1.95	2.34		ug/L		120	70 - 130
beta-BHC	<0.097		1.95	2.02		ug/L		104	70 - 130
Bis(2-ethylhexyl) phthalate	<0.58		1.95	2.21		ug/L		113	70 - 130
Bromacil	<0.097		1.95	2.42		ug/L		124	70 - 130
Butachlor	<0.049	^3+	1.95	1.48		ug/L		76	70 - 130
Butylbenzylphthalate	<0.49		1.95	2.21		ug/L		114	70 - 130
Chlorobenzilate	<0.097		1.95	2.42		ug/L		124	70 - 130
Chloroneb	<0.097		1.95	2.08		ug/L		107	70 - 130
Chlorothalonil (Draconil, Bravo)	<0.097		1.95	2.20		ug/L		113	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-216945-B-4-A MS**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Chlorpyrifos	<0.049	^3+	1.95	1.95		ug/L		100	70 - 130
Chrysene	<0.019		1.95	2.05		ug/L		105	70 - 130
delta-BHC	<0.097		1.95	1.99		ug/L		102	70 - 130
Di(2-ethylhexyl)adipate	<0.58		1.95	2.05		ug/L		105	70 - 130
Dibenz(a,h)anthracene	<0.049		1.95	2.45		ug/L		126	70 - 130
Diclorvos (DDVP)	<0.049		1.95	2.24		ug/L		115	70 - 130
Dieldrin	<0.0097		1.95	2.18		ug/L		112	70 - 130
Diethylphthalate	<0.49		1.95	2.23		ug/L		110	70 - 130
Dimethylphthalate	<0.49		1.95	2.09		ug/L		107	70 - 130
Di-n-butyl phthalate	<0.97		3.89	3.93		ug/L		101	70 - 130
Di-n-octyl phthalate	<0.097		1.95	2.11		ug/L		109	70 - 130
Endosulfan I (Alpha)	<0.097		1.95	1.96		ug/L		101	70 - 130
Endosulfan II (Beta)	<0.097		1.95	1.88		ug/L		97	70 - 130
Endosulfan sulfate	<0.097		1.95	2.18		ug/L		112	70 - 130
Endrin	<0.0097		1.95	2.22		ug/L		114	70 - 130
Endrin aldehyde	<0.097		1.95	1.34		ug/L		69	60 - 130
EPTC	<0.097		1.95	2.06		ug/L		106	70 - 130
Fluoranthene	<0.097		1.95	1.93		ug/L		99	70 - 130
Fluorene	<0.049		1.95	1.92		ug/L		99	70 - 130
gamma-Chlordane	<0.049	F1	1.95	1.38		ug/L		71	70 - 130
Heptachlor	<0.0097		1.95	1.50		ug/L		77	70 - 130
Heptachlor epoxide (isomer B)	<0.0097	^3+	1.95	2.07		ug/L		107	70 - 130
Hexachlorobenzene	<0.049		1.95	1.71		ug/L		88	70 - 130
Hexachlorocyclopentadiene	<0.049		1.95	1.87		ug/L		96	70 - 130
Indeno[1,2,3-cd]pyrene	<0.049	F1	1.95	2.68	F1	ug/L		138	70 - 130
Isophorone	<0.097		1.95	1.95		ug/L		98	70 - 130
Lindane	<0.0097		1.95	2.25		ug/L		116	70 - 130
Malathion	<0.097		1.95	2.07		ug/L		106	70 - 130
Methoxychlor	<0.049	^3+	1.95	2.38		ug/L		122	70 - 130
Metolachlor	<0.049		1.95	2.08		ug/L		107	70 - 130
Molinate	<0.097		1.95	2.12		ug/L		109	70 - 130
Naphthalene	<0.097		1.95	1.88		ug/L		97	70 - 130
Parathion	<0.097	F1	1.95	2.50		ug/L		129	70 - 130
Pendimethalin (Penoxaline)	<0.097		1.95	2.46		ug/L		126	70 - 130
Phenanthrene	<0.039		1.95	1.89		ug/L		97	70 - 130
Propachlor	<0.049	^3+	1.95	2.30		ug/L		118	70 - 130
Pyrene	<0.049		1.95	1.90		ug/L		97	70 - 130
Simazine	<0.049	^3+	1.95	1.85		ug/L		95	70 - 130
Terbacil	<0.097	F1	1.95	2.42		ug/L		124	70 - 130
Terbutylazine	<0.097		1.95	2.02		ug/L		104	70 - 130
Thiobencarb	<0.097		1.95	1.98		ug/L		102	70 - 130
trans-Nonachlor	<0.049		1.95	2.04		ug/L		105	70 - 130
Trifluralin	<0.097	F1	1.95	2.54	F1	ug/L		131	70 - 130

Surrogate	MS MS		Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	100		70 - 130
Perylene-d12	103		70 - 130
Triphenylphosphate	111		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-216586-1  
 SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-216945-B-4-B MSD**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Sample	Sample Qualifier	Spike Added	MSD	MSD Qualifier	Unit	D	%Rec	%Rec	RPD	Limit
	Result			Result					Limits		
1-Methylnaphthalene	<0.097		1.97	1.95		ug/L		98	70 - 130	2	20
2,4'-DDD	<0.097		1.97	2.06		ug/L		105	70 - 130	4	20
2,4'-DDE	<0.097	F1	1.97	0.851	F1	ug/L		43	70 - 130	3	20
2,4'-DDT	<0.097		1.97	2.07		ug/L		105	70 - 130	5	20
2,4-Dinitrotoluene	<0.097		1.97	2.46		ug/L		125	70 - 130	7	20
2,6-Dinitrotoluene	<0.097		1.97	2.43		ug/L		123	70 - 130	4	20
2-Methylnaphthalene	<0.097		1.97	1.93		ug/L		97	70 - 130	2	20
4,4'-DDD	<0.097		1.97	2.02		ug/L		102	70 - 130	4	20
4,4'-DDE	<0.097		1.97	1.75		ug/L		89	70 - 130	4	20
4,4'-DDT	<0.097		1.97	2.09		ug/L		106	70 - 130	7	20
Acenaphthene	<0.097		1.97	2.00		ug/L		102	70 - 130	3	20
Acenaphthylene	<0.097		1.97	2.01		ug/L		102	70 - 130	3	20
Acetochlor	<0.097		1.97	2.20		ug/L		112	70 - 130	1	20
Alachlor	<0.049		1.97	2.21		ug/L		112	70 - 130	2	20
alpha-BHC	<0.097		1.97	2.01		ug/L		102	70 - 130	3	20
alpha-Chlordane	<0.049		1.97	1.92		ug/L		98	70 - 130	1	20
Anthracene	<0.019		1.97	1.98		ug/L		100	70 - 130	6	20
Atrazine	<0.049		1.97	2.17		ug/L		110	70 - 130	7	20
Benz(a)anthracene	<0.049		1.97	2.28		ug/L		116	70 - 130	4	20
Benzo[a]pyrene	<0.019		1.97	2.54		ug/L		129	70 - 130	3	20
Benzo[b]fluoranthene	<0.019		1.97	2.48		ug/L		126	70 - 130	5	20
Benzo[g,h,i]perylene	<0.049		1.97	2.45		ug/L		125	70 - 130	0	20
Benzo[k]fluoranthene	<0.019		1.97	2.27		ug/L		115	70 - 130	3	20
beta-BHC	<0.097		1.97	2.17		ug/L		110	70 - 130	7	20
Bis(2-ethylhexyl) phthalate	<0.58		1.97	2.29		ug/L		116	70 - 130	4	20
Bromacil	<0.097		1.97	2.48		ug/L		126	70 - 130	3	20
Butachlor	<0.049	^3+	1.97	1.46		ug/L		74	70 - 130	1	20
Butylbenzylphthalate	<0.49		1.97	2.30		ug/L		117	70 - 130	4	20
Chlorobenzilate	<0.097		1.97	2.46		ug/L		125	70 - 130	2	20
Chloroneb	<0.097		1.97	2.08		ug/L		106	70 - 130	0	20
Chlorothalonil (Draconil, Bravo)	<0.097		1.97	2.25		ug/L		114	70 - 130	2	20
Chlorpyrifos	<0.049	^3+	1.97	2.01		ug/L		102	70 - 130	3	20
Chrysene	<0.019		1.97	2.13		ug/L		108	70 - 130	4	20
delta-BHC	<0.097		1.97	2.02		ug/L		103	70 - 130	2	20
Di(2-ethylhexyl)adipate	<0.58		1.97	2.18		ug/L		111	70 - 130	6	20
Dibenz(a,h)anthracene	<0.049		1.97	2.44		ug/L		124	70 - 130	1	20
Diclorvos (DDVP)	<0.049		1.97	2.38		ug/L		121	70 - 130	6	20
Dieldrin	<0.0097		1.97	2.13		ug/L		108	70 - 130	2	20
Diethylphthalate	<0.49		1.97	2.29		ug/L		112	70 - 130	3	20
Dimethylphthalate	<0.49		1.97	2.13		ug/L		108	70 - 130	2	20
Di-n-butyl phthalate	<0.97		3.94	3.75		ug/L		95	70 - 130	5	20
Di-n-octyl phthalate	<0.097		1.97	2.19		ug/L		111	70 - 130	4	20
Endosulfan I (Alpha)	<0.097		1.97	1.87		ug/L		95	70 - 130	5	20
Endosulfan II (Beta)	<0.097		1.97	1.96		ug/L		100	70 - 130	4	20
Endosulfan sulfate	<0.097		1.97	2.25		ug/L		114	70 - 130	3	20
Endrin	<0.0097		1.97	2.24		ug/L		114	70 - 130	1	20
Endrin aldehyde	<0.097		1.97	1.37		ug/L		70	60 - 130	2	20
EPTC	<0.097		1.97	2.12		ug/L		108	70 - 130	3	20

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-216945-B-4-B MSD**  
**Matrix: Water**  
**Analysis Batch: 231376**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 231139**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	RPD	Limit
	Result	Qualifier	Added	Result	Qualifier				Limits		
Fluoranthene	<0.097		1.97	1.96		ug/L		100	70 - 130	1	20
Fluorene	<0.049		1.97	2.00		ug/L		101	70 - 130	4	20
gamma-Chlordane	<0.049	F1	1.97	1.28	F1	ug/L		65	70 - 130	8	20
Heptachlor	<0.0097		1.97	1.49		ug/L		76	70 - 130	1	20
Heptachlor epoxide (isomer B)	<0.0097	^3+	1.97	2.16		ug/L		110	70 - 130	4	20
Hexachlorobenzene	<0.049		1.97	1.80		ug/L		92	70 - 130	5	20
Hexachlorocyclopentadiene	<0.049		1.97	2.09		ug/L		106	70 - 130	11	20
Indeno[1,2,3-cd]pyrene	<0.049	F1	1.97	2.80	F1	ug/L		142	70 - 130	4	20
Isophorone	<0.097		1.97	2.10		ug/L		104	70 - 130	7	20
Lindane	<0.0097		1.97	2.32		ug/L		118	70 - 130	3	20
Malathion	<0.097		1.97	2.16		ug/L		110	70 - 130	4	20
Methoxychlor	<0.049	^3+	1.97	2.52		ug/L		128	70 - 130	6	20
Metolachlor	<0.049		1.97	2.13		ug/L		108	70 - 130	2	20
Molinate	<0.097		1.97	2.22		ug/L		113	70 - 130	5	20
Naphthalene	<0.097		1.97	1.91		ug/L		97	70 - 130	2	20
Parathion	<0.097	F1	1.97	2.60	F1	ug/L		132	70 - 130	4	20
Pendimethalin (Penoxaline)	<0.097		1.97	2.56		ug/L		130	70 - 130	4	20
Phenanthrene	<0.039		1.97	1.97		ug/L		100	70 - 130	4	20
Propachlor	<0.049	^3+	1.97	2.48		ug/L		126	70 - 130	8	20
Pyrene	<0.049		1.97	1.93		ug/L		98	70 - 130	2	20
Simazine	<0.049	^3+	1.97	2.08		ug/L		106	70 - 130	12	20
Terbacil	<0.097	F1	1.97	2.61	F1	ug/L		133	70 - 130	8	20
Terbutylazine	<0.097		1.97	2.20		ug/L		112	70 - 130	8	20
Thiobencarb	<0.097		1.97	2.07		ug/L		105	70 - 130	4	20
trans-Nonachlor	<0.049		1.97	2.10		ug/L		106	70 - 130	3	20
Trifluralin	<0.097	F1	1.97	2.66	F1	ug/L		135	70 - 130	5	20

Surrogate	MSD	MSD	Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	98		70 - 130
Perylene-d12	101		70 - 130
Triphenylphosphate	115		70 - 130

## Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 570-746664/1-A**  
**Matrix: Water**  
**Analysis Batch: 749030**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Tentatively Identified Compound	MB	MB	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
	Est. Result	Qualifier							
Tentatively Identified Compound	None		ug/L			N/A	05/31/26 08:21	06/04/26 11:39	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
2,4,6-Tribromophenol (Surr)	106		33 - 139	05/31/26 08:21	06/04/26 11:39	1
2-Fluorobiphenyl (Surr)	89		33 - 126	05/31/26 08:21	06/04/26 11:39	1
2-Fluorophenol (Surr)	60		12 - 120	05/31/26 08:21	06/04/26 11:39	1
Nitrobenzene-d5 (Surr)	108		36 - 120	05/31/26 08:21	06/04/26 11:39	1
Phenol-d6 (Surr)	37		10 - 120	05/31/26 08:21	06/04/26 11:39	1
p-Terphenyl-d14 (Surr)	92		47 - 131	05/31/26 08:21	06/04/26 11:39	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

**Lab Sample ID: MB 570-746664/1-A**  
**Matrix: Water**  
**Analysis Batch: 748206**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
2-Methylnaphthalene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Acenaphthene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Acenaphthylene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Anthracene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Benzo[a]anthracene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Benzo[a]pyrene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Benzo[b]fluoranthene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Benzo[g,h,i]perylene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Benzo[k]fluoranthene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Chrysene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Dibenz(a,h)anthracene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Fluoranthene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Fluorene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Indeno[1,2,3-cd]pyrene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Naphthalene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Phenanthrene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1
Pyrene	<0.20		0.20	ug/L		05/31/26 08:21	06/03/26 07:55	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	85		28 - 127	05/31/26 08:21	06/03/26 07:55	1
2-Fluorobiphenyl (Surr)	83		31 - 120	05/31/26 08:21	06/03/26 07:55	1
2-Fluorophenol (Surr)	49		17 - 120	05/31/26 08:21	06/03/26 07:55	1
Nitrobenzene-d5 (Surr)	80		27 - 120	05/31/26 08:21	06/03/26 07:55	1
Phenol-d6 (Surr)	31		10 - 120	05/31/26 08:21	06/03/26 07:55	1
p-Terphenyl-d14 (Surr)	78		45 - 120	05/31/26 08:21	06/03/26 07:55	1

**Lab Sample ID: LCS 570-746664/2-A**  
**Matrix: Water**  
**Analysis Batch: 748206**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	20.0	13.7		ug/L		69	47 - 120
2-Methylnaphthalene	20.0	12.3		ug/L		61	43 - 120
Acenaphthene	20.0	15.6		ug/L		78	60 - 132
Acenaphthylene	20.0	15.5		ug/L		77	54 - 126
Anthracene	20.0	15.2		ug/L		76	43 - 120
Benzo[a]anthracene	20.0	14.3		ug/L		71	42 - 133
Benzo[a]pyrene	20.0	15.1		ug/L		75	32 - 148
Benzo[b]fluoranthene	20.0	14.1		ug/L		71	42 - 140
Benzo[g,h,i]perylene	20.0	13.9		ug/L		69	1 - 195
Benzo[k]fluoranthene	20.0	14.2		ug/L		71	25 - 146
Chrysene	20.0	13.5		ug/L		68	44 - 140
Dibenz(a,h)anthracene	20.0	15.2		ug/L		76	1 - 200
Fluoranthene	20.0	15.7		ug/L		78	43 - 121
Fluorene	20.0	15.7		ug/L		78	70 - 120
Indeno[1,2,3-cd]pyrene	20.0	14.6		ug/L		73	1 - 151
Naphthalene	20.0	12.9		ug/L		65	36 - 120

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

**Lab Sample ID: LCS 570-746664/2-A**  
**Matrix: Water**  
**Analysis Batch: 748206**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Phenanthrene	20.0	15.2		ug/L		76	65 - 120
Pyrene	20.0	14.4		ug/L		72	70 - 120

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2,4,6-Tribromophenol (Surr)	73		28 - 127
2-Fluorobiphenyl (Surr)	75		31 - 120
2-Fluorophenol (Surr)	50		17 - 120
Nitrobenzene-d5 (Surr)	64		27 - 120
Phenol-d6 (Surr)	34		10 - 120
p-Terphenyl-d14 (Surr)	70		45 - 120

**Lab Sample ID: LCSD 570-746664/3-A**  
**Matrix: Water**  
**Analysis Batch: 748206**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit
1-Methylnaphthalene	20.0	13.0		ug/L		65	47 - 120	6	20
2-Methylnaphthalene	20.0	11.1		ug/L		56	43 - 120	10	20
Acenaphthene	20.0	15.2		ug/L		76	60 - 132	3	29
Acenaphthylene	20.0	14.9		ug/L		75	54 - 126	4	45
Anthracene	20.0	14.4		ug/L		72	43 - 120	5	40
Benzo[a]anthracene	20.0	15.4		ug/L		77	42 - 133	8	32
Benzo[a]pyrene	20.0	16.3		ug/L		82	32 - 148	8	43
Benzo[b]fluoranthene	20.0	15.6		ug/L		78	42 - 140	10	43
Benzo[g,h,i]perylene	20.0	15.7		ug/L		79	1 - 195	12	61
Benzo[k]fluoranthene	20.0	15.7		ug/L		79	25 - 146	10	38
Chrysene	20.0	15.1		ug/L		76	44 - 140	11	53
Dibenz(a,h)anthracene	20.0	17.0		ug/L		85	1 - 200	11	75
Fluoranthene	20.0	15.0		ug/L		75	43 - 121	4	40
Fluorene	20.0	15.2		ug/L		76	70 - 120	3	23
Indeno[1,2,3-cd]pyrene	20.0	16.0		ug/L		80	1 - 151	9	60
Naphthalene	20.0	11.9		ug/L		60	36 - 120	8	39
Phenanthrene	20.0	15.2		ug/L		76	65 - 120	0	24
Pyrene	20.0	15.6		ug/L		78	70 - 120	8	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
2,4,6-Tribromophenol (Surr)	69		28 - 127
2-Fluorobiphenyl (Surr)	72		31 - 120
2-Fluorophenol (Surr)	47		17 - 120
Nitrobenzene-d5 (Surr)	61		27 - 120
Phenol-d6 (Surr)	33		10 - 120
p-Terphenyl-d14 (Surr)	74		45 - 120

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

**Lab Sample ID: 380-216586-1 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 748206**

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
1-Methylnaphthalene	<0.19		19.6	14.6		ug/L		74	36 - 120
2-Methylnaphthalene	<0.19		19.6	13.0		ug/L		66	32 - 124
Acenaphthene	<0.19		19.6	16.6		ug/L		85	47 - 145
Acenaphthylene	<0.19		19.6	16.5		ug/L		84	33 - 145
Anthracene	<0.19		19.6	15.4		ug/L		78	27 - 133
Benzo[a]anthracene	<0.19		19.6	16.3		ug/L		83	33 - 143
Benzo[a]pyrene	<0.19		19.6	17.7		ug/L		90	17 - 163
Benzo[b]fluoranthene	<0.19		19.6	16.6		ug/L		84	24 - 159
Benzo[g,h,i]perylene	<0.19		19.6	16.6		ug/L		85	1 - 219
Benzo[k]fluoranthene	<0.19		19.6	16.4		ug/L		84	11 - 162
Chrysene	<0.19		19.6	16.5		ug/L		84	17 - 168
Dibenz(a,h)anthracene	<0.19		19.6	17.8		ug/L		90	1 - 227
Fluoranthene	<0.19		19.6	16.2		ug/L		83	26 - 137
Fluorene	<0.19		19.6	16.9		ug/L		86	59 - 121
Indeno[1,2,3-cd]pyrene	<0.19		19.6	17.2		ug/L		88	1 - 171
Naphthalene	<0.19		19.6	13.6		ug/L		69	21 - 133
Phenanthrene	<0.19		19.6	16.5		ug/L		84	54 - 120
Pyrene	<0.19		19.6	17.1		ug/L		87	52 - 120

Surrogate	MS	MS	Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	80		28 - 127
2-Fluorobiphenyl (Surr)	79		31 - 120
2-Fluorophenol (Surr)	54		17 - 120
Nitrobenzene-d5 (Surr)	69		27 - 120
Phenol-d6 (Surr)	37		10 - 120
p-Terphenyl-d14 (Surr)	78		45 - 120

**Lab Sample ID: 380-216586-1 MSD**  
**Matrix: Drinking Water**  
**Analysis Batch: 748206**

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
	Result	Qualifier	Added	Result	Qualifier						
1-Methylnaphthalene	<0.19		19.6	14.8		ug/L		75	36 - 120	1	30
2-Methylnaphthalene	<0.19		19.6	13.1		ug/L		67	32 - 124	1	30
Acenaphthene	<0.19		19.6	16.6		ug/L		85	47 - 145	0	48
Acenaphthylene	<0.19		19.6	16.7		ug/L		85	33 - 145	1	74
Anthracene	<0.19		19.6	15.8		ug/L		81	27 - 133	3	66
Benzo[a]anthracene	<0.19		19.6	17.6		ug/L		90	33 - 143	8	53
Benzo[a]pyrene	<0.19		19.6	18.6		ug/L		95	17 - 163	5	72
Benzo[b]fluoranthene	<0.19		19.6	17.7		ug/L		90	24 - 159	7	71
Benzo[g,h,i]perylene	<0.19		19.6	17.3		ug/L		88	1 - 219	4	97
Benzo[k]fluoranthene	<0.19		19.6	17.5		ug/L		89	11 - 162	6	63
Chrysene	<0.19		19.6	16.8		ug/L		86	17 - 168	2	87
Dibenz(a,h)anthracene	<0.19		19.6	18.6		ug/L		95	1 - 227	4	126
Fluoranthene	<0.19		19.6	16.3		ug/L		83	26 - 137	0	66
Fluorene	<0.19		19.6	16.7		ug/L		85	59 - 121	1	38
Indeno[1,2,3-cd]pyrene	<0.19		19.6	17.7		ug/L		90	1 - 171	2	99
Naphthalene	<0.19		19.6	13.7		ug/L		70	21 - 133	0	65

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

**Lab Sample ID: 380-216586-1 MSD**  
**Matrix: Drinking Water**  
**Analysis Batch: 748206**

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**Prep Type: Total/NA**  
**Prep Batch: 746664**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	RPD	Limit
	Result	Qualifier	Added	Result	Qualifier						
Phenanthrene	<0.19		19.6	17.1		ug/L		87	54 - 120	4	39
Pyrene	<0.19		19.6	18.1		ug/L		92	52 - 120	6	49
<b>MSD MSD</b>											
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>								
2,4,6-Tribromophenol (Surr)	77		28 - 127								
2-Fluorobiphenyl (Surr)	83		31 - 120								
2-Fluorophenol (Surr)	57		17 - 120								
Nitrobenzene-d5 (Surr)	69		27 - 120								
Phenol-d6 (Surr)	39		10 - 120								
p-Terphenyl-d14 (Surr)	81		45 - 120								

## Method: 8015B GRO LL - Gasoline Range Organics - (GC)

**Lab Sample ID: MB 570-750190/6**  
**Matrix: Water**  
**Analysis Batch: 750190**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac			
	Result	Qualifier									
GRO (C6-C10)	<10		10	ug/L			06/06/26 13:51	1			
<b>MB MB</b>											
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>						<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
4-Bromofluorobenzene (Surr)	102		38 - 134						06/06/26 13:51	1	

**Lab Sample ID: LCS 570-750190/3**  
**Matrix: Water**  
**Analysis Batch: 750190**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike	LCS	LCS	Unit	D	%Rec	%Rec	Limit
Gasoline Range Organics (C4-C13)	400	404		ug/L		101	78 - 120	
<b>LCS LCS</b>								
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>					
4-Bromofluorobenzene (Surr)	98		38 - 134					

**Lab Sample ID: LCSD 570-750190/4**  
**Matrix: Water**  
**Analysis Batch: 750190**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**

Analyte	Spike	LCSD	LCSD	Unit	D	%Rec	%Rec	Limit	RPD	Limit
Gasoline Range Organics (C4-C13)	400	424		ug/L		106	78 - 120	5	10	
<b>LCSD LCSD</b>										
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>							
4-Bromofluorobenzene (Surr)	98		38 - 134							

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

**Lab Sample ID: MRL 570-750190/5**  
**Matrix: Water**  
**Analysis Batch: 750190**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	10.0	14.0		ug/L		140	50 - 150
<b>Surrogate</b>		<b>MRL %Recovery</b>	<b>MRL Qualifier</b>				<b>Limits</b>
4-Bromofluorobenzene (Surr)		99					38 - 134

**Lab Sample ID: 380-216586-1 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 750190**

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	<10		400	402		ug/L		101	68 - 122
<b>Surrogate</b>		<b>MS %Recovery</b>		<b>MS Qualifier</b>					<b>Limits</b>
4-Bromofluorobenzene (Surr)		100							38 - 134

**Lab Sample ID: 380-216586-1 MSD**  
**Matrix: Drinking Water**  
**Analysis Batch: 750190**

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	<10		400	399		ug/L		100	68 - 122	1	18
<b>Surrogate</b>		<b>MSD %Recovery</b>		<b>MSD Qualifier</b>					<b>Limits</b>		
4-Bromofluorobenzene (Surr)		89							38 - 134		

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

**Lab Sample ID: MB 570-747016/1-A**  
**Matrix: Water**  
**Analysis Batch: 750788**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 747016**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<25		25	ug/L		06/01/26 09:03	06/08/26 11:46	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		06/01/26 09:03	06/08/26 11:46	1
C8-C18	<25		25	ug/L		06/01/26 09:03	06/08/26 11:46	1
<b>Surrogate</b>		<b>MB %Recovery</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
n-Octacosane (Surr)		103				06/01/26 09:03	06/08/26 11:46	1

**Lab Sample ID: LCS 570-747016/2-A**  
**Matrix: Water**  
**Analysis Batch: 750788**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 747016**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	1600	1720		ug/L		108	56 - 127

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

**Lab Sample ID: LCS 570-747016/2-A**  
**Matrix: Water**  
**Analysis Batch: 750788**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 747016**

	LCS %Recovery	LCS Qualifier	Limits
<i>n-Octacosane (Surr)</i>	108		60 - 130

**Lab Sample ID: LCSD 570-747016/3-A**  
**Matrix: Water**  
**Analysis Batch: 750788**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 747016**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit		
C10-C28	1600	1680		ug/L		105	56 - 127	2	23		
<i>n-Octacosane (Surr)</i>											
Surrogate	%Recovery	LCSD Qualifier	Limits								
<i>n-Octacosane (Surr)</i>	109		60 - 130								

**Lab Sample ID: MRL 570-747016/4-A**  
**Matrix: Water**  
**Analysis Batch: 750788**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 747016**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits		
C10-C28	0.0200	0.0311	^3+	mg/L		156	50 - 150		
<i>n-Octacosane (Surr)</i>									
Surrogate	%Recovery	MRL Qualifier	Limits						
<i>n-Octacosane (Surr)</i>	98		60 - 130						

**Lab Sample ID: 380-216586-1 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 750788**

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**Prep Type: Total/NA**  
**Prep Batch: 747016**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	<25	^3+	1650	1800		ug/L		109	70 - 130
<i>n-Octacosane (Surr)</i>									
Surrogate	%Recovery	MS Qualifier	Limits						
<i>n-Octacosane (Surr)</i>	110		60 - 130						

**Lab Sample ID: 380-216586-1 MSD**  
**Matrix: Drinking Water**  
**Analysis Batch: 750788**

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**  
**Prep Type: Total/NA**  
**Prep Batch: 747016**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	<25	^3+	1660	1720		ug/L		104	70 - 130
<i>n-Octacosane (Surr)</i>									
Surrogate	%Recovery	MSD Qualifier	Limits						
<i>n-Octacosane (Surr)</i>	102		60 - 130						

# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## GC/MS Semi VOA

### Prep Batch: 231139

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	525.2	
MB 380-231139/21-A	Method Blank	Total/NA	Water	525.2	
LCS 380-231139/23-A	Lab Control Sample	Total/NA	Water	525.2	
MRL 380-231139/22-A	Lab Control Sample	Total/NA	Water	525.2	
380-216945-B-4-A MS	Matrix Spike	Total/NA	Water	525.2	
380-216945-B-4-B MSD	Matrix Spike Duplicate	Total/NA	Water	525.2	

### Analysis Batch: 231376

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	525.2	231139
MB 380-231139/21-A	Method Blank	Total/NA	Water	525.2	231139
LCS 380-231139/23-A	Lab Control Sample	Total/NA	Water	525.2	231139
MRL 380-231139/22-A	Lab Control Sample	Total/NA	Water	525.2	231139
380-216945-B-4-A MS	Matrix Spike	Total/NA	Water	525.2	231139
380-216945-B-4-B MSD	Matrix Spike Duplicate	Total/NA	Water	525.2	231139

### Prep Batch: 746664

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	625.1	
MB 570-746664/1-A	Method Blank	Total/NA	Water	625.1	
LCS 570-746664/2-A	Lab Control Sample	Total/NA	Water	625.1	
LCSD 570-746664/3-A	Lab Control Sample Dup	Total/NA	Water	625.1	
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	625.1	
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	625.1	

### Analysis Batch: 748206

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	625.1 SIM	746664
MB 570-746664/1-A	Method Blank	Total/NA	Water	625.1 SIM	746664
LCS 570-746664/2-A	Lab Control Sample	Total/NA	Water	625.1 SIM	746664
LCSD 570-746664/3-A	Lab Control Sample Dup	Total/NA	Water	625.1 SIM	746664
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	625.1 SIM	746664
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	625.1 SIM	746664

### Analysis Batch: 749030

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	625.1	746664
MB 570-746664/1-A	Method Blank	Total/NA	Water	625.1	746664

## GC VOA

### Analysis Batch: 750190

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	8015B GRO LL	
380-216586-2	TB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	8015B GRO LL	
MB 570-750190/6	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-750190/3	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-750190/4	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	
MRL 570-750190/5	Lab Control Sample	Total/NA	Water	8015B GRO LL	
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	8015B GRO LL	
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	8015B GRO LL	

Eurofins Pomona

# QC Association Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-216586-1  
 SDG: Weekly: Halawa Wells P1 (MS/MSD)

## GC Semi VOA

### Prep Batch: 747016

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	3510C	
MB 570-747016/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-747016/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-747016/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-747016/4-A	Lab Control Sample	Total/NA	Water	3510C	
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	3510C	
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	3510C	

### Analysis Batch: 750788

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	8015B	747016
MB 570-747016/1-A	Method Blank	Total/NA	Water	8015B	747016
LCS 570-747016/2-A	Lab Control Sample	Total/NA	Water	8015B	747016
LCSD 570-747016/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	747016
MRL 570-747016/4-A	Lab Control Sample	Total/NA	Water	8015B	747016
380-216586-1 MS	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	8015B	747016
380-216586-1 MSD	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	8015B	747016



# Lab Chronicle

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-216586-1  
 SDG: Weekly: Halawa Wells P1 (MS/MSD)

**Client Sample ID: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-1**

**Date Collected: 05/26/26 10:22**

**Matrix: Drinking Water**

**Date Received: 05/28/26 10:10**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			231139	IQ42	EA POM	06/02/26 15:52
Total/NA	Analysis	525.2		1	231376	UPAC	EA POM	06/03/26 18:54
Total/NA	Prep	625.1			746664	KLZQ	EET CAL 4	05/31/26 08:21
Total/NA	Analysis	625.1		1	749030	PQS1	EET CAL 4	06/04/26 12:27
Total/NA	Prep	625.1			746664	KLZQ	EET CAL 4	05/31/26 08:21
Total/NA	Analysis	625.1 SIM		1	748206	PQS1	EET CAL 4	06/03/26 11:54
Total/NA	Analysis	8015B GRO LL		1	750190	A9VE	EET CAL 4	06/06/26 14:47
Total/NA	Prep	3510C			747016	TVD6	EET CAL 4	06/01/26 09:04
Total/NA	Analysis	8015B		1	750788	NR	EET CAL 4	06/08/26 13:54

**Client Sample ID: TB: HALAWA WELLS P1 (331-023-WL065)**

**Lab Sample ID: 380-216586-2**

**Date Collected: 05/26/26 10:22**

**Matrix: Water**

**Date Received: 05/28/26 10:10**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	750190	A9VE	EET CAL 4	06/06/26 14:23

**Laboratory References:**

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Laboratory: Eurofins Pomona

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *
The following analytes are included in this report, but the laboratory is not certified by Hawaii State CA00006. This list may include analytes for which the agency does not offer certification:			
Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	1-Methylnaphthalene
525.2	525.2	Drinking Water	2,4'-DDD
525.2	525.2	Drinking Water	2,4'-DDE
525.2	525.2	Drinking Water	2,4'-DDT
525.2	525.2	Drinking Water	2,4-Dinitrotoluene
525.2	525.2	Drinking Water	2,6-Dinitrotoluene
525.2	525.2	Drinking Water	2-Methylnaphthalene
525.2	525.2	Drinking Water	4,4'-DDD
525.2	525.2	Drinking Water	4,4'-DDE
525.2	525.2	Drinking Water	4,4'-DDT
525.2	525.2	Drinking Water	Acetochlor
525.2	525.2	Drinking Water	alpha-BHC
525.2	525.2	Drinking Water	alpha-Chlordane
525.2	525.2	Drinking Water	beta-BHC
525.2	525.2	Drinking Water	Chlorobenzilate
525.2	525.2	Drinking Water	Chloroneb
525.2	525.2	Drinking Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Drinking Water	Chlorpyrifos
525.2	525.2	Drinking Water	delta-BHC
525.2	525.2	Drinking Water	Diclorvos (DDVP)
525.2	525.2	Drinking Water	Endosulfan I (Alpha)
525.2	525.2	Drinking Water	Endosulfan II (Beta)
525.2	525.2	Drinking Water	Endosulfan sulfate
525.2	525.2	Drinking Water	Endrin aldehyde
525.2	525.2	Drinking Water	EPTC
525.2	525.2	Drinking Water	gamma-Chlordane
525.2	525.2	Drinking Water	Isophorone
525.2	525.2	Drinking Water	Malathion
525.2	525.2	Drinking Water	Parathion
525.2	525.2	Drinking Water	Pendimethalin (Penoxaline)
525.2	525.2	Drinking Water	Terbacil
525.2	525.2	Drinking Water	Terbutylazine
525.2	525.2	Drinking Water	Total Permethrin (mixed isomers)
525.2	525.2	Drinking Water	trans-Nonachlor

## Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
A2LA	Dept. of Defense ELAP	7296.01	11-30-26
A2LA	ISO/IEC 17025	7296.01	11-30-26
Alaska (UST)	State	25-005	03-02-27
Arizona	State	AZ0830	11-17-26
California	Los Angeles County Sanitation Districts	9257304	07-31-26
California	SCAQMD LAP	17LA0919	11-30-26

\* Accreditation/Certification renewal pending - accreditation/certification considered valid.

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

## Laboratory: Eurofins Calscience (Continued)

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	3082	07-31-26
Kansas	NELAP	E-10420	07-31-26
Nevada	State	CA00111	07-31-26
Oregon	NELAP	4175	02-02-27
USDA	US Federal Programs	525-23-159-97150	09-30-26
Utah	NELAP	CA00111	02-28-27
Washington	State	C916	10-12-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA POM
625.1	Semivolatile Organic Compounds (GC/MS)	EPA	EET CAL 4
625.1 SIM	Semivolatile Organic Compounds GC/MS (SIM)	EPA	EET CAL 4
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4
525.2	Extraction of Semivolatile Compounds	EPA	EA POM
625.1	Liquid-Liquid Extraction	40CFR136A	EET CAL 4

#### Protocol References:

40CFR136A = "Methods for Organic Chemical Analysis of Municipal Industrial Wastewater", 40CFR, Part 136, Appendix A, October 26, 1984 and subsequent revisions.

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

#### Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-216586-1  
SDG: Weekly: Halawa Wells P1 (MS/MSD)

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-216586-1	HALAWA WELLS P1 (331-023-WL065)	Drinking Water	05/26/26 10:22	05/28/26 10:10	HI0000331
380-216586-2	TB: HALAWA WELLS P1 (331-023-WL065)	Water	05/26/26 10:22	05/28/26 10:10	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

**Chain of Custody Record**



380-216586 COC

**Client Information**  
 Client Contact: Kirk Iwamoto  
 Phone: +1 808 748 5940  
 City & County of Honolulu  
 Address: 630 South Beretania Street Chemistry Lab  
 Honolulu HI, 96843  
 State, Zip: HI, 96843  
 Phone: 808-748-5840 (Tel)  
 Email: kiwamoto@hbws.org  
 Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill  
 Site: Hawaii

**Company:**  
 Lab Pkt: Lopez, Maria  
 E-Mail: Maria.Lopez@et.eurofins.com  
 PWSID:  
 Date Date Requested:  
 TAT Requested (days):  
 Compliance Project:  Yes  No  
 PO #: C20525101 exp 05312023  
 WO #:  
 Project #: 38001111  
 SSOW#:

Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (Water, Specific, Organics, Other)	Preservation Code	Field Filtered Sample (Yes or No)		Perform Method (Yes or No)		RA	Q	QA	Y	I	Special Instructions/Notes:
						Field Filtered Sample (Yes or No)	Perform Method (Yes or No)	RA	Q						
Halawa Wells P1 (331-023-WL065)	26-May-2026	1022	G	Water		X	X	4	5	4	2				
Halawa Wells P1 (331-023-WL065) (Matrix Spike)				Water		X	X								
Halawa Wells P1 (331-023-WL065) (Matrix Spike Duplicate)				Water		X	X								
TB: Halawa Wells P1 (331-023-WL065)	26-May-2026	1022		Water				2							

**Possible Hazard Identification**  
 Non-Hazard  Flammable  Skin Irritant  Poison B  Unknown  Radiological  
 Deliverable Requested: I, II, III, IV, Other (specify)

**Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)**  
 Return To Client  Disposal By Lab  Archive For \_\_\_\_\_ Months

**Special Instructions/QC Requirements:**

Method of Shipment: Fedex 977308375588  
 Date/Time: 5/25/26 1010  
 Company: ELK  
 Received by: Maria Lopez  
 Date/Time: \_\_\_\_\_  
 Company: \_\_\_\_\_  
 Received by: \_\_\_\_\_  
 Date/Time: \_\_\_\_\_  
 Company: \_\_\_\_\_

Cooler Temperature(s) °C and Other Remarks: (331A) 5-8x0-0-58 gal - frozen  
 Ver: 04/02/2024



**Chain of Custody Record**



380-216586 COC

**Client Information**  
 Client Contact: Kirk Iwamoto  
 Phone: +1 808 748 5940  
 City & County of Honolulu  
 Address: 630 South Beretania Street Chemistry Lab  
 Honolulu HI, 96843  
 Phone: 808-748-5840 (Tel)  
 Email: kiwamoto@hbws.org  
 Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill  
 Site: Hawaii

**Company:**  
 Lab Pkt: Lopez, Maria  
 E-Mail: Maria.Lopez@et.eurofins.com  
 State of Origin: Hawaii

**Analysis Requested**  
 Preservation Codes:  
 R - NaHSO4  
 RA - NaThioHCl  
 G - Na2SO3  
 CA - Na2SO3/HCl  
 Y - Triams  
 I - NH4 Acetate

Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Preservation Code	Matrix (Water, Specific, Organics, Other)	Field Filtered Sample (Yes or No)	Perform Method (Yes or No)	RA	QA	Y	I	Special Instructions/Notes:
Halawa Wells P1 (331-023-WL065)	26-May-2026	1022	G		Water	X	X	4	5	4	2	
Halawa Wells P1 (331-023-WL065) (Matrix Spike)					Water			X	X	X		
Halawa Wells P1 (331-023-WL065) (Matrix Spike Duplicate)					Water			X	X	X		
TB: Halawa Wells P1 (331-023-WL065)	26-May-2026	1022			Water			2				

**Possible Hazard Identification**  
 Non-Hazard  Flammable  Skin Irritant  Poison B  Unknown  Radiological  
 Deliverable Requested: I, II, III, IV, Other (specify)

**Empty Kit Relinquished by:** \_\_\_\_\_ Date: \_\_\_\_\_  
**Relinquished by:** \_\_\_\_\_ Date/Time: \_\_\_\_\_  
**Relinquished by:** \_\_\_\_\_ Date/Time: \_\_\_\_\_  
**Relinquished by:** \_\_\_\_\_ Date/Time: \_\_\_\_\_

**Custody Seals Intact:** Custody Seal No. \_\_\_\_\_  
 Δ Yes Δ No

**Special Instructions/QC Requirements:**  
 Return To Client  Disposal By Lab  Archive For \_\_\_\_\_ Months

**Method of Shipment:** Fedex 977308375588  
 Date/Time: 5/25/26 1010  
 Company: ELK  
 Received by: \_\_\_\_\_ Date/Time: \_\_\_\_\_  
 Received by: \_\_\_\_\_ Date/Time: \_\_\_\_\_  
 Received by: \_\_\_\_\_ Date/Time: \_\_\_\_\_

**Cooler Temperature(s) °C and Other Remarks:** (331A) 5-8x0-0-58 gal - frozen  
 Ver: 04/02/2024





ORIGIN ID HIK  
BWS CHEMLAB  
HONOLULU BOARD OF WATER SUPPLY  
630 S. BERETANIA ST.  
CHEMICAL LABORATORY  
HONOLULU, HI 96843  
UNITED STATES US

SHIP DATE 27MAY26  
ACTWGT 62.00 LB  
CAD 258050552/INET4535


BILL RECIPIENT

TO EUROFINS RECEIVING DEPARTMENT  
EUROFINS DRINKING WATER TESTING  
941 CORPORATE CENTER DR  
POMONA CA 91768

REF: (626) 386-1100  
INV# PO# DEPT#



1 of 6  
TRK# 8723 0837 5555  
# MASTER # 91768  
WM ONTA CA-US  
THU - 28 MAY 10:30A  
PRIORITY OVERNIGHT



(631A) 4.6 x 0.0 x 4.6 961-470261  
New Mark Urtatin 5/29/26 1003

After printing this label  
CONSIGNEE COPY - PLEASE PLACE IN FRONT OF POUCH  
1 Fold the printed page along the horizontal line  
2 Place label in shipping pouch and affix it to your shipment

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on  
fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage,  
delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document  
your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from  
FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and  
other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized  
declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g.  
jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed  
within strict time limits, see current FedEx Service Guide

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

# Chain of Custody Record



380-216586 COC

**Client Information**  
 Client Contact: Kirk Iwamoto  
 Phone: +1 808 748 5840  
 Lab PM: Lopez, Maria  
 State of Origin:   
 Carrier Tracking No(s):   
 Page: Page 1 of 1  
 Job #:

**Company**  
 City & County of Honolulu  
 Address: 630 South Beretania Street Chemistry Lab  
 City: Honolulu  
 State: HI, 96843  
 Phone: 808-748-5840 (Tel)  
 Email: kiwamoto@hbws.org  
 Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill  
 Site: Hawaii

**Analysis Requested**  
 Due Date Requested:   
 TAT Requested (days):   
 Compliance Project:   
 PO #: C20525101 exp 05312023  
 WO #:   
 Project #: 38001111  
 SSON#:   
 Matrix (Water, Soil, Sediment, Other):   
 Matrix (Water, Soil, Sediment, Other):   
 Matrix (Water, Soil, Sediment, Other):   
 Matrix (Water, Soil, Sediment, Other):

Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Preservation Code:	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	R	RA	Q	QA	Y	Special Instructions/Note:
Halawa Wells P1 (331-023-WL065)	26-May-2026	1022	G		X		4	5	4	2		
Halawa Wells P1 (331-023-WL065) (Matrix Spike)					X		X	X				
Halawa Wells P1 (331-023-WL065)(Matrix Spike Duplicate)					X		X	X				
TB: Halawa Wells P1 (331-023-WL065)	26-May-2026	1022					2					

**Possible Hazard Identification**  
 Non-Hazard  Flammable  Skin Irritant  Unknown  Radiological  
 Deliverable Requested: I, II, III, IV, Other (specify)   
 Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)  
 Return To Client  Disposal By Lab  Archive For   
 Months

**Empty Kit Relinquished by:**  
 Relinquished by:   
 Date/Time:   
 Relinquished by:   
 Date/Time:   
 Relinquished by:   
 Date/Time:   
 Custody Seal No.:   
 Δ Yes Δ No

**Method of Shipment:** FedEx 8723 0837 5555  
 Received by:   
 Date/Time: 5/29/26 1003  
 Company: CEAR  
 Received by:   
 Date/Time:   
 Company:   
 Received by:   
 Date/Time:   
 Company:   
 Cooler Temperature(s) °C and Other Remarks: (31A) 46-40.0 4.6 981-F0026  
 Ver 04/02/2024



ORIGIN ID HIK  
BWS CHEMLAB  
HONOLULU BOARD OF WATER SUPPLY  
630 S. BERETANIA ST.  
CHEMICAL LABORATORY  
HONOLULU, HI 96843  
UNITED STATES US



SHIP DATE 27MAY26  
ACTWGT 62.00 LB  
CAD 258050552/INET4535

BILL RECIPIENT

TO EUROFINS RECEIVING DEPARTMENT  
EUROFINS DRINKING WATER TESTING  
941 CORPORATE CENTER DR  
POMONA CA 91768

REF: (626) 386-1100

DEPT:



THU - 28 MAY 10:30A  
PRIORITY OVERNIGHT

1 of 6

TRK# 8723 0837 5555

0201  
## MASTER #

WM ONTA

91768

CA-US



(631A) 4.6 x 0.0 x 4.6 961-470261  
New Mark Urtatin 5/29/26 1003

After printing this label  
CONSIGNEE COPY - PLEASE PLACE IN FRONT OF POUCH  
1 Fold the printed page along the horizontal line  
2 Place label in shipping pouch and affix it to your shipment

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on  
fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage,  
delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document  
your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from  
FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and  
other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized  
declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g.  
jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed  
within strict time limits, see current FedEx Service Guide

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16









## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-216586-1  
SDG Number: Weekly: Halawa Wells P1 (MS/MSD)

**Login Number: 216586**  
**List Number: 1**  
**Creator: Tran, Kristine**

**List Source: Eurofins Pomona**

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	



## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-216586-1  
SDG Number: Weekly: Halawa Wells P1 (MS/MSD)

**Login Number: 216586**

**List Number: 2**

**Creator: Khana, Piyush**

**List Source: Eurofins Calscience**

**List Creation: 05/29/26 03:51 PM**

Question	Answer	Comment
Radioactivity wasn't checked or is </= background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	1.4
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	fgf5
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	



## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-216586-1  
SDG Number: Weekly: Halawa Wells P1 (MS/MSD)

**Login Number: 216586**

**List Number: 3**

**Creator: Khana, Piyush**

**List Source: Eurofins Calscience**

**List Creation: 05/29/26 07:11 PM**

Question	Answer	Comment
Radioactivity wasn't checked or is <math>\leq</math> background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	1.9
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <math><6\text{mm}</math> (1/4").	True	fgf5
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	