

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Erwin Kawata  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 310  
Honolulu, Hawaii 96843

Generated 5/27/2026 11:08:18 AM

## JOB DESCRIPTION

RED-HILL  
PFAS: Ka'amilo Wells P2

## JOB NUMBER

380-215228-1

# Eurofins Pomona

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

## Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

## Authorization



Authorized for release by  
Maria Lopez, Project Manager  
[Maria.Lopez@et.eurofinsus.com](mailto:Maria.Lopez@et.eurofinsus.com)  
(626)386-1100

Generated  
5/27/2026 11:08:18 AM



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	11
Surrogate Summary . . . . .	12
Isotope Dilution Summary . . . . .	13
QC Sample Results . . . . .	14
QC Association Summary . . . . .	25
Lab Chronicle . . . . .	26
Certification Summary . . . . .	27
Method Summary . . . . .	28
Sample Summary . . . . .	29
Chain of Custody . . . . .	30
Receipt Checklists . . . . .	32

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Qualifiers

### LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: City & County of Honolulu  
Project: RED-HILL

Job ID: 380-215228-1

**Job ID: 380-215228-1**

**Eurofins Pomona**

## Job Narrative 380-215228-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

### Receipt

The samples were received on 5/20/2026 9:27 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.2°C.

### PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-1**

**PWSID Number: HI0000331**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid (PFBS)	3.0		2.0	ng/L	1		533	Total/NA
Perfluoroheptanoic acid (PFHpA)	2.0		2.0	ng/L	1		533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.7		2.0	ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	4.3		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.7		2.0	ng/L	1		533	Total/NA
Perfluorooctanoic acid (PFOA)	3.9		2.0	ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	4.5		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	5.1		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanoic acid (PFHxA)	4.4		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	4.5		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	4.3		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorobutanesulfonic acid (PFBS)	3.4		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluoroheptanoic acid (PFHpA)	2.4		2.0	ng/L	1		EPA 537.1 V2	Total/NA

**Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-2**

**PWSID Number: HI0000331**

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-1**

Date Collected: 05/18/26 12:44

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Perfluorobutanesulfonic acid (PFBS)</b>	<b>3.0</b>		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Perfluoroheptanoic acid (PFHpA)</b>	<b>2.0</b>		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.7</b>		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Perfluorohexanoic acid (PFHxA)</b>	<b>4.3</b>		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>4.7</b>		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Perfluorooctanoic acid (PFOA)</b>	<b>3.9</b>		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Perfluoropentanoic acid (PFPeA)</b>	<b>4.5</b>		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:37	1
<b>Isotope Dilution</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>			<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
13C3 HFPO-DA	103		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C6 PFDA	101		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C5 PFHxA	104		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C4 PFHpA	108		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C8 PFOA	108		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C9 PFNA	108		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C7 PFUnA	110		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C2 PFDoA	103		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C4 PFBA	111		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C5 PFPeA	105		50 - 200			05/21/26 18:50	05/23/26 11:37	1
13C3 PFBS	109		50 - 200			05/21/26 18:50	05/23/26 11:37	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-1**

Date Collected: 05/18/26 12:44

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 PFHxS	112		50 - 200	05/21/26 18:50	05/23/26 11:37	1
13C8 PFOS	112		50 - 200	05/21/26 18:50	05/23/26 11:37	1
13C2-4:2-FTS	111		50 - 200	05/21/26 18:50	05/23/26 11:37	1
13C2-6:2-FTS	104		50 - 200	05/21/26 18:50	05/23/26 11:37	1
13C2-8:2-FTS	101		50 - 200	05/21/26 18:50	05/23/26 11:37	1

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>5.1</b>		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
N-methylperfluorooctanesulfonamide cetic acid (NMeFOSAA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
<b>Perfluorohexanoic acid (PFHxA)</b>	<b>4.4</b>		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
<b>Perfluorooctanoic acid (PFOA)</b>	<b>4.5</b>		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>4.3</b>		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
<b>Perfluorobutanesulfonic acid (PFBS)</b>	<b>3.4</b>		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
<b>Perfluoroheptanoic acid (PFHpA)</b>	<b>2.4</b>		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:10	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	112		70 - 130	05/21/26 10:42	05/22/26 20:10	1
13C2 PFHxA	113		70 - 130	05/21/26 10:42	05/22/26 20:10	1
13C2 PFDA	114		70 - 130	05/21/26 10:42	05/22/26 20:10	1
13C3-GenX	107		70 - 130	05/21/26 10:42	05/22/26 20:10	1

**Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-2**

Date Collected: 05/18/26 12:44

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-2**

Date Collected: 05/18/26 12:44

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		05/21/26 18:50	05/23/26 14:29	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	106		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C6 PFDA	104		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C5 PFHxA	116		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C4 PFHpA	116		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C8 PFOA	114		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C9 PFNA	119		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C7 PFUnA	109		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C2 PFDoA	105		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C4 PFBA	112		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C5 PFPeA	113		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C3 PFBS	110		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C3 PFHxS	112		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C8 PFOS	112		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C2-4:2-FTS	117		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C2-6:2-FTS	106		50 - 200	05/21/26 18:50	05/23/26 14:29	1
13C2-8:2-FTS	105		50 - 200	05/21/26 18:50	05/23/26 14:29	1

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-2**

Date Collected: 05/18/26 12:44

Matrix: Water

Date Received: 05/20/26 09:27

PWSID Number: HI0000331

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
N-methylperfluorooctanesulfonamide acetic acid (NMeFOSAA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/21/26 10:42	05/22/26 20:19	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>			<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
d5-NEtFOSAA	109		70 - 130			05/21/26 10:42	05/22/26 20:19	1
13C2 PFHxA	115		70 - 130			05/21/26 10:42	05/22/26 20:19	1
13C2 PFDA	119		70 - 130			05/21/26 10:42	05/22/26 20:19	1
13C3-GenX	106		70 - 130			05/21/26 10:42	05/22/26 20:19	1

## Action Limit Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-1**

**PWSID Number: HI0000331**

### Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.7		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>4.7</b>		ng/L	<b>4</b>	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	3.9		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>5.1</b>		ng/L	<b>4</b>	2.0	EPA 537.1 V2	Total/NA
<b>Perfluorooctanoic acid (PFOA)</b>	<b>4.5</b>		ng/L	<b>4</b>	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	4.3		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

**Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-2**

**PWSID Number: HI0000331**

### Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-215228-1  
 SDG: PFAS: Ka'amilo Wells P2

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

**Matrix: Water**

**Prep Type: Total/NA**

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-214936-D-2-A MSD	Matrix Spike Duplicate	105	113	110	106
380-214936-E-2-A MS	Matrix Spike	105	114	113	114
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	112	113	114	107
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	109	115	119	106
LCS 380-228643/21-A	Lab Control Sample	103	111	118	113
MBL 380-228643/19-A	Method Blank	105	104	109	96
MRL 380-228643/20-A	Lab Control Sample	110	109	107	100

**Surrogate Legend**

d5NEFOS = d5-NEtFOSAA  
 PFHxA = 13C2 PFHxA  
 PFDA = 13C2 PFDA  
 GenX = 13C3-GenX



# Isotope Dilution Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

**Matrix: Water**

**Prep Type: Total/NA**

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	103	101	104	108	108	108	110	103
380-215228-1 MS	Ka'amilo Wells P2 (331-600-WL085)	118	105	107	108	113	114	114	111
380-215228-1 MSD	Ka'amilo Wells P2 (331-600-WL085)	109	102	107	105	104	111	112	107
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	106	104	116	116	114	119	109	105
LCS 380-228770/22-A	Lab Control Sample	112	106	110	105	108	111	114	113
MBL 380-228770/20-A	Method Blank	86	87	94	91	92	97	98	96
MRL 380-228770/21-A	Lab Control Sample	88	96	99	97	99	103	103	101

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	111	105	109	112	112	111	104	101
380-215228-1 MS	Ka'amilo Wells P2 (331-600-WL085)	109	106	107	109	113	107	107	109
380-215228-1 MSD	Ka'amilo Wells P2 (331-600-WL085)	112	104	108	106	111	104	107	107
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	112	113	110	112	112	117	106	105
LCS 380-228770/22-A	Lab Control Sample	81	104	112	109	112	105	110	107
MBL 380-228770/20-A	Method Blank	87	91	105	108	115	116	105	103
MRL 380-228770/21-A	Lab Control Sample	87	98	107	110	115	114	107	107

**Surrogate Legend**

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

**Lab Sample ID: MBL 380-228770/20-A**  
**Matrix: Water**  
**Analysis Batch: 229158**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 228770**

Analyte	MBL	MBL	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.25		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		05/21/26 18:50	05/23/26 11:08	1

Isotope Dilution	MBL	MBL	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C3 HFPO-DA	86		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C6 PFDA	87		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C5 PFHxA	94		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C4 PFHpA	91		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C8 PFOA	92		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C9 PFNA	97		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C7 PFUnA	98		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C2 PFDoA	96		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C4 PFBA	87		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C5 PFPeA	91		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C3 PFBS	105		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C3 PFHxS	108		50 - 200	05/21/26 18:50	05/23/26 11:08	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MBL 380-228770/20-A**  
**Matrix: Water**  
**Analysis Batch: 229158**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 228770**

Isotope Dilution	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C8 PFOS	115		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C2-4:2-FTS	116		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C2-6:2-FTS	105		50 - 200	05/21/26 18:50	05/23/26 11:08	1
13C2-8:2-FTS	103		50 - 200	05/21/26 18:50	05/23/26 11:08	1

**Lab Sample ID: LCS 380-228770/22-A**  
**Matrix: Water**  
**Analysis Batch: 229158**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 228770**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	120	102		ng/L		85	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	112		ng/L		93	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	106		ng/L		88	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	108		ng/L		90	70 - 130
Perfluorodecanoic acid (PFDA)	120	112		ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	120	106		ng/L		88	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	113		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	111		ng/L		92	70 - 130
Perfluorohexanoic acid (PFHxA)	120	111		ng/L		93	70 - 130
Perfluorononanoic acid (PFNA)	120	108		ng/L		90	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	109		ng/L		91	70 - 130
Perfluorooctanoic acid (PFOA)	120	110		ng/L		91	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	110		ng/L		92	70 - 130
Perfluorobutanoic acid (PFBA)	120	108		ng/L		90	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	108		ng/L		90	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	116		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	104		ng/L		87	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	108		ng/L		90	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	120	113		ng/L		94	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	111		ng/L		92	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	109		ng/L		91	70 - 130
Perfluoropentanoic acid (PFPeA)	120	110		ng/L		92	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	111		ng/L		92	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: LCS 380-228770/22-A**  
**Matrix: Water**  
**Analysis Batch: 229158**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 228770**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	111		ng/L		92	70 - 130
<b>LCS LCS</b>							
<b>Isotope Dilution</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				
13C3 HFPO-DA	112		50 - 200				
13C6 PFDA	106		50 - 200				
13C5 PFHxA	110		50 - 200				
13C4 PFHpA	105		50 - 200				
13C8 PFOA	108		50 - 200				
13C9 PFNA	111		50 - 200				
13C7 PFUnA	114		50 - 200				
13C2 PFDoA	113		50 - 200				
13C4 PFBA	81		50 - 200				
13C5 PFPeA	104		50 - 200				
13C3 PFBS	112		50 - 200				
13C3 PFHxS	109		50 - 200				
13C8 PFOS	112		50 - 200				
13C2-4:2-FTS	105		50 - 200				
13C2-6:2-FTS	110		50 - 200				
13C2-8:2-FTS	107		50 - 200				

**Lab Sample ID: MRL 380-228770/21-A**  
**Matrix: Water**  
**Analysis Batch: 229158**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 228770**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.55	J	ng/L		78	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	1.71	J	ng/L		85	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	1.84	J	ng/L		92	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	1.87	J	ng/L		94	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.78	J	ng/L		89	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	1.97	J	ng/L		98	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.00	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	1.87	J	ng/L		93	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.02	J	ng/L		101	50 - 150
Perfluorononanoic acid (PFNA)	2.00	1.85	J	ng/L		92	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	1.86	J	ng/L		93	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	1.93	J	ng/L		97	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	1.99	J	ng/L		99	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	2.04	J	ng/L		102	50 - 150

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MRL 380-228770/21-A**

**Matrix: Water**

**Analysis Batch: 229158**

**Client Sample ID: Lab Control Sample**

**Prep Type: Total/NA**

**Prep Batch: 228770**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	1.93	J	ng/L		96	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.14	J	ng/L		107	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	1.72	J	ng/L		86	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.00	1.93	J	ng/L		96	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.00	2.07	J	ng/L		103	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	1.80	J	ng/L		90	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	1.71	J	ng/L		85	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.03	J	ng/L		101	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	1.77	J	ng/L		88	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	1.88	J	ng/L		94	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	88		50 - 200
13C6 PFDA	96		50 - 200
13C5 PFHxA	99		50 - 200
13C4 PFHpA	97		50 - 200
13C8 PFOA	99		50 - 200
13C9 PFNA	103		50 - 200
13C7 PFUnA	103		50 - 200
13C2 PFDoA	101		50 - 200
13C4 PFBA	87		50 - 200
13C5 PFPeA	98		50 - 200
13C3 PFBS	107		50 - 200
13C3 PFHxS	110		50 - 200
13C8 PFOS	115		50 - 200
13C2-4:2-FTS	114		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	107		50 - 200

**Lab Sample ID: 380-215228-1 MS**

**Matrix: Water**

**Analysis Batch: 229158**

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Prep Type: Total/NA**

**Prep Batch: 228770**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		120	99.5		ng/L		83	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		120	104		ng/L		86	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		120	113		ng/L		94	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-215228-1 MS**

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Matrix: Water**

**Prep Type: Total/NA**

**Analysis Batch: 229158**

**Prep Batch: 228770**

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Hexafluoropropylene Oxide	<2.0		120	109		ng/L		91	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	3.0		120	118		ng/L		96	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		120	118		ng/L		98	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		120	107		ng/L		89	70 - 130
Perfluoroheptanoic acid (PFHpA)	2.0		120	112		ng/L		91	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	3.7		120	114		ng/L		91	70 - 130
Perfluorohexanoic acid (PFHxA)	4.3		120	117		ng/L		93	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		120	105		ng/L		88	70 - 130
Perfluorooctanesulfonic acid (PFOS)	4.7		120	108		ng/L		86	70 - 130
Perfluorooctanoic acid (PFOA)	3.9		120	112		ng/L		90	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		120	109		ng/L		91	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		120	113		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		120	106		ng/L		88	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		120	117		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		120	106		ng/L		88	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		120	115		ng/L		96	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		120	123		ng/L		102	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		120	122		ng/L		102	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		120	116		ng/L		97	70 - 130
Perfluoropentanoic acid (PFPeA)	4.5		120	117		ng/L		94	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		120	105		ng/L		88	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		120	110		ng/L		91	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	118		50 - 200
13C6 PFDA	105		50 - 200
13C5 PFHxA	107		50 - 200
13C4 PFHpA	108		50 - 200
13C8 PFOA	113		50 - 200
13C9 PFNA	114		50 - 200
13C7 PFUnA	114		50 - 200
13C2 PFDoA	111		50 - 200
13C4 PFBA	109		50 - 200
13C5 PFPeA	106		50 - 200
13C3 PFBS	107		50 - 200
13C3 PFHxS	109		50 - 200
13C8 PFOS	113		50 - 200

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-215228-1 MS**

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Matrix: Water**

**Prep Type: Total/NA**

**Analysis Batch: 229158**

**Prep Batch: 228770**

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C2-4:2-FTS	107		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	109		50 - 200

**Lab Sample ID: 380-215228-1 MSD**

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Matrix: Water**

**Prep Type: Total/NA**

**Analysis Batch: 229158**

**Prep Batch: 228770**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD MSD		Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
				Result	Qualifier						
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		120	103		ng/L		85	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		120	104		ng/L		87	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		120	115		ng/L		95	70 - 130	1	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		120	115		ng/L		95	70 - 130	5	30
Perfluorobutanesulfonic acid (PFBS)	3.0		120	118		ng/L		96	70 - 130	0	30
Perfluorodecanoic acid (PFDA)	<2.0		120	120		ng/L		100	70 - 130	2	30
Perfluorododecanoic acid (PFDoA)	<2.0		120	109		ng/L		91	70 - 130	2	30
Perfluoroheptanoic acid (PFHpA)	2.0		120	115		ng/L		94	70 - 130	3	30
Perfluorohexanesulfonic acid (PFHxS)	3.7		120	117		ng/L		94	70 - 130	3	30
Perfluorohexanoic acid (PFHxA)	4.3		120	115		ng/L		92	70 - 130	1	30
Perfluorononanoic acid (PFNA)	<2.0		120	109		ng/L		91	70 - 130	3	30
Perfluorooctanesulfonic acid (PFOS)	4.7		120	112		ng/L		89	70 - 130	4	30
Perfluorooctanoic acid (PFOA)	3.9		120	119		ng/L		95	70 - 130	5	30
Perfluoroundecanoic acid (PFUnA)	<2.0		120	114		ng/L		95	70 - 130	5	30
Perfluorobutanoic acid (PFBA)	<2.0		120	110		ng/L		90	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		120	110		ng/L		92	70 - 130	4	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		120	117		ng/L		97	70 - 130	0	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		120	110		ng/L		91	70 - 130	3	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		120	117		ng/L		98	70 - 130	2	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		120	126		ng/L		104	70 - 130	2	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		120	125		ng/L		104	70 - 130	2	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		120	126		ng/L		105	70 - 130	8	30
Perfluoropentanoic acid (PFPeA)	4.5		120	124		ng/L		99	70 - 130	6	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		120	108		ng/L		89	70 - 130	2	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		120	114		ng/L		94	70 - 130	3	30

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	109		50 - 200
13C6 PFDA	102		50 - 200
13C5 PFHxA	107		50 - 200
13C4 PFHpA	105		50 - 200
13C8 PFOA	104		50 - 200
13C9 PFNA	111		50 - 200
13C7 PFUnA	112		50 - 200
13C2 PFDoA	107		50 - 200
13C4 PFBA	112		50 - 200
13C5 PFPeA	104		50 - 200
13C3 PFBS	108		50 - 200
13C3 PFHxS	106		50 - 200
13C8 PFOS	111		50 - 200
13C2-4:2-FTS	104		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	107		50 - 200

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

**Lab Sample ID: MBL 380-228643/19-A**  
**Matrix: Water**  
**Analysis Batch: 229059**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 228643**

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
Perfluorotridecanoic acid (PFTTrDA)	<0.36		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		05/21/26 10:42	05/22/26 16:56	1

Surrogate	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
d5-NEtFOSAA	105		70 - 130	05/21/26 10:42	05/22/26 16:56	1
13C2 PFHxA	104		70 - 130	05/21/26 10:42	05/22/26 16:56	1
13C2 PFDA	109		70 - 130	05/21/26 10:42	05/22/26 16:56	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

**Lab Sample ID: MBL 380-228643/19-A**  
**Matrix: Water**  
**Analysis Batch: 229059**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 228643**

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	96		70 - 130	05/21/26 10:42	05/22/26 16:56	1

**Lab Sample ID: LCS 380-228643/21-A**  
**Matrix: Water**  
**Analysis Batch: 229059**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 228643**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	50.1	46.7		ng/L		93	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	50.1	49.2		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	50.1	51.6		ng/L		103	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.1	49.5		ng/L		99	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.1	46.8		ng/L		93	70 - 130
Perfluorohexanoic acid (PFHxA)	50.1	51.3		ng/L		102	70 - 130
Perfluorododecanoic acid (PFDoA)	50.1	50.9		ng/L		102	70 - 130
Perfluorooctanoic acid (PFOA)	50.1	50.4		ng/L		101	70 - 130
Perfluorodecanoic acid (PFDA)	50.1	51.4		ng/L		103	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	50.1	49.6		ng/L		99	70 - 130
Perfluorobutanesulfonic acid (PFBS)	50.1	48.6		ng/L		97	70 - 130
Perfluoroheptanoic acid (PFHpA)	50.1	54.8		ng/L		109	70 - 130
Perfluorononanoic acid (PFNA)	50.1	51.6		ng/L		103	70 - 130
Perfluorotetradecanoic acid (PFTA)	50.1	44.3		ng/L		88	70 - 130
Perfluorotridecanoic acid (PFTrDA)	50.1	52.7		ng/L		105	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	50.1	48.3		ng/L		96	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.1	49.1		ng/L		98	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.1	49.3		ng/L		98	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	103		70 - 130
13C2 PFHxA	111		70 - 130
13C2 PFDA	118		70 - 130
13C3-GenX	113		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

**Lab Sample ID: MRL 380-228643/20-A**  
**Matrix: Water**  
**Analysis Batch: 229059**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 228643**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	2.00	1.85	J	ng/L		93	50 - 150
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	2.00	2.10	J	ng/L		105	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.08	J	ng/L		104	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	2.14	J	ng/L		107	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.97	J	ng/L		98	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.17	J	ng/L		108	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.17	J	ng/L		108	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.81	J	ng/L		90	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.09	J	ng/L		104	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.32	J	ng/L		116	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	1.98	J	ng/L		99	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	2.20	J	ng/L		110	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	1.91	J	ng/L		95	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.79	J	ng/L		89	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.02	J	ng/L		101	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	110		70 - 130
13C2 PFHxA	109		70 - 130
13C2 PFDA	107		70 - 130
13C3-GenX	100		70 - 130

**Lab Sample ID: 380-214936-D-2-A MSD**  
**Matrix: Water**  
**Analysis Batch: 229059**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 228643**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide	<2.0		50.1	46.8		ng/L		93	70 - 130	3	30
Dimer Acid (HFPO-DA/GenX)											
Perfluorooctanesulfonic acid (PFOS)	5.4		50.1	55.6		ng/L		100	70 - 130	3	30
Perfluoroundecanoic acid (PFUnA)	<2.0		50.1	50.6		ng/L		101	70 - 130	2	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.1	49.3		ng/L		98	70 - 130	6	30

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

**Lab Sample ID: 380-214936-D-2-A MSD**

**Client Sample ID: Matrix Spike Duplicate**

**Matrix: Water**

**Prep Type: Total/NA**

**Analysis Batch: 229059**

**Prep Batch: 228643**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	RPD	RPD
	Result	Qualifier	Added	Result	Qualifier				Limits		
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		50.1	51.3		ng/L		102	70 - 130	1	30
Perfluorohexanoic acid (PFHxA)	<2.0		50.1	51.0		ng/L		101	70 - 130	2	30
Perfluorododecanoic acid (PFDoA)	<2.0		50.1	51.0		ng/L		102	70 - 130	3	30
Perfluorooctanoic acid (PFOA)	<2.0		50.1	51.7		ng/L		102	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		50.1	52.3		ng/L		104	70 - 130	2	30
Perfluorohexanesulfonic acid (PFHxS)	2.1		50.1	54.0		ng/L		104	70 - 130	4	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		50.1	49.5		ng/L		97	70 - 130	4	30
Perfluoroheptanoic acid (PFHpA)	<2.0		50.1	51.0		ng/L		102	70 - 130	2	30
Perfluorononanoic acid (PFNA)	<2.0		50.1	51.0		ng/L		102	70 - 130	3	30
Perfluorotetradecanoic acid (PFTA)	<2.0		50.1	39.5		ng/L		79	70 - 130	15	30
Perfluorotridecanoic acid (PFTrDA)	<2.0		50.1	51.8		ng/L		103	70 - 130	4	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		50.1	50.0		ng/L		100	70 - 130	2	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		50.1	49.2		ng/L		98	70 - 130	6	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		50.1	48.9		ng/L		98	70 - 130	2	30
<b>Surrogate</b>		<b>MSD</b>	<b>MSD</b>								
		<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>							
d5-NEtFOSAA		105		70 - 130							
13C2 PFHxA		113		70 - 130							
13C2 PFDA		110		70 - 130							
13C3-GenX		106		70 - 130							

**Lab Sample ID: 380-214936-E-2-A MS**

**Client Sample ID: Matrix Spike**

**Matrix: Water**

**Prep Type: Total/NA**

**Analysis Batch: 229059**

**Prep Batch: 228643**

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec	RPD	RPD
	Result	Qualifier	Added	Result	Qualifier				Limits		
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.2	48.3		ng/L		96	70 - 130		
Perfluorooctanesulfonic acid (PFOS)	5.4		50.2	54.2		ng/L		97	70 - 130		
Perfluoroundecanoic acid (PFUnA)	<2.0		50.2	49.8		ng/L		99	70 - 130		
N-methylperfluorooctanesulfonamide acetic acid (NMeFOSAA)	<2.0		50.2	52.1		ng/L		104	70 - 130		
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		50.2	50.9		ng/L		101	70 - 130		
Perfluorohexanoic acid (PFHxA)	<2.0		50.2	51.9		ng/L		102	70 - 130		
Perfluorododecanoic acid (PFDoA)	<2.0		50.2	52.7		ng/L		105	70 - 130		
Perfluorooctanoic acid (PFOA)	<2.0		50.2	52.3		ng/L		103	70 - 130		
Perfluorodecanoic acid (PFDA)	<2.0		50.2	53.4		ng/L		106	70 - 130		

Eurofins Pomona



# QC Association Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-215228-1  
 SDG: PFAS: Ka'amilo Wells P2

## LCMS

### Prep Batch: 228643

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	537.1 DW	
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	537.1 DW	
MBL 380-228643/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-228643/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-228643/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-214936-D-2-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	
380-214936-E-2-A MS	Matrix Spike	Total/NA	Water	537.1 DW	

### Prep Batch: 228770

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	
MBL 380-228770/20-A	Method Blank	Total/NA	Water	533	
LCS 380-228770/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-228770/21-A	Lab Control Sample	Total/NA	Water	533	
380-215228-1 MS	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	
380-215228-1 MSD	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	

### Analysis Batch: 229059

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	EPA 537.1 V2	228643
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	EPA 537.1 V2	228643
MBL 380-228643/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	228643
LCS 380-228643/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	228643
MRL 380-228643/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	228643
380-214936-D-2-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	228643
380-214936-E-2-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	228643

### Analysis Batch: 229158

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	228770
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	228770
MBL 380-228770/20-A	Method Blank	Total/NA	Water	533	228770
LCS 380-228770/22-A	Lab Control Sample	Total/NA	Water	533	228770
MRL 380-228770/21-A	Lab Control Sample	Total/NA	Water	533	228770
380-215228-1 MS	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	228770
380-215228-1 MSD	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	228770

# Lab Chronicle

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-215228-1  
 SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-1**

Date Collected: 05/18/26 12:44

Matrix: Water

Date Received: 05/20/26 09:27

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			228770	E2HD	EA POM	05/21/26 18:50
Total/NA	Analysis	533		1	229158	Y5FM	EA POM	05/23/26 11:37
Total/NA	Prep	537.1 DW			228643	L9UA	EA POM	05/21/26 10:42
Total/NA	Analysis	EPA 537.1 V2		1	229059	SZ9R	EA POM	05/22/26 20:10

**Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)**

**Lab Sample ID: 380-215228-2**

Date Collected: 05/18/26 12:44

Matrix: Water

Date Received: 05/20/26 09:27

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			228770	E2HD	EA POM	05/21/26 18:50
Total/NA	Analysis	533		1	229158	Y5FM	EA POM	05/23/26 14:29
Total/NA	Prep	537.1 DW			228643	L9UA	EA POM	05/21/26 10:42
Total/NA	Analysis	EPA 537.1 V2		1	229059	SZ9R	EA POM	05/22/26 20:19

**Laboratory References:**

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

## Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

\* Accreditation/Certification renewal pending - accreditation/certification considered valid.

# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

**Protocol References:**

EPA = US Environmental Protection Agency

**Laboratory References:**

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-215228-1  
SDG: PFAS: Ka'amilo Wells P2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-215228-1	Ka'amilo Wells P2 (331-600-WL085)	Water	05/18/26 12:44	05/20/26 09:27	HI0000331
380-215228-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Water	05/18/26 12:44	05/20/26 09:27	HI0000331

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

**Monrovia, CA (Suite 100)**

750 Royal Oaks Drive Suite 100  
 Monrovia, CA 91016  
 Phone (626) 386-1100

**Chain of Custody Record**



Enviro  
Amer



<b>Client Information</b>	Sampler bailey	Lab PM Lopez, Maria	Carrier Tracking No(s):	COC No:
Client Contact kirk iwamoto	Phone +1 808 748 5840	E-Mail Maria.Lopez@et.euronisus.com	State of Origin:	Page: 80-215228 COC

Company City & County of Honolulu	PWSID.	<b>Analysis Requested</b>	Job #:
--------------------------------------	--------	---------------------------	--------

Address: 630 South Beretania Street; Chemistry Lab	Due Date Requested:	Field Filled Sample (Yes or No) Perform MS/MSD (Yes or No) SUBCONTRACT - 625 PAH Physis LL (EAL) + TICs 8015B_GRO_LL - (MOD) GRO 8015B_DRO_LL_CS - HNL Ranges: C10-C24/C24-C36/C8-C18 525.2_PREC - (MOD) 525plus PLUS TICs 537.1_DW_PREC - 537.1 Full List 533 - All Analytes	Preservation Codes: A - HCL B - NaOH C - Zn Acetate D - Nitric Acid E - NaHSO4 F - MeOH G - Amchlor H - Ascorbic Acid I - Ice J - DI Water K - EDTA L - EDA M - Hexane N - None O - AsNaO2 P - Na2O4S Q - Na2SO3 R - Na2S2O3 S - H2SO4 T - TSP Dodecahydrate U - Acetone V - MCAA W - pH 4-5 Y - Trizma Z - other (specify)
City: Honolulu	TAT Requested (days): <b>RUSH</b>		
State, Zip: HI, 96843	Compliance Project: <input type="checkbox"/> No		
Phone: 808-748-5840 (tel)	PO #: C20525101 exp 05312023		
Email: kiwamoto@hbws.org	WO #:		
Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill	Project #: 38001111		
Site:	SSOW#:	Total Number of Containers:	

Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, S=solid, O=waste/oil, DT=TISSUE, A=AL)	Field Filled Sample (Yes or No)	Perform MS/MSD (Yes or No)	Subcontract	8015B_GRO_LL - (MOD) GRO	8015B_DRO_LL_CS - HNL Ranges: C10-C24/C24-C36/C8-C18	525.2_PREC - (MOD) 525plus PLUS TICs	537.1_DW_PREC - 537.1 Full List	533 - All Analytes	Other:	Special Instructions/Note:
Ka'amilo Wells P2 (331-600-WL085)	18-May-2026	1244	G	Water	X	X	R	A	Q	QA	Y	I		
FB: Ka'amilo Wells P2 (331-600-WL085)	18-May-2026	1244		Water							1	1		

<b>Possible Hazard Identification</b> <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological	<b>Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)</b> <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months
Deliverable Requested I, II, III, IV, Other (specify)	Special Instructions/QC Requirements

Empty Kit Relinquished by	Date	Time	Method of Shipment
			Fed X 8720 0324 5047
Relinquished by	Date/Time	Company	Received by
	19 MAY 2026 1400	HBWS	Mark Urwita
Relinquished by	Date/Time	Company	Received by

Custody Seals Intact: <input type="checkbox"/> Yes <input type="checkbox"/> No	Custody Seal No.	Cooler Temperature(s) °C and Other Remarks (631A) 2-2+0.0 2.2 gel-frozen
---	------------------	---



ORIGIN ID:HIKA (808) 748-5840  
BWS CHEMLAB  
HONOLULU BOARD OF WATER SUPPLY  
630 S. BERETANIA ST  
CHEMICAL LABORATORY  
HONOLULU, HI 96843  
UNITED STATES US

SHIP DATE 19MAY26  
ACTWGT 62.00 LB  
CAD 258050552/INET4535

BILL RECIPIENT

TO EUROFINS RECEIVING DEPARTMENT  
EUROFINS DRINKING WATER TESTING  
941 CORPORATE CENTER DR

POMONA CA 91768


(626) 386-1100

REF

INV  
PO:

DEPT.

58KJ3/A906484B



FedEx Express



J281828T201V

6 of 8

WED - 20 MAY 10:30A  
PRIORITY OVERNIGHT

MPS#

0263

8720 0324 5847


Mstr# 8720 0324 5799

0201

91768

WM ONTA

CA-US ONT



(631A) 2 2+00 22 901-frozen

Manana Blank Urotin 5/20/26 927

After printing this label  
CONSIGNEE COPY - PLEASE PLACE IN FRONT OF POUCH  
1. Fold the printed page along the horizontal line  
2. Place label in shipping pouch and affix it to your shipment.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-215228-1  
SDG Number: PFAS: Ka'amilo Wells P2

**Login Number: 215228**

**List Number: 1**

**Creator: Avila, Ivan**

**List Source: Eurofins Pomona**

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

