

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 5/12/2026 4:47:20 PM

JOB DESCRIPTION

RED-HILL
PFAS: Halawa Wells P1

JOB NUMBER

380-210974-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
5/12/2026 4:47:20 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-210974-1

Job ID: 380-210974-1

Eurofins Pomona

Job Narrative 380-210974-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 4/29/2026 9:47 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 3.6°C.

PFAS

Method EPA 537.1 V2: PFAS results by 537.1 HALAWA WELLS P1 (331-023-WL065) (380-210974-1) and FB: HALAWA WELLS P1 (331-023-WL065) (380-210974-2) collected on 04/27/26 are a resample for Halawa Wells P1 (331-023-WL065) job # 380-209881-1 collected on 04/20/26. (XWB4)

EPA 537.1 and EPA 533 are two distinct methods for the analysis of PFAS in drinking water. The analyses are conducted on differing instrumentation, with calibrations, extraction solvents and sample preservatives being dissimilar among the two methods. Therefore it is probable and not unexpected to see the methods having slight variations in analytical results:HALAWA WELLS P1 (331-023-WL065) (380-210974-1). (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)
PWSID Number: HI0000331

Lab Sample ID: 380-210974-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	2.4		2.0	ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	2.9		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.3		2.0	ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	4.0		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.7		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanoic acid (PFHxA)	3.3		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	2.0		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.9		2.0	ng/L	1		EPA 537.1 V2	Total/NA

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)
PWSID Number: HI0000331

Lab Sample ID: 380-210974-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-1

Date Collected: 04/27/26 10:39

Matrix: Drinking Water

Date Received: 04/29/26 09:47

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorohexanesulfonic acid (PFHxS)	2.4		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorohexanoic acid (PFHxA)	2.9		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorooctanesulfonic acid (PFOS)	2.3		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoropentanoic acid (PFPeA)	4.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 19:51	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	104		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C6 PFDA	107		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C5 PFHxA	108		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C4 PFHpA	109		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C8 PFOA	112		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C9 PFNA	110		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C7 PFUnA	105		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C2 PFDoA	109		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C4 PFBA	109		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C5 PFPeA	109		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C3 PFBS	107		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C3 PFHxS	110		50 - 200	05/08/26 06:09	05/08/26 19:51	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-1

Date Collected: 04/27/26 10:39

Matrix: Drinking Water

Date Received: 04/29/26 09:47

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	108		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C2-4:2-FTS	109		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C2-6:2-FTS	103		50 - 200	05/08/26 06:09	05/08/26 19:51	1
13C2-8:2-FTS	102		50 - 200	05/08/26 06:09	05/08/26 19:51	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorooctanesulfonic acid (PFOS)	2.7		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorohexanoic acid (PFHxA)	3.3		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorooctanoic acid (PFOA)	2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorohexanesulfonic acid (PFHxS)	2.9		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:11	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	112		70 - 130			05/01/26 09:00	05/01/26 19:11	1
13C2 PFHxA	106		70 - 130			05/01/26 09:00	05/01/26 19:11	1
13C2 PFDA	114		70 - 130			05/01/26 09:00	05/01/26 19:11	1
13C3-GenX	106		70 - 130			05/01/26 09:00	05/01/26 19:11	1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-2

Date Collected: 04/27/26 10:39

Matrix: Water

Date Received: 04/29/26 09:47

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-2

Date Collected: 04/27/26 10:39

Matrix: Water

Date Received: 04/29/26 09:47

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		05/08/26 06:09	05/08/26 20:01	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	96		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C6 PFDA	104		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C5 PFHxA	101		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C4 PFHpA	105		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C8 PFOA	110		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C9 PFNA	107		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C7 PFUnA	106		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C2 PFDoA	107		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C4 PFBA	101		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C5 PFPeA	101		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C3 PFBS	102		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C3 PFHxS	108		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C8 PFOS	107		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C2-4:2-FTS	104		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C2-6:2-FTS	98		50 - 200			05/08/26 06:09	05/08/26 20:01	1
13C2-8:2-FTS	95		50 - 200			05/08/26 06:09	05/08/26 20:01	1

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-2

Date Collected: 04/27/26 10:39

Matrix: Water

Date Received: 04/29/26 09:47

PWSID Number: HI0000331

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		05/01/26 09:00	05/01/26 19:21	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	102		70 - 130			05/01/26 09:00	05/01/26 19:21	1
13C2 PFHxA	98		70 - 130			05/01/26 09:00	05/01/26 19:21	1
13C2 PFDA	111		70 - 130			05/01/26 09:00	05/01/26 19:21	1
13C3-GenX	100		70 - 130			05/01/26 09:00	05/01/26 19:21	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-1

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.4		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.3		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.7		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.9		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-2

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-210974-1
 SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-210974-1	HALAWA WELLS P1 (331-023-V	112	106	114	106

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-210960-B-1-A MS	Matrix Spike	107	112	115	112
380-210960-C-1-A MSD	Matrix Spike Duplicate	111	109	115	110
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	102	98	111	100
LCS 380-224319/21-A	Lab Control Sample	111	107	115	105
MBL 380-224319/19-A	Method Blank	111	102	112	98
MRL 380-224319/20-A	Lab Control Sample	107	102	110	96

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-210974-1	HALAWA WELLS P1 (331-023-V)	104	107	108	109	112	110	105	109

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-210974-1	HALAWA WELLS P1 (331-023-V)	109	109	107	110	108	109	103	102

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-210960-E-1-A MS	Matrix Spike	111	109	109	112	114	112	109	111
380-210960-F-1-A MSD	Matrix Spike Duplicate	117	111	110	113	114	114	111	116
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	96	104	101	105	110	107	106	107
LCS 380-225731/22-A	Lab Control Sample	116	114	114	115	114	115	113	118
MBL 380-225731/20-A	Method Blank	99	99	107	104	105	106	99	102
MRL 380-225731/21-A	Lab Control Sample	102	106	113	109	109	109	106	104

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-210960-E-1-A MS	Matrix Spike	108	115	110	108	109	99	95	95
380-210960-F-1-A MSD	Matrix Spike Duplicate	113	117	108	110	112	100	97	97
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	101	101	102	108	107	104	98	95
LCS 380-225731/22-A	Lab Control Sample	113	115	110	111	112	101	95	94
MBL 380-225731/20-A	Method Blank	101	100	101	100	102	97	97	88
MRL 380-225731/21-A	Lab Control Sample	104	105	108	112	111	103	98	103

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA

Isotope Dilution Summary

Client: City & County of Honolulu

Project/Site: RED-HILL

13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDoA = 13C2 PFDoA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-225731/20-A
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 225731

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		05/08/26 06:09	05/08/26 16:09	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	99		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C6 PFDA	99		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C5 PFHxA	107		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C4 PFHpA	104		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C8 PFOA	105		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C9 PFNA	106		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C7 PFUnA	99		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C2 PFDoA	102		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C4 PFBA	101		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C5 PFPeA	100		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C3 PFBS	101		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C3 PFHxS	100		50 - 200	05/08/26 06:09	05/08/26 16:09	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-225731/20-A
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 225731

<i>Isotope Dilution</i>	<i>MBL %Recovery</i>	<i>MBL Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C8 PFOS	102		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C2-4:2-FTS	97		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C2-6:2-FTS	97		50 - 200	05/08/26 06:09	05/08/26 16:09	1
13C2-8:2-FTS	88		50 - 200	05/08/26 06:09	05/08/26 16:09	1

Lab Sample ID: LCS 380-225731/22-A
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 225731

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	120	102		ng/L		85	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	120	101		ng/L		84	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	103		ng/L		86	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	102		ng/L		84	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	104		ng/L		86	70 - 130
Perfluorodecanoic acid (PFDA)	120	100		ng/L		83	70 - 130
Perfluorododecanoic acid (PFDoA)	120	103		ng/L		85	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	102		ng/L		84	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	104		ng/L		87	70 - 130
Perfluorohexanoic acid (PFHxA)	120	101		ng/L		84	70 - 130
Perfluorononanoic acid (PFNA)	120	101		ng/L		84	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	101		ng/L		84	70 - 130
Perfluorooctanoic acid (PFOA)	120	104		ng/L		86	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	103		ng/L		86	70 - 130
Perfluorobutanoic acid (PFBA)	120	101		ng/L		84	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	105		ng/L		88	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	99.0		ng/L		82	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	109		ng/L		90	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	106		ng/L		88	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	120	106		ng/L		88	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	98.5		ng/L		82	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	100		ng/L		83	70 - 130
Perfluoropentanoic acid (PFPeA)	120	99.0		ng/L		82	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	103		ng/L		86	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-225731/22-A
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 225731

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	105		ng/L		88	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	116		50 - 200				
13C6 PFDA	114		50 - 200				
13C5 PFHxA	114		50 - 200				
13C4 PFHpA	115		50 - 200				
13C8 PFOA	114		50 - 200				
13C9 PFNA	115		50 - 200				
13C7 PFUnA	113		50 - 200				
13C2 PFDoA	118		50 - 200				
13C4 PFBA	113		50 - 200				
13C5 PFPeA	115		50 - 200				
13C3 PFBS	110		50 - 200				
13C3 PFHxS	111		50 - 200				
13C8 PFOS	112		50 - 200				
13C2-4:2-FTS	101		50 - 200				
13C2-6:2-FTS	95		50 - 200				
13C2-8:2-FTS	94		50 - 200				

Lab Sample ID: MRL 380-225731/21-A
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 225731

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.82	J	ng/L		90	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	1.75	J	ng/L		87	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.87	J	ng/L		93	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.88	J	ng/L		94	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.92	J	ng/L		96	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	1.94	J	ng/L		97	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.99	J	ng/L		99	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	1.95	J	ng/L		97	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	1.84	J	ng/L		92	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	1.91	J	ng/L		95	50 - 150
Perfluorononanoic acid (PFNA)	2.01	1.89	J	ng/L		94	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.82	J	ng/L		90	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	1.91	J	ng/L		95	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.77	J	ng/L		88	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	1.95	J	ng/L		97	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-225731/21-A
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 225731

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	1.89	J	ng/L		94	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	2.06	J	ng/L		103	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	2.13	J	ng/L		106	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	1.69	J	ng/L		84	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.01	1.77	J	ng/L		88	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	1.95	J	ng/L		97	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.88	J	ng/L		94	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	2.04	J	ng/L		102	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	1.86	J	ng/L		92	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	1.82	J	ng/L		90	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	102		50 - 200
13C6 PFDA	106		50 - 200
13C5 PFHxA	113		50 - 200
13C4 PFHpA	109		50 - 200
13C8 PFOA	109		50 - 200
13C9 PFNA	109		50 - 200
13C7 PFUnA	106		50 - 200
13C2 PFDoA	104		50 - 200
13C4 PFBA	104		50 - 200
13C5 PFPeA	105		50 - 200
13C3 PFBS	108		50 - 200
13C3 PFHxS	112		50 - 200
13C8 PFOS	111		50 - 200
13C2-4:2-FTS	103		50 - 200
13C2-6:2-FTS	98		50 - 200
13C2-8:2-FTS	103		50 - 200

Lab Sample ID: 380-210960-E-1-A MS
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 225731

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		120	103		ng/L		86	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		120	105		ng/L		87	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		120	105		ng/L		87	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-210960-E-1-A MS
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 225731

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		120	105		ng/L		87	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	2.2		120	106		ng/L		86	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		120	107		ng/L		89	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		120	108		ng/L		90	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		120	108		ng/L		88	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		120	106		ng/L		87	70 - 130
Perfluorohexanoic acid (PFHxA)	3.1		120	106		ng/L		85	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		120	106		ng/L		88	70 - 130
Perfluorooctanesulfonic acid (PFOS)	2.5		120	108		ng/L		88	70 - 130
Perfluorooctanoic acid (PFOA)	2.7		120	103		ng/L		84	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		120	102		ng/L		85	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		120	106		ng/L		87	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		120	110		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		120	102		ng/L		84	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		120	112		ng/L		93	70 - 130
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		120	106		ng/L		88	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		120	105		ng/L		87	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		120	107		ng/L		88	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		120	102		ng/L		85	70 - 130
Perfluoropentanoic acid (PFPeA)	3.2		120	104		ng/L		83	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		120	108		ng/L		89	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		120	104		ng/L		87	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	111		50 - 200
13C6 PFDA	109		50 - 200
13C5 PFHxA	109		50 - 200
13C4 PFHpA	112		50 - 200
13C8 PFOA	114		50 - 200
13C9 PFNA	112		50 - 200
13C7 PFUnA	109		50 - 200
13C2 PFDoA	111		50 - 200
13C4 PFBA	108		50 - 200
13C5 PFPeA	115		50 - 200
13C3 PFBS	110		50 - 200
13C3 PFHxS	108		50 - 200
13C8 PFOS	109		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-210960-E-1-A MS
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 225731

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	99		50 - 200
13C2-6:2-FTS	95		50 - 200
13C2-8:2-FTS	95		50 - 200

Lab Sample ID: 380-210960-F-1-A MSD
Matrix: Water
Analysis Batch: 225892

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 225731

<i>Analyte</i>	<i>Sample Result</i>	<i>Sample Qualifier</i>	<i>Spike Added</i>	<i>MSD Result</i>	<i>MSD Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>	<i>RPD</i>	<i>RPD Limit</i>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		120	104		ng/L		87	70 - 130	1	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		120	104		ng/L		86	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		120	107		ng/L		89	70 - 130	1	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		120	107		ng/L		89	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	2.2		120	111		ng/L		90	70 - 130	5	30
Perfluorodecanoic acid (PFDA)	<2.0		120	107		ng/L		89	70 - 130	1	30
Perfluorododecanoic acid (PFDoA)	<2.0		120	108		ng/L		90	70 - 130	0	30
Perfluoroheptanoic acid (PFHpA)	<2.0		120	111		ng/L		91	70 - 130	3	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		120	107		ng/L		88	70 - 130	0	30
Perfluorohexanoic acid (PFHxA)	3.1		120	109		ng/L		88	70 - 130	3	30
Perfluorononanoic acid (PFNA)	<2.0		120	107		ng/L		89	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	2.5		120	109		ng/L		88	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	2.7		120	107		ng/L		87	70 - 130	4	30
Perfluoroundecanoic acid (PFUnA)	<2.0		120	107		ng/L		89	70 - 130	5	30
Perfluorobutanoic acid (PFBA)	<2.0		120	108		ng/L		88	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		120	112		ng/L		93	70 - 130	2	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		120	107		ng/L		89	70 - 130	6	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		120	109		ng/L		91	70 - 130	2	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		120	119		ng/L		99	70 - 130	12	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		120	114		ng/L		95	70 - 130	8	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		120	108		ng/L		90	70 - 130	2	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		120	106		ng/L		88	70 - 130	3	30
Perfluoropentanoic acid (PFPeA)	3.2		120	105		ng/L		84	70 - 130	1	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		120	108		ng/L		90	70 - 130	1	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		120	103		ng/L		85	70 - 130	1	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

<i>Isotope Dilution</i>	<i>MSD</i>	<i>MSD</i>	<i>Limits</i>
	<i>%Recovery</i>	<i>Qualifier</i>	
13C3 HFPO-DA	117		50 - 200
13C6 PFDA	111		50 - 200
13C5 PFHxA	110		50 - 200
13C4 PFHpA	113		50 - 200
13C8 PFOA	114		50 - 200
13C9 PFNA	114		50 - 200
13C7 PFUnA	111		50 - 200
13C2 PFDoA	116		50 - 200
13C4 PFBA	113		50 - 200
13C5 PFPeA	117		50 - 200
13C3 PFBS	108		50 - 200
13C3 PFHxS	110		50 - 200
13C8 PFOS	112		50 - 200
13C2-4:2-FTS	100		50 - 200
13C2-6:2-FTS	97		50 - 200
13C2-8:2-FTS	97		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-224319/19-A
Matrix: Water
Analysis Batch: 224453

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 224319

<i>Analyte</i>	<i>MBL</i>	<i>MBL</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>Result</i>	<i>Qualifier</i>						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		05/01/26 09:00	05/01/26 17:42	1
<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>%Recovery</i>	<i>Qualifier</i>						
d5-NEtFOSAA	111		70 - 130			05/01/26 09:00	05/01/26 17:42	1
13C2 PFHxA	102		70 - 130			05/01/26 09:00	05/01/26 17:42	1
13C2 PFDA	112		70 - 130			05/01/26 09:00	05/01/26 17:42	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-224319/19-A
Matrix: Water
Analysis Batch: 224453

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 224319

<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3-GenX	98	Qualifier	70 - 130	05/01/26 09:00	05/01/26 17:42	1

Lab Sample ID: LCS 380-224319/21-A
Matrix: Water
Analysis Batch: 224453

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 224319

<i>Analyte</i>	<i>Spike</i>	<i>LCS</i>	<i>LCS</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i>	<i>Limits</i>
<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>					
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	50.1	47.1	ng/L	94	70 - 130			
Perfluorooctanesulfonic acid (PFOS)	50.1	51.1	ng/L	102	70 - 130			
Perfluoroundecanoic acid (PFUnA)	50.1	51.2	ng/L	102	70 - 130			
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.1	49.4	ng/L	99	70 - 130			
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.1	49.6	ng/L	99	70 - 130			
Perfluorohexanoic acid (PFHxA)	50.1	49.4	ng/L	99	70 - 130			
Perfluorododecanoic acid (PFDoA)	50.1	49.7	ng/L	99	70 - 130			
Perfluorooctanoic acid (PFOA)	50.1	48.2	ng/L	96	70 - 130			
Perfluorodecanoic acid (PFDA)	50.1	51.6	ng/L	103	70 - 130			
Perfluorohexanesulfonic acid (PFHxS)	50.1	48.9	ng/L	98	70 - 130			
Perfluorobutanesulfonic acid (PFBS)	50.1	50.4	ng/L	101	70 - 130			
Perfluoroheptanoic acid (PFHpA)	50.1	48.4	ng/L	97	70 - 130			
Perfluorononanoic acid (PFNA)	50.1	51.7	ng/L	103	70 - 130			
Perfluorotetradecanoic acid (PFTA)	50.1	41.2	ng/L	82	70 - 130			
Perfluorotridecanoic acid (PFTrDA)	50.1	50.5	ng/L	101	70 - 130			
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	50.1	50.7	ng/L	101	70 - 130			
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.1	50.7	ng/L	101	70 - 130			
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.1	49.2	ng/L	98	70 - 130			
<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>					
d5-NEtFOSAA	111		70 - 130					
13C2 PFHxA	107		70 - 130					
13C2 PFDA	115		70 - 130					
13C3-GenX	105		70 - 130					

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-224319/20-A
Matrix: Water
Analysis Batch: 224453

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 224319

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	1.80	J	ng/L		90	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.08	J	ng/L		104	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.19	J	ng/L		109	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	2.02	J	ng/L		101	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	2.03	J	ng/L		102	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.95	J	ng/L		97	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.03	J	ng/L		102	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	1.99	J	ng/L		99	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	1.97	J	ng/L		98	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.17	J	ng/L		108	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.04	J	ng/L		102	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.02	J	ng/L		101	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.11	J	ng/L		106	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	1.86	J	ng/L		93	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	2.10	J	ng/L		105	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.07	J	ng/L		103	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.03	J	ng/L		101	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.01	J	ng/L		100	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	107		70 - 130
13C2 PFHxA	102		70 - 130
13C2 PFDA	110		70 - 130
13C3-GenX	96		70 - 130

Lab Sample ID: 380-210960-B-1-A MS
Matrix: Water
Analysis Batch: 224453

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 224319

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.2	50.6		ng/L		101	70 - 130
Perfluorooctanesulfonic acid (PFOS)	2.6		50.2	53.4		ng/L		101	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		50.2	53.5		ng/L		107	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.2	49.7		ng/L		99	70 - 130

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

LCMS

Prep Batch: 224319

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-210974-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	537.1 DW	
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	537.1 DW	
MBL 380-224319/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-224319/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-224319/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-210960-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-210960-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 224453

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-210974-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	EPA 537.1 V2	224319
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	EPA 537.1 V2	224319
MBL 380-224319/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	224319
LCS 380-224319/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	224319
MRL 380-224319/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	224319
380-210960-B-1-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	224319
380-210960-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	224319

Prep Batch: 225731

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-210974-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	533	
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	533	
MBL 380-225731/20-A	Method Blank	Total/NA	Water	533	
LCS 380-225731/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-225731/21-A	Lab Control Sample	Total/NA	Water	533	
380-210960-E-1-A MS	Matrix Spike	Total/NA	Water	533	
380-210960-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 225892

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-210974-1	HALAWA WELLS P1 (331-023-WL065)	Total/NA	Drinking Water	533	225731
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	Total/NA	Water	533	225731
MBL 380-225731/20-A	Method Blank	Total/NA	Water	533	225731
LCS 380-225731/22-A	Lab Control Sample	Total/NA	Water	533	225731
MRL 380-225731/21-A	Lab Control Sample	Total/NA	Water	533	225731
380-210960-E-1-A MS	Matrix Spike	Total/NA	Water	533	225731
380-210960-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	225731

Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-210974-1
 SDG: PFAS: Halawa Wells P1

Client Sample ID: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-1

Date Collected: 04/27/26 10:39

Matrix: Drinking Water

Date Received: 04/29/26 09:47

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			225731	XTD8	EA POM	05/08/26 06:09
Total/NA	Analysis	533		1	225892	Y5FM	EA POM	05/08/26 19:51
Total/NA	Prep	537.1 DW			224319	L9UA	EA POM	05/01/26 09:00
Total/NA	Analysis	EPA 537.1 V2		1	224453	Y5FM	EA POM	05/01/26 19:11

Client Sample ID: FB: HALAWA WELLS P1 (331-023-WL065)

Lab Sample ID: 380-210974-2

Date Collected: 04/27/26 10:39

Matrix: Water

Date Received: 04/29/26 09:47

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			225731	XTD8	EA POM	05/08/26 06:09
Total/NA	Analysis	533		1	225892	Y5FM	EA POM	05/08/26 20:01
Total/NA	Prep	537.1 DW			224319	L9UA	EA POM	05/01/26 09:00
Total/NA	Analysis	EPA 537.1 V2		1	224453	Y5FM	EA POM	05/01/26 19:21

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-210974-1
SDG: PFAS: Halawa Wells P1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-210974-1	HALAWA WELLS P1 (331-023-WL065)	Drinking Water	04/27/26 10:39	04/29/26 09:47	HI0000331
380-210974-2	FB: HALAWA WELLS P1 (331-023-WL065)	Water	04/27/26 10:39	04/29/26 09:47	HI0000331

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Monrovia, CA (Suite 100)
 750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone (626) 386-1100

Chain of Custody Record



Environment Testing
 America



380-210974 COC

Client Information
 Client Contact: kirik iwamoto
 Phone: +1 808 748 5840
 Company: City & County of Honolulu
 Address: 630 South Beretania Street, Chemistry Lab
 City: Honolulu
 State, Zip: HI, 96843
 Phone: 808-748-5840 (tel)
 Email: kiwamoto@hibws.org
 Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill
 Site:

Analysis Requested
 Lab PII: Lopez, Maria
 E-Mail: Maria.Lopez@et.eurofins.com
 Camer Tracking No(s):
 State of Origin:
 COC No:
 Page: Page 1 of 1
 Job #:

Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (W=water, S=solid, O=soil, C=contaminant)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	Subcontract - 625 PAH Phys (LL (EAL) + TICs)	8016B GRO_LL - (MOD) GRO	8016B DRO_LL_CS - HNL Ranges: C10-C24/C24-C36/C8-C18	525.2_PRC - (MOD) 625PLU PLUS TICs	527.1_DW_PRC - 527.1 Full List	633 - All Analytes	Total Number of Containers	Special Instructions/Note:
Halawa Wells P1 (331-023-WL065)	27-Apr-2026	1039	G	Water	X	X								
FB: Halawa Wells P1 (331-023-WL065)	27-Apr-2026	1039												

Possible Hazard Identification
 Non-Hazard Flammable Skin Irritant Poison B Unknown Radiological
 Deliverable Requested: I, II, III, IV, Other (specify)

Empty Kit Relinquished by
 Date/Time: 28/04/2026
 Date/Time: 4/29/16
 Date/Time:
 Relinquished by:
 Relinquished by:
 Custody Seals Intact:
 Δ Yes Δ No
 Custody Seal No.

Method of Shipment: **FedEx 5711 6593 366**
 Date/Time: **4/29/16 9:47**
 Date/Time:
 Date/Time:
 Company: **FEA**
 Company:
 Company:
 Cooler Temperature(s) °C and Other Remarks:
79A 3.4 4.2 = 3.6 sel Frelca



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-210974-1
SDG Number: PFAS: Halawa Wells P1

Login Number: 210974

List Number: 1

Creator: Tran, Kristine

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

